

# PUBLISHED VERSION

Shah M Azarisman, Andrew Li, James D Richardson, Dennis T Wong, Seng Keong Chua, Michael Cursaro, Vince Schirripa, Kerry Williams, Ben Koschade, Mitra Shirazi, Julie Bradley, Karen Teo, Matthew Worthley, Stephen G Worthley

**Optimal planimetry location for MRI-derived mitral inflow velocity assessment of diastolic function**  
Abstracts from the 16th Annual SCMR Scientific Sessions, as published in Journal of Cardiovascular Magnetic Resonance, 2013 / vol.15, iss.Suppl. 1, pp.P90

© 2013 Azarisman et al; licensee BioMed Central Ltd. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/2.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Originally published at:

<http://doi.org/10.1186/1532-429X-15-S1-P90>

## PERMISSIONS

<http://creativecommons.org/licenses/by/2.0/>



Attribution 2.0 Generic (CC BY 2.0)

This is a human-readable summary of (and not a substitute for) the [license](#).

[Disclaimer](#)



### You are free to:

**Share** — copy and redistribute the material in any medium or format

**Adapt** — remix, transform, and build upon the material

for any purpose, even commercially.

The licensor cannot revoke these freedoms as long as you follow the license terms.

### Under the following terms:



**Attribution** — You must give **appropriate credit**, provide a link to the license, and **indicate if changes were made**. You may do so in any reasonable manner, but not in any way that suggests the licensor endorses you or your use.

**No additional restrictions** — You may not apply legal terms or **technological measures** that legally restrict others from doing anything the license permits.

<http://hdl.handle.net/2440/90005>

POSTER PRESENTATION

Open Access

# Optimal planimetry location for MRI-derived mitral inflow velocity assessment of diastolic function

Shah M Azarisman\*, Andrew Li, James D Richardson, Dennis T Wong, Seng Keong Chua, Michael Cursorso, Vince Schirripa, Kerry Williams, Ben Koschade, Mitra Shirazi, Julie Bradley, Karen Teo, Matthew Worthley, Stephen G Worthley

From 16th Annual SCMR Scientific Sessions  
San Francisco, CA, USA. 31 January - 3 February 2013

## Background

Diastolic function is almost exclusively assessed using transthoracic echocardiography (TTE), however velocity-encoded phase-contrast imaging permits diastolic evaluation with cardiac magnetic resonance (CMR). However, previous studies have utilized heterogeneous planimetric contour locations to measure mitral valve (MV) inflow velocities and the optimal contour is uncertain. We therefore evaluated CMR MV inflow velocities measured at various regions against TTE to identify the optimal method.

## Methods

Patients with revascularized acute MI and preserved LV systolic function were assessed by 1.5T CMR and TTE at 24-72 hours after presentation. Early and late peak diastolic mitral in-flow velocities were determined at 3 planimetric contour locations: (i) annulus, (ii) MV leaflet orifice, (iii) mid-MV inflow region and the E/A ratio and deceleration times (DT) compared to TTE measurements.

## Results

Twenty-one patients were analyzed and mean LVEF was  $56.9 \pm 6.9\%$  (TTE) and  $58.6 \pm 8.3\%$  (CMR). Peak E and A velocities was underestimated by CMR, however E/A showed high correlation with TTE with  $r^2$  values of 0.70, 0.69 and 0.74 for the annulus, leaflet and inflow region contours respectively (all  $p < 0.001$ ). Correlation of DT was very strong with  $r^2$  values of 0.89, 0.84 for leaflet and inflow region contours respectively (all  $p < 0.001$ ).

However the annulus contour poorly correlated with TTE ( $r^2=0.17$ ,  $p=0.12$ ). Bland-Altman analysis showed the MV leaflet contour to have the best agreement between modalities.

## Conclusions

CMR evaluation of diastolic function is readily achievable and demonstrates high correlation with TTE. However measurements vary according to planimetric contour location, with the greatest correlation noted with MV leaflet whereas the annulus contour demonstrated the weakest relationship.

## Funding

Cardiovascular Investigation Unit, Royal Adelaide Hospital, North Terrace, Adelaide, SA 5000.

Published: 30 January 2013

doi:10.1186/1532-429X-15-S1-P90

**Cite this article as:** Azarisman et al.: Optimal planimetry location for MRI-derived mitral inflow velocity assessment of diastolic function. *Journal of Cardiovascular Magnetic Resonance* 2013 **15**(Suppl 1):P90.

Cardiovascular Investigation Unit, Royal Adelaide Hospital, Adelaide, SA, Australia