



A novel filtration system for point of care washing of cellular therapy products

Journal:	<i>Journal of Tissue Engineering and Regenerative Medicine</i>
Manuscript ID	TERM-16-0050.R1
Wiley - Manuscript type:	Research Article
Date Submitted by the Author:	n/a
Complete List of Authors:	Tostoes, Rui; University College London, Biochemical Engineering Dodgson, John; Dodgson Associates LTD Weil, Benjamin; University College London, Biochemical Engineering Gerontas, Spyridon; University College London, Biochemical Engineering Mason, Chris; University College London, Department of Biochemical Engineering Veraitch, Farlan; University College London, Biochemical Engineering
Keywords:	filtration, cell therapy, cell processing, stem cell processing, cell washing, point of care, cell manufacturing, cell delivery

SCHOLARONE™
Manuscripts

1
2
3 **1 A novel filtration system for point of care washing of cellular**
4
5
6 **2 therapy products**
7

8
9 Rui Tostoes¹, John R. Dodgson², Ben Weil¹, Spyridon Gerontas¹, Chris Mason¹, Farlan
10 Veraitch¹
11

12
13
14 1 – The Advanced Centre for Biochemical Engineering, Department of Biochemical
15 Engineering, University College London, Bernard Katz Building, Gower Street, London
16 WC1H 0AH, UK
17
18

19
20 2 – Dodgson Associates Ltd., London, UK
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

1
2
3 23**Abstract**4
5 24

6
7 25 The cell therapy industry would greatly benefit from a simple point of care solution to remove
8
9 26 Dimethyl Sulfoxide (DMSO) from small volume thawed cell suspensions prior to injection. We have
10
11 27 designed and validated a novel dead-end filtration device, which takes advantage of the higher density
12
13 28 of thawed cell suspensions to remove the DMSO and protein impurities from the cell suspension
14
15 29 without fouling the filter membrane. The filter was designed to avoid fluid circuits and minimize the
16
17 30 surface area that is contacted by the cell suspension, thus reducing cell losses by design.

18
19 31 The filtration process was established through optimization of the fluid flow configuration, backflush
20
21 32 cycles and filter geometry. Overall, this novel filtration device allows for a 1 mL of thawed
22
23 33 cryopreserved cell suspensions, containing 10^7 cells of a foetal lung fibroblast cell line (MRC-5), to
24
25 34 be washed in less than 30 minutes. More than 95% of the DMSO and up to 94% of the Albumin-
26
27 35 Fluorescein-Isothiocyanate content can be removed while the viable cell recovery is higher than 80%.
28
29 36 We have also demonstrated that this system can be used for bone marrow-derived human
30
31 37 mesenchymal stem cells with more than 73% cell recovery and 85% DMSO reduction. This is the first
32
33 38 time that a dead end (normal) filtration process has been used to successfully wash high density
34
35 39 human cell suspensions. In practice, this novel solid-liquid separation technology fills the need for
36
37 40 small volume washing in closed processing systems for cellular therapies.

38
39 4140
41
42 42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

1. Introduction

The cryopreservation of cellular therapy products is usually performed by controlled-rate freezing (1-5°C/min) in the presence of DMSO, which is the preferred cryopreservation agent (CPA) in virtually every cryopreservation solution (Rowley 2009). This is due to its efficient diffusion across the cell membrane coupled with the ability to prevent intracellular ice formation and osmotic stress during slow rate freezing (Mazur 1984). In the treatment of blood disorders the cell product may be infused to patients without any reduction in DMSO concentration even though the presence of this CPA has been shown to cause nausea, vomiting, cardiac arrhythmia (Cox et al. 2012) and, less frequently, more severe adverse reactions such as neurotoxicity (Abdelkefi et al. 2009), cardiac arrest or epileptic seizure (Cox et al. 2012). Thus, safety concerns are raised when the cellular product consists of a small volume (<10 mL) cell suspension which will be injected into a solid tissue such as the brain or heart (FDA 1997). Such an injection is likely to expose the patient's cells to a much higher DMSO concentration than a blood stream infused product. Although there is no safety data for concentrated DMSO injections in humans, there is published work in animals (Galvao et al. 2014) and brain tissue culture (Tamagnini et al. 2014). These data strongly suggest that a reduction in DMSO concentration is required for cell therapies which are administered via injection into solid tissues and should be performed even in infused cell therapies.

To reduce the DMSO concentration in thawed cell suspensions the most common process is open lab centrifugation. This process is only acceptable in a laminar flow chamber placed in a grade B cleanroom background, which represents an unacceptable financial burden for the current Point of Care (POC) sites such as hospitals or clinics that do not have a cell-processing facility. Another solution is the use of closed centrifugal systems such as the COBE 2991 (Terumo BCT, Colorado, USA) or Sepax S-100 devices (Biosafe S.A., Eysins, Switzerland). These can be used to separate mononuclear cells from red blood cells or as cell washing devices, reaching 97% removal of the DMSO content (Rodríguez et al. 2004) and a total nucleated cell recovery of 75% in cryopreserved peripheral blood progenitor cell products for autologous infusion (Sánchez-Salinas et al. 2012). While these centrifuge-based systems are efficient, they are expensive and not amenable to small scale cell

1
2
3 70 processing. Hanna and colleagues have described a microfluidic device which removes up to 70% of
4
5 71 the initial DMSO concentration from fresh cell suspensions spiked with the cryoprotectant in a three-
6
7 72 stream diffusion system (Hanna et al. 2012). The post-wash viable cell recovery reported by these
8
9 73 authors was above 95%. Recently, a dilution-filtration system has been reported by Zhou *et al.* This
10
11 74 system relies on a tangential flow filtration system (TFF) which enables the removal of 93% of the
12
13 75 CPA (glycerol in this instance) from thawed human red blood cells with a viable cell recovery of 91%
14
15 76 (Zhou et al. 2011). The main disadvantage of TFF based methods is that they are mostly used for
16
17 77 volumes starting at hundreds of mL or litres, and any scale down design of a TFF will include an
18
19 78 external circuit where cells are likely to be lost either by liquid adsorption to the material or shear
20
21 79 (Jaouen et al. 1999; Kuo et al. 2010). Overall, there is an unmet need for a small scale solid-liquid
22
23 80 separation technologies which are cost-effective, simple to operate and enable a high cell recovery.

24
25 81 In this work, we have designed and validated a novel dead end filtration device for the removal of
26
27 82 DMSO from 1 mL thawed cell suspensions. A typical dead-end filtration operation produces an
28
29 83 exponential pressure increase in the retentate chamber due to the progressive membrane clogging by
30
31 84 the retained cells (Rushton et al. 2000; Xu and Chellam 2005). However, it is also a simple system
32
33 85 with no moving parts and thus amenable to become a cost-efficient closed point of care solution for
34
35 86 cell washing. The location of the filter membrane on top of the cell suspension and the non-turbulent
36
37 87 flow allowed the thawed cell suspension to form a bed at the bottom of the filter device and limit the
38
39 88 contact of the cells with the filter membrane, thus enabling the dead end filtration to progress for long
40
41 89 enough to remove more than 95% of the DMSO. This device was also designed to have a minimal
42
43 90 contact area with cells, to avoid cell losses due to liquid adsorption into long fluidic circuits. With this
44
45 91 minimal surface area approach the viable cell recoveries were above 80%, after the filtration of
46
47 92 thawed 1mL cell suspensions.

49 93 **2. Materials and Methods**

50 94 **2.1. Cell culture and cryopreservation**

51
52 95 The human lung fibroblast cell line MRC-5 (ATCC Catalog No. CCL-171) was used as a model cell
53
54 96 line throughout this work. These cells were cultured for a maximum of 40 population doublings (PD)
55
56 97 as these cells can reach 42 PD without senescence, according to the supplier. The cells were cultured
57
58
59
60

1
2
3 98 in Eagle's Minimum Essential Medium (EMEM, M5650 Sigma-Aldrich UK) supplemented with 10%
4
5 99 Foetal Bovine Serum (FBS, Seralab UK), 1% Glutamine and 1% MEM Non-Essential Amino Acids
6
7 100 (NEAA, GIBCO, UK) using Nunc T-flasks with 175 cm² surface area (Thermo Scientific, UK). Bone
8
9 101 Marrow-derived human Mesenchymal Stromal Cells (hMSC) from 2 different donors were purchased
10
11 102 from Lonza (PoieticsTM human mesenchymal stem cells) and cultured in complete growth medium:
12
13 103 DMEM low glucose (21885-025, Life Technologies, UK) supplemented with 10% FBS, 1% NEAA
14
15 104 and 1 ng/mL recombinant human basic Fibroblast Growth Factor (bFGF, R&D Systems, MN, US).
16
17 105 The hMSCs were cultured for expansion up to passage 7, to minimize the senescence of the
18
19 106 population and used as a cell therapy-relevant model to validate the filtration system developed in this
20
21 107 work. The hMSC were differentiated using GIBCO® StemPro reagents (GIBCO,UK): For
22
23 108 adipogenic, chondrogenic and osteogenic differentiation, hMSCs were seeded into 12-well plates at 1
24
25 109 $\times 10^4$ cells/cm², 1.6×10^7 cells/mL 10 μ L droplets and 5×10^3 cells/cm², respectively. After 24h in
26
27 110 culture in complete growth medium, the media was discarded and 1 ml of Adipogenesis,
28
29 111 Chondrogenesis or Osteogenesis Differentiation Medium (StemPro Differentiaton Kit, GIBCO,
30
31 112 Invitrogen) were added to each well. Medium was fully exchanged every 3 days and at day 21 cells
32
33 113 were washed in 1X PBS and fixed at room temperature, with 4% paraformaldehyde (PFA), for 10
34
35 114 minutes for adipogenesis cultures and 30 minutes for chondrogenesis and osteogenesis cultures. For
36
37 115 adipogenesis cultures, 60% isopropanol was added to the cells and let sit for 5 minutes. Cells were
38
39 116 then stained with a filtered solution of 0.3% w/v Oil Red O/isopropanol in distilled water 3:2 for 5
40
41 117 minutes, rinsed under tap and counterstained with hematoxilin for 1 minute. For chondrogenesis
42
43 118 cultures, 1% Alcian Blue solution prepared in 0.1M HCl was used to stain the chondrocytes. After 30
44
45 119 minutes, cells were rinsed twice in 0.1M HCL and then under running tap water. For osteogenesis, .
46
47
48 120 2% Alizarin Red S solution (pH 4.2) was added for 5 minutes and then the wells were carefully rinsed
49
50 121 with distilled water to avoid the micromasses disruption.
51
52 122 For cryopreservation, cells were exposed to Trypsin (Sigma-Aldrich, UK) for 5 min at 37 °C and the
53
54 123 reaction was stopped by adding 2 volumes of EMEM (supplemented with 10% FBS) per volume of
55
56 124 trypsin. The resulting cell suspension was centrifuged at 300g for 5 min at room temperature, washed
57
58 125 with PBS and centrifuged again under the same conditions. The washed cell pellet was resuspended in
59
60

1
2
3 126 cold (2-8 °C) CryoStor™ containing 5% v/v of DMSO (CS5, BioLife Solutions, WA, USA) to yield 1
4
5 127 mL single cell suspension containing 10-20 million cells/mL. This cell suspension was dispensed into
6
7 128 3 mL Crystal™ vials (Aseptic Technologies, Belgium), incubated for 10 minutes on ice and then
8
9 129 transferred to a -80 °C freezer where it was cooled at an approximate rate of -1 °C/min, using a
10
11 130 CoolCell container (BioCision, CA, US) and kept up to a maximum of 1 month before use.

13 131 2.2. Device design and operation

14
15 132 The filter devices were designed using AutoCad 2011 (Autodesk, CA 94903, USA) and manufactured
16
17 133 by machining in UCL's Biochemical Engineering workshop, using either Poly(methyl methacrylate)
18
19 134 (PMMA) or Stainless Steel 316 (SS) as a the base material, and by injection molding using the
20
21 135 Protomold service from Protolabs, UK. The base material used for injection molding was Cyclic
22
23 136 Olefin Co-polymer (COC).

24
25 137 All the devices were designed to allow the 1 mL thawed cell suspension to form an undisturbed bed
26
27 138 while the washing buffer flowed through the cell suspension. This configuration minimized and
28
29 139 delayed the progressive filter clogging typical in a dead end filtration process. The design shown in
30
31 140 Figure 1A enabled the formation of this cell bed by localizing the filter membrane above the cell
32
33 141 suspension and ensuring that the lower chamber had a higher volume than that same loaded cell
34
35 142 suspension, such that during the wash operation there was a decreasing cell concentration from the top
36
37 143 surface of the cell bed to the filter membrane (Figure 1B, top). The wash step was performed by
38
39 144 flowing wash buffer (PBS) from the bottom and lateral inlets of the lower chamber (Figure 1B, top).
40
41 145 The semi-conical shape of this chamber was designed to enable the post wash gas phase volume
42
43 146 reduction of the cell suspension: this shape stabilizes an injected air bubble which occupied the top
44
45 147 part of the bottom chamber, leaving behind a volume reduced cell suspension (Figure 1B, middle) that
46
47 148 was collected directly into a syringe (Figure 1B, bottom); these process steps are shown in the
48
49 149 supplementary movies S1-3 (A with dye and B with cells) and S4. The fixed process inputs and
50
51 150 parameters are listed in Table I; the operating temperature was kept at $21 \pm 2^\circ\text{C}$.

52 151 2.3. Cell counts

53
54
55 152 The cell numbers and viability were counted using a Vi-CELL Cell Viability Analyser (Beckman
56
57 153 Coulter, UK) before and after the cell suspension processing. The results are shown as percent viable

1
2
3 154 cell recovery (VCR) and correspond to the ratio between total viable cells after processing and total
4
5 155 viable cells before processing.

6 7 156 2.4. DMSO measurement

8
9 157 The concentration of DMSO in the filtrate and retentate was measured by reverse phase high
10
11 158 performance liquid chromatography (HPLC) using an Agilent 1200 series device (Agilent
12
13 159 Technologies UK Limited) under the control of the Agilent Chemstation software. A 4 mm x 125 mm
14
15 160 (internal diameter x length) Nucleosil C18 column with a particle size of 5 μm (Sigma-Aldrich, UK)
16
17 161 was used as the stationary phase, while the mobile phase was based on a previously published
18
19 162 protocol (Thumm et al. 1991). Briefly, an aliquot was taken from the filtrate or retentate sample and
20
21 163 diluted 1:20 in 0.1% (v/v) Trifluoro Acetic Acid (TFA, Sigma-Aldrich UK). Calibration curves of CS
22
23 164 5 were prepared at the appropriate dilution and analyzed in parallel with the samples. The elution
24
25 165 profile consisted of 100% distilled water in the first 5 min followed by a linear gradient of
26
27 166 water/acetonitrile until 100% of acetonitrile was reached at 7 min; the reverse gradient was applied to
28
29 167 reach 100% water at 9 min and maintained until 15 min. The DMSO peak was detected (210 nm
30
31 168 detection wavelength) at 3-5 min of elution time. All results are presented as the ratio $\frac{C}{C_0}$, where C and
32
33
34 169 C_0 are the final and initial concentration of DMSO, respectively.

35 36 170 2.5. FITC-Albumin experiments and quantification

37
38 171 To assess the removal of larger molecular weight molecules, 20 μL of conjugated Albumin-
39
40 172 Fluorescein Isothiocyanate (FITC) (A9971, Sigma UK) at a concentration of 5 mg/mL in CS 5 were
41
42 173 added to the 1 mL cell suspension immediately after thaw and before the filtration process. The
43
44 174 Albumin-FITC concentrations were measured using a FluoStar Optima microplate reader (excitation
45
46 175 450 nm, emission 544 nm). Calibration curves of Albumin FITC in CS 5 were prepared at the
47
48 176 appropriate dilution and analyzed in parallel with the samples. All results are presented as the ratio $\frac{C}{C_0}$,
49
50
51 177 where C and C_0 are the final and initial concentrations of FITC-Albumin, respectively.

52 53 178 2.6. Pressure Measurements

54
55 179 The pressure in the lower chamber of the system was measured against the atmospheric pressure using
56
57 180 either a manual or a digital manometer with data-logging capabilities (Sper Scientific, AZ US).

181 2.7. Filtration model

182 A modified form of the diafiltration equation was used to model and predict solute removal from the
183 cell suspensions:

$$\frac{C}{C_0} = e^{-\frac{F \times t}{V_c} \times k} = e^{-N \times k} \quad \text{Equation 1}$$

184 where F is the flow rate in mL/min, t is time in min, V_c is the lower (cell-containing) chamber
185 volume, N is the number of diavolumes, defined as the quotient between the total volume used for
186 filtration and the volume of the filtration chamber, and k is a wash efficiency parameter which is
187 dependent on the geometry of the device.

188 2.8. Computational Fluid Dynamics

189 The commercial software package COMSOL Multiphysics 5.0 (Hatfield, Comsol, UK) was used to
190 simulate the velocity, pressure and DMSO profiles in the filter device. The laminar flow application
191 mode within the COMSOL Multiphysics software was chosen to simulate the velocity and pressure,
192 whereas the mass transfer for diluted species application mode was chosen to simulate the DMSO
193 profile. To lower the computer memory requirements of the overall simulation, the flow field is
194 simulated first as the viscosity and densities of the culture medium containing DMSO are considered
195 constant (relatively low species concentration). The solved flow field is then saved and coupled with a
196 mass transfer model, which describes DMSO convection/diffusion in the filter device.

197 In all simulations, the boundaries of the filtering device were meshed with 52155 triangular elements
198 (extra fine mesh; default settings) and its domain with 376906 quadrilateral elements (normal mesh
199 option; default settings) using the free mesh option of COMSOL Multiphysics's mesh generator. The
200 number of degrees of freedom was 377702. The convergence was tested by a two-fold increase in the
201 number of elements. The velocity, pressure and DMSO concentration simulated by using the new
202 mesh were agreed at the three significant figures level with the velocity, pressure and DMSO
203 concentration simulated by using the default mesh. The simulation results are presented as contour
204 plots, blue represents areas of low DMSO concentration and red represents areas of high values for
205 DMSO concentration.

206 2.9. Flow Cytometry

1
2
3 207 The post filtration multipotency of the hMSCs was assessed using an hMSC phenotyping kit
4
5 208 (Miltenyi Biotec, Surrey, UK) flow cytometry assay. Briefly, single cell suspensions, following wash
6
7 209 and volume reduction using the novel filtration device, were equally divided into 7 samples.
8
9 210 Antibodies were diluted 1:11 in 110ul buffer per 1×10^6 cells and incubated with the samples on ice
10
11 211 for 30 minutes. The 7 samples for each MSC test consisted of 4 single fluorophore controls (CD105-
12
13 212 PE, CD90-FITC, CD73-APC, CD14-/CD20-/CD34-/CD45-PerCP) for compensation, a cell only
14
15 213 sample (incubated with 110ul FACS buffer, also for 30 minutes on ice), the Isotype Control Cocktail
16
17 214 and the MSC Phenotyping Cocktail provided in the MSC phenotyping kit (Miltenyi). After antibody
18
19 215 labelling, cells were washed and re-suspended in 0.5ml FACS buffer in flow analysis tubes (BD
20
21 216 Bioscience, Oxford, UK). Flow cytometry was performed using a BD LSR II flow cytometer. For
22
23 217 analysis, at least 50,000 events were initially gated based on forward and side scatter and forward
24
25 218 scatter height against area was used for doublet removal. The isotype control cocktail was run with the
26
27 219 same gating strategy and MSCs defined from gating the top 1% of isotype expression. To confirm an
28
29 220 MSC phenotype, greater than 95% expression of CD105, CD90, CD73, and less than 2% expression
30
31 221 of CD14, CD20, CD34 and CD45 must be observed (Dominici et al. 2006). N=3 samples were run
32
33 222 and the standard deviation calculated.

35 223 2.10. Statistics

36
37 224 The statistical analysis was performed using the Excel data analysis add-on and the GraphPad
38
39 225 software. The p-values presented were calculated for the two tails of the normal distribution and when
40
41 226 more than 2 groups were compared Tukey's multicomparison test was used. The notation "n"
42
43 227 (lowercase, not to be mistaken by uppercase "N", the diavolumes) was used to denote the number of
44
45 228 independent replicates.

47 229 **3. Results**

48
49 230 To test this new filtration device it was necessary to choose a filter medium and a pore size which
50
51 231 minimized the rate of pressure increase during the wash operation; initial tests demonstrated that a
52
53 232 hydrophilic Polyvinylidene fluoride (hPVDF) with a nominal pore size of $0.65 \mu\text{m}$ was a feasible
54
55 233 starting point (data not shown).

1
2
3 234 To maximize the number of cells recovered, it was hypothesized that a backflush step between the
4
5 235 wash and concentration steps would dislodge cells from the filter membrane and increase the post
6
7 236 process VCR. The wash step was performed using a flow rate of 2 mL/min via the bottom inlet and 5
8
9 237 mL/min through the lateral inlet (2/5 flow). While no significant difference in VCR was observed
10
11 238 when this backflush step was present (compared to its absence), there were significant effects on the
12
13 239 amount of solute removed. Figure 2A depicts the DMSO concentration in the filtrate over time, and it
14
15 240 is visible that after the backflush step there is an increase in the DMSO concentration in the filtrate. It
16
17 241 was hypothesized that the overall DMSO removal of the retentate could be enhanced if several
18
19 242 backflush steps were applied during the wash phase. Figure 2B depicts the experimental design used
20
21 243 to test this hypothesis, where 3 backflush steps were performed during a 2/5 wash for 10 min (3x
22
23 244 backflush) and compared with a 10 min 2/5 wash with one backflush step after the wash operation (1x
24
25 245 backflush). Figure 2C shows that the 3x backflush wash yielded a lower DMSO concentration than
26
27 246 the 1x backflush wash, with 0.136 ± 0.003 and 0.20 ± 0.01 of the initial DMSO concentration,
28
29 247 respectively ($n=3$, $p<0.001$). This increase in DMSO removal was also associated with a higher lower
30
31 248 chamber pressure in the 3x backflush wash when compared with the 1x backflush wash, after 10
32
33 249 minutes, as shown in Figure 2D.

34
35 250 The individual contributions of the bottom and lateral flows for the DMSO removal from the cell
36
37 251 suspension were investigated. Figure 3 shows that, after a 10 minutes wash period, the lateral flow
38
39 252 reduces the initial amount of DMSO to 0.66 ± 0.09 ($n=5$) of its initial concentration. There was no
40
41 253 significant difference between the concentration of DMSO after the bottom flow wash and the
42
43 254 lateral+bottom (2/5) flow wash (0.25 ± 0.01 , $n=3$ and 0.17 ± 0.06 , $n=3$), for the 10 min wash period.
44
45 255 However, the presence of a bottom flow results in a significantly higher DMSO reduction when
46
47 256 compared to using lateral flow only ($p<0.01$).

48
49 257 The combinatorial effect of the lateral and bottom flows (2/5 flow) on DMSO reduction was further
50
51 258 investigated as a function of the diavolumes N (Bottom Flow x Time), as shown in Figure 4A. The
52
53 259 curve corresponding to the 2/5 flow is in agreement with the diafiltration equation, with the wash
54
55 260 efficiency value k being constant ($R^2=0.998$); without the lateral flow, the bottom flow curve deviates
56
57 261 from Equation 1, yielding lower DMSO reduction values. These differences correlate with different
58
59
60

1
2
3 262 pressure profiles (Figure 4B): the lateral flow of PBS leads to an increased pressure ending in 80
4
5 263 mbar, whereas pressure in the bottom flow configuration is kept at 20 mbar or less.

6
7 264 To be useful as part of a point of care device, this filter will be manufactured as an injection molded
8
9 265 part, like most medical devices or bioprocessing consumables. The panels C and D in Figure 4 depict
10
11 266 a PMMA filter device which houses an injection molded (IM) COC lower chamber whose shape is
12
13 267 identical to the one in Figure 1A. This IM device was used to process thawed MRC-5 cell suspensions
14
15 268 at 2 mL/min bottom flow, without lateral flow, for 24 min and the volumes, cell concentration and
16
17 269 DMSO removal are shown in Table 2. The DMSO reduction for this data was $91 \pm 2\%$ for $N=32$, as
18
19 270 shown in Figure 4A. The cell recovery and volume reduction was $84 \pm 1\%$ and 2.7 ± 0.5 fold,
20
21 271 respectively. These results demonstrate that this filtration system can be successfully used for small
22
23 272 volume cell suspension wash and volume reduction, with a cell recovery of more than 80%.
24
25 273 Moreover, the use of a thermoplastic material indicates that the performance of the PMMA device is
26
27 274 close to the performance of the consumable.

28
29 275 The filtration device was redesigned to increase the DMSO reduction while avoiding pressure build-
30
31 276 up in the lower chamber, i.e. increasing the wash without using the lateral flow. Figure 5A shows the
32
33 277 3 geometries of the filtration device, where M1 (mark 1) is the original design, with a lower chamber
34
35 278 height (bottom inlet to filter membrane) of 0.8 cm and a filtration area of 3.5 cm^2 . To minimize the
36
37 279 diffusional distances and increase the fluid velocity, the M2 geometry was designed to increase the
38
39 280 height of the chamber by 50% (compared to M1) while maintaining a constant volume, thus
40
41 281 decreasing the cross sectional area of the bottom flow chamber. This was achieved by changing the
42
43 282 chamber from a round bottom to a V-bottom shape. Computational fluid dynamics (CFD) was used to
44
45 283 model DMSO removal in M2 under the 2/5 flow configuration; the surface plot in figure 5B depicts
46
47 284 the top view of the M2 device and it indicates that there is accumulation of DMSO in the outer area of
48
49 285 the membrane. Again, these areas are the furthest away from the bottom flow jet and they could be
50
51 286 designed out (or “trimmed” out) of the M2 geometry thus yielding the M3 geometry, with a 2.5 cm^2
52
53 287 area (30% decrease vs M2) and a lower chamber height of 1.5 cm (25% increase vs M2). These
54
55 288 geometry changes had significant effects on the DMSO removal rate, as demonstrated in Figure 5C.
56
57 289 When using only the bottom inlet to flow wash buffer through the cell suspension (solid lines) it is
58
59
60

290 visible that the amount of DMSO removed per diavolumes of buffer increases from M1 to M3. Unlike
 291 the M1 and M2 bottom flow curves, the M3 curve fits a single exponential line as described in
 292 Equation 1 (solid line, $R^2=0.97$); likewise, when the lateral flow is used for the M1 and M2
 293 geometries, their DMSO reduction curves also fit a single exponential decay (solid lines, $R^2=0.97$ for
 294 both). However, the M3 geometry, operating with bottom flow only, sustains the lower chamber
 295 pressure below 15 mbar, whereas the M1 and M2 geometries, operating with the 2/5 flow, can reach
 296 pressures of 100 and 400 mbar, respectively (Figure 5D, $n=3$). These data indicate that the M3
 297 geometry enables an efficient DMSO reduction while operating at low pressures.

298 To further validate the hypothesis that the diffusional distance is the critical factor behind these
 299 increases in efficiency from M1 to M3, the differences in these geometries were factored into
 300 Equation 1. By keeping the volume of the chamber constant and increasing the height of the chamber,
 301 the whole shape becomes “thinner” along its length thereby reducing the diffusional distance and
 302 increasing the fluid velocity. As such, the increase in height is proportional to the increase in the
 303 amount of DMSO removed. To incorporate this into equation 1 it should be noted that the effect of the
 304 height implies that for the same level of DMSO removal, less diavolumes (N) are needed as the height
 305 is increased. In this way, N should be replaced by the product $N \times \text{height}$. However, since this is an
 306 exponential term it must be dimensionless. Since N is dimensionless so should the height be; to
 307 achieve this it was noted that $\text{height} = \sin(\alpha) \times \text{longer slope}$ (as per figure 5E) and the $\sin(\alpha)$ was used
 308 instead of the height. Thus, Equation 1 can be modified to:

$$\frac{C}{C_0} = e^{-\frac{Ft}{V_c} k \sin \alpha} = e^{-Nk \sin \alpha} \quad \text{Equation 2}$$

309 The semi-logarithmic plot of $\frac{C}{C_0}$ vs $N \times \sin \alpha$ is depicted in Figure 5F. This plot shows that the data
 310 points derived from Equation 1 in Figure 5C (M1 and M2 with lateral flow and M3 with bottom flow)
 311 can be modelled by a single curve using Equation 2 ($R^2=0.98$).

312 The M3 geometry was further characterized by its ability to reduce the amount of a model protein,
 313 albumin-FITC; After a 32 diavolumes wash the initial concentration of this protein was reduced by $94 \pm 1\%$
 314 ($\pm 1\%$) (Figure 5G). Furthermore, the viable cell recovery using M3 was $94 \pm 5\%$, thus confirming that
 315 this geometry also enables a high cell recovery, while the volume reduction was 1.55 ± 0.08 fold

1
2
3 316 (Table 2). To assess the functionality and long-term viability of the cells after being processed, the
4
5 317 M3 geometry was manufactured in stainless steel 316. Figure 5H demonstrates that, after a 16
6
7 318 diavolumes wash, the MRC-5 cell line still grows at the same specific rate as fresh cells or cells which
8
9 319 have also been thawed but processed using a bench top centrifuge (see methods). Importantly, cell
10
11 320 counts after a 24 hour culture showed that after thawing, processing and plating $62 \pm 15\%$ of the cells
12
13 321 (post thaw counts) survived; when no cell processing was performed and cells were directly plated in
14
15 322 adherent flasks, the cell recovery was $58 \pm 5\%$ (n=4).

16
17 323 To validate this geometry for cell therapy applications, hMSCs were processed using the M3
18
19 324 geometry filtration device. The post filtration characterization of these cells was performed according
20
21 325 to the ISCT criteria (Dominici et al. 2006) and the 3 cell runs (generated from 2 donors) showed that
22
23 326 hMSCs retained their characteristic surface markers (Figure 6 A and B), while retaining their ability
24
25 327 to adhere to plastic, proliferate (Figure 6E and F) and the capability for chondrogenic, osteogenic and
26
27 328 adipogenic differentiation (Figure 6G, H and I). The viable cell recovery after these 3 runs ranged
28
29 329 from 73 to 99%, while the DMSO removal ranged from 86-88% (Figure 6 C and D).

30 31 32 330 **4. Discussion**

33
34 331 In this work, we have designed and characterized a filtration system to process thawed cell
35
36 332 suspensions which removes more than 90% of the DMSO and recovers more than 80% of the viable
37
38 333 cells. Most importantly, this filtration system does not require an outside loop like a TFF system, thus
39
40 334 minimizing the shear stress that cells are exposed to and the area where liquid can be lost by
41
42 335 adsorption.

43
44 336 The central element of the filter housing design (Figure 1A) is the laminar flow that is established at
45
46 337 these dimensions and flow velocities. While it is visible that the cell suspension has settled at the end
47
48 338 of the filtration process, it is unlikely that cell settling can be faster than the flow rate of wash buffer
49
50 339 from the bottom inlet. The average linear velocity of wash buffer crossing the membrane from the
51
52 340 bottom flow is 0.57 cm/min while the average cell settling velocity is at least one order of magnitude
53
54 341 below this velocity (Chalmers et al. 1999). Another possibility is that the settling velocity of the cells
55
56 342 is increased in the Cryostor media; while this is true in absolute terms, because the CS5 is hypertonic,

1
2
3 343 it is the density difference between the cells in CS5 and the CS5 medium that is relevant for the
4
5 344 settling velocity. Thus, it is expected that the density differences between fresh cells in PBS and
6
7 345 thawed cells in CS5 are the same; this was confirmed by an isopycnic centrifugation using a Percoll
8
9 346 gradient (Supplementary Figure 1) which confirmed that the density of the MRC-5 cells in CS5 is
10
11 347 between 1.04 and 1.06 g/mL, whereas the literature reports a similar density for these cells in PBS
12
13 348 (1.05 g/mL) (Birnie and Rickwood 1980).

14
15 349 Using the lateral flow configuration, the DMSO concentration was reduced to 0.66 of its initial value
16
17 350 (Figure 4); interestingly, this value is the same as the ratio of the cell suspension volume to the lower
18
19 351 chamber volume (1mL/1.5mL), which strongly suggests that the lateral flow alone does not remove
20
21 352 any significant amount of DMSO after 10 minutes. Nevertheless, the addition of lateral flow to the
22
23 353 bottom flow (2/5 flow) seems to provide a more efficient mixing for both M1 and M2 geometries, and
24
25 354 thus a better reduction in DMSO concentration, when compared with using the bottom flow only
26
27 355 (Figure 5C). These results follow the same trend as the use of increased backflush cycles to enhance
28
29 356 DMSO reduction (Figure 2), also at the expense of higher pressure differences. Taken together, these
30
31 357 two data sets strongly suggest that both the lateral flow and the backflush cycles increase the DMSO
32
33 358 removal, by dispersing the cell bed that is established if bottom flow only is used. The design changes
34
35 359 that lead to the M3 geometry had the goal of avoiding the dispersion of the cell bed that results in
36
37 360 increased lower chamber pressures This was accomplished by prioritizing the minimization of the
38
39 361 average diffusional length and the maximization of the fluid velocity. The M3 geometry, while having
40
41 362 a lower volume reduction capability when compared to M1, does not require lateral flow to have a
42
43 363 single exponential wash profile, as described by equation 1, while washing more than 95% of the
44
45 364 DMSO using less diavolumes; using only one inlet to flow the wash buffer makes the operation
46
47 365 simpler and will decrease the manufacturing cost of the POC device. The empirical model described
48
49 366 by Equation 2 validates the aforementioned design hypothesis, as depicted in Figure 5F; when the
50
51 367 DMSO removal curve fits equation 1, the different geometries can be normalized to the distance
52
53 368 between the filter membrane and the bottom flow inlet. Both the M1 and M2 geometries DMSO
54
55 369 removal data points shift to the right of the curve in Figure 5F when only bottom flow is used, a result
56
57 370 which can be explained by cell settling in a low shear environment. The M3 filtration device was also

1
2
3 371 shown to remove up to 94% of a model protein, Albumin-FITC (Figure 5 G); this ability to removes
4
5 372 higher molecular weight contaminants from cell suspensions can be of high value when the whole cell
6
7 373 therapy process is considered.

8
9 374 The processing of hMSCs (Figure 6) demonstrated the capability of this novel filtration device for
10
11 375 small volume cell wash, using cells relevant for cell therapy. Despite the DMSO removal efficiency
12
13 376 being lower than for the MRC-5 cells wash, 88% of the DMSO was removed from hMSC cell
14
15 377 suspensions. This lower washing efficiency can be attributed to different biophysical characteristics,
16
17 378 namely cell size, and it is a clear area of research in the development of this technology. Another
18
19 379 point for future research is how to have sterility as a release criterion for the cellular therapies, since
20
21 380 this takes about 24h to be tested and the therapies need to be administered within minutes.

22
23 381 In a broader sense, a simple solid-liquid separation device for small volumes is a key enabling
24
25 382 technology to make the distributed manufacturing of cellular therapies possible.

26 27 383 **Acknowledgments**

28
29 384 The authors thank John Vincent, Sarah Callens (eXmoor Pharma, UK), Jesus Zurdo (Lonza UK) and
30
31 385 Behzad Mahdavi (Lonza Walkersville, US) for the valuable technical discussions, John Langlon and
32
33 386 Erich Herrmann (UCL Department of Biochemical engineering workshop) for the useful discussions
34
35 387 and device machining.

36
37 388 This work was supported by Innovate UK, under the collaborative R&D project CPREP and by the
38
39 389 Department of Biochemical Engineering, University College London, UK. The CPREP project was a
40
41 390 collaboration between Lonza UK, UCL, Amercare Ltd and eXmoor Ltd and the resulting technology
42
43 391 has been licensed to Closed Cell Systems Ltd (CCS) (Supplementary Figure 2 shows the processor
44
45 392 and consumable which use this filtration device for the solid-liquid separation).

46 47 393 **Competing financial interests**

48
49 394 RT and FV are authors and have financial interests in 2 patents related to this work, which were
50
51 395 licensed to CCS. None of the manuscript's authors have any other financial interest in the company
52
53 396 CCS. The authors have no other competing financial interests.

54 55 397 **5. Bibliography**

- 1
2
3 398 Abdelkefi A, Lakhal A, Moojat N, Hamed LB, Fekih J, Ladeb S, Torjman L, Othman TB. 2009.
4
5 399 Severe neurotoxicity associated with dimethyl sulphoxide following PBSCT. Bone Marrow
6
7 400 Transplant. England. p 323-4.
8
9 401 Birnie G, Rickwood D. 1980. Centrifugal Separations in Molecular and Cell Biology. Journal of Basic
10
11 402 Microbiology 20(2):153.
12
13 403 Chalmers JJ, Haam S, Zhao Y, McCloskey K, Moore L, Zborowski M, Williams PS. 1999.
14
15 404 Quantification of cellular properties from external fields and resulting induced velocity:
16
17 405 cellular hydrodynamic diameter. Biotechnol Bioeng 64(5):509-18.
18
19 406 Cox MA, Kastrup J, Hrubisko M. 2012. Historical perspectives and the future of adverse reactions
20
21 407 associated with haemopoietic stem cells cryopreserved with dimethyl sulfoxide. Cell Tissue
22
23 408 Bank 13(2):203-15.
24
25 409 Dominici M, Le Blanc K, Mueller I, Slaper-Cortenbach I, Marini F, Krause D, Deans R, Keating A,
26
27 410 Prockop D, Horwitz E. 2006. Minimal criteria for defining multipotent mesenchymal stromal
28
29 411 cells. The International Society for Cellular Therapy position statement. Cytotherapy
30
31 412 8(4):315-7.
32
33 413 FDA. 1997. Food and Drug Administration Guidance for Industry - Q3C Impurities: Residual
34
35 414 Solvents.
36
37 415 Galvao J, Davis B, Tilley M, Normando E, Duchon MR, Cordeiro MF. 2014. Unexpected low-dose
38
39 416 toxicity of the universal solvent DMSO. Faseb j 28(3):1317-30.
40
41 417 Hanna J, Hubel A, Lemke E. 2012. Diffusion-based extraction of DMSO from a cell suspension in a
42
43 418 three stream, vertical microchannel. Biotechnol Bioeng 109(9):2316-24.
44
45 419 Jaouen P, Vandanjon L, Quéméneur F. 1999. The shear stress of microalgal cell suspensions
46
47 420 (Tetraselmis suecica) in tangential flow filtration systems: the role of pumps. 68(2):149-154.
48
49 421 Kuo JS, Zhao Y, Schiro PG, Ng L, Lim DS, Shelby JP, Chiu DT. 2010. Deformability considerations
50
51 422 in filtration of biological cells. Lab Chip 10(7):837-42.
52
53 423 Mazur P. 1984. Freezing of living cells: mechanisms and implications. Am J Physiol 247(3 Pt
54
55 424 1):C125-42.
56
57
58
59
60

- 1
2
3 425 Rodríguez L, Azqueta C, Azzalin S, García J, Querol S. 2004. Washing of cord blood grafts after
4
5 426 thawing: high cell recovery using an automated and closed system*. *Vox Sanguinis*
6
7 427 87(3):165-172.
- 8
9 428 Rowley S. 2009. Hematopoietic Stem Cell Cryopreservation: A Review of Current Techniques.
10
11 429 *Journal of Hematotherapy* 1(3):233-250.
- 12
13 430 Rushton A, Ward A, Holdich R. 2000. Solid-liquid filtration and separation technology. New York:
14
15 431 Wiley-VCH: Weinheim. 587 p.
- 16
17 432 Sánchez-Salinas A, Cabañas-Perianes V, Blanquer M, Majado M, Insausti C, Monserrat J, Sánchez-
18
19 433 Ibáñez M, Menchón P, García-Hernández A, Gómez-Espuch J and others. 2012. An
20
21 434 automatic wash method for dimethyl sulfoxide removal in autologous hematopoietic stem cell
22
23 435 transplantation decreases the adverse effects related to infusion. *Transfusion* 52(11):2382-
24
25 436 2386.
- 26
27 437 Tamagnini F, Scullion S, Brown JT, Randall AD. 2014. Low Concentrations of the Solvent Dimethyl
28
29 438 Sulphoxide Alter Intrinsic Excitability Properties of Cortical and Hippocampal Pyramidal
30
31 439 Cells. *PLoS One*.
- 32
33 440 Thumm W, Freitag D, Kettrup A. 1991. Determination and quantification of dimethyl sulphoxide by
34
35 441 HPLC. *Chromatographia* 32(9):461-462.
- 36
37 442 Xu W, Chellam S. 2005. Initial Stages of Bacterial Fouling during Dead-End Microfiltration†.
38
39 443 *Environmental Science & Technology* 39(17):6470-6476.
- 40
41 444 Zhou X, Liu Z, Shu Z, Ding W, Du P, Chung J, Liu C, Heimfeld S, Gao D. 2011. A dilution-filtration
42
43 445 system for removing cryoprotective agents. *J Biomech Eng* 133(2):021007.
44
45
46 446
47
48 447
49
50
51
52
53
54
55
56
57
58
59
60

1
2
3 448**Table Legends**4
5 449 **Table 1** – Fixed inputs and parameters for the operation of the filtration device.
6

7 450

8
9 451 **Table 2** – Viable cell recovery, volume reduction and cell suspension viability after the filtration
10 operation, under 2mL/min bottom flow for 24 minutes, without lateral flow, for the injection molded
11 version of the filtration device.
12
13 453

14 454

15
16
17 455 **Table 3** – Viable cell recovery, volume reduction and cell suspension viability using the M3
18 geometry, under 2 mL/min bottom flow for 24 min, without lateral flow.
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Figure Legends

1
2
3 457
4
5 458 **Figure 1 – Filtration device design and fluid flow.** A)- CAD drawing of the filtration device housing,
6
7 459 divided in a top and bottom part (A, top); when the device is assembled (A, bottom) the filter is
8
9 460 located between the two parts and the cell suspension is introduced into the bottom, where it is
10
11 461 washed, concentrated and loaded into a syringe (B, Supplementary videos). Panel C) depicts the fluid
12
13 462 flow in the device: for visualization purposes, a red dyed PBS solution was used to wash 1 mL of a
14
15 463 (transparent) cell-free CryoStor5 solution (the full video is available in Supplementary Video S2A).
16
17 464 This frame corresponds to $t=1$ s in the wash process.

18
19 465 **Figure 2 – DMSO removal is enhanced by backflush frequency.** The filter device was operated for 10
20
21 466 min in the wash mode with a 2 mL/min bottom flow rate and a 5 mL/min cross flow rate; after this
22
23 467 wash step and before the concentration step a backflush was performed to dislodge cells from the
24
25 468 filter membrane (see methods). The concentration of DMSO in the collected filtrate fractions
26
27 469 increases after backflush (A). The backflush step was repeated 3 times throughout the wash period
28
29 470 (3xbackflush) and compared to a single backflush after the wash step (1xbackflush) (B) for the
30
31 471 amount of DMSO left in the cell suspension (lower chamber, which contains the retentate) (C) and the
32
33 472 pressure difference generated in the lower chamber (D). ***, $p<0.001$.

34
35 473 **Figure 3 – Lateral flow wash alone shows a limited DMSO reduction.** The filter device was operated
36
37 474 for 10 minutes in the wash mode with a 2 mL/min bottom flow rate without cross flow (20 mL, $n=3$),
38
39 475 with a 5 mL/min lateral flow rate without bottom flow (Lateral 5 mL/min, $n=3$) and with a 2 mL/min
40
41 476 bottom flow rate and a 5 mL/min cross flow rate (2/5 flow, $n=6$). ***, $p<0.001$ (Tukey's
42
43 477 multicomparison test).

44
45 478 **Figure 4 – Synergistic effect of the lateral and bottom flows.** The presence of lateral flow during cell
46
47 479 washing yields a DMSO reduction profile described by a single exponential function, as predicted by
48
49 480 equation 1 (A, closed squares, $n\geq 3$ for all data points), whereas in the absence of the lateral flow, the
50
51 481 bottom flow wash does not fit the equation 1 model (A, closed circles, $n\geq 3$ for all data points). The
52
53 482 typical pressure profiles for these 2 wash modes show that the presence of cross flow also leads to
54
55 483 pressure build-up in the lower chamber, unlike the bottom flow only operation mode (B). To mimic
56
57
58
59
60

1
2
3 484 the final thermoplastic product the bottom flow chamber geometry was injection molded (C, purple)
4
5 485 to fit a PMMA housing (C, gray).

6
7 486 **Figure 5 – Redesigning the filtration device for improved DMSO removal.** The original filtration
8
9 487 device was redesigned by changing the geometry of the bottom chamber which holds the cell
10
11 488 suspension, which led to changes in the distance between the bottom inlet and the filter membrane,
12
13 489 and filtration area (A). A CFD analysis on the M2 geometry revealed “dead zones” where the DMSO
14
15 490 was being removed with less efficiency (B). The 3 different geometries were evaluated for their
16
17 491 DMSO reduction capability (C, $n \geq 3$ for all data points), operated with (open symbols) and without
18
19 492 (closed symbols) lateral flow, except for M3, where lateral flow lead to clogging (data not shown); the
20
21 493 final pressure in the lower chamber, after 24 min, also revealed differences between geometries and
22
23 494 wash modes (D, $n=3$ for all data points). The sine of the angle α , formed between the horizontal plane
24
25 495 and the lower chamber slope that connects the bottom and lateral inlets (E), is different between
26
27 496 devices; when the number of diavolumes is multiplied by the $\sin(\alpha)$, specific for each device, the
28
29 497 curves that obey equation 1 in C follow a single exponential decay empirical model, as described by
30
31 498 equation 2 (F). The M3 geometry was assessed for its ability to reduce albumin-FITC by washing a
32
33 499 cell suspension for 24 minutes at 2mL/min bottom flow (G); furthermore, the cellular proliferation
34
35 500 after filtration processing, using the M3 geometry ($n=4$), was compared with post-centrifugation
36
37 501 processing ($n=5$) and fresh cell culture processing (i.e. cell counts during passaging, $n=5$) by
38
39 502 calculating the specific growth rate μ (day^{-1}).

40
41
42 503 **Figure 6 – hMSC post filtration recovery, proliferation and characterization.** hMSCs were processed
43
44 504 using the M3 geometry filtration device. The post filtration characterization after 3 cell runs
45
46 505 (generated from 2 donors) was performed according to the ISCT criteria: cell surface markers (Figure
47
48 506 A, individual flow cytometry histograms, and B, aggregate results from the 3 cell runs), adherence
49
50 507 to plastic (Figure 6E, scale bar=400 μm) and the capability for chondrogenic, osteogenic and
51
52 508 adipogenic differentiation (Figure 6G, H, scale bar=100 μm , and I, scale bar=50 μm , respectively).
53
54 509 The viable cell recovery, DMSO removal and hMSC growth rates after filtration are displayed in
55
56 510 Figure 6 C, D and F, respectively).

1
2
3 511 **Supplementary Figure 1** – Isopycnic gradient centrifugation of MRC-5 cells in CryoStor 5. A cell
4
5 512 suspension of 10^7 MRC-5 cells and a set of density marker beads (Cospherix, CA, USA) were
6
7 513 centrifuged in parallel in a 5-layer CS5/Percoll (GE Healthcare) gradient. This gradient was
8
9 514 established by overlaying CS5/Percoll solutions, in two 50 mL tubes, at a decreasing concentration of
10
11 515 Percoll (50, 40, 30, 20 and 10%); after overlaying the cell or marker bead CS5 suspension, the two
12
13 516 tubes were centrifuged at 4°C for 30 min without acceleration or brake.

14
15 517 **Supplementary Figure 2** – Processor with the docked consumable (left) and the consumable only
16
17 518 (right) for the commercial version of the CPrep device. The frozen cellular therapy is delivered from a
18
19 519 centralized manufacturing facility to the hospital. When a patient is ready the frozen vial is placed
20
21 520 into the CPREP consumable (yellow area), which is already docked to the processor. Here it is
22
23 521 thawed before being pumped into the novel filter device where the DMSO is washed out. After the
24
25 522 washing step the cells are drained into a syringe ready for administration. This entire fluidic circuit is
26
27 523 an integral part of the consumable.

28
29
30 524 **Supplementary Video S1A and S1B** – Loading using the novel filtration device. Video of the
31
32 525 loading process for CS5 with Alizarin Red (Supplementary Video 1A) and 10^7 MRC-5 cells
33
34 526 (Supplementary Video 1B).

35
36 527 **Supplementary Video S2A and S2B** – Washing using the novel filtration device. Video of the
37
38 528 washing process for PBS with Alizarin Red (Supplementary Video 2A, the dye has been dissolved in
39
40 529 PBS to visualize the jet flow from the bottom inlet to the membrane) and 10^7 MRC-5 cells
41
42 530 (Supplementary Video 2B).

43
44 531 **Supplementary Video S3A and S3B** – Backflush using the novel filtration device. Video of the
45
46 532 backflush process for CS5 with Alizarin Red (Supplementary Video 3A) and 10^7 MRC-5 cells
47
48 533 (Supplementary Video 3B).

49
50 534 **Supplementary Video S4** – Volume reduction using the novel filtration device. Video of the volume
51
52 535 reduction process for CS5 containing 10^7 MRC-5 cells.

53
54
55 536

Table 1

Parameter/input	Value
Cell type	MRC-5 (lung fibroblast cell line)
Wash buffer	Phosphate buffered saline (PBS)
Cryopreservation solution	CryoStor-5 (5% DMSO)
Initial cell suspension volume	1 mL
Initial cell suspension concentration	10^7 /mL
Cell chamber volume	1.5 mL
Total process time (w/o priming or manual steps)	≤ 30 min
Cell loading flow rate	2 mL/min
Lateral flow rate	5 mL/min
Backflush flow rate and time	5 mL/min for 10 sec (between wash and concentration)
Air flow rate (volume reduction)	1 mL/min

Fixed inputs and parameters for the operation of the filtration device.
254x190mm (96 x 96 DPI)

Table 2

Run#	V_in (mL)	V_out (mL)	pre-process viable cell concentration (10 ⁶ /mL)	Input viability (%)	post-process viable cell concentration (10 ⁶ /mL)	Output viability (%)	Fold volume reduction	Total viable cell recovery (%)
1	1.086	0.393	15.5	89	36.1	92	2.8	84
2	0.821	0.408	12.1	88	20.1	91	2	83
3	1.052	0.334	10.8	91	28.8	94	3.1	85
4	0.952	0.322	10.9	88	27.5	92	3	85
Average	1.0	0.36	12	89	28	92	2.7	84
stdev	0.1	0.04	2	1	7	1	0.5	1
relstdev (%)	10	11	17	1	25	1	19	1


Viable cell recovery, volume reduction and cell suspension viability after the filtration operation, under 2mL/min bottom flow for 24 minutes, without lateral flow, for the injection molded version of the filtration device.

254x190mm (96 x 96 DPI)

Table 3


M3 Run #	V _{in} (mL)	V _{out} (mL)	pre-process viable cell concentration (10 ⁶ /mL)	Input viability (%)	post-process viable cell concentration (10 ⁶ /mL)	Output viability (%)	Fold volume reduction	Total viable cell recovery (%)
1	1.03	0.68	9.7	70.2	13.6	89.1	1.52	92
2	1.00	0.66	10.2	86.6	14.1	89.1	1.52	91
3	0.96	0.65	10.2	89.0	15.7	89.0	1.49	103
4	1.04	0.62	9.3	83.4	14.1	85.9	1.69	89
5	0.92	0.61	8.8	89.5	12.4	87.8	1.51	93
Average	1.01	0.65	10	84	14	88	1.55	94
Std. dev.	0.04	0.03	1	8	1	1	0.08	5
C.V.(%)	4	5	10	10	7	1	5	5

Viable cell recovery, volume reduction and cell suspension viability using the M3 geometry, under 2 mL/min bottom flow for 24 min, without lateral flow.
254x190mm (96 x 96 DPI)




Filtration device design and fluid flow. A)- CAD drawing of the filtration device housing, divided in a top and bottom part (A, top); when the device is assembled (A, bottom) the filter is located between the two parts and the cell suspension is introduced into the bottom, where it is washed, concentrated and loaded into a syringe (B, Supplementary videos). Panel C) depicts the fluid flow in the device: for visualization purposes, a red dyed PBS solution was used to wash 1 mL of a (transparent) cell-free CryoStor5 solution (the full video is available in Supplementary Video S2A). This frame corresponds to $t=1s$ in the wash process.


115x127mm (300 x 300 DPI)




DMSO removal is enhanced by backflush frequency. The filter device was operated for 10 min in the wash mode with a 2 mL/min bottom flow rate and a 5 mL/min cross flow rate; after this wash step and before the concentration step a backflush was performed to dislodge cells from the filter membrane (see methods). The concentration of DMSO in the collected filtrate fractions increases after backflush (A). The backflush step was repeated 3 times throughout the wash period (3xbackflush) and compared to a single backflush after the wash step (1xbackflush) (B) for the amount of DMSO left in the cell suspension (lower chamber, which contains the retentate) (C) and the pressure difference generated in the lower chamber (D). ***, $p < 0.001$. 160x94mm (300 x 300 DPI)



Lateral flow wash alone shows a limited DMSO reduction. The filter device was operated for 10 minutes in the wash mode with a 2 mL/min bottom flow rate without cross flow (20 mL, n=3), with a 5 mL/min lateral flow rate without bottom flow (Lateral 5 mL/min, n=3) and with a 2 mL/min bottom flow rate and a 5 mL/min cross flow rate (2/5 flow, n=6). ***, p<0.001 (Tukey's multicomparison test).
254x190mm (96 x 96 DPI)




Synergistic effect of the lateral and bottom flows. The presence of lateral flow during cell washing yields a DMSO reduction profile described by a single exponential function, as predicted by equation 1 (A, closed squares, $n \geq 3$ for all data points), whereas in the absence of the lateral flow, the bottom flow wash does not fit the equation 1 model (A, closed circles, $n \geq 3$ for all data points). The typical pressure profiles for these 2 wash modes show that the presence of cross flow also leads to pressure build-up in the lower chamber, unlike the bottom flow only operation mode (B). To mimic the final thermoplastic product the bottom flow chamber geometry was injection molded (C, purple) to fit a PMMA housing (C, gray).
85x56mm (300 x 300 DPI)



– Redesigning the filtration device for improved DMSO removal. The original filtration device was redesigned by changing the geometry of the bottom chamber which holds the cell suspension, which led to changes in the distance between the bottom inlet and the filter membrane, and filtration area (A). A CFD analysis on the M2 geometry revealed “dead zones” where the DMSO was being removed with less efficiency (B). The 3 different geometries were evaluated for their DMSO reduction capability (C, $n \geq 3$ for all data points), operated with (open symbols) and without (closed symbols) lateral flow, except for M3, where lateral flow lead to clogging (data not shown); the final pressure in the lower chamber, after 23 min, also revealed differences between geometries and wash modes (D, $n=3$ for all data points). The sine of the angle α , formed between the horizontal plane and the lower chamber slope that connects the bottom and lateral inlets (E), is different between devices; when the number of diavolumes is multiplied by the $\sin(\alpha)$, specific for each device, the curves that obey equation 1 in C follow a single exponential decay empirical model, as described by equation 2 (F). The M3 geometry was assessed for its ability to reduce albumin-FITC by washing a cell suspension for 24 minutes at 2mL/min bottom flow (G); furthermore, the cellular proliferation

1
2
3 after filtration processing, using the M3 geometry (n=4), was compared with post-centrifugation processing
4 (n=5) and fresh cell culture processing (i.e. cell counts during passaging, n=5) by calculating the specific
5 growth rate μ (day⁻¹).
6 119x150mm (300 x 300 DPI)
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

For Peer Review



hMSC post filtration recovery, proliferation and characterization. hMSCs were processed using the M3 geometry filtration device. The post filtration characterization after 3 cell runs (generated from 2 donors) was performed according to the ISCT criteria: cell surface markers (Figure 6 A, individual flow cytometry histograms, and B, aggregate results from the 3 cell runs), adherence to plastic (Figure 6E, scale bar=400 μ m) and the capability for chondrogenic, osteogenic and adipogenic differentiation (Figure 6G, H, scale bar=100 μ m, and I, scale bar=50 μ m, respectively). The viable cell recovery, DMSO removal and hMSC growth rates after filtration are displayed in Figure 6 C, D and F, respectively).

119x134mm (300 x 300 DPI)