

Toronto Western Hospital–UHN, University of Toronto and Krembil Research Institute (A.F., A.M.L.), Toronto, Ontario, Canada.

Author contributions: Alfonso Fasano: drafting/ revising the manuscript, study concept or design, analysis or interpretation of data, accepts responsibility for conduct of research and will give final approval, acquisition of data, study supervision. Francesco Sammartino: drafting/ revising the manuscript, study concept or design, analysis or interpretation of data, accepts responsibility for conduct of research and will give final approval, acquisition of data, statistical analysis. Mabeleth Llinas: drafting/ revising the manuscript, analysis or interpretation of data, accepts responsibility for conduct of research and will give final approval, statistical analysis, study supervision. Andres M. Lozano: drafting/ revising the manuscript, study concept or design, analysis or interpretation of data, accepts responsibility for conduct of research and will give final approval, contribution of vital reagents/tools/ patients, acquisition of data, statistical analysis, study supervision, obtaining funding.

Study funding: No targeted funding reported.

Disclosure: A.M.L. is a consultant to the Focused Ultrasound Foundation and has received research support from InSightec in the past. Dr. Fasano has received grant support from the University of Toronto, the McLaughlin Centre, and the Michael J. Fox Foundation; he received speaking honoraria from UCB Pharma, Medtronic, Boston Scientific, AbbVie, Novartis, Chiesi Pharmaceutical, and TEVA; he is on an advisory board for AbbVie and provided consultancies for UCB Pharma, Medtronic, Boston Scientific, and AbbVie. Ms. Llinas and Dr. Sammartino report no disclosures. Dr. Lozano holds stock ownership in medically related fields (Functional Neuromodulation) and provided consultancies for Medtronic St. Jude, Boston Scientific, and InSightec. Go to Neurology.org for full disclosures.

© 2016 American Academy of Neurology

1. Fasano A, Deuschl G. Therapeutic advances in tremor. *Mov Disord* 2015;30:1557–1565.
2. Lipsman N, Schwartz ML, Huang Y, et al. MR-guided focused ultrasound thalamotomy for essential tremor: a proof-of-concept study. *Lancet Neurol* 2013;12:462–468.
3. Elias WJ, Huss D, Voss T, et al. A pilot study of focused ultrasound thalamotomy for essential tremor. *N Engl J Med* 2013;369:640–648.
4. Chang WS, Jung HH, Kweon EJ, Zadicario E, Rachmilevitch I, Chang JW. Unilateral magnetic resonance guided focused ultrasound thalamotomy for essential tremor: practices and clinicoradiological outcomes. *J Neurol Neurosurg Psychiatry* 2015;86:257–264.
5. Sammartino F, Krishna V, Kon Kam King N, Lozano A, Hodaie M. Tractography-based ventral intermediate nucleus targeting: novel methodology and intraoperative validation. *Mov Disord Epub* 2016 May 23.
6. Huss DS, Dallapiazza RF, Shah BB, Harrison MB, Diamond J, Elias WJ. Functional assessment and quality of life in essential tremor with bilateral or unilateral DBS and focused ultrasound thalamotomy. *Mov Disord* 2015;30:1937–1943.
7. Jung HH, Kim SJ, Roh D, et al. Bilateral thermal capsulotomy with MR-guided focused ultrasound for patients with treatment-refractory obsessive-compulsive disorder: a proof-of-concept study. *Mol Psychiatry* 2015;20:1205–1211.

Francis B. Panosyan, MD,
PhD
Joan M. Mountain, RN
Mary M. Reilly, MD
Michael E. Shy, MD
David N. Herrmann,
MBBCh

RYDEL-SEIFFER FORK REVISITED: BEYOND A SIMPLE CASE OF BLACK AND WHITE

In 1903, A. Rydel and F.W. Seiffer published the use of what is now commonly known as the Rydel-Seiffer tuning fork.¹ In their paper, they described an adaptation of the standard tuning fork to make it semiquantitative on an 8-point scale in measuring vibration sensing. The Rydel-Seiffer tuning fork has been used extensively over the past century and normative values have been published.² It has also been validated with the use of more modern sensory nerve action potentials.³ Vibration threshold measurement with the Rydel-Seiffer tuning fork is therefore increasingly used in outcome measures assessing neuropathy from diverse etiologies: chemotherapy-induced peripheral neuropathy,⁴ Charcot-Marie-Tooth disease,⁵ multifocal motor neuropathy,⁶ and immune-mediated neuropathies.⁷

The modern Rydel-Seiffer tuning fork has a black triangle on one prong and a white triangle on the second prong (figure, A). After the fork is snapped into motion, the prongs start to oscillate and the illusion of 2 triangles becomes visible on each damper. The intersect between the 2 virtual triangles moves up the scale with gradual attenuation of the amplitude of the

oscillation. The position of the intersect is recorded on an arbitrary scale from 0 to 8 once the subject is no longer perceiving vibration.


We have observed consistent discrepancy between the values obtained from the black triangle in comparison to the white triangle on the Rydel-Seiffer fork. Values obtained from the virtual intersect of the white triangle are consistently lower in comparison to that from the black triangle (figure, C and D). This discrepancy can be as much as 25% or 2 units in the scale of 0 to 8. This is best appreciated during the motion of the prongs (video on the *Neurology*[®] Web site at Neurology.org).

There is currently no report distinguishing between the black and white scales and both are used interchangeably.^{2,3} Of note, the original publication from Rydel and Seiffer depicts the tuning fork with 2 black triangles on the dampers (figure, B). It is unclear as to why or when a modification was made and one of the black triangles was substituted with a white triangle.

Remarkably, the observed discrepancy between the 2 scales can be eliminated by simply darkening the white triangle on the fork with a marker (figure, E and F). It is of interest that, upon introducing color to the white triangle (purple or red), the readings obtained from the black triangle showed lower values in comparison

Supplemental data
at Neurology.org

Figure Rydel-Seiffer tuning fork



(A) An example of a modern-day Rydel-Seiffer tuning fork. (B) Original design of the tuning fork as published by Rydel and Seiffer in 1903 (reprinted with permission of Springer).¹ (C) Rydel-Seiffer tuning fork at rest. (D) Rydel-Seiffer tuning fork in motion with the black triangle showing intersect between 4 and 5 (on the 0 to 8 scale) and white triangle showing intersect between 3 and 4 (discrepancy is better appreciated in the video). (E) Same Rydel-Seiffer tuning fork with the white triangle blacked out at rest. (F) Rydel-Seiffer tuning fork with 2 black triangles in motion showing equivalent reading with intersect at 4.

to the colored triangle (video). The perceived intersection between the triangles seems to be therefore influenced by the contrast and the color of the triangles. This is likely attributable to visual afterimage processing, although a more tangible explanation remains elusive.

To maintain an acceptable interrater reliability in the use the Rydel-Seiffer tuning fork, we suggest that the original design with the 2 black triangles is considered. Alternatively, consensus should be established on which triangle to use in the future for clinical practice and clinical research.

From the Department of Neurology (F.B.P., J.M.M., D.N.H.), University of Rochester Medical Center, NY; MRC Centre for Neuromuscular Diseases (M.M.R.), UCL Institute of Neurology, London, UK; and Department of Neurology (M.E.S.), University of Iowa Hospitals and Clinics, Iowa City.

Author contributions: Francis B. Panosyan: concept and design, acquisition of data, analysis and interpretation of data. Joan M. Mountain: revising the manuscript for content, acquisition of data. Mary M. Reilly: analysis and interpretation of data, revising the manuscript for content. Michael E. Shy: analysis and interpretation of data, revising the manuscript for content. David N. Herrmann: analysis and interpretation of data, revising the manuscript for content, supervisor.

Study funding: Supported by NIH-RDCRN U54 (U54NS065712).

Disclosure: F. Panosyan, J. Mountain, M. Reilly, and M. Shy report no disclosures relevant to the manuscript. D. Herrmann received grant support from NINDS (U54NS065712), FARA, MDA, and has received funds for consulting or adjudication activities from Medpace, Guidepoint Global, Voyager, Inc., and LAM Therapeutics. Go to Neurology.org for full disclosures.

Received March 9, 2016. Accepted in final form May 5, 2016.

Correspondence to Dr. Panosyan: Francis_Panosyan@urmc.rochester.edu

© 2016 American Academy of Neurology

- Rydel A, Seiffer W. Examinations of feelings of vibrations or so-called "bone sensitivity" (pallaesthesia). *Arch Psychiatr Nervenkr* 1903;37:488–536.
- Martina IS, van Koningsveld R, Schmitz PI, van der Meché FG, van Doorn PA. Measuring vibration threshold with a graduated tuning fork in normal aging and in patients with polyneuropathy. *European Inflammatory Neuropathy Cause and Treatment (INCAT) Group. J Neurol Neurosurg Psychiatry* 1998;65:743–747.
- Pestronk A, Florence J, Levine T, et al. Sensory exam with a quantitative tuning fork: rapid, sensitive and predictive of SNAP amplitude. *Neurology* 2004;62:461–464.
- Cavaletti G, Cornblath DR, Merkies IS, et al. The chemotherapy-induced peripheral neuropathy outcome measures standardization study: from consensus to the first validity and reliability findings. *Ann Oncol* 2013;24:454–462.
- Murphy SM, Herrmann DN, McDermott MP, et al. Reliability of the CMT neuropathy score (second version) in Charcot-Marie-Tooth disease. *J Peripher Nerv Syst* 2011;16:191–198.
- Cats EA, van der Pol WL, Piepers S, et al. Correlates of outcome and response to IVIg in 88 patients with multifocal motor neuropathy. *Neurology* 2010;75:818–825.
- Merkies ISJ, Schmitz PIM, van der Meche FGA, van Doorn PA. Reliability and responsiveness of a graduated tuning fork in immune mediated polyneuropathies. *The Inflammatory Neuropathy Cause and Treatment (INCAT) Group. J Neurol Neurosurg Psychiatry* 2000;68:669–671.

Look What's New at the 2016 AAN Fall Conference

Visit AAN.com/view/fall to register for the 2016 AAN Fall Conference, set for October 14–16, at The Cosmopolitan of Las Vegas. This is your year-end destination for acquiring the latest clinical advances in key disease states, improving your practice's efficiency and bottom line, and earning up to 15.75 CME credits.

Look What's New!

- All-inclusive registration rate offers greater value than ever
- Flexible meeting format lets you build your own tailored schedule and select the programs of most interest—while on the go on-site
- New Topics: Update in Stroke, AAN Leadership University Course: Challenges of Leadership in Private Practice, Headache Skills Workshop (*pre-registration required*)

MACRA Changes How Neurologists Are Paid

What Is MACRA, and Why Is It Necessary?

The **Medicare Access and CHIP Reauthorization Act of 2015—MACRA**—replaced the Medicare Sustainable Growth Rate formula for calculating physician payments. The rapidly rising costs of health care in the US are unsustainable. Changes in the health care system are essential and must happen now.

How Will MACRA Affect Physician Payment?

Our health care system is moving from "fee-for-service" payments to a wider array of "value-based" payment models that put the patient at the center of care by calling for the improvement of the quality, safety, and overall experience of patient care while demonstrating cost-effectiveness by providing care that is less expensive and delivers similar or improved clinical outcomes.

Learn more at AAN.com/view/MACRA.

Neurology[®]

Rydel-Seiffer fork revisited: Beyond a simple case of black and white

Francis B. Panosyan, Joan M. Mountain, Mary M. Reilly, et al.

Neurology 2016;87;738-740 Published Online before print July 13, 2016

DOI 10.1212/WNL.0000000000002991

This information is current as of July 13, 2016

Updated Information & Services	including high resolution figures, can be found at: http://www.neurology.org/content/87/7/738.full.html
Supplementary Material	Supplementary material can be found at: http://www.neurology.org/content/suppl/2016/07/13/WNL.0000000000002991.DC1
References	This article cites 7 articles, 4 of which you can access for free at: http://www.neurology.org/content/87/7/738.full.html##ref-list-1
Subspecialty Collections	This article, along with others on similar topics, appears in the following collection(s): All Neuromuscular Disease http://www.neurology.org/cgi/collection/all_neuromuscular_disease Clinical neurology examination http://www.neurology.org/cgi/collection/clinical_neurology_examination Clinical neurology history http://www.neurology.org/cgi/collection/clinical_neurology_history
Permissions & Licensing	Information about reproducing this article in parts (figures, tables) or in its entirety can be found online at: http://www.neurology.org/misc/about.xhtml#permissions
Reprints	Information about ordering reprints can be found online: http://www.neurology.org/misc/addir.xhtml#reprintsus

Neurology® is the official journal of the American Academy of Neurology. Published continuously since 1951, it is now a weekly with 48 issues per year. Copyright © 2016 American Academy of Neurology. All rights reserved. Print ISSN: 0028-3878. Online ISSN: 1526-632X.

