provided by UCL Discover

RATE

Special Collection: Affordable Rehabilitation and Assistive Technologies

Identifying key experience-related differences in over-ground manual wheelchair propulsion biomechanics

Journal of Rehabilitation and Assistive Technologies Engineering Volume 3: 1–10 © The Author(s) 2016 Reprints and permissions: sagepub.co.uk/journalsPermissions.nav DOI: 10.1177/2055668316678362 irt.sagepub.com

Andrew Symonds¹, Catherine Holloway², Tatsuto Suzuki³, Peter Smitham¹, Angela Gall⁴ and Stephen JG Taylor¹

Abstract

Objectives: The purpose of this study was to investigate technique differences between expert and novice manual wheelchair users during over-ground wheelchair propulsion.

Method: Seven experts (spinal cord injury level between T5 and L1) and six novices (non-wheelchair users) pushed a manual wheelchair over level ground, a 2.5% cross slope and up a 6.5% incline (7.2 m length) and 12% incline (1.5 m length). Push rim kinetics, trunk and shoulder kinematics and muscle activity level were measured.

Results: During the level and cross slope tasks, the experts completed the tasks with fewer pushes by applying a similar push rim moment over a greater push arc, demonstrating lower muscle activity. During the incline tasks, the experts required fewer pushes and maintained a greater average velocity, generating greater power by applying a similar push rim moment over a greater push arc with greater angular velocity, demonstrating greater trunk flexion and higher shoulder muscle activity.

Conclusions: This study identifies experience-related differences during over-ground manual wheelchair propulsion. These differences are particularly evident during incline propulsion, with the experts generating significantly greater power to maintain a higher velocity.

Keywords

Spinal cord injury, manual wheelchair propulsion, biomechanics, experience, technique

Introduction

Sustained manual wheelchair propulsion can lead to shoulder injury, with rotator cuff degeneration most commonly reported. Rotator cuff injuries are associated with increasing age and time as a manual wheelchair user, and can lead to secondary degenerative complications. Published guidelines suggest manual wheelchair users should aim to minimise repetition and peak forces experienced during tasks to reduce upper limb injury risk. 4,5

Previous research has examined differences in propulsion technique between novice and expert wheelchair users. Rodgers et al. examined experts and novices propelling on a wheelchair ergometer when both fresh and fatigued.⁶ The experts applied a lower hand rim moment to maintain the required velocity, with a significantly higher push rate and lower contact time. Another study examined biomechanical differences between novices and experts during propulsion

at different speeds on a dynamometer. The expert users maintained a greater average velocity than the novices, generating greater power without an increase in application of torque, achieved in part by application of force over a greater push arc. A further study examined muscle activity levels of experts and novices, demonstrating higher muscle activity levels in the

Corresponding author:

Andrew Symonds, University College London Institute of Orthopaedics and Musculoskeletal Science, Royal National Orthopaedic Hospital Stanmore, Brockley Hill, Stanmore HA7 4LP, UK. Email: andrew.symonds.12@ucl.ac.uk

¹Aspire Centre for Rehabilitation Engineering and Assistive Technology, University College London, United Kingdom

²University College London Interaction Centre

³Department of Civil, Environmental and Geomatic Engineering, University College London, United Kingdom

⁴London Spinal Cord Injury Centre, Royal National Orthopaedic Hospital NHS Trust, United Kingdom

expert paraplegic users, who chose to propel at a higher velocity than the novices.⁸

These ergometer-based studies demonstrate that the expert users are able to propel more effectively than novices, either by applying torque to the push rim over a greater push arc at a greater angular velocity or at a higher push rate, but that higher muscle activity levels may be required to achieve this. Further investigation is required to examine whether such differences in propulsion technique are evident during over-ground propulsion, particularly when tasks become more challenging. In particular, it would be useful to examine whether expert users are able to maintain the suggested optimal technique to complete more challenging propulsion tasks at a greater velocity with fewer pushes, and what impact this has on muscle activity level.

Aims

The aim of this study was to compare manual wheelchair propulsion technique between experts and novices during a variety of over-ground tasks, by examining push rim kinetics, trunk and upper limb kinematics and also shoulder muscle activity level. It was hypothesised that the expert users would be able to achieve each task with fewer pushes and maintain a higher average velocity.

Methods

Participants

Thirteen participants were recruited, seven experienced wheelchair users with a history of spinal cord injury (SCI) and six novices without mobility impairment (Table 1). The study was approved by the London Stanmore Research Ethics committee and the University College London (UCL) Ethics committee. The SCI participants were recruited if they used a wheelchair as a primary form of mobility and had a history of SCI below T1 with no previous history of

Table 1. Participant characteristics.

	SCI participants	Non-SCI participants		
Participants (number)	7	6		
Mean age \pm SD (years)	$\textbf{42.71} \pm \textbf{13.26}$	$\textbf{34.67} \pm \textbf{8.56}$		
Mean time since injury \pm SD (years)	$\textbf{8.85} \pm \textbf{4.67}$	n/a		
Sex (M/F)	7/0	5/1		
Injury level (range)	T5 – LI	n/a		
Mean weight \pm SD (kg)	83.14 ± 8.05	71.25 ± 12.29		

shoulder pain or major shoulder surgery. The ablebodied participants were recruited if they reported no history of shoulder pain or surgery. Participants provided written informed consent in advance of study participation. There were no statistically significant differences in age or weight between the two groups.

Experimental protocol

participants attended UCL's Pedestrian Accessibility and Movement Environment Laboratory (PAMELA) for a single visit. Participants transferred into the test wheelchair, the Vanos Excel G6 High Active 'Sport Edition'. The chair was adjusted to ensure an elbow joint flexion angle of 100-130° when the hand was placed at the top dead centre of the push rim. The participants performed four pushing tasks: level surface (8.4 m), 2.5% cross slope (7.2 m, instrumented side on the down slope), 6.5% incline (7.2 m) and 12% incline (1.5 m ramp). During each of the tasks, push rim kinetics, trunk and upper limb kinematics and surface electromyography (EMG) were recorded.

Push rim kinetics

Push rim kinetics were recorded using the Sensewheel Mark 1 (Movement Metrics, London, UK), a low-cost and lightweight instrumented wheelchair wheel measuring three-dimensional forces applied to the push rim, torque about the wheel axle and angular velocity of the wheel. The Sensewheel was positioned on the left side of the wheelchair; data were sampled at 50 Hz and analysed using Matlabr2012b (Mathworks Inc, MA, USA).

The push phase of the propulsion cycle was defined by measurement of the application of a positive moment about the wheel axle. The number of pushes to complete the task was calculated from detection of the first push phase, until detection of the braking phase. Mean velocity and push rate were calculated for the same time period. Power was calculated using measurement of the moment applied to the wheel (tangential force × wheel radius) and angular velocity of the wheel, and mean peak and the mean value for the whole task were calculated.⁹

Power (W) = Moment (N · m) ·
$$\omega$$
 (rad · s⁻¹)

The mean work per cycle was calculated using the mean power and push rate.

Work (J) = Power (W)/Average push rate (
$$S^{-1}$$
)

The mean moment value was calculated from the whole task. Mean angular velocity, percentage push

phase and push arc were calculated as an average of each push phase from the whole task.

Trunk and upper limb kinematics

Trunk and left thoraco-humeral kinematics were measured using the XSens MTw inertial measurement system (XSens Technologies, NL). XSens units were attached to the thorax and humerus. The participant was positioned in a 'neutral' position of thoraco-humeral angle, to align the sensors with the 'anatomical' coordinate system. Data were sampled at 50 Hz, and the rotation matrix for each unit exported to Matlab for post processing. Matrix multiplication to calculate the relative position of the XSens unit on the thorax with respect to the neutral start position (trunk kinematics) and the XSens unit on the humerus with respect to the XSens unit on the thorax (thoraco-humeral kinematics) was completed, 10 and then the Euler angles from the rotation matrix were calculated. For each push of each task, maximum, minimum and change in trunk flexion and thoraco-humeral extension, abduction and internal rotation were calculated. Average values of each measurement were calculated for statistical analysis.

Surface EMG

Surface EMG was recorded from the anterior deltoid (AD), pectoralis major (PM) and infraspinatus (IS) muscles using the Delsys Trigno TM Wireless System (Delsys Inc, MA, USA). Each sensor contains four contacts, 99.9% silver, dimensions 5×1 mm, with two active contacts and two stabilising references. Skin surface EMG is amplified by 1000, with a signal bandwidth ranging from 20 to 450 Hz. Baseline noise is reported as $<750\,\mathrm{nV}$ RMS, with a Common Mode Rejection Ratio of $>80\,\mathrm{db}$. Data were sampled at 2000 Hz.

Sensors were attached to the left upper limb, in accordance with the SENIAM guidelines for sensor placement.¹¹ Data were recorded from maximal voluntary isometric contractions (MVIC) for each of the muscles using the functional tests described,¹² and then during each of the pushing tasks.

The data were exported to Matlabr2012b (Mathworks Inc, MA, USA) for analysis. All data were full wave rectified, and low-pass filtered using a fourth-order Butterworth filter with a cut-off frequency of 5 Hz. The pushing tasks data for each muscle were normalised using the values obtained from the MVIC tests. The peak and mean values for each muscle were obtained for each push phase of each propulsion cycle for each of the tasks. A mean value for peak and mean muscle activity level for each muscle was calculated for each of the pushing tasks, using the peak and mean value from every push of each task.

Statistical analysis

Statistical analysis was completed using IBM SPSS Statistics version 22 (IBM Corp., NY, USA). Homogeneity of variance was analysed in advance of the between-group comparisons using Levene's test. A split plot ANOVA with two groups (novice and expert) and four repeated measures (level, cross slope, 6.5% incline, 12% incline) was performed for each push rim parameter and kinematic and surface EMG variable. For the repeat measures component of the analysis, when Mauchly's test indicated that the assumption of sphericity had been violated, degrees of freedom were corrected using the Greenhouse-Geisser estimates. Between-group differences for each outcome measure during each of the tasks were assessed using the independent samples t-test. Significance level was set at p < 0.05.

Results

Push rim kinetics

Table 2 summarises the push rim kinetics measured using the Sensewheel. Analysis of the Sensewheel data revealed a significant experience level by task interaction for a number of the push rim parameters, with the two groups adopting significantly different techniques to negotiate the more challenging incline tasks.

The expert group required fewer pushes compared with the novice group when negotiating each of the propulsion tasks. Although not significantly different, the expert users required fewer pushes to complete the level and cross slope tasks by applying a similar moment over a greater push arc, using a greater percentage of the push cycle. The reduction in the number of pushes required by the experts was significant during the 6.5% incline task (6.29 pushes vs. 8.83 pushes, p = 0.028) and also lower during the 12% incline task (3.57 pushes vs. 5.17 pushes, p = 0.060). The expert group were also able to maintain a higher average velocity during the incline tasks, 6.5% incline (0.81 m·s⁻¹ vs. 0.61 m·s⁻¹, p = 0.009) and 12% incline (0.59 m·s⁻¹ vs. 0.41 m·s⁻¹, p = 0.012).

The expert group completed the incline tasks with fewer pushes while maintaining a higher average velocity by applying a significantly greater average power, 6.5% incline (30.45 W vs. 21.98 W, p = 0.033) and 12% incline (33.27 W vs. 17.67 W, p = 0.003). The expert group generated this significantly greater average power by applying a similar moment to the push rim as the novice group at a greater angular velocity, 6.5% incline (150.48°·s⁻¹ vs. 110.46°·s⁻¹, P = 0.007) and 12% incline (107.37°·s⁻¹ vs. 77.67°·s⁻¹, p = 0.023). The expert group also applied the moment over a greater push arc, 6.5% incline (94.32° vs. 65.15°, p = 0.002) and 12% incline (78.23° vs. 59.74°,

Table 2. Push rim kinetics, data are mean (SD), statistically significant results in bold.

	Task					Between-group comparisons			
	Level	2.5% cross slope	6.5% incline	12% incline	ANOVA	Level	2.5% cross slope	6.5% incline	12% incline
	Number of pus	hes							
Novice	8.67 (1.63)	7.83 (1.33)	8.83 (2.14)	5.17 (1.47)	0.111	0.098	0.516	0.028	0.060
Expert	7.29 (1.11)	7.14 (2.19)	6.29 (1.50)	3.57 (1.27)					
	Mean velocity (m·s ⁻¹)							
Novice	0.87 (0.11)	0.91 (0.07)	0.61 (0.10)	0.41 (0.11)	0.006	0.945	0.952	0.009	0.012
Expert	0.87 (0.15)	0.91 (0.15)	0.81 (0.12)	0.59 (0.11)					
	Mean power (V	V)							
Novice	10.76 (2.54)	19.55 (4.89)	21.98 (5.29)	17.67 (4.18)	0.000	0.984	0.942	0.033	0.003
Expert	10.80 (4.06)	19.35 (4.97)	30.45 (6.96)	33.27 (9.26)					
	Maximum power	er (W)							
Novice	55.13 (13.35)	85.21 (14.19)	74.28 (16.07)	55.49 (13.48)	0.000	0.570	0.159	0.014	0.004
Expert	49.50 (19.96)	71.90 (17.06)	114.55 (30.18)	111.14 (35.43)					
	Mean work per	cycle (J)							
Novice	12.07 (3.88)	19.51 (4.64)	22.43 (4.28)	18.51 (3.26)	0.005	0.647	0.559	0.034	0.005
Expert	13.22 (4.75)	21.64 (7.49)	32.20 (9.05)	32.59 (9.47)					
	Wheel moment	t (N.m)							
Novice	4.39 (0.82)	6.92 (1.44)	10.87 (2.08)	14.26 (4.67)	0.328	0.867	0.828	0.773	0.357
Expert	4.30 (1.04)	6.78 (0.81)	11.19 (1.76)	16.19 (2.36)					
	Mean angular v	elocity (°·s ⁻¹)							
Novice	162.91 (18.54)	176.75 (10.32)	110.46 (19.06)	77.67 (20.62)	0.001	0.849	0.506	0.007	0.023
Expert	165.63 (29.46)	169.48 (23.93)	150.48 (23.61)	107.37 (19.94)					
	Push rate (s ⁻¹)								
Novice	0.92 (0.10)	1.00 (0.10)	0.98 (0.13)	0.95 (0.11)	0.075	0.292	0.277	0.872	0.165
Expert	0.82 (0.18)	0.92 (0.14)	0.97 (0.15)	1.02 (0.06)					
	Percentage push	h phase (%)							
Novice	41.28 (5.25)	48.02 (5.73)	58.09 (3.81)	74.83 (7.73)	0.104	0.065	0.176	0.225	0.433
Expert	47.21 (5.14)	51.93 (4.00)	60.43 (2.72)	72.14 (3.89)					
	Push arc (°)								
Novice	71.04 (14.83)	80.75 (10.05)	65.15 (10.03)	59.74 (26.05)	0.309	0.013	0.188	0.002	0.137
Expert	93.83 (13.01)	91.25 (15.71)	94.32 (14.34)	78.23 (14.86)					

p = 0.137). There were no significant differences between each of the groups across all tasks in terms of push rate and percentage of the push phase. The key differences in propulsion technique between the novices and experts are presented in Figure 1.

Upper limb and trunk kinematics

Tables 3 and 4 demonstrate the kinematic results calculated using the XSens inertial measurement system. The kinematic analysis did not demonstrate a significant experience level by task interaction for maximum, minimum or change in thoraco-humeral angle (in each of the three planes of movement). The results of the betweengroup comparisons demonstrated that the novice

group propelled the chair with a greater maximum and minimum abduction angle than the expert group, but there was no difference in change in abduction.

The results demonstrated a significant experience level by task interaction for change in trunk flexion angle. During both incline tasks, the expert group demonstrated a significantly greater change in trunk flexion angle compared with the novice group, 6.5% incline (19.96° vs. 7.85° , p = 0.020) and 12% incline (21.26° vs. 8.99° , p = 0.006).

Surface EMG

For each of the trials, data from one participant were excluded due to anomalous results. In these results,




Figure 1. Sensewheel data during different propulsion tasks.

normalised peak muscle activity was significantly in excess of 100% MVIC, indicating that that the MVIC test was not completed effectively. The results demonstrated a significant experience level by task interaction for each of the muscles tested (Table 5). During the level and 2.5% cross slope tasks, the expert group demonstrated lower muscle activity levels than the novice group for each of the muscles tested, although there were no significant differences between the groups. During the incline tasks, the expert group demonstrated higher muscle activity levels than the novice group for each of the muscles tested, with a significantly greater peak activity of the AD muscle during the 12% incline task (65.73% vs. 30.24%, p = 0.039). Peak EMG results are presented in Figure 2.

Discussion

Push rim kinetics

The results demonstrated a significant experience level by task interaction, with the expert users demonstrating a significantly different propulsion technique to the novices during the incline propulsion tasks. During both incline propulsion tasks, the expert group required fewer pushes, and maintained a significantly higher velocity. They achieved this by generating greater power, by applying a similar moment over a greater push arc, at a significantly greater angular velocity, similar to findings during ergometer-based testing. There were

no significant differences in push rate, or percentage push phase. Although not statistically significant, the experts also required fewer pushes to complete the level and cross slope tasks. The expert group closely followed the suggested guidelines in terms of push rate and push arc, while minimising push force.¹³

Trunk and upper limb kinematics

The expert users demonstrated a significantly greater change in trunk flexion angle during both of the incline propulsion tasks in comparison with the novices. An increase in trunk flexion angle has been previously reported with progressive increases of incline. 14,15 Increasing trunk flexion angle enabled the expert users to apply force to the push rim over a greater arc, without a significant increase in thoraco-humeral flexion angle. Increased trunk flexion has also been previously reported as a mechanism of force production for wheelchair propulsion. 16 These results are of interest, as the expert group demonstrated greater trunk flexion than the novice group despite not having full activity of the trunk and hip flexor muscles.

The only experience level-related kinematic difference in thoraco-humeral angle was that the novice group propelled at a greater abduction angle than the expert group, although this difference was not influenced by change in task. Excessive abduction should be avoided, as the combined posture of extreme shoulder joint extension, abduction and internal rotation at

Table 3. Thoraco-humeral kinematics, data are mean (SD), statistically significant results in bold.

	Task					Between-group comparisons			
	Level	2.5% cross slope	6.5% incline	12% incline	ANOVA	Level	2.5% cross slope	6.5% incline	12% incline
	Maximum exten	sion (°)							
Novice	41.58 (8.20)	44.54 (10.13)	40.10 (6.99)	35.45 (8.44)	0.235	0.566	0.552	0.718	0.365
Expert	43.92 (6.09)	41.02 (10.50)	41.48 (6.45)	39.36 (6.45)					
	Minimum extens	sion (°)							
Novice	-2.46 (8.23)	-2.43 (6.61)	-1.91 (6.67)	-1.07 (10.09)	0.492	0.422	0.570	0.417	0.834
Expert	0.63 (4.96)	-5.17 (9.67)	-5.69 (9.06)	-2.07 (6.60)					
	Change in exten	nsion (°)							
Novice	44.04 (8.97)	46.97 (5.46)	42.01 (6.18)	36.53 (10.23)	0.257	0.868	0.837	0.141	0.305
Expert	43.29 (6.74)	46.19 (7.55)	47.17 (5.57)	41.43 (5.97)					
	Maximum abduc	ction (°)							
Novice	41.12 (6.76)	43.59 (6.76)	43.44 (16.04)	39.11 (17.47)	0.635	0.033	0.026	0.03 I	0.114
Expert	29.23 (10.15)	27.81 (9.11)	25.92 (9.07)	24.67 (12.83)					
	Minimum abduct	tion (°)							
Novice	21.44 (4.81)	19.23 (9.54)	21.34 (10.70)	18.87 (12.03)	0.504	0.006	0.040	0.024	0.068
Expert	11.87 (5.19)	9.45 (5.35)	7.57 (8.25)	6.73 (9.67)					
	Change in abdu	ction (°)							
Novice	19.69 (5.11)	24.36 (12.58)	22.10 (16.77)	20.24 (15.31)	0.554	0.476	0.267	0.573	0.707
Expert	17.36 (6.09)	18.35 (4.90)	18.36 (3.25)	17.94 (3.99)					
	Maximum intern	nal rotation (°)							
Novice	16.44 (8.11)	28.75 (16.50)	16.05 (27.64)	14.13 (25.28)	0.468	0.751	0.127	0.566	0.518
Expert	14.83 (9.48)	16.02 (11.19)	9.22 (12.39)	7.01 (11.79)					
	Minimum interna	al rotation (°)							
Novice	-26.01 (10.23)	-21.42 (12.38)	-28.09 (17.19)	-25.19 (11.60)	0.433	0.483	0.768	0.858	0.693
Expert	-22.31 (8.16)	-23.27 (9.74)	-29.54 (11.21)	-27.72 (10.84)					
	Change in rotati	ion (°)							
Novice	42.44 (8.55)	50.17 (11.04)	44.13 (13.17)	39.32 (14.91)	0.652	0.277	0.078	0.386	0.516
Expert	37.14 (8.15)	39.29 (9.19)	38.76 (8.10)	34.73 (9.62)					

Table 4. Trunk kinematics, data are mean (SD), statistically significant results in bold.

	Task					Between-group comparisons			
	Level	2.5% cross slope	6.5% incline	12% incline	ANOVA	Level	2.5% cross slope	6.5% incline	12% incline
	Minimum trun	k flexion (°)							
Novice	1.12 (6.76)	1.32 (9.38)	3.48 (12.81)	9.77 (13.58)	0.671	0.547	0.466	0.953	0.507
Expert	-0.91 (4.99)	-I.69 (4.49)	3.18 (4.18)	6.16 (3.18)					
	Maximum trun	k flexion (°)							
Novice	8.78 (4.91)	8.49 (7.32)	11.34 (14.68)	18.76 (16.21)	0.062	0.468	0.960	0.126	0.262
Expert	6.46 (6.00)	8.66 (4.09)	23.14 (11.04)	27.42 (9.97)					
	Change in trui	nk flexion (°)							
Novice	7.66 (3.00)	7.17 (3.18)	7.85 (2.18)	8.99 (3.27)	0.003	0.874	0.197	0.020	0.006
Expert	7.37 (3.26)	10.34 (4.80)	19.96 (10.64)	21.26 (8.37)					

Table 5. Surface EMG, data are mean (SD), statistically significant results in bold.

	Task					Between-group comparisons			
	Level	2.5% cross slope	6.5% incline	12% incline	ANOVA	Level	2.5% cross slope	6.5% incline	12% incline
'	Peak anterior	deltoid (% MVIC)							
Novice	19.77 (4.87)	28.05 (8.29)	30.97 (7.46)	30.24 (11.23)	0.029	0.099	0.261	0.105	0.039
Expert	15.39 (3.30)	22.25 (8.59)	57.34 (35.49)	65.73 (34.17)					
·	Mean anterior	deltoid (% MVIC)	, ,					
Novice	12.99 (3.74)	17.48 (3.95)	18.67 (5.20)	17.14 (7.16)	0.025	0.051	0.083	0.174	0.071
Expert	8.94 (2.46)	12.72 (4.58)	29.08 (16.62)	34.20 (19.36)					
	Peak pectoralis	major (% MVIC)						
Novice	28.81 (17.04)	40.31 (24.25)	34.37 (24.54)	37.01 (29.52)	0.012	0.071	0.075	0.645	0.573
Expert	15.03 (5.74)	20.60 (9.37)	40.39 (19.58)	45.45 (21.06)					
	Mean pectorali	s major (% MVIC	()						
Novice	14.80 (9.09)	20.79 (13.69)	18.87 (12.76)	19.51 (11.78)	0.001	0.132	0.133	0.684	0.339
Expert	8.80 (3.19)	11.65 (5.17)	21.49 (9.03)	26.41 (11.68)					
	Peak infraspina	tus (% MVIC)	, ,	, ,					
Novice	49.34 (22.92)	58.02 (19.91)	63.60 (16.93)	57.57 (24.41)	0.065	0.583	0.468	0.509	0.329
Expert	42.70 (12.77)	48.40 (22.23)	72.11 (24.03)	72.57 (23.51)					
·	Mean infraspin	atus (% MVIC)	, ,	,					
Novice	26.63 (10.32)	32.43 (11.69)	33.63 (5.74)	28.52 (11.19)	0.020	0.939	0.785	0.193	0.050
Expert	26.21 (6.02)	30.60 (9.47)	40.93 (11.15)	44.79 (12.68)					




Figure 2. Peak muscle activity levels during the propulsion tasks.

the start of the push phase has been identified as a potential cause of injury.⁴ Different propulsion styles have previously been examined, and the semi-circular style of propulsion has been advised to minimise the risk of injury.¹⁷ This study only measured trunk and

thoraco-humeral kinematics. In the future, it would be beneficial to examine full upper limb kinematics to analyse the association between propulsion styles, push rim kinetics and muscle activity level during over-ground propulsion.

Surface EMG

The results demonstrated a significant experience level by task interaction for both peak and mean muscle activity level of AD and PM and for mean muscle activity level of IS. For each muscle, during the level and cross slope tasks, the expert group demonstrated lower muscle activity level than the novice group, although there were not significant differences between the groups. The results differ to previous results, which reported significantly greater muscle activity in paraplegic versus able-bodied participants during level ergometer propulsion.8 These results may differ, as in this study the two groups travelled at the same velocity during the level and cross slope tasks, whereas in the previous study the paraplegic group travelled at a significantly greater velocity. For each muscle during the 6.5% and 12% incline tasks, the expert group demonstrated higher muscle activity levels than the novice group, significantly so for the AD during the 12% incline task. During the incline tasks, the expert group maintained a higher average velocity by applying a similar push rim moment at a greater angular velocity over a greater push arc. Pushing at faster speed has been shown to require higher levels of muscle activity level in both propulsive and recovery muscles. 18

Propulsion technique and injury risk

The guidelines for preservation of upper limb function following SCI suggest minimisation of task repetition and peak forces. The expert group demonstrated a propulsion technique that enabled completion of each of the propulsion tasks with fewer repetitions than the novices. During the less demanding tasks, this technique was also associated with lower peak muscle activity levels than the novices. During the more demanding incline tasks, this technique was associated with significantly higher power output and higher peak muscle activity level than the novices. This highlights the difficulty of informing optimal technique during overground wheelchair propulsion. Rotator cuff degeneration is the most common shoulder injury in manual wheelchair users. 1 Animal models have suggested that overuse is one of multiple factors involved in rotator cuff degeneration and injury, 19 and it is theorised that overload of the tendon can lead to micro trauma.²⁰ It is apparent that when modifying propulsion technique, the complex interaction between task repetition and muscle force requirement should be considered. In future research, it would be useful to calculate how altered propulsion technique influences joint contact forces, to further inform the optimal balance between repetition and peak force. Further investigation is also required to determine how expert wheelchair users are able to generate greater power during challenging tasks. A previous study, investigating ergometer propulsion at different speeds, reported a correlation between muscle strength and force imparted at the push rim.²¹ Further research to examine correlation between muscle strength and push rim parameters during challenging over-ground propulsion could be used to inform physical training for manual wheelchair users.

Real-time feedback for wheelchair propulsion training

Previous research has demonstrated the beneficial effect of real-time feedback on wheelchair propulsion biomechanics. During ergometer-based studies, both real-time visual feedback 22-26 and real-time haptic feedback 27 have been used to influence wheelchair propulsion biomechanics. A low-cost and lightweight tool such as the Sensewheel has potential to integrate with other systems to provide real-time visual, auditory or haptic feedback during daily functional propulsion tasks. Further research is required to determine whether real-time feedback could be used to train novice wheelchair users in more effective over-ground wheelchair propulsion technique, as demonstrated by the experts during this study.

Limitations

Although the sample size is small, statistically significant differences in propulsion technique between the experts and novices were identified. The expert user group only included paraplegic participants with SCI below T1 and it is highly likely that tetraplegic subjects would demonstrate significantly different technique.²⁸ The results presented can therefore only be applied to manual wheelchair users with full use of the upper limbs. The generalisability of the findings are also limited as the study only assessed one female and no older participants, who may have demonstrated differences in propulsion technique. The study only measured propulsion biomechanics on the left side. It would be beneficial to measure bilaterally, considering asymmetry in propulsion technique has been previously reported.²⁹ The upper limb kinematic analysis did not include elbow and wrist joint motion, and thoraco-humeral rather than gleno-humeral motion was measured and reported, which excludes the influence of differences in scapula motion.30

Conclusion

The purpose of this study was to identify experiencerelated biomechanical differences during over-ground manual wheelchair propulsion. The results demonstrated that expert users employed a propulsion technique during over-ground tasks requiring fewer pushes than

novices. During less challenging tasks, this technique was associated with reduced peak muscle activity levels than the technique used by the novices. During more challenging incline propulsion tasks, this technique was associated with greater muscle activity levels than the technique used by the novices. Further research is required to determine whether real-time feedback during over-ground propulsion could be used to improve propulsion technique in novice wheelchair users.

Acknowledgements

The authors would like to thank all of the study participants and also Nikos Papadosifos and Derrick Boampong for their valuable assistance in preparing the PAMELA Laboratory for the experiment.

Declaration of Conflicting Interests

The author(s) declared no potential conflicts of interest with respect to the research, authorship, and/or publication of this article.

Funding

The author(s) disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: This project formed part of the UCL Healthcare Engineering for an Ageing Population Studentship. This work was supported by The Engineering and Physical Sciences Research Council and Aspire.

References

- Akbar M, Balean G, Brunner M, et al. Prevalence of rotator cuff tear in paraplegic patients compared with controls. J Bone Joint Surg Am 2010; 92A(1): 23–30.
- 2. Akbar M, Brunner M, Balean G, et al. A cross-sectional study of demographic and morphologic features of rotator cuff disease in paraplegic patients. *J Shoulder Elbow Surg* 2011; 20(7): 1108–1113.
- Mercer JL, Boninger M, Koontz A, et al. Shoulder joint kinetics and pathology in manual wheelchair users. *Clin Biomech* 2006; 21(8): 781–789.
- 4. Boninger ML, Waters RL, Chase T, et al. Preservation of upper limb function following spinal cord injury: A clinical practice guideline for health-care professionals. *J Spinal Cord Med* 2005; 28(5): 433–470.
- Cowan RE, Boninger ML, Sawatzky BJ, et al. Preliminary outcomes of the SmartWheel Users' Group database: A proposed framework for clinicians to objectively evaluate manual wheelchair propulsion. *Arch Phys Med Rehabil* 2008; 89(2): 260–268.
- Rodgers MM, McQuade KJ, Rasch EK, et al. Upper-limb fatigue-related joint power shifts in experienced wheelchair users and nonwheelchair users. *J Rehabil Res Dev* 2003; 40(1): 27–37.
- Hwang S, Kim S, Son J, et al. Upper limb joint motion of two different user groups during manual wheelchair propulsion. *J Korean Phys Soc* 2013; 62(4): 648–656.

8. Louis N and Gorce P. Surface electromyography activity of upper limb muscle during wheelchair propulsion: Influence of wheelchair configuration. *Clin Biomech* 2010; 25(9): 879–885.

- Mason BS, Van der Woude LHV, Tolfrey K, et al. Effects of wheel and hand-rim size on submaximal propulsion in wheelchair athletes. *Med Sci Sports Exerc* 2012; 44(1): 126–134.
- Kobrick R, Carr C, Meyen F, et al. Using inertial measurement units for measuring spacesuit mobility and work envelope capability for intravehicular and extravehicular activities. In: 63rd International Astronautical Conference, Naples, Italy, 1 Oct-5 Oct 2012, paper no. IAC-12-A1.6.6.
- Hermens HJ, Freriks B, Disselhorst-Klug C, et al. Development of recommendations for SEMG sensors and sensor placement procedures. *J Electromyogr Kinesiol* 2000; 10(5): 361–374.
- Boettcher CE, Ginn KA and Cathers I. Standard maximum isometric voluntary contraction tests for normalizing shoulder muscle EMG. *J Orthop Res* 2008; 26(12): 1591–1597.
- 13. Sawatzky B, DiGiovine C, Berner T, et al. The need for updated clinical practice guidelines for preservation of upper extremities in manual wheelchair users a position paper. *Am J Phys Med Rehabil* 2015; 94(4): 313–324.
- 14. Gagnon D, Babineau A-C, Champagne A, et al. Trunk and shoulder kinematic and kinetic and electromyographic adaptations to slope increase during motorized treadmill propulsion among manual wheelchair users with a spinal cord injury. *Biomed Res Int*, Epub 2015. DOI: 10.1155/2015/636319.
- Chow JW, Millikan TA, Carlton LG, et al. Kinematic and electromyographic analysis of wheelchair propulsion on ramps of different slopes for young men with paraplegia. *Arch Phys Med Rehabil* 2009; 90(2): 271–278.
- Rodgers MM, Keyser RE, Gardner ER, et al. Influence of trunk flexion on biomechanics of wheelchair propulsion. J Rehabil Res Dev 2000; 37(3): 283–295.
- Koontz AM, Roche BM, Collinger JL, et al. Manual wheelchair propulsion patterns on natural surfaces during start-up propulsion. *Arch Phys Med Rehabil* 2009; 90(11): 1916–1923.
- 18. Qi L, Wakeling J, Grange S, et al. Changes in surface electromyography signals and kinetics associated with progression of fatigue at two speeds during wheelchair propulsion. *J Rehabil Res Dev* 2012; 49(1): 23–34.
- Soslowsky LJ, Thomopoulos S, Tun S, et al. Neer Award 1999 – Overuse activity injures the supraspinatus tendon in an animal model: A histologic and biomechanical study. J Shoulder Elbow Surg 2000; 9(2): 79–84.
- Nho SJ, Yadav H, Shindle MK, et al. Rotator cuff degeneration Etiology and pathogenesis. Am J Sports Med 2008; 36(5): 987–993.
- Ambrosio F, Boninger ML, Souza AL, et al. Biomechanics and strength of manual wheelchair users. J Spinal Cord Med 2005; 28(5): 407–414.
- 22. Rice IM, Pohlig RT, Gallagher JD, et al. Handrim wheelchair propulsion training effect on overground

- propulsion using biomechanical real-time visual feedback. Arch Phys Med Rehabil 2013; 94(2): 256–263.
- 23. de Groot S, Veeger HEJ, Hollander AP, et al. Consequence of feedback-based learning of an effective hand rim wheelchair force production on mechanical efficiency. *Clin Biomech* 2002; 17(3): 219–226.
- 24. DeGroot KK, Hollingsworth HH, Morgan KA, et al. The influence of verbal training and visual feedback on manual wheelchair propulsion. *Disabil Rehabil Assist Technol* 2009; 4(2): 86–94.
- Kotajarvi BR, Basford JR, An KN, et al. The effect of visual biofeedback on the propulsion effectiveness of experienced wheelchair users. *Arch Phys Med Rehabil* 2006; 87(4): 510–55.
- Richter WM, Kwarciak AM, Guo L, et al. Effects of single-variable biofeedback on wheelchair handrim biomechanics. Arch Phys Med Rehabil 2011; 92(4): 572–577.

- 27. Blouin M, Lalumiere M, Gagnon DH, et al. Characterization of the immediate effect of a training session on a manual wheelchair simulator with haptic biofeedback: Towards more effective propulsion. *IEEE Trans Neural Syst Rehabil Eng* 2015; 23(1): 104–115.
- 28. Newsam CJ, Mulroy SJ, Gronley JK, et al. Temporal-spatial characteristics of wheelchair propulsion Effects of level of spinal cord injury, terrain, and propulsion rate. *Am J Phys Med Rehabil* 1996; 75(4): 292–299.
- 29. Hurd WJ, Morrow MM, Kaufman KR, et al. Biomechanic evaluation of upper-extremity symmetry during manual wheelchair propulsion over varied terrain. *Arch Phys Med Rehabil* 2008; 89(10): 1996–2002.
- Raina S, McNitt-Gray JL, Mulroy S, et al. Effect of increased load on scapular kinematics during manual wheelchair propulsion in individuals with paraplegia and tetraplegia. *Hum Mov Sci* 2012; 31(2): 397–407.