



## University of Dundee

### Active MRI tracking for robotic assisted FUS

Xiao, Xu; Huang, Zhihong; Melzer, Andreas

*Published in:*

Proceedings from the 13th International Symposium on Therapeutic Ultrasound

*DOI:*

[10.1063/1.4976607](https://doi.org/10.1063/1.4976607)

*Publication date:*

2017

*Document Version*

Publisher's PDF, also known as Version of record

[Link to publication in Discovery Research Portal](#)

*Citation for published version (APA):*

Xiao, X., Huang, Z., & Melzer, A. (2017). Active MRI tracking for robotic assisted FUS. In Proceedings from the 13th International Symposium on Therapeutic Ultrasound (Vol. 1816). [080003] AIP Publishing.  
<https://doi.org/10.1063/1.4976607>

#### General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

#### Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

# Active MRI tracking for robotic assisted FUS

Cite as: AIP Conference Proceedings **1816**, 080003 (2017); <https://doi.org/10.1063/1.4976607>  
Published Online: 03 March 2017

Xu Xiao, Zhihong Huang, and Andreas Melzer



View Online



Export Citation

## ARTICLES YOU MAY BE INTERESTED IN

[Motion tracing system for ultrasound guided HIFU](#)

AIP Conference Proceedings **1816**, 030001 (2017); <https://doi.org/10.1063/1.4976593>

[Ultrasonography-based motion tracking for MRgFUS](#)

AIP Conference Proceedings **1816**, 050001 (2017); <https://doi.org/10.1063/1.4976599>

[Full acoustic and thermal characterization of HIFU field in the presence of a ribcage model](#)

AIP Conference Proceedings **1816**, 020002 (2017); <https://doi.org/10.1063/1.4976592>

This article may be downloaded for personal use only. Any other use requires prior permission of the author and AIP Publishing. Xiao, X, Huang, Z & Melzer, A 2017, Active MRI tracking for robotic assisted FUS. in Proceedings from the 13th International Symposium on Therapeutic Ultrasound. vol. 1816, 080003, AIP Publishing, 13th International Symposium on Therapeutic Ultrasound, ISTU 2013, Shanghai, China, 12/05/13. <https://doi.org/10.1063/1.4976607>

**AIP** | Conference Proceedings

Get **30% off** all  
print proceedings!

Enter Promotion Code **PDF30** at checkout



# Active MRI Tracking For Robotic Assisted FUS

Xu Xiao<sup>1, 2, a</sup>, Zhihong Huang<sup>1, b</sup>, Andreas Melzer<sup>2, c</sup>

<sup>1</sup>*Bioinstrumentation Research Center, University of Dundee, UK*

<sup>2</sup>*Institute for Medical Science and Technology IMSaT, University of Dundee, Dundee, UK*

<sup>a</sup>Corresponding author: x.xiao@dundee.ac.uk

<sup>b</sup>z.y.huang@dundee.ac.uk

<sup>c</sup>a.melzer@dundee.ac.uk

**Abstract:** MR guided FUS is a noninvasive method producing thermal necrosis at the position of tumors with high accuracy and temperature control. Because the typical size of the ultrasound focus is smaller than the area of interested treatment tissues, focus repositioning become necessary to achieve multiple sonications to cover the whole targeted area. Using MR compatible mechanical actuators could help the ultrasound beam to reach a wider treatment range than using electrical beam steering technique and more flexibility in position the transducer. An active MR tracking technique was combined into the MRgFUS system to help locating the position of the mechanical actuator and the FUS transducer. For this study, a precise agar reference model was designed and fabricated to test the performance of the active tracking technique when it was used on the MR-compatible robotics InnoMotion™ (IBSMM, Engineering spol. s r.o. / Ltd, Czech Republic). The precision, tracking range and positioning speed of the combined robotic FUS system were evaluated in this study. Compared to the existing MR guided HIFU systems, the combined robotic system with active tracking techniques provides a potential that allows the FUS treatment to operate in a larger spatial range and with a faster speed, which is one of the main challenges for organ motion tracking.

## I. INTRODUCTION

Focused ultrasound surgery (FUS) has been proved to be a precise and non-invasive modality for tissue ablation in experimental and clinical studies. Focus precision and energy efficiency of the FUS therapy highly benefit from guidance and monitoring by magnetic resonance (MR) imaging<sup>1</sup>.

As the typical spatial volumes of the ultrasound focus are smaller than the target volume, a complete HIFU intervention requires multiple sonications and focus repositioning<sup>2</sup>. Such focal spot scanning techniques can be realized via mechanical actuators. We propose an MRgFUS system that combines an US transducer with a fixed focal length and a robotic assistance system to steer the US beam. The MRI active tracking technique is integrated with the robotic FUS system and the MRI to enable fast systems localization.


This combined approach would provide a wide range of spatial flexibility for positioning of the high power ultrasound transducer within the magnet bore. The combined robotic arm assisted FUS setup might offer additional acoustic treatment pathways which could, for example, avoid bone structures during sonication without taking the patients off from the scanner. In our previous work<sup>3</sup>, an active tracking technique was applied on the robotic arm to precisely detect the position of the system. The tracking accuracy was tested without considering the system error caused by the robotic arm itself. Here in this paper, a new agar phantom structure is designed to be as the external reference model. Furthermore, a MRI compatible web camera is fixed in the MRI suite to monitor the relative

position relationship between the robotic arm and the reference model to make sure the robot arm could move to the right position. A higher tracking accuracy result has been achieved with technique described in this paper.

## II. METHODS

The Mechanical device used for moving US transducer is a MRI-compatible robotic arm (IBSMM Engineering spol. S r.o./Ltd., Czech Republic) (FIGURE. 1a). This robotic system is driven pneumatically in 5 degrees of freedom with a translational and an angular positioning accuracy of  $\pm 1$  mm and  $\pm 1^\circ$ , respectively<sup>4</sup>.


During the motion of the robotic arm, a way of detecting the coordinates of the robot within the MR core, which is named active tracking is used in this work. The tracking technique is based on a fundamental principle of MR imaging that the magnetic field<sup>5</sup>. To realize the tracking, a tracking plate made of PVA (FIGURE 1b) is designed on which there are 4 trackers, each of them contains a volume of fluid in a dedicated coil provided by InSightec Ltd. (Tirat Carmel, Israel). The size of every coil is about 2mm in diameter and 5mm in length. This device allows mechanical connection between the robotic arm and the FUS transducer.



**FIGURE 1.**(a) View of an Innomotion robotic arm guided MR intervention (GE 1.5 Tesla HDx) (b) MR tracking device with 4 active micro coils (Insightec)


To prove the efficiency of active tracking method, a calibration experiment is built up which can test the tracking accuracy when using micro-coils to get the positions of the robotic arm. To this end, a 3D reference model is designed to provide reference points for the robotic arm.

These reference points are precisely located on the reference frame (FIGURE 2a) which is made of 5% proportion agar. The model is 140\*140\*140mm in dimension which is much larger than normal beam steering space of phased array transducers. If the robotic arm can precisely locate a HIFU transducer in this wide space volume, a conclusion can be made that the mechanical guiding method is worth for HIFU applications in which the ultrasound beam has to steer in a wide range or the beam has to be adjusted to avoid interference rib bone. The 35 reference holes are located uniformly on seven different layers, the height distance between two adjacent layers is 20mm and the lateral distance between two adjacent holes is 30mm.



**FIGURE 2.** (a) Agar phantom for providing reference points; (b) Sagittal and axial MR scan (imaging data) of agar phantom and Gd (1 Gd: 100 water) filled head of application part

In the calibration experiment, a needle shaped application part (FIGURE. 2) was fixed at the center of the tracking device. The head of the application part was filled with Gadolinium (Gd) with a ratio to gelatin of 1:50. This application part will appear brighter than the agar phantom, which can help the observer to confirm the position of the robotic arm (FIGURE 2b).



**FIGURE 3.** (a) Robotic arm guides the application part inserting into the reference holes of the agar phantom; (b) Using external camera to monitor the position of the robotic arm in the MR scanner ( GE 1.5 Tesla HDx, DuoFlex Interventional coil by MRInstruments)

The robotic arm guides the needle to align with the reference points of the phantom (FIGURE. 3). At every reference position, the robotic arm position could be computed from the coordinate readout of the four micro coils. Comparing the computed coordinates with the reference coordinates, the tracking reliability of the micro coils methods could be tested.

### III. RESULTS

The whole tracking area is defined as the robotic arm movement range which is 180 mm in axial direction, 160mm in sagittal direction and 180mm in coronal direction. The coordinates of the micro coils in the MRI frame are recorded in less than 0.02 second. The computed in plane center of the four micro coils represents the position of the robotic arm. By comparing the position with the reference data, the tracking accuracy results have been calculated.

Table 1 is the summary of the tracking error in the whole range. In the more homogeneous magnetic area, the tracking error is smaller than 0.41mm, almost one third of the error in the area far from the iso-centre of the MR scanner. The error (in the coronal direction is larger than in the other two directions, which might come from two reasons. Firstly, it is more difficult to ensure the application part to touch the bottom of the reference holes of the phantom. Second, during the long period of tracking experiment, the agar and water based phantom shrinks most in the coronal direction because of its self-weight.

**TABLE 1** Summary of the tracking result in the area close to the iso-centre of the MR

Direction & Distance to isocenter	Error (mm)	Direction & Distance to isocenter	Error (mm)
Axial, <120mm	30.03±0.28	Axial, ≥120, ≤180mm	30.03±0.60
Sagittal, <120mm	20.04±0.28	Sagittal, ≥120, ≤160mm	20.04±0.66
Coronal, <140mm	19.76±0.41	Coronal, ≥120, ≤180mm	19.60±1.30

### IV. DISCUSSION AND CONCLUSION

After replace the needle shaped application part with a HIFU transducer, the system supposed to be able to guide the transducer to steer the US beam in a wide range. The combination of a commercial robotic arm and a customized

active tracking device was proposed as a new type of experimental setup for validation of future MRgFUS therapy. The newly designed agar phantom modal provide more reliable calibration result for the tracking technique.

Compared to existing MR guided HIFU systems, the combined robotic HIFU with the active tracking technique facilitates the HIFU treatment to operate in a larger spatial range (at least 14×14×14 cm) and at faster speed. This MR compatible active tracking device should provide a significant reduction the treatment time as it obviates the need for time consuming manual repositioning.

## V. ACKNOWLEDGEMENT

The authors gratefully acknowledge Martin Rube (IMSaT, University of Dundee, Dundee, UK) for his assistance of programming of tracking sequence and Tingyi Jiang for his design of phantom box. This work is supported by CSC (China Scholarship Council, China). The robotic arm was supported by IBSMM (IBSMM Engineering spol. S r.o./Ltd., Czech Republic).

The DouFlex coild was granted free for the experiment by MR Instruments (MR Instruments, Hopkins, MN, USA)

InSightec kindly provided the micro coils and Alexander Volovick supported the the implementation of the tracking.

The authors acknowledge support by European Commission FP7 Program under grant agreement Nr 238802 (IIIOS) and 230674 (NANOPORATION)

## VI. REFERENCES

1. G. ter Haar and C. Coussios, High intensity focused ultrasound: Physical principles and devices, [Int. J. Hyperthermia](#), 2007, pp. 89-104
2. Chen L, Bouley D, Harris B, Butts K., MRI study of immediate cell viability in focused ultrasound lesions in the rabbit brain. [Journal of Magnetic Resonance Imaging](#). 2001, 13:23-30,.
3. Xu Xiao, Zhihong Huang, Alexander Volovick, and Andreas Melzer, Robotic active positioning for magnetic resonance-guided high-intensity focused ultrasound, [AIP Conf. Proc.](#) 1503, 217 (2012);
4. A. Melzer, B. Gutmann, T. Remmele, R. Wolf, A. Lukoscheck, M. Bock, H. Bardenheuer, and H. Fischer, “INNOMOTION for percutaneous image-guided interventions: Principles and evaluation of this MR- and CT-compatible robotic system”, [IEEE Eng. Med. Biol. Mag.](#), 2008, pp. 66-73
5. Dumoulin, C. L., Souza, S. P. and Darrow, R. D. (1993), Real-time position monitoring of invasive devices using magnetic resonance. [Magnetic Resonance in Medicine](#), 29: 411–415.