

Modified limb lead system: Its effects on wave amplitudes and axis in surface ECG

Modified limb lead (MLL) electrocardiogram (ECG) system may be used during rest or exercise ECG test, or in atrial activity enhancement. Due to modifications to limb electrode placement, changes are likely to happen in ECG wave amplitudes and frontal plane axis, which may alter the clinical limits of normality and ECG diagnostic criteria (1). There are also several other alternative lead systems that are placed on the human torso to record and study the electrical activity of the atria (2–4). A few recent studies have proposed modification to the standard 12-lead ECG system of placing the limb electrodes closer to the atria in order to enhance atrial ECG components (5–7).

Sivaraman et al. (3) recently proposed a MLL system. In their subsequent study, they reported on the normal limits of the MLL system and documented the changes in P wave amplitudes and frontal plane P wave axis (8). It was found that P wave amplitude increased in all the modified leads compared with the standard leads, which led the MLL system to be seen as optimal lead system to study the electrical activity of atrial ECG components. Seen in the light of their previous findings using MLL system, observing the changes that occurred in ECG wave amplitudes (R, S, T), ST segment amplitudes (STa), and frontal plane axis shift due to the MLL system as described in the paper of Sivaraman et al. (9) published in this issue of the Anatolian Journal of Cardiology was of interest. In this study, they also examined the magnitude of ECG wave amplitudes and STa differences between standard limb lead (SLL) and MLL ECG systems, which may lead to deviations beyond clinical limits of normality. The observational study included 60 male patients with sinus rhythm of mean age 38.85 ± 8.76 (SD) years, with range of 25 to 58 years. The authors comprehensively measured and evaluated STa from the J point to 80 milliseconds after the onset of the J point (J+80ms). STa deviations were analyzed for 20 milliseconds after the onset of the J point to validate whether the MLL system had any effect on the ST segments (Table 3 of the article). In general, ST segment elevation greater than $100 \mu\text{V}$ is defined as clinical threshold level in the frontal plane (10). The documented STa values from this study are of importance in understanding the effects of the MLL system on STa deviations.

As the authors (9) themselves pointed out, the QRS and T axis measurements were generally more affected when the electrodes were moved from the limb to the torso of the patients, and concluded that the deviations in the frontal plane meant QRS axis had an effect on clinical specificity. This is due to the

fact that MLL system records the ECG in a principally different way from the standard 12-lead ECG. In summary, this study on MLL ECG is useful in observing changes in ECG amplitudes and frontal plane axis shift and helps the clinicians/researchers to come to the conclusion that this MLL ECG may not be used to diagnose with SLL threshold values.

Vimal Prabhu Pandiyan

Department of Biomedical Engineering, Indian Institute of Technology, Hyderabad; Telangana-India

References

1. Jowett N, Turner AM, Cole A, Jones P. Modified electrode placement must be recorded when performing 12-lead electrocardiograms. *Postgrad Med J* 2005; 81: 122-5. [Crossref](#)
2. Szekely P. Chest leads for the demonstration of auricular activity. *Br Heart J* 1944; 6: 238-46. [Crossref](#)
3. Sivaraman J, Uma G, Venkatesan S, Umapathy M, Dhandapani VE. A novel approach to determine atrial repolarization in electrocardiograms. *J Electrocardiol* 2013; 46: e1. [Crossref](#)
4. Sivaraman J, Uma G, Venkatesan S, Umapathy M, Ravi MS. Unmasking of atrial repolarization waves using a simple modified limb lead system. *Anatol J Cardiol* 2015; 15: 605-10. [Crossref](#)
5. Kennedy A, Finlay DD, Guldenring D, Bond RR, McLaughlin J. Detecting the elusive P-wave: a new ECG lead to improve the recording of atrial activity. *IEEE Trans Biomed Eng* 2016; 63: 243-9. [Crossref](#)
6. Petrenas A, Marozas V, Jarusevicius G, Sornmo L. A modified Lewis ECG lead system for ambulatory monitoring of atrial arrhythmias. *J Electrocardiol* 2015; 48: 157-63. [Crossref](#)
7. Sivaraman J, Uma G, Venkatesan S, Umapathy M, Keshav Kumar N. A study on atrial Ta wave morphology in healthy subjects: An approach using P wave signal-averaging method. *J Med Imaging Health Inf* 2014; 4: 675-80. [Crossref](#)
8. Sivaraman J, Uma G, Venkatesan S, Umapathy M, Dhandapani VE. Normal limits of ECG measurements related to atrial activity using a modified limb lead system. *Anatol J Cardiol* 2015; 15: 2-6. [Crossref](#)
9. Sivaraman J, Venkatesan S, Periyasamy R, Justin Joseph, Ravi MS. Modified limb lead ECG system effects on electrocardiographic wave amplitudes and frontal plane axis in sinus rhythm subjects. *Anatol J Cardiol* 2017; 17: 46-54.
10. Sheppard JP, Barker TA, Ranasinghe AM, Clutton-Brock TH, Frenneaux MP, Parkes MJ. Does modifying electrode placement of the 12 lead ECG matter in healthy subjects? *Int J Cardiol* 2011; 152: 184-91. [Crossref](#)

Address for correspondence: Vimal Prabhu Pandiyan, M.E., Ph.D, Department of Biomedical Engineering
Indian Institute of Technology, Hyderabad
Kandi, sangareddy-502285, Telangana-India
E-mail: vimalbme@gmail.com

Accepted Date: 11.05.2016

©Copyright 2017 by Turkish Society of Cardiology - Available online at www.anatoljcardiol.com
DOI:10.14744/AnatolJCardiol.2016.21995

