

Lab-on-a-Tip (LOT): Where Nanotechnology can Revolutionize Fibre Optics

Perspective Article

Qimin Quan^{1*} and Yiying Zhang^{2*}

¹ Rowland Institute at Harvard University, Cambridge, MA, USA

² Geriatric Anesthesia Research Unit, Department of Anesthesia, Critical Care and Pain Medicine, Massachusetts General Hospital and Harvard Medical School, Charlestown, MA, USA

* Corresponding author(s) E-mail: quan@rowland.harvard.edu, yizhang37@mgh.harvard.edu

Received 28 October 2014; Accepted 02 March 2015

DOI: 10.5772/60518

© 2015 The Author(s). Licensee InTech. This is an open access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Abstract

Recently developed lab-on-a-chip technologies integrate multiple traditional assays on a single chip with higher sensitivity, faster assay time, and more streamlined sample operation. We discuss the prospects of the lab-on-a-tip platform, where assays can be integrated on a miniaturized tip for *in situ* and *in vivo* analysis. It will resolve some of the limitations of available lab-on-a-chip platforms and enable next generation multifunctional *in vivo* sensors, as well as analytical techniques at the single cell or even sub-cellular levels.

Keywords Lab-on-a-tip (LOT), Nanotechnology, Biosensor, Endoscopy, Single cell analysis

1. Introduction

Biochemical and immunological assays have been the major analytical methods in biological research, clinical diagnostics, and the drug and food industries. Recently developed lab-on-a-chip (LOC) technologies brought the traditional assays to a higher level of integration, leading to more streamlined processing. They also yielded im-

proved sensitivity and faster assay time by speeding up the chemical reaction and physical processes (thermo-cycling and diffusion) in a confined miniaturized space. As extensive research has already been focused on LOC platforms [1 - 4], this perspective aims to present the prospects for a complementary platform – lab-on-a-tip (LOT) – that integrates one or several laboratory functions on a single miniaturized tip. Although the assays could be performed by optical, electrical, or mechanical means, we will focus our discussions on optical measurements in this article. Early examples of LOT platforms include fibre optic intracranial pressure sensors, thermometers, and endoscopes. However, LOT devices are at an early stage – devices are typically mono-functional, fabrication methods are not standardized, and most devices are macroscopic. We envision more comprehensive lab functions to be integrated on LOT (Fig. 1). In particular, devices will measure multiple physical and chemical properties (parallel integration), and will integrate imaging, diagnostics, and therapeutic functions (vertical integration). We also anticipate novel and scalable LOT fabrication techniques will be developed and standardized, and further development of nano-LOT devices will enable new research at single cell or sub-cellular levels.

2. Multifunctional LOT devices

The great success of fibre optic technology in the telecommunication industry led to the development of their applications in the biomedical and pharmaceutical industries. Optical fibres offer natural "lab-on-a-tip" operation, and were extensively developed during 1980s and 1990s [5]. The fibre optic intracranial pressure sensor is one of the earliest developed LOT devices. It is used in continuously monitoring the intracranial pressure of head trauma patients, and works by sending light to the tip of the fibre where it is reflected back from a diaphragm that deflects with external pressure. Non-optical piezo-resistive or capacitive transducers also exist, but they suffer from the risk of electrical-shock hazard and long-term drifts, while fibre optic intracranial pressure sensors are electrical shock hazard free. They are also small, flexible, and have long-term reliability. First commercialized by Camino Labs in the 1980s, fibre optic intracranial pressure sensors are now widely used in diagnosing potential elevation of intracranial pressure and monitoring its progression in the clinical neurological examination.

Another LOT success is the fibre optic thermometer used in hyperthermia therapy, which is used to kill cancer cells or to make cancer cells more sensitive to radiation therapy by exposing body tissue to a slightly higher temperature. The temperature of the tissue is controlled by the application of microwave or radiofrequency radiation, and must be precisely monitored. Fibre optic thermometers offer the unique advantages of natural electrical isolation and freedom from electromagnetic interference. Different measurement modalities exist [6], the most well-known ones are temperature sensitive fluorescent phosphors and GaAs crystals patterned on the fibre tip that provides spectrum features sensitive to the temperature. It was first commercialized by Luxtron Inc. in the 1980s, and subsequently developed by many other companies.

Fibre optic biosensors were also used to measure physical parameters such as pH, blood flow rate, blood oxygen levels, radiation dosage, and biting force in dentistry. New materials, especially metals and semiconductors have been integrated into glass or polymer fibres to achieve optoelectronic properties [7]. Composite geometries, such as photonic crystal fibres [8] and hollow core Bragg fibres [9], have been developed. They have the advantage of guiding light in their hollow cores, thus allowing high power applications (e.g., laser surgery [10]).

One future development is likely to be the multifunctional fibre optic sensors. An interesting analogy can be made between fibre sensors in the medical industry and the drills in the oil and gas industry. A large number of sensors are equipped along the drill and perform comprehensive measurement while drilling, including depth, pressure, gas composition, resistivity, porosity, gamma ray, drilling system orientation, and wellbore geometry [11]. We envision more functions will be integrated into fibre optics and catheters, and thus allowing for more powerful and comprehensive monitoring during diagnostic or therapeutic

operations. Since fibre optic sensors have much smaller dimensions than drills, more powerful and scalable techniques will need to be developed. Nanofabrication techniques are already well developed for planarized surfaces (such as chips and wafers), but much less so for non-traditional surfaces. Several strategies are currently being explored, including post-transfer chips onto fibres [12, 13], and direct fabrication on fibre facets by focused ion beam (FIB) [14] and electron beam lithography (ebeam) [15]. Traditional fibre drawing techniques are also extended to new materials, including metals, semiconductors, and polymers [16].

3. Next generation of endoscopy

Endoscopy allows diagnostics inside the human body by bringing an optical system in proximity to the tissues of interest. It also revolutionized surgery, allowing minimally invasive procedures to be performed inside the body. Philipp Bozzini (1773–1809) was credited with the invention of the first endoscope about 200 years ago, and Max Nitze (1849–1906) was among the first to take endoscopic operations inside the bladder. Since Basil Hirschowitz (1925–2013) developed the first fibre optic endoscope, many improvements were made, including using charge-coupled devices (CCD) for imaging and rod-lens systems to improve the quality of imaging. In 1997, optical coherence tomography (OCT) endoscopy demonstrated improvement of imaging resolution from the previous 100 μm to 10 μm with large penetration depth [17]. Subsequently, two-photon and multi-photon imaging (MPM) [18, 19] further improved the imaging resolution ($\sim 3 \mu\text{m}$), but they have inferior penetration depth compared to OCT in general. Optical frequency-domain imaging (OFDI) [20] and more recently Doppler OFDI [21] improved the imaging speed by several orders of magnitude over OCT. Its deep penetration made it possible to image significantly larger tumour volume than MPM. Furthermore, scanning fibre wide-field fluorescence imaging [22] provides information about fluorescent labelled molecules or auto-fluorescent proteins. Coherent Raman scanning fibre endoscopy [23] provides some chemical information, such as CH_2 or CH_3 bonds, without labels. Efforts to combine multiple imaging modalities that reveal both morphology and molecular information are currently being actively pursued. For example, recently developed OCT and fluorescent imaging [24], OFDI and near-infrared fluorescence imaging [25], provide both imaging and molecular information.

For more than a hundred years, endoscopists have relied on the visualization of the morphology of tissue with subsequent biopsy procedure for immunological and molecular diagnostics. The next generation endoscopy will extend its capability from imaging to diagnostics and therapy, which we call vertical integration. Fluorescent labelling [24, 25] or other label-free biosensors can be integrated to the endoscope and bring immunological and molecular diagnostics *in situ*. Therapeutic capabilities can be integrated as well. For example, gold nanoparticle

photothermal therapy with near-infrared light [26] or visible laser [27] has been demonstrated in mice model. However, more controlled methods to provide irradiation and alternative methods to enhance the *in vivo* delivery are required [28]. The next generation endoscopy will integrate high-resolution three-dimensional imaging, immunological and molecular diagnostics, and subsequently photothermal therapy with targeted delivery of nanoparticles on the malignant cells.

4. Nano-LOT devices for single cell analysis

Most of our knowledge in biology so far has been based on ensemble measurements on cells. Patch clamp technology made it possible to record the currents of single ion channels for the first time [29], for which Erwin Neher and Bert Sakmann shared the Nobel Prize in Physiology or Medicine in 1991. More recent development of nanoscale patch clamp techniques made it possible to record single ion channels at specific locations such as the presynaptic terminal, when used in combination with super-resolution (below optical diffraction limit) imaging techniques [30]. Electrophysiology measurement is also achieved with less invasive systems, such as Pt nanowires pulled in glass capillaries [31], carbon coated nanopipes [32], and free-standing kinked nanowire transistor nanoprobes [33].

Other physical and chemical parameters of single cells can also be measured with optical methods. In 1992, Tan et al. fabricated a sub-micron fibre tip which was coated with pH-sensitive fluorescent polymers to detect pH levels in a

rat embryo [34]. Vo-Dinh et al. further miniaturized the tip and demonstrated detection of fluorescent molecules inside single cells [35, 36]. Calcium level [37] and lactate concentration [38] measurements were also demonstrated by using fluorescent dyes coated on the fibre nanotips. Raman spectra were measured inside single cells by coating gold nanoparticles on the glass capillary tip in order to enhance the Raman signals [39]. Recently, Hong et al. measured the dynamics of intracellular p53 proteins using a nano-fibre tip coated with a single gold nanoparticle [40], extending its capability to detect non-fluorescent proteins in single live cells. Other materials, for example carbon nanotubes, have been used to deliver quantum dots into cells [41], as well as to study the force and indentation depth during cell penetration [42]. Boron nitride nanotubes were also used to deliver quantum dots [43]. When the process of inserting these nanotips into cells was studied using confocal imaging [44], it was found that the dimension of the tip is of critical importance: sub-100 nm dimension will not damage the cell membrane or disturb the cellular energetic system [45]. Nano-LOT devices offer opportunities to study single cells and their biochemical compositions at high spatial resolution, as well as their temporal response to external stimuli in their natural living conditions [46]-[48]. We envision that versatile nano-LOT devices will be proposed and developed, and that more delicate structures and sophisticated mechanisms will be invented for precise monitoring and controlling at the single cell level.




Figure 1. Illustration of integrative LOT devices. I. Parallel integration. An optical fibre LOT that comprises multiple lab functions. Fluorescence detection channel (1), carbon dioxide laser surgery channel (2), white light illumination channel (3), and pressure monitoring channel with a deformable membrane (4) are present on the fibre tip. Other channels, e.g., temperature, pH, oxygen level, etc., can also be integrated along the fibre (5). II. Vertical integration. An endoscope LOT that has two parallel channels: one for endoscopic imaging (1) and the other for molecular diagnostics and therapeutics (2). In the imaging channel, lenses (3) and gratings (or microelectromechanical systems) can be used to steer the light. As suspicious malignant tissues (5) are identified, biologically functional nanosensors (represented by yellow dots) can be injected from the diagnostic/therapeutic channel (2) to perform molecular diagnostics, followed by subsequent *in situ* therapeutic operations (e.g., using gold nanoparticle photothermal therapy). III. Nanoscale LOT device. Nanosensors can be integrated to LOT, and perform analysis at sub-cellular components inside the living cells: e.g., the optical fibre can be fabricated down to a nanoscale tip (1) and nanosensors (2) can be integrated on its tip to perform nucleic acid or protein detection in the nucleus of the cell (3).

5. Summary

We reviewed the early development of LOT devices, including fibre optic sensors, endoscopes, and more recently developed nanoscale bioprobes. With the rapid

advance of nanotechnology, LOT devices will become multi-functional, will integrate imaging, diagnostics, and therapeutic capabilities, and will become more powerful in analysing smaller biological components such as single cells. LOT platforms will be a great complement to LOC

platforms, and together they will offer comprehensive analysis and control of biological systems both *in vitro* and *in vivo*.

6. Conflict of Interest

The authors declare that there are no conflict of interest.

7. Acknowledgements

The authors thank helpful discussions with Michael Burns and Howard Berg at Rowland Institute at Harvard University.

8. References

- [1] Whitesides, G. M. The origins and the future of microfluidics. *Nature* 442, 368–373, doi:10.1038/nature05058 (2006).
- [2] Haeberle, S. and Zengerle, R. Microfluidic platforms for lab-on-a-chip applications. *Lab on a chip* 7, 1094–1110, doi:10.1039/b706364b (2007).
- [3] Mark, D., Haeberle, S., Roth, G., von Stetten, F. and Zengerle, R. Microfluidic lab-on-a-chip platforms: requirements, characteristics and applications. *Chemical Society Reviews* 39, 1153–1182, doi:10.1039/b820557b (2010).
- [4] Whitesides, G. The lab finally comes to the chip! *Lab on a chip* 14, 3125–3126, doi:10.1039/c4lc90072c (2014).
- [5] Mignani, A. G. and Baldini, F. Biomedical sensors using optical fibres. *Rep Prog Phys* 59, 1–28, doi:10.1088/0034-4885/59/1/001 (1996).
- [6] Rubio, M. F. J. C., Hernández, A. V., Salas, L. L. *Fiber Optics for Thermometry in Hyperthermia Therapy*. (2012).
- [7] Abouraddy, A. F. et al. Towards multimaterial multifunctional fibres that see, hear, sense and communicate. *Nature Materials* 6, 336–347, doi:10.1038/nmat1889 (2007).
- [8] Russell, P. Photonic crystal fibers. *Science* 299, 358–362, doi:10.1126/science.1079280 (2003).
- [9] Temelkuran, B., Hart, S. D., Benoit, G., Joannopoulos, J. D. and Fink, Y. Wavelength-scalable hollow optical fibres with large photonic bandgaps for CO₂ laser transmission. *Nature* 420, 650–653, doi:10.1038/nature01275 (2002).
- [10] Ryan, R. W. et al. Application of a flexible CO₂ laser fiber for neurosurgery: laser-tissue interactions Laboratory investigation. *J Neurosurg* 112, 434–443, doi:10.3171/2009.7.Jns09356 (2010).
- [11] Dowell, I., Mills, A., Ridgway, M., Lora, M. *Petroleum Engineering Handbook, Volume II – Drilling Engineering, Chapter 15* (ed R. F. Mitchell) (2006).
- [12] Smythe, E. J., Dickey, M. D., Whitesides, G. M. and Capasso, F. A technique to transfer metallic nano-scale patterns to small and non-planar surfaces. *ACS Nano* 3, 59–65, doi:10.1021/nn800720r (2009).
- [13] Lipomi, D. J. et al. Patterning the tips of optical fibers with metallic nanostructures using nanoskiving. *Nano Lett* 11, 632–636, doi:10.1021/nl103730g (2011).
- [14] Lin, Y. B., Zou, Y. and Lindquist, R. G. A reflection-based localized surface plasmon resonance fiber-optic probe for biochemical sensing. *Biomed Opt Express* 2, 478–484, doi:10.1364/Boe.2.000478 (2011).
- [15] Consales, M. et al. Lab-on-fiber technology: toward multifunctional optical nanoprobe. *ACS nano* 6, 3163–3170, doi:10.1021/nn204953e (2012).
- [16] Orf, N. D. et al. Fiber draw synthesis. *P Natl Acad Sci USA* 108, 4743–4747, doi:10.1073/pnas.1101160108 (2011).
- [17] Tearney, G. J. et al. *In vivo* endoscopic optical biopsy with optical coherence tomography. *Science* 276, 2037–2039 (1997).
- [18] Brown, E. B. et al. *In vivo* measurement of gene expression, angiogenesis and physiological function in tumors using multiphoton laser scanning microscopy. *Nature Medicine* 7, 864–868, doi:10.1038/89997 (2001).
- [19] Bird, D. and Gu, M. Two-photon fluorescence endoscopy with a micro-optic scanning head. *Opt Lett* 28, 1552–1554 (2003).
- [20] Yun, S. H. et al. Comprehensive volumetric optical microscopy *in vivo*. *Nature medicine* 12, 1429–1433, doi:10.1038/nm1450 (2006).
- [21] Vakoc, B. J. et al. Three-dimensional microscopy of the tumor microenvironment *in vivo* using optical frequency domain imaging. *Nature medicine* 15, 1219–1223, doi:10.1038/nm.1971 (2009).
- [22] Lee, C. M., Engelbrecht, C. J., Soper, T. D., Helmchen, F. and Seibel, E. J. Scanning fiber endoscopy with highly flexible, 1 mm catheterscopes for wide-field, full-color imaging. *J Biophotonics* 3, 385–407, doi:10.1002/jbio.200900087 (2010).
- [23] Saar, B. G., Johnston, R. S., Freudiger, C. W., Xie, X. S. and Seibel, E. J. Coherent Raman scanning fiber endoscopy. *Opt Lett* 36, 2396–2398 (2011).
- [24] Mavadia, J., Xi, J. F., Chen, Y. P. and Li, X. D. An all-fiber-optic endoscopy platform for simultaneous OCT and fluorescence imaging. *Biomed Opt Express* 3, 2851–2859 (2012).
- [25] Yoo, H. et al. Intra-arterial catheter for simultaneous microstructural and molecular imaging *in vivo*. *Nature medicine* 17, 1680–1684, doi:10.1038/nm.2555 (2011).
- [26] Hirsch, L. R. et al. Nanoshell-mediated near-infrared thermal therapy of tumors under magnetic resonance guidance. *P Natl Acad Sci USA* 100, 13549–13554, doi:10.1073/pnas.2232479100 (2003).
- [27] El-Sayed, I. H., Huang, X. H. and El-Sayed, M. A. Selective laser photo-thermal therapy of epithelial

- carcinoma using anti-EGFR antibody conjugated gold nanoparticles. *Cancer Lett* 239, 129–135, doi: 10.1016/j.canlet.2005.07.035 (2006).
- [28] Kennedy, L. C. et al. A New Era for Cancer Treatment: Gold-Nanoparticle-Mediated Thermal Therapies. *Small* 7, 169–183, doi: 10.1002/smll.201000134 (2011).
- [29] Neher, E. and Sakmann, B. Single-channel currents recorded from membrane of denervated frog muscle fibres. *Nature* 260, 799–802 (1976).
- [30] Novak, P. et al. Nanoscale-targeted patch-clamp recordings of functional presynaptic ion channels. *Neuron* 79, 1067–1077, doi:10.1016/j.neuron.2013.07.012 (2013).
- [31] Sun, P. et al. Nanoelectrochemistry of mammalian cells. *P Natl Acad Sci USA* 105, 443–448, doi: 10.1073/pnas.0711075105 (2008).
- [32] Schrlau, M. G., Dun, N. J. and Bau, H. H. Cell electrophysiology with carbon nanopipettes. *ACS nano* 3, 563–568, doi:10.1021/nn800851d (2009).
- [33] Qing, Q. et al. Free-standing kinked nanowire transistor probes for targeted intracellular recording in three dimensions. *Nat Nanotechnol* 9, 142–147, doi: 10.1038/Nnano.2013.273 (2014).
- [34] Tan, W., Shi, Z. Y., Smith, S., Birnbaum, D. and Kopelman, R. Submicrometer intracellular chemical optical fiber sensors. *Science* 258, 778–781 (1992).
- [35] Vo-Dinh, T., Alarie, J. P., Cullum, B. M. and Griffin, G. D. Antibody-based nanoprobe for measurement of a fluorescent analyte in a single cell. *Nature Biotechnology* 18, 764–767, doi:10.1038/77337 (2000).
- [36] Kasili, P. M., Song, J. M. and Vo-Dinh, T. Optical sensor for the detection of caspase-9 activity in a single cell. *Journal of the American Chemical Society* 126, 2799–2806, doi:10.1021/ja037388t (2004).
- [37] Wang, S. Q. et al. Silver-coated near field optical scanning microscope probes fabricated by silver mirror reaction. *Appl Phys B-Lasers O* 92, 49–52, doi: 10.1007/s00340-008-3054-y (2008).
- [38] Zheng, X. T., Yang, H. B. and Li, C. M. Optical Detection of Single Cell Lactate Release for Cancer Metabolic Analysis. *Anal Chem* 82, 5082–5087, doi: 10.1021/AC100074n (2010).
- [39] Vitol, E. A. et al. *In situ* intracellular spectroscopy with surface enhanced Raman spectroscopy (SERS)-enabled nanopipettes. *ACS Nano* 3, 3529–3536, doi: 10.1021/nn9010768 (2009).
- [40] Hong, W., Liang, F., Schaak, D., Loncar, M. and Quan, Q. Nanoscale Label-free Bioprobes to Detect Intracellular Proteins in Single Living Cells. *Sci Rep* 4, 6179, doi:10.1038/srep06179 (2014).
- [41] Chen, X., Kis, A., Zettl, A. and Bertozzi, C. R. A cell nanoinjector based on carbon nanotubes. *P Natl Acad Sci USA* 104, 8218–8222, doi:10.1073/pnas.0700567104 (2007).
- [42] Vakarelski, I. U., Brown, S. C., Higashitani, K. and Moudgil, B. M. Penetration of living cell membranes with fortified carbon nanotube tips. *Langmuir* 23, 10893–10896, doi: 10.1021/La701878n (2007).
- [43] Yum, K., Na, S., Xiang, Y., Wang, N. and Yu, M. F. Mechanochemical delivery and dynamic tracking of fluorescent quantum dots in the cytoplasm and nucleus of living cells. *Nano Lett* 9, 2193–2198, doi: 10.1021/nl901047u (2009).
- [44] Obataya, I., Nakamura, C., Han, S., Nakamura, N. and Miyake, J. Nanoscale operation of a living cell using an atomic force microscope with a nanoneedle. *Nano Lett* 5, 27–30, doi: 10.1021/Nl0485399 (2005).
- [45] Singhal, R. et al. Multifunctional carbon-nanotube cellular endoscopes. *Nat Nanotechnol* 6, 57–64, doi: 10.1038/Nnano.2010.241 (2011).
- [46] Yum, K., Wang, N. and Yu, M. F. Nanoneedle: a multifunctional tool for biological studies in living cells. *Nanoscale* 2, 363–372, doi:10.1039/b9nr00231f (2010).
- [47] Zheng, X. T. and Li, C. M. Single cell analysis at the nanoscale. *Chemical Society Reviews* 41, 2061–2071, doi: 10.1039/C1cs15265c (2012).
- [48] Gao, Y. et al. One-dimensional nanoprobe for single-cell studies. *Nanomedicine (Lond)* 9, 153–168, doi:10.2217/nnm.13.196 (2014).