



LUND UNIVERSITY

PIXE Analysis of Environmental Samples

Akselsson, Roland; Bohgard, Mats; Carlsson, Lars-Eric; Hansson, Hans Christen; Johansson, Eva-Marta; Johansson, Gerd; Lannefors, Hans; Malmqvist, Klas; Pallon, Jan

1979

[Link to publication](#)

Citation for published version (APA):

Akselsson, R., Bohgard, M., Carlsson, L-E., Hansson, H. C., Johansson, E-M., Johansson, G., Lannefors, H., Malmqvist, K., & Pallon, J. (1979). *PIXE Analysis of Environmental Samples*. Abstract from XXI Colloquium Spectroscopium Internationale, Cambridge, United Kingdom.

Total number of authors:

9

General rights

Unless other specific re-use rights are stated the following general rights apply:

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: <https://creativecommons.org/licenses/>

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00

110362, PIXE ANALYSIS OF ENVIRONMENTAL SAMPLES

K. R. AKSELSSON, M. BOHGARD, L. E. CARLSSON, H. C. HANSSON, E. M. JOHANSSON,
G. JOHANSSON, H. LANNEFORS, K. MALMQVIST and J. PALLON.

Department of Nuclear Physics, Lund Institute of Technology,
Sölvegatan 14, S-223 62 Lund, Sweden.

Particle induced X-ray emission (PIXE-)analysis constitutes a fast reliable method for multielemental analysis with low detection limits. The method is especially suited for small samples (1 μg - 1 mg total mass). Its mass calibration is checked within a minute or two. The possibility of simultaneous applications of other nuclear techniques for the analysis of lighter elements extends the merits of PIXE for some types of environmental samples considerably.

At our laboratory routinely 2.5 MeV protons are used to induce X-rays, which are detected in 135° in a 80 mm^2 Si(Li)-detector. Work is in progress to optimize the arrangement and the parameters for simultaneous PIXE and low Z backscattering analysis. A sample typically requires 1 - 4 minutes for analysis. Before and after the analysis of every container holding 40 samples, a two-element standard is run. The arrangement is very stable and the results of the standard runs are routinely reproducible at the 5 % level.

The PIXE method is well suited for studies of elemental composition of airborne particulates. Due to its capability to analyse small samples, small and convenient samplers may be used. Due to its speed and low cost many samples may be analysed. Thus PIXE offers the possibility of good time, size and site resolution. These features will be illustrated by results from two investigations performed in Sweden. In the first of them a 6-stage single orifice cascade impactor was used. It was e.g. found that air masses coming from the south-west contained 10 to 30 times more of the antropogenic small particle elements sulfur, lead and copper than those coming from the north-west. In the second investigation a sampler of a special continuous time sequence design (J.W.Nelson, Dept of Physics, FSU, Tallahassee) was used. This filter sampler gives a time resolution of about 2 hours and a sample from more than one week is contained in a 5 mm wide and about 170 mm long streak.

Detection limits for trace elements in e.g. sea and rain water below the 0.1 ppb level are obtained by using a technique involving complex-binding and adsorption on activated carbon.