



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Patient-specific Foot Orthotics Effect on Lower Extremity Joint Loading for Rheumatoid Arthritis Patients Using MRI-based Musculoskeletal Models

Simonsen, Morten Bilde; Næsborg-Andersen, Ketil; Kowalski, Marcin Ryszard; Petersen, Kim Hørslev; Andersen, Michael Skipper

Publication date:
2017

Document Version
Accepted author manuscript, peer reviewed version

[Link to publication from Aalborg University](#)

Citation for published version (APA):

Simonsen, M. B., Næsborg-Andersen, K., Kowalski, M. R., Petersen, K. H., & Andersen, M. S. (2017). *Patient-specific Foot Orthotics Effect on Lower Extremity Joint Loading for Rheumatoid Arthritis Patients Using MRI-based Musculoskeletal Models*. Abstract from XXVI Congress of the International Society of Biomechanics, Brisbane, Australia.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- ? Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- ? You may not further distribute the material or use it for any profit-making activity or commercial gain
- ? You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Patient-Specific Foot Orthotics Effect on Lower Extremity Joint Loading for Rheumatoid Arthritis Patients using MRI-Based Musculoskeletal Models

¹ Morten B Simonsen, ² Ketil Næsberg-Andersen, ³ Marcin R Kowalski, ⁴ Kim Hørslev-Petersen and ¹ Michael S Andersen

¹ Dept. of Mechanical and Manufacturing Engineering, Aalborg University

² Bandagist-Centret A/S

³ Department of Rheumatology, Hjørring Hospital

⁴ King Christian 10th Hospital for Rheumatic Diseases

Corresponding author email: mortenbildes@gmail.com

INTRODUCTION

Rheumatoid arthritis (RA) patients often have pain in the lower extremity, which in some cases can be caused by foot deformity and foot pain. With the intention to stabilize and align the foot patient-specific foot orthotics (FO) are often prescribed to this patient group. A limitation of the previous literature on FO to treat RA and related diseases is that it has focused exclusively on clinical outcomes of FO such as pain and physical function, while overlooking the biomechanical principles on which the rationales for FO is based [1].

The aim of this study was to investigate the effect of patient-specific FO on ankle, knee and hip loading during gait. This was accomplished by developing patient-specific (PS) musculoskeletal models (MS) capable of estimating joint mechanics with and without the FO.

METHODS

Four early stage RA patients were recruited for this study. A pair of FO was developed for each patient using a weight bearing casting technique. PS bone geometry was obtained from magnetic resonance imaging (MRI) images and segmented in an image analysis package (Mimics 19, Materialise, Belgium). Motion capture was performed with an eight-camera setup (Qualysis, Sweden) with reflective markers together with three force plates (AMTI, USA) sampling at 100 and 1000 Hz respectively. The gait trial consisted of two conditions: one with the PS FO and one with a control insole (C).

PS MS models of each patient were developed using the AnyBody Modeling System (AnyBody Technology, Denmark), Figure 1B. Muscle attachments were made PS based on the Twente Lower Extremity Model version 2.0 dataset using advanced morphing to customize a generic cadaver-based model with respect to PS morphology acquired from MRI [2].

Accurate joint centres and axes were calculated with analytical surface fits to the segmented MRI bones. Joint reference systems were defined based on the ISB recommendations.

RESULTS AND DISCUSSION

Resultant joint forces for the ankle, knee and hip joints are presented in (Figure 1) for the C and FO. The results showed that the ankle force is reduced with FO. Meanwhile, the resultant knee and hip joint forces increased with FO. Additional data from the model showed that, despite an increase in the total resultant forces in the knee and hip, change in the different joint moments and muscle forces still occur. These changes may potentially contribute to the reduction in pain. Further studies are needed to confirm this.




Figure 1: A: Mean peak resultant joint forces for the Ankle-, knee- and hip joint during the stance phase of gait with \pm one standard deviation. B: Patient-specific AnyBody model

CONCLUSIONS

The results of this study indicate that FO can change the load distribution in the lower extremity. MS modeling is appealing for studying biomechanics due to the challenging nature of directly measuring the internal loading of the complex structure of the human body that occurs during motion. Further studies are needed to investigate if there is a relationship between changed loading in MS models and pain for RA patients.

ACKNOWLEDGEMENTS

The study is financially supported by The Danish Rheumatism Association [R142-A4113].

REFERENCES

1. Lewinson R. et al., Annals of biomedical engineering, **44**: 3173-3185, 2016.
2. Marra M, et al., Journal of Biomechanical Engineering. **137**: BIO-14-1490, 2015.