



Open Archive TOULOUSE Archive Ouverte (OATAO)

OATAO is an open access repository that collects the work of Toulouse researchers and makes it freely available over the web where possible.

This is an author-deposited version published in : <http://oatao.univ-toulouse.fr/>
Eprints ID : 16868

To link to this article : DOI : 10.1109/TMI.2015.2493241
URL : <http://dx.doi.org/10.1109/TMI.2015.2493241>

<p>To cite this version : Chen, Zhouye and Basarab, Adrian and Kouamé, Denis <i>Compressive deconvolution in medical ultrasound imaging</i>. (2016) IEEE Transactions on Medical Imaging, vol. 35 (n° 3). pp. 728-737. ISSN 0278-0062</p>
--

Any correspondence concerning this service should be sent to the repository administrator: staff-oatao@listes-diff.inp-toulouse.fr

Compressive Deconvolution in Medical Ultrasound Imaging

Zhouye Chen*, *Student Member, IEEE*, Adrian Basarab, *Member, IEEE*, and Denis Kouamé, *Member, IEEE*

Abstract—The interest of compressive sampling in ultrasound imaging has been recently extensively evaluated by several research teams. Following the different application setups, it has been shown that the RF data may be reconstructed from a small number of measurements and/or using a reduced number of ultrasound pulse emissions. Nevertheless, RF image spatial resolution, contrast and signal to noise ratio are affected by the limited bandwidth of the imaging transducer and the physical phenomenon related to US wave propagation. To overcome these limitations, several deconvolution-based image processing techniques have been proposed to enhance the ultrasound images. In this paper, we propose a novel framework, named compressive deconvolution, that reconstructs enhanced RF images from compressed measurements. Exploiting an unified formulation of the direct acquisition model, combining random projections and 2D convolution with a spatially invariant point spread function, the benefit of our approach is the joint data volume reduction and image quality improvement. The proposed optimization method, based on the Alternating Direction Method of Multipliers, is evaluated on both simulated and *in vivo* data.

Index Terms—Compressive sampling, deconvolution, ultrasound imaging, alternating direction method of multipliers.

I. INTRODUCTION

ULTRASOUND (US) medical imaging has the advantages of being noninvasive, harmless, cost-effective and portable over other imaging modalities such as X-ray Computed Tomography or Magnetic Resonance Imaging [1].

Despite its intrinsic rapidity of acquisition, several US applications such as cardiac, Doppler, elastography or 3D imaging may require higher frame rates than those provided by conventional acquisition schemes (e.g., ultrafast imaging [2]) or may suffer from the high amount of acquired data. In this context, a few research teams have recently evaluated the application of compressive sampling (CS) to 2D and 3D US imaging (e.g., [3]–[8]) or to duplex Doppler [9]. CS is a mathematical framework allowing to recover, via non linear optimization routines, an image from few linear measurements (below the limit standardly imposed by the Shannon-Nyquist theorem) [10], [11]. The CS acquisition model is given by

$$\mathbf{y} = \Phi \mathbf{r} + \mathbf{n} \quad (1)$$

Manuscript received September 11, 2015; revised October 14, 2015; accepted October 17, 2015. Date of publication October 26, 2015; date of current version March 01, 2016. This work was partially supported by ANR-11-LABX-0040-CIMI within the program ANR-11-IDEX-0002-02 of the University of Toulouse and CSC (Chinese Scholarship Council). *Asterisk indicates corresponding author.*

*Z. Chen is with IRIT UMR CNRS 5505, University of Toulouse, 31062 Toulouse, France (e-mail: zhouye.chen@irit.fr).

A. Basarab and D. Kouamé are with IRIT UMR CNRS 5505, University of Toulouse, 31062 Toulouse, France (e-mail: adrian.basarab@irit.fr; denis.kouame@irit.fr).

Digital Object Identifier 10.1109/TMI.2015.2493241

where $\mathbf{y} \in \mathbb{R}^M$ corresponds to the M compressed measurements of the image $\mathbf{r} \in \mathbb{R}^N$ (one US radiofrequency (RF) image in our case), $\Phi \in \mathbb{R}^{M \times N}$ represents the CS acquisition matrix composed for example of M random Gaussian vectors with $M \ll N$ and $\mathbf{n} \in \mathbb{R}^M$ stands for a zero-mean additive white Gaussian noise.

The CS theory demonstrates that the N pixels of image \mathbf{r} may be recovered from the M measurements in \mathbf{y} provided two conditions: i) the image must have a sparse representation in a known basis or frame and ii) the measurement matrix and sparsifying basis must be incoherent [12]. In US imaging, despite the difficulties of sparsifying the data because of the speckle noise, it has been shown that RF images may be recovered in basis such as 2D Fourier [4], wavelets [8], waveatoms [7] or learning dictionaries [13], considering Bernoulli Gaussian [14] or α -stable statistical assumptions [15] and using various acquisition schemes such as plane-wave [8], Xampling [5] or projections on Gaussian [3] or Bernoulli random vectors [4].

However, the existing methods of CS in US have been shown to be able to recover images with a quality at most equivalent to those acquired using standard schemes. Nevertheless, the spatial resolution, the signal-to-noise ratio and the contrast of standard US images (\mathbf{r} in (1)) are affected by the limited bandwidth of the imaging transducer, the physical phenomenon related to US wave propagation such as the diffraction and the imaging system. In order to overcome these issues, one of the research tracks extensively explored in the literature is the deconvolution of US images [16]–[21]. Based on the first order Born approximation, these methods assume that the US RF images follow a 2D convolution model between the point spread function (PSF) and the tissue reflectivity function (TRF), i.e., the image to be recovered [22]. Specifically, this results in $\mathbf{r} = H\mathbf{x}$, where $H \in \mathbb{R}^{N \times N}$ is a block circulant with circulant block (BCCB) matrix related to the 2D PSF of the system and $\mathbf{x} \in \mathbb{R}^N$ represents the lexicographically ordered tissue reflectivity function. We emphasize that this convolution model is based on the assumption of spatially invariant PSF. Although the PSF of an ultrasound system varies spatially and mostly in the axial direction of the images, many settings of the ultrasound imaging system allow to attenuate this spatial variation in practice, such as the dynamic focalization of the received echoes, the multiple focusing in emission and the time gain compensation (TGC). For this reason, the PSF is considered spatially invariant in our work, as it is the case in part of the existing works on image deconvolution in ultrasound imaging (e.g., [17], [20]).

The objective of our paper is to propose a novel technique which is able to jointly achieve US data volume reduction and image quality improvement. In other words, the main idea is to

combine the two frameworks of CS and deconvolution applied to US imaging, resulting in the so-called compressive deconvolution (or CS deblurring) problem [23]–[27]. The combined direct model of joint CS and deconvolution is as follows:

$$\mathbf{y} = \Phi H \mathbf{x} + \mathbf{n} \quad (2)$$

where the variables \mathbf{y} , Φ , H , \mathbf{x} and \mathbf{n} have the same meaning as defined above. Inverting the model in (2) will allow us to estimate the TRF \mathbf{x} from the compressed RF measurements \mathbf{y} .

To our knowledge, our work is the first attempt of addressing the compressive deconvolution problem in US imaging. In the general-purpose image processing literature, a few methods have been already proposed aiming at solving (2) [23]–[29]. In [25], [28], [29], the authors assumed \mathbf{x} was sparse in the direct or image domain and the PSF was unknown. In [28], [29], a study on the number of measurements lower bound is presented, together with an algorithm to estimate the PSF and \mathbf{x} alternatively. The authors in [25] solved the compressive deconvolution problem using an ℓ_1 -norm minimization algorithm by making use of the “all-pole” model of the autoregressive process. In [23], [24], \mathbf{x} was considered sparse in a transformed domain and the PSF was supposed known. An algorithm based on Poisson singular integral and iterative curvelet thresholding was shown in [23]. The authors in [24] further combined the curvelet regularization with total variation to improve the performance in [23]. Finally, the methods in [26], [27] supposed the blurred signal $\mathbf{r} = H\mathbf{x}$ was sparse in a transformed domain and the PSF unknown. They proposed a compressive deconvolution framework that relies on a constrained optimization technique allowing to exploit existing CS reconstruction algorithms.

In this paper, we propose a compressive deconvolution technique adapted to US imaging. Our solution is based on the alternating direction method of multipliers (ADMM) [30], [31] and exploits two constraints. The first one, inspired from CS, imposes via an l_1 -norm the sparsity of the RF image \mathbf{r} in the transformed domain (Fourier domain, wavelet domain, etc). The second one imposes *a priori* information for the TRF \mathbf{x} . Gaussian and Laplacian statistics have been extensively explored in US imaging (see e.g., [17], [32], [33]). Moreover, recent results show that the Generalized Gaussian Distributed (GGD) is well adapted to model the TRF. Consequently, we employ herein the minimization of an l_p -norm of \mathbf{x} , covering all possible cases ranging from 1 to 2 [19], [20], [34]. Similar to all existing frameworks, we consider the CS sampling matrix Φ known. In US imaging, the PSF is unknown in practical applications. However, its estimation from the RF data as an initialization step for the deconvolution has been extensively explored in US imaging. In this paper, we adopted the approach in [35] in order to estimate the PSF further used to construct the matrix H .

This paper is organized as follows. In Section II, we formulate our problem as a convex optimization routine and propose an ADMM-based method to efficiently solve it. Supporting simulated and experimental results are provided in Section III showing the contribution of our approach compared to existing methods and its efficiency in recovering the TRF from compressed US data. The conclusions are drawn in IV.

II. PROPOSED ULTRASOUND COMPRESSIVE DECONVOLUTION ALGORITHM

A. Optimization Problem Formulation

1) *Sequential Approach*: In order to estimate the TRF \mathbf{x} from the compressed and blurred measurements \mathbf{y} , an intuitive idea to invert the direct model in (2) is to proceed through two sequential steps. The aim of the first step is to recover the blurred US RF image $\mathbf{r} = H\mathbf{x}$ from the compressed measurements \mathbf{y} by solving the following optimization problem:

$$\min_{\mathbf{a} \in \mathbb{R}^N} \|\mathbf{a}\|_1 + \frac{1}{2\mu} \|\mathbf{y} - \Phi\Psi\mathbf{a}\|_2^2 \quad (3)$$

where \mathbf{a} is the sparse representation of the US RF image \mathbf{r} in the transformed domain Ψ , that is, $\mathbf{r} = H\mathbf{x} = \Psi\mathbf{a}$. Different basis have been shown to provide good results in the application of CS in US imaging, such as wavelets, waveatoms or 2D Fourier basis [6]. In this paper the wavelet transform has been employed.

Once the blurred RF image, denoted by $\hat{\mathbf{r}}$, is recovered by solving the convex problem in (3), one can restore the TRF \mathbf{x} by minimizing:

$$\min_{\mathbf{x} \in \mathbb{R}^N} \alpha \|\mathbf{x}\|_p^p + \|\hat{\mathbf{r}} - H\mathbf{x}\|_2^2 \quad (4)$$

The first term in (4) aims at regularizing the TRF by a GGD statistical assumption, where p is related to the shape parameter of the GGD. In this paper, we focus on shape parameters ranging from 1 to 2 ($1 \leq p \leq 2$), allowing us to generalize the existing works in US image deconvolution mainly based on Laplacian or Gaussian statistics [16]–[18].

2) *Proposed Approach*: While the sequential approach represents the most intuitive way to solve the compressive deconvolution problem, dividing a single problem into two separate subproblems will inevitably generate larger estimation errors as shown by the results in Section III. Therefore, we propose herein a method to solve the CS and deconvolution problem simultaneously. Similarly to [26], we formulate the reconstruction process into a constrained optimization problem exploiting the relationship between (3) and (4).

$$\begin{aligned} \min_{\mathbf{x} \in \mathbb{R}^N, \mathbf{a} \in \mathbb{R}^N} \quad & \|\mathbf{a}\|_1 + \alpha \|\mathbf{x}\|_p^p + \frac{1}{2\mu} \|\mathbf{y} - \Phi\Psi\mathbf{a}\|_2^2 \\ \text{s.t.} \quad & H\mathbf{x} = \Psi\mathbf{a} \end{aligned} \quad (5)$$

However, since our goal is to recover enhanced US imaging by estimation the TRF \mathbf{x} , we further reformulate the problem above into a unconstrained optimization problem:

$$\min_{\mathbf{x} \in \mathbb{R}^N} \|\Psi^{-1}H\mathbf{x}\|_1 + \alpha \|\mathbf{x}\|_p^p + \frac{1}{2\mu} \|\mathbf{y} - \Phi H\mathbf{x}\|_2^2 \quad (6)$$

The objective function in (6) contains, in addition to the data fidelity term, two regularization terms. The first one aims at imposing the sparsity of the RF data $H\mathbf{x}$ (i.e., minimizing the ℓ_1 -norm of the target image \mathbf{x} convolved with a bandlimited function) in a transformed domain Ψ . We should note that such an assumption has been extensively used in the application of CS in US imaging, see e.g., [4], [6]–[8], [13], [34]. Transformations such as 2D Fourier, wavelet or wave atoms have been shown to provide good results in US imaging. The second term

aims at regularizing the TRF \mathbf{x} and is related to the GGD statistical assumption of US images, see e.g., [19], [20], [34].

We notice that our regularized reconstruction problem based on the objective function in (6) is different from a typical CS reconstruction. Specifically, the objective function of a standard CS technique applied to our model would only contain the classical data fidelity term and an ℓ_1 -norm penalty similar to the first term in (6) but without the operator H . However, such a CS reconstruction is not adapted to compressive deconvolution, mainly because the requirements of CS theory such as the restricted isometry property might not be guaranteed [26].

To solve the optimization problem in (6), we propose hereafter an algorithm based on the alternating direction method of multipliers (ADMM).

B. Basics of Alternating Direction Method of Multipliers

Before going into the details of our algorithm, we report in this paragraph the basics of ADMM. ADMM has been extensively studied in the areas of convex programming and variational inequalities, e.g., [30]. The general optimization problem considered in ADMM framework is as follows:

$$\begin{aligned} \min_{u,v} f(u) + g(v) \\ \text{s.t. } Bu + Cv = b, u \in \mathcal{U}, v \in \mathcal{V} \end{aligned} \quad (7)$$

where $\mathcal{U} \subseteq \mathbb{R}^s$ and $\mathcal{V} \subseteq \mathbb{R}^t$ are given convex sets, $f : \mathcal{U} \rightarrow \mathbb{R}$ and $g : \mathcal{V} \rightarrow \mathbb{R}$ are closed convex functions, $B \in \mathbb{R}^{r \times s}$ and $C \in \mathbb{R}^{r \times t}$ are given matrices and $\mathbf{b} \in \mathbb{R}^r$ is a given vector.

By attaching the Lagrangian multiplier $\lambda \in \mathbb{R}^r$ to the linear constraint, the Augmented Lagrangian (AL) function of (7) is

$$\begin{aligned} \mathcal{L}(u, v, \lambda) = f(u) + g(v) - \lambda^t (Bu + Cv - b) \\ + \frac{\beta}{2} \| Bu + Cv - b \|^2_2 \end{aligned} \quad (8)$$

where $\beta > 0$ is the penalty parameter for the linear constraints to be satisfied. The standard ADMM framework follows the three steps iterative process:

$$\begin{cases} u^{k+1} \in \underset{u \in \mathcal{U}}{\operatorname{argmin}} \mathcal{L}(u, v^k, \lambda^k) \\ v^{k+1} \in \underset{v \in \mathcal{V}}{\operatorname{argmin}} \mathcal{L}(u^{k+1}, v, \lambda^k) \\ \lambda^{k+1} = \lambda^k - \beta (Bu^{k+1} + Cv^{k+1} - b) \end{cases} \quad (9)$$

The main advantage of ADMM, in addition to the relative ease of implementation, is its ability to split awkward intersections and objectives to easy subproblems, resulting into iterations comparable to those of other first-order methods.

C. Proposed ADMM Parameterization for Ultrasound Compressive Deconvolution

In this subsection, we propose an ADMM method for solving the ultrasound compressive deconvolution problem in (6).

Using a trivial variable change, the minimization problem in (6) can be rewritten as:

$$\min_{\mathbf{x} \in \mathbb{R}^N} \|\mathbf{w}\|_1 + \alpha \|\mathbf{x}\|_p^p + \frac{1}{2\mu} \|\mathbf{y} - A\mathbf{a}\|_2^2 \quad (10)$$

where $\mathbf{a} = \Psi^{-1}H\mathbf{x}$, $\mathbf{w} = \mathbf{a}$ and $A = \Phi\Psi$. Let us denote $\mathbf{z} = \begin{bmatrix} \mathbf{w} \\ \mathbf{x} \end{bmatrix}$. The reformulated problem in (10) can fit the general ADMM framework in (7) by choosing: $f(\mathbf{a}) = (1)/(2\mu) \|\mathbf{y} - A\mathbf{a}\|_2^2$, $g(\mathbf{z}) = \|\mathbf{w}\|_1 + \alpha \|\mathbf{x}\|_p^p$, $B = \begin{bmatrix} I_N \\ \Psi \end{bmatrix}$, $C = \begin{bmatrix} -I_N & \mathbf{0} \\ \mathbf{0} & -H \end{bmatrix}$ and $\mathbf{b} = \mathbf{0}$. $I_N \in \mathbb{R}^{N \times N}$ is the identity matrix.

The augmented Lagrangian function of (10) is given by

$$\begin{aligned} \mathcal{L}(\mathbf{a}, \mathbf{z}, \boldsymbol{\lambda}) = f(\mathbf{a}) + g(\mathbf{z}) - \boldsymbol{\lambda}^t (B\mathbf{a} + C\mathbf{z}) \\ + \frac{\beta}{2} \| B\mathbf{a} + C\mathbf{z} \|^2_2 \end{aligned} \quad (11)$$

where $\boldsymbol{\lambda} \in \mathbb{R}^{2N}$ stands for $\boldsymbol{\lambda} = \begin{bmatrix} \boldsymbol{\lambda}_1 \\ \boldsymbol{\lambda}_2 \end{bmatrix}$, $\boldsymbol{\lambda}_i \in \mathbb{R}^N$ ($i = 1, 2$). According to the standard ADMM iterative scheme, the minimizations with respect to \mathbf{a} and \mathbf{z} will be performed alternatively, followed by the update of $\boldsymbol{\lambda}$.

D. Implementation Details

In this subsection, we detail each of the three steps of our ADMM-based compressive deconvolution method. While the following mathematical developments are given for the case when the regularization term for TRF \mathbf{x} is equal to $\|\mathbf{x}\|_p^p$ (adapted to US images), our approach using a generalized total variation regularization is also detailed in Appendix A and may be useful for other (medical) applications.

Step 1 consists in solving the \mathbf{z} -problem, since $\mathbf{z} = \begin{bmatrix} \mathbf{w} \\ \mathbf{x} \end{bmatrix}$, this problem can be further divided into two subproblems.

Step 1.1 aims at solving:

$$\begin{aligned} \mathbf{w}^k &= \underset{\mathbf{w} \in \mathbb{R}^N}{\operatorname{argmin}} \|\mathbf{w}\|_1 - \left(\boldsymbol{\lambda}_1^{k-1} \right)^t (\mathbf{a}^{k-1} - \mathbf{w}) \\ &\quad + \frac{\beta}{2} \|\mathbf{a}^{k-1} - \mathbf{w}\|_2^2 \\ \Leftrightarrow \mathbf{w}^k &= \underset{\mathbf{w} \in \mathbb{R}^N}{\operatorname{argmin}} \|\mathbf{w}\|_1 + \frac{\beta}{2} \left\| \mathbf{a}^{k-1} - \mathbf{w} - \frac{\boldsymbol{\lambda}_1^{k-1}}{\beta} \right\|_2^2 \\ \Leftrightarrow \mathbf{w}^k &= \operatorname{prox}_{\|\cdot\|_1/\beta} \left(\mathbf{a}^{k-1} - \frac{\boldsymbol{\lambda}_1^{k-1}}{\beta} \right) \end{aligned} \quad (12)$$

where prox stands for the proximal operator as proposed in [36]–[38]. The proximal operators of various kinds of functions including $\|\mathbf{x}\|_p^p$ have been given explicitly in the literature (see e.g., [39]). Basics about the proximal operator of $\|\mathbf{x}\|_p^p$ are reminded in Appendix B.

Step 1.2 consists in solving:

$$\begin{aligned} \mathbf{x}^k &= \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} \alpha \|\mathbf{x}\|_p^p - \boldsymbol{\lambda}_2^{k-1} (\Psi\mathbf{a}^{k-1} - H\mathbf{x}) \\ &\quad + \frac{\beta}{2} \|\Psi\mathbf{a}^{k-1} - H\mathbf{x}\|_2^2 \end{aligned} \quad (13)$$

For p equal to 2, the minimization in (13) can be easily solved in the Fourier domain, as follows:

$$\mathbf{x}^k = [\beta H^t H + 2\alpha I_N]^{-1} \times \left[\beta H^t \Psi \mathbf{a}^{k-1} - H^t \boldsymbol{\lambda}_2^{k-1} \right] \quad (14)$$

For $1 \leq p < 2$, we propose to use the proximal operator to solve (13). In this case, (13) will be equivalent to

$$\mathbf{x}^k = \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} \alpha \|\mathbf{x}\|_p^p + \frac{\beta}{2} \left\| \Psi \mathbf{a}^{k-1} - H\mathbf{x} - \frac{\lambda_2^{k-1}}{\beta} \right\|_2^2 \quad (15)$$

Denoting $h(\mathbf{x}) = (1)/(2) \|\Psi \mathbf{a}^{k-1} - H\mathbf{x} - (\lambda_2^{k-1})/(\beta)\|_2^2$, we can further approximate $h(\mathbf{x})$ by

$$h'(\mathbf{x}^{k-1})(\mathbf{x} - \mathbf{x}^{k-1}) + \frac{1}{2\gamma} \|\mathbf{x} - \mathbf{x}^{k-1}\|_2^2 \quad (16)$$

where $\gamma > 0$ is a parameter related to the Lipschitz constant [40] and $h'(\mathbf{x}^{k-1})$ is the gradient of $h(\mathbf{x})$ when $\mathbf{x} = \mathbf{x}^{k-1}$.

By plugging (16) into (15), we obtain:

$$\begin{aligned} \mathbf{x}^k &\approx \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} \alpha \|\mathbf{x}\|_p^p + \beta h'(\mathbf{x}^{k-1})(\mathbf{x} - \mathbf{x}^{k-1}) \\ &\quad + \frac{\beta}{2\gamma} \|\mathbf{x} - \mathbf{x}^{k-1}\|_2^2 \\ \Leftrightarrow \mathbf{x}^k &\approx \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} \alpha \|\mathbf{x}\|_p^p \\ &\quad + \frac{\beta}{2\gamma} \|\mathbf{x} - \mathbf{x}^{k-1} + \gamma h'(\mathbf{x}^{k-1})\|_2^2 \end{aligned} \quad (17)$$

According to the definition of the proximal operator, we can finally get

$$\mathbf{x}^k \approx \operatorname{prox}_{\alpha\gamma \|\cdot\|_p/\beta} \{\mathbf{x}^{k-1} - \gamma h'(\mathbf{x}^{k-1})\} \quad (18)$$

We should note that (18) provides an approximate solution, thus resulting into an inexact ADMM scheme. However, the convergence of such inexact ADMM has been already established in [30], [41], [42].

Step 2 aims at solving:

$$\begin{aligned} \mathbf{a}^k &= \underset{\mathbf{a} \in \mathbb{R}^N}{\operatorname{argmin}} \frac{1}{2\mu} \|\mathbf{y} - A\mathbf{a}\|_2^2 - (\lambda^{k-1})^t (B\mathbf{a} + C\mathbf{z}^k) \\ &\quad + \frac{\beta}{2} \|B\mathbf{a} + C\mathbf{z}^k\|_2^2 \\ \Leftrightarrow \mathbf{a}^k &= \left(\frac{1}{\mu} A^t A + \beta I_N + \beta \Psi^t \Psi \right)^{-1} \\ &\quad \times \left(\frac{1}{\mu} A^t \mathbf{y} + \lambda_1^{k-1} + \Psi^t \lambda_2^{k-1} + \beta \mathbf{w}^k + \beta \Psi^t H \mathbf{x}^k \right) \end{aligned} \quad (19)$$

The formula above is equivalent to solving an $N \times N$ linear system or inverting an $N \times N$ matrix. However, since the sparse basis Ψ considered is orthogonal (e.g., the wavelet transform), it can be reduced to solving a smaller $M \times M$ linear system or inverting an $M \times M$ matrix by exploiting the Sherman-Morrison-Woodbury inversion matrix lemma [43]:

$$\begin{aligned} (\beta_1 I_N + \beta_2 A^t A)^{-1} &= \frac{1}{\beta_1} I_N \\ &\quad - \frac{\beta_2}{\beta_1} A^t (\beta_1 I_M + \beta_2 A A^t)^{-1} A \end{aligned} \quad (20)$$

In this paper, without loss of generality, we considered the compressive sampling matrix Φ as a Structurally Random Ma-

trix (SRM) [44]. Therefore, A was formed by randomly taking a subset of rows from orthonormal transform matrices, that is, $AA^t = I_M$. As a consequence, there is no need to solve a linear system and the main computational cost consists into two matrix-vector multiplications per iteration.

Step 3 consists in solving:

$$\lambda^k = \lambda^{k-1} - \beta(B\mathbf{a}^k + C\mathbf{z}^k) \quad (21)$$

The proposed optimization routine is summarized in Algorithm 1.

Algorithm 1 ADMM algorithm for Solving (6)

Input: $\mathbf{a}^0, \lambda^0, \alpha, \mu, \beta$

- 1: **while** not converged **do**
- 2: $\mathbf{w}^k \leftarrow \mathbf{a}^{k-1}, \lambda^{k-1}$ ▷ update \mathbf{w}^k using (12)
- 3: $\mathbf{x}^k \leftarrow \mathbf{a}^{k-1}, \lambda^{k-1}$ ▷ update \mathbf{x}^k using (14) or (18)
- 4: $\mathbf{a}^k \leftarrow \mathbf{w}^k, \mathbf{x}^k, \lambda^{k-1}$ ▷ update \mathbf{a}^k using (19)
- 5: $\lambda^k \leftarrow \mathbf{w}^k, \mathbf{x}^k, \mathbf{a}^k, \lambda^{k-1}$ ▷ update λ^k using (21)
- 6: **end while**

Output: \mathbf{x}

III. RESULTS

The performance of the proposed compressive deconvolution method are evaluated on several simulated and experimental data sets. First, we test our algorithm on a modified Shepp-Logan phantom containing speckle noise to confirm that the l_p -norm regularization term is more adapted to US images than the generalized TV used in [26]. The approach in [26] is referred as CD_Amizic hereafter. Second, we give the results of our algorithm for different l_p -norm optimizations on simulated US images, showing the superiority of our approach over the intuitive sequential method explained in Section II. Finally, compressive deconvolution results on two *in vivo* ultrasound images are presented. Moreover, a comparison between our approach and CD_Amizic on the standard Shepp-Logan phantom is provided in Appendix C.

A. Results on Modified Shepp-Logan Phantom

We modified the Shepp-Logan phantom in order to simulate the speckle noise that degrades in practice the US images. For this, we followed the procedure classically used in US imaging [45]. First, scatterers at uniformly random locations have been generated, with amplitudes distributed according to a zero-mean generalized Gaussian distribution (GGD) with the shape parameter set to 1.3 and the scale parameter equal to 1. The scatterer amplitudes were further multiplied by the values of the original Shepp-Logan phantom pixels located at the closest positions to the scatterers. The resulting image, mimicking the tissue reflectivity function (TRF) in US imaging, is shown in Fig. 1(a). The blurred image in Fig. 1(b) was obtained by 2D convolution between the TRF and a spatially invariant PSF generated with Field II [46], a state-of-the-art simulator in US imaging. It corresponds to a 3.5 MHz linear probe, sampled in the axial direction at 20 MHz. The compressive measurements were obtained by projecting the blurred image onto SRM and by adding a Gaussian noise corresponding to a SNR of 40 dB.




Fig. 1. Reconstruction results for SNR = 40 dB and a CS ratio of 0.6. (a) Modified Shepp-Logan phantom containing random scatterers (TRF), (b) Degraded image by convolution with a simulated US PSF, (c) Reconstruction result with CD_Amizic, (d) Reconstruction result with the proposed method using a generalized TV prior (ADMM_GTV), (e), (f), (g) Reconstruction results with the proposed method using an l_p -norm prior, for p equal to 1.5, 1.3 and 1 (ADMM_L1.5, ADMM_L1.3 and ADMM_L1).

TABLE I
QUANTITATIVE RESULTS FOR THE MODIFIED SHEPP-LOGAN PHANTOM WITH US SPECKLE (SNR = 40 dB)

CS ratios		CD_Amizic	ADMM_GTV	ADMM_L1.5	ADMM_L1.3	ADMM_L1
80%	PSNR	30.82	31.11	32.23	32.32	32.05
	SSIM	83.24	85.03	86.44	88.77	87.70
60%	PSNR	29.68	29.83	31.27	31.50	31.32
	SSIM	74.58	77.83	82.26	86.03	85.64
40%	PSNR	26.76	28.11	29.58	30.04	30.12
	SSIM	43.43	61.46	73.88	79.95	81.75
20%	PSNR	20.22	21.53	26.81	27.29	28.20
	SSIM	8.35	12.77	51.70	62.93	72.34

The results were quantitatively evaluated using the standard peak signal-to-noise ratio (PSNR) and the Structural Similarity (SSIM) [47]. PSNR is defined as

$$\text{PSNR} = 10 \log_{10} \frac{NL^2}{\|\mathbf{x} - \hat{\mathbf{x}}\|^2} \quad (22)$$

where \mathbf{x} and $\hat{\mathbf{x}}$ are the original and reconstructed images, and the constant L represents the maximum intensity value in \mathbf{x} . SSIM is extensively used in US imaging and defined as

$$\text{SSIM} = \frac{(2\mu_x\mu_{\hat{x}} + c_1)(2\sigma_{x\hat{x}} + c_2)}{(\mu_x^2 + \mu_{\hat{x}}^2 + c_1)(\sigma_x^2 + \sigma_{\hat{x}}^2 + c_2)} \quad (23)$$

where x and \hat{x} are the original and reconstructed images, $\mu_x, \mu_{\hat{x}}, \sigma_x$ and $\sigma_{\hat{x}}$ are the mean and variance values of x and \hat{x} , $\sigma_{x\hat{x}}$ is the covariance between x and \hat{x} ; $c_1 = (k_1L)^2$ and $c_2 = (k_2L)^2$ are two variables aiming at stabilizing the division with weak denominator, L is the dynamic range of the pixel-values and k_1, k_2 are constants. Herein, $L = 1, k_1 = 0.01$ and $k_2 = 0.03$.

Reconstruction results for a CS ratio of 0.6 are shown in Fig. 1. They were obtained with: the recent compressive deconvolution technique reported in [26] (referred as CD_Amizic), the proposed method using the generalized TV prior (denoted by ADMM_GTV) and the proposed method using the l_p -norm for p equal to 1.5, 1.3 and 1 (denoted respectively by ADMM_L1.5, ADMM_L1.3 and ADMM_L1). All the hyperparameters were set to their best possible values by cross-validation. For CD_Amizic, $\{\beta, \alpha, \eta, \tau\} = \{10^7, 1, 10^4, 10^2\}$. For ADMM_GTV

$\{\mu, \alpha, \beta\} = \{10^{-5}, 2 \times 10^{-1}, 10^2\}$ and for the proposed method with l_p -norms, $\{\mu, \alpha, \beta, \gamma\} = \{10^{-5}, 2 \times 10^{-1}, 10^1, 3 \times 10^{-2}\}$. The quantitative results for different CS ratios are regrouped in Table I. They confirm that the l_p -norm is better adapted to recover the TRF in US imaging than the generalized TV. The difference between the two priors is further accentuated when the CS ratio decreases.

Keeping in mind that the generalized TV prior is not well suited to model the TRF in US imaging, we did not use CD_Amizic in the following sections dealing with simulated and experimental US images. Moreover, the proposed method was only evaluated in its l_p -norm minimization form.

B. Results on Simulated Ultrasound Images

In this section, we compared the compressive deconvolution results with our method to those obtained with a sequential approach. The latter recovers in a first step the blurred US image from the CS measurements and reconstructs in a second step the TRF by deconvolution.

Two ultrasound data sets were generated, as shown in Figs. 2 and 3. They were obtained by 2D convolution between spatially invariant PSFs and the TRFs. For the first simulated image, the same PSF as in the previous section was simulated and the TRF corresponds to a simple medium representing a round hypoechoic inclusion into a homogeneous medium. The scatterer amplitudes were random variables distributed according to a GGD with the shape parameter set to 1. The second data set is one of the examples proposed by the Field II simulator [46], mimicking a kidney tissue. The PSF was also




Fig. 2. Simulated US image and its compressive deconvolution results for a CS ratio of 0.4 and a SNR of 40 dB. (a) Original tissue reflectivity function, (b) Simulated US image, (c) Results using the sequential method, (d), (e), (f) Results with the proposed method for p equal to 1.5, 1.3 and 1 respectively.




Fig. 3. Simulated kidney image and its compressive deconvolution results for a CS ratio of 0.2 and a SNR of 40 dB. (a) Original tissue reflectivity function, (b) Simulated US image, (c) Results using the sequential method, (d), (e), (f) Results with the proposed method for p equal to 1.5, 1.3 and 1 respectively.

generated with Field II corresponding to a 4 MHz central frequency and an axial sampling frequency of 40 MHz. It corresponds to a focalized emission (the PSF was measured at the focal point) with a simulated linear probe containing 128 elements. The shape parameter of the GGD used to generate the scatterer amplitudes was set to 1.5 and the number of scatterers was considered sufficiently large (10^6) to ensure fully developed speckle. In both experiments, the compressed measurements were obtained by projecting the RF images on SRM, aiming at reducing the amount of data available.

With the sequential approach, YALL1 [42] was used to process the CS reconstruction following the minimization in (3). The deconvolution step was processed using the Forward-Backward Splitting method [48], [49]. Both the CS reconstruction and the deconvolution procedures were performed with the same priors as the proposed compressive deconvolution approach.

The algorithm stops when the convergence criterion $\| \mathbf{x}^k - \mathbf{x}^{k-1} \| / \| \mathbf{x}^{k-1} \| < 1e^{-3}$ is satisfied. In order to highlight the influence of these hyperparameters on the reconstruction results, we consider the simulated US image in Fig. 2. The PSNR values obtained while varying the values of these hyperparameters are shown in Fig. 4. From Fig. 4, one can observe that the best results are obtained for small values of μ , corresponding to an important weight given to the data attachment term. The




Fig. 4. The impact of hyperparameters on the performance of proposed algorithm on Fig. 2.

best value of α is the one providing the best compromise between the two prior terms considered in (6), promoting minimal ℓ_1 -norm of $H\mathbf{x}$ in the wavelet domain and GGD statistics for \mathbf{x} . The choice of β and γ parameters, used in the augmented Lagrangian function and in the approximation of the ℓ_p -norm proximal operator, have an important impact on the algorithm convergence. Moreover, one may observe that for a given range of values, the choice of γ has less impact on the quality of the results than the other three hyperparameters. Despite different optimal values for each CS ratio, in the results presented through the paper, we considered their values fixed for all the CS ratios. The hyperparameters with our approach were set to $\{\mu, \alpha, \beta, \gamma\} = \{10^{-5}, 2 \times 10^{-1}, 1, 10^{-2}\}$ for the round cyst image and $\{\mu, \alpha, \beta, \gamma\} = \{10^{-5}, 2 \times 10^{-1}, 1 \times 10^3, 10^{-4}\}$ for the simulated kidney image.

The quantitative results in Table II show that the proposed method outperforms the sequential approach, for all the CS ratios and values of p considered. They confirm the visual impression given by Figs. 2 and 3. We should remark that for the first simulated data set, the ℓ_1 -norm gives the best result. This may be explained by the simple geometry of the simulated TRF, namely its sparse appearance. The second data set, more realistic and more representative of experimental situations, shows the interest of using different values of p . It confirms the generality interest of the proposed method, namely its flexibility in the choice of TRF priors.

C. In vivo Study

In this section, we tested our method with two *in vivo* data sets. The experimental data were acquired with a 20 MHz single-element US probe on a mouse bladder (first example) and kidney (second example). Unlike the simulated cases studied previously, the PSF is not known in these experiments and has to be estimated from the data. In this paper, the PSF estimation method presented in [35] has been adopted. The PSF estimation adopted is not iterative and the computational




Fig. 5. From left to right, the original *in vivo* image and its compressive deconvolution results for CS ratios of 1, 0.8, 0.6 and 0.4 respectively with $p = 1$.




Fig. 6. From left to right, the original *in vivo* image and its compressive deconvolution results for CS ratios of 1, 0.8, 0.6 and 0.4 respectively with $p = 1.5$.

TABLE II
QUANTITATIVE RESULTS FOR SIMULATED US IMAGES (SNR = 40 dB)

CS Ratios		Sequential	Proposed ($l_{1.5}$)	Proposed ($l_{1.3}$)	Proposed (l_1)
Figure 2					
80%	PSNR	26.50	24.74	25.29	26.82
	SSIM	75.01	73.91	77.66	79.45
60%	PSNR	25.96	24.44	24.74	26.03
	SSIM	68.59	69.37	74.72	76.26
40%	PSNR	23.38	24.21	24.57	25.28
	SSIM	47.60	62.58	71.86	72.78
20%	PSNR	21.10	23.72	24.42	24.77
	SSIM	36.07	50.34	66.48	70.44
Figure 3					
80%	PSNR	26.06	26.71	26.72	26.69
	SSIM	45.99	56.81	56.84	56.71
60%	PSNR	25.44	26.38	26.31	26.29
	SSIM	38.86	54.14	53.90	53.80
40%	PSNR	25.37	25.89	25.95	25.97
	SSIM	34.61	50.22	50.51	50.61
20%	PSNR	24.96	25.22	25.20	25.12
	SSIM	30.89	41.41	41.32	40.97

TABLE III
CNR ASSESSMENT FOR *IN VIVO* DATA

Figure	Original CNR	p values	CS ratios			
			100%	80%	60%	40%
Fig.6	1.106	$p = 1$	1.748	1.546	1.367	1.333
		$p = 1.5$	1.690	1.424	1.304	1.287
Fig.7	1.316	$p = 1$	2.373	2.162	1.895	1.434
		$p = 1.5$	2.317	2.082	1.905	1.451

time for this pre-processing step is negligible compared to the reconstruction process. The compressive deconvolution results are shown in Figs. 5 and 6 for different CS ratios.

Given that the true TRF is not known in experimental conditions, the quality of the reconstruction results is evaluated using the contrast-to-noise ratio (CNR) [50], defined as

$$\text{CNR} = \frac{|\mu_1 - \mu_2|}{\sqrt{\sigma_1^2 + \sigma_2^2}} \quad (24)$$

where μ_1 and μ_2 are the mean of pixels located in two regions extracted from the image while σ_1 and σ_2 are the standard deviations of the same blocks. The two regions selected for the computation of the CNR are highlighted by the two red rectangles in Figs. 5(a) and 6(a). Table III gives the CNR assessment for these two *in vivo* data sets with different CS ratios and p values. Given

the sparse appearance of the bladder image in Fig. 5(a), the best result was obtained for p equal to 1. However, the complexity of the tissue structures in the kidney image in Fig. 6 results into better results for p larger than 1. Nevertheless, both the visual impression and the CNR results show the ability of our method to both recover the image from compressive measurements and to improve its contrast compared to the standard US image. In particular, we may remark the improved contrast of the structures inside the kidney on our reconstructed images compared to the original one.

IV. CONCLUSION

This paper introduced an ADMM-based compressive deconvolution framework for ultrasound imaging systems. The main benefit of our approach is its ability to reconstruct enhanced ultrasound RF images from compressed measurements, by inverting a linear model combining random projections and 2D convolution. Compared to a standard compressive sampling reconstruction that operates in the sparse domain, our method solves a regularized inverse problem combining the data attachment and two prior terms. One of the regularizers promotes minimal ℓ_1 -norm of the target image transformed by 2D convolution with a bandlimited ultrasound PSF. The second one is seeking for imposing GGD statistics on the tissue reflectivity function to be reconstructed. Simulation results in Appendix C on the standard Shepp-Logan phantom show the superiority of our method, both in accuracy and in computational time, over a recently

published compressive deconvolution approach. Moreover, we show that the proposed joint CS and deconvolution approach is more robust than an intuitive technique consisting of first reconstructing the RF data and second deconvolving it. Finally, promising results on *in vivo* data demonstrate the effectiveness of our approach in practical situations. We emphasize that the 2D convolution model may not be valid over the entire image because of the spatially variant PSF. While in this paper we focused on compressive image deconvolution based on spatially invariant PSF, a more complicated global model combining several local shift invariant PSFs represents an interesting perspective of our approach. Our future work includes: I) the consideration of blind deconvolution techniques able to estimate (update) the spatially variant or invariant PSF during the reconstruction process, II) automatic techniques for choosing the optimal parameter p used to regularize the tissue reflectivity function, III) extend our method to 3D ultrasound imaging, IV) evaluate other existing setups to obtain the random compressed measurements further adapted to accelerate the frame rate instead of only reducing the amount of acquired data, V) consider the compressed deconvolution of temporal image sequences by taking advantage of the information redundancy and by including in our model the PSF frame-to-frame variation caused by strong clutters in *in vivo* scenarios, VI) evaluate our approach on more experimental data.

APPENDIX A

PROPOSED COMPRESSIVE DECONVOLUTION WITH GENERALIZED TV

Although the generalized total variation (TV) used in [26] is not suitable for ultrasound images, it may have great interest in other (medical) application dealing with piecewise constant images. As suggested in [26], the generalized TV is given by:

$$\sum_{d \in D} 2^{1-o(d)} \sum_i |\Delta_i^d(\mathbf{x})|^p \quad (25)$$

where $o(d) \in \{1, 2\}$ denotes the order of the difference operator $\Delta_i^d(\mathbf{x})$, $0 < p < 1$, and $d \in D = \{h, v, hh, vv, hv\}$. $\Delta_i^h(\mathbf{x})$ and $\Delta_i^v(\mathbf{x})$ correspond, respectively, to the horizontal and vertical first order differences, at pixel i , that is, $\Delta_i^h(\mathbf{x}) = u_i - u_{l(i)}$ and $\Delta_i^v(\mathbf{x}) = u_i - u_{a(i)}$, where $l(i)$ and $a(i)$ denote the nearest neighbors of i , to the left and above, respectively. The operators $\Delta_i^{hh}(\mathbf{x})$, $\Delta_i^{vv}(\mathbf{x})$, $\Delta_i^{hv}(\mathbf{x})$ correspond, respectively, to horizontal, vertical and horizontal-vertical second order differences, at pixel i .

Replacing the ℓ_p -norm by the generalized TV in our compressive deconvolution scheme results in a modified \mathbf{x} update step, that turns in solving:

$$\mathbf{x}^k = \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} \alpha \sum_{d \in D} 2^{1-o(d)} \sum_i |\Delta_i^d(\mathbf{x})|^p - \lambda_2^{k-1} (\Psi \mathbf{a}^{k-1} - H \mathbf{x}) + \frac{\beta}{2} \|\Psi \mathbf{a}^{k-1} - H \mathbf{x}\|_2^2$$

Similarly to the first step of the method in [26], the equation above can be solved iteratively by:

$$\mathbf{x}^{k,l} = \left[\beta H^t H + \alpha p \sum_d 2^{1-o(d)} (\Delta^d)^t B_d^{k,l} (\Delta^d) \right]^{-1} \times \left[\beta H^t \Psi \mathbf{a}^{k-1} - H^t \lambda_2^{k-1} \right] \quad (26)$$

where l is the iteration number in the process of updating \mathbf{x} , $B_d^{k,l}$ is a diagonal matrix with entries Δ^d is the convolution matrix (BCCB matrix) of the difference operator $\Delta_i^d(\cdot)$ and $B_d^{k,l}(i, i) = (v_{d,i}^{k,l})$, which is updated iteratively by:

$$v_{d,i}^{k,l+1} = [\Delta_i^d(\mathbf{x}^{k,l})]^2 \quad (27)$$

When a stopping criterion is met, we can finally get an update of \mathbf{x} .

APPENDIX B

PROXIMAL OPERATOR

The proximal operator of a function f is defined for $x^0 \in \mathbb{R}^N$ by:

$$\operatorname{prox}_f(x^0) = \underset{\mathbf{x} \in \mathbb{R}^N}{\operatorname{argmin}} f(\mathbf{x}) + \frac{1}{2} \|\mathbf{x} - x^0\|_2^2 \quad (28)$$

When $f = K|x|^p$, the corresponding proximal operator has been given by [39]:

$$\operatorname{prox}_{K|x|^p}(x^0) = \operatorname{sign}(x^0)q \quad (29)$$

where $q \geq 0$ and

$$q + pKq^{p-1} = |x^0| \quad (30)$$

It is obvious that the proximal operator of $K|x|$ is a soft thresholding, which is equal to:

$$\operatorname{prox}_{K|x|}(x^0) = \max\{|x^0| - K, 0\} \frac{x^0}{|x^0|} \quad (31)$$

When $p \neq 1$, we used Newton's method to obtain its numerical solution, i.e., the value of q .

APPENDIX C

COMPARISON WITH THE COMPRESSIVE DECONVOLUTION METHOD IN [26]

In this Appendix we show an experiment aiming to evaluate the performance of the proposed approach compared to the one in [26], denoted by CD_Amizic herein. The comparison results are obtained on the standard 256×256 Shepp-Logan phantom. The measurements have been generated in a similar manner as in [26], i.e., the original image was normalized, degraded by a 2D Gaussian PSF with a 5-pixel variance, projected onto a structured random matrix (SRM) to generate the CS measurements and corrupted by an additive Gaussian noise. We should remark that in [26] the compressed measurements were originally generated using a Gaussian random matrix. However, we




Fig. 7. Shepp-logan image and its compressive deconvolution results for a SNR of 40 dB. (a) Original image, (e) Blurred image, (b), (c), (d) Compressive deconvolution results with CD_Amizic for CS ratios of 0.8, 0.6 and 0.4, (f), (g), (h) Compressive deconvolution results with the proposed method for CS ratios of 0.8, 0.6 and 0.4.

TABLE IV
PSNR ASSESSMENT FOR SHEPP-LOGAN PHANTOM

SNR	CS ratios	20%	40%	60%	80%
40dB	CD_Amizic	23.04	24.88	25.30	25.51
	Proposed method	24.09	25.38	26.26	26.91
30dB	CD_Amizic	22.61	24.05	24.40	24.55
	Proposed method	23.92	25.12	25.82	26.33




Fig. 8. Mean reconstruction running time for 10 experiments conducted for each CS ratio for a SNR of 40 dB.

have found that the reconstruction results with CD_Amizic are slightly better when using a SRM compared to the PSNR results reported in [26]. Both methods were based on the generalized TV to model the image to be estimated and the 3-level Haar wavelet transform as sparsifying basis Ψ . With our method, the hyperparameters were set to $\{\alpha, \mu, \beta\} = \{10^{-1}, 10^{-5}, 10^2\}$. The same hyperparameters as reported in [26] were used for CD_Amizic. Both algorithms based on the non-blind deconvolution (PSF is supposed to be known) and used the same stopping criteria.

Fig. 7 shows the original Shepp-Logan image, its blurred version and a series of compressive deconvolution reconstructions using both our method and CD_Amizic for CS ratios running from 0.4 to 0.8 and a SNR of 40 dB. Table IV regroups

the PSNRs obtained with our method and with CD_Amizic for two SNRs and for four CS ratios from 0.2 to 0.4. In each case, the reported PSNRs are the mean values of 10 experiments. We may observe that our method outperforms CD_Amizic in all the cases, allowing a PSNR improvement in the range of 0.5 to 2 dB. Moreover, Fig. 8 shows the computational times with CD_Amizic and the proposed method, obtained with Matlab implementations (for CD_Amizic, the original code provided by the authors of [26] has been employed on a standard desktop computer (Intel Xeon CPU E5620@2.40 GHz, 4.00G RAM). We notice that our approach is less time consuming than CD_Amizic for all the CS ratios considered.

ACKNOWLEDGMENT

The authors would like to thank Prof. Rafael Molina for providing the compressive deconvolution code used for comparison purpose in this paper.

REFERENCES

- [1] T. L. Szabo, *Diagnostic Ultrasound Imaging: Inside Out*. New York: Academic, 2004.
- [2] M. Tanter and M. Fink, "Ultrafast imaging in biomedical ultrasound," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 61, no. 1, pp. 102–119, Jan. 2014.
- [3] A. Achim, B. Buxton, G. Tzagkarakis, and P. Tsakalides, "Compressive sensing for ultrasound RF echoes using a-stable distributions," in *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc.*, 2010, pp. 4304–4307.
- [4] C. Quinsac, A. Basarab, and D. Kouamé, "Frequency domain compressive sampling for ultrasound imaging," *Adv. Acoust. Vibrat.* vol. 2012, 2012 [Online]. Available: <http://dx.doi.org/10.1155/2012/231317>
- [5] T. Chernyakova and Y. C. Eldar, "Fourier-domain beamforming: The path to compressed ultrasound imaging," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 61, no. 8, pp. 1252–1267, Aug. 2014.
- [6] H. Liebgott, A. Basarab, D. Kouamé, O. Bernard, and D. Friboulet, "Compressive sensing in medical ultrasound," in *Proc. IEEE Int. Ultrason. Symp.*, 2012, pp. 1–6.
- [7] H. Liebgott, R. Prost, and D. Friboulet, "Pre-beamformed RF signal reconstruction in medical ultrasound using compressive sensing," *Ultrasonics*, vol. 53, no. 2, pp. 525–533, 2013.
- [8] M. F. Schiffner and G. Schmitz, "Pulse-echo ultrasound imaging combining compressed sensing and the fast multipole method," in *Proc. IEEE Int. Ultrason. Symp.*, 2014, pp. 2205–2208.
- [9] J. Richy, H. Liebgott, R. Prost, and D. Friboulet, "Blood velocity estimation using compressed sensing," in *Proc. IEEE Int. Ultrason. Symp.*, 2011, pp. 1427–1430.
- [10] D. L. Donoho, "Compressed sensing," *IEEE Trans. Inf. Theory*, vol. 52, no. 4, pp. 1289–1306, 2006.

- [11] E. J. Candès, J. Romberg, and T. Tao, "Robust uncertainty principles: Exact signal reconstruction from highly incomplete frequency information," *IEEE Trans. Inf. Theory*, vol. 52, no. 2, pp. 489–509, 2006.
- [12] E. J. Candès and M. B. Wakin, "An introduction to compressive sampling," *IEEE Signal Process. Mag.*, vol. 25, no. 2, pp. 21–30, 2008.
- [13] O. Lortintiu, H. Liebgott, M. Alessandrini, O. Bernard, and D. Friboulet, "Compressed sensing reconstruction of 3D ultrasound data using dictionary learning and line-wise subsampling," *IEEE Trans. Med. Imag.*, to be published.
- [14] N. Dobigeon, A. Basarab, D. Kouamé, and J.-Y. Tourneret, "Regularized Bayesian compressed sensing in ultrasound imaging," in *Proc. 20th Eur. Signal Process. Conf.*, 2012, pp. 2600–2604.
- [15] A. Basarab, A. Achim, and D. Kouamé, "Medical ultrasound image reconstruction using compressive sampling and l_p -norm minimization," in *SPIE Med. Imag.*, 2014, pp. 90401H–90401H.
- [16] T. Taxt and J. Strand, "Two-dimensional noise-robust blind deconvolution of ultrasound images," *IEEE Trans. Ultrason., Ferroelectr. Freq. Control*, vol. 48, no. 4, pp. 861–866, Jul. 2001.
- [17] O. Michailovich and A. Tannenbaum, "Blind deconvolution of medical ultrasound images: A parametric inverse filtering approach," *IEEE Trans. Image Process.*, vol. 16, no. 12, p. 3005, Dec. 2007.
- [18] R. Morin, S. Bidon, A. Basarab, and D. Kouame, "Semi-blind deconvolution for resolution enhancement in ultrasound imaging," in *Proc. 20th IEEE Int. Conf. Image Process.*, Sep. 2013, pp. 1413–1417.
- [19] N. Zhao, A. Basarab, D. Kouamé, and J.-Y. Tourneret, "Restoration of ultrasound images using a hierarchical Bayesian model with a generalized Gaussian prior," in *Int. Conf. Image Process.*, 2014, pp. 4577–4581.
- [20] M. Alessandrini *et al.*, "A restoration framework for ultrasonic tissue characterization," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 58, no. 11, pp. 2344–2360, Nov. 2011.
- [21] M. Alessandrini, "Statistical methods for analysis and processing of medical ultrasound: Applications to segmentation and restoration," Ph.D. dissertation, Univ. Bologna, Bologna, Italy, 2011.
- [22] J. A. Jensen, J. Mathorne, T. Gravesen, and B. Stage, "Deconvolution of in vivo ultrasound B-mode images," *Ultrason. Imag.*, vol. 15, no. 2, pp. 122–133, 1993.
- [23] J. Ma and F.-X. Le Dimet, "Deblurring from highly incomplete measurements for remote sensing," *IEEE Trans. Geosci. Remote Sens.*, vol. 47, no. 3, pp. 792–802, Mar. 2009.
- [24] L. Xiao, J. Shao, L. Huang, and Z. Wei, "Compounded regularization and fast algorithm for compressive sensing deconvolution," in *Proc. 6th Int. Conf. Image Graph.*, 2011, pp. 616–621.
- [25] M. Zhao and V. Saligrama, "On compressed blind deconvolution of filtered sparse processes," in *Proc. IEEE Int. Conf. Acoust. Speech Signal Process.*, 2010, pp. 4038–4041.
- [26] B. Amizic, L. Spinoulas, R. Molina, and A. K. Katsaggelos, "Compressive blind image deconvolution," *IEEE Trans. Image Process.*, vol. 22, no. 10, pp. 3994–4006, Oct. 2013.
- [27] L. Spinoulas, B. Amizic, M. Vega, R. Molina, and A. K. Katsaggelos, "Simultaneous Bayesian compressive sensing and blind deconvolution," in *Proc. 20th Eur. Signal Process. Conf.*, 2012, pp. 1414–1418.
- [28] C. Hegde and R. G. Baraniuk, "Compressive sensing of streams of pulses," in *Proc. 47th Annu. Allerton Conf. Commun., Control, Comput.*, 2009, pp. 44–51.
- [29] C. Hegde and R. Baraniuk, "Sampling and recovery of pulse streams," *IEEE Trans. Signal Process.*, vol. 59, no. 4, pp. 1505–1517, Apr. 2011.
- [30] S. Boyd, N. Parikh, E. Chu, B. Peleato, and J. Eckstein, "Distributed optimization and statistical learning via the alternating direction method of multipliers," *Found. Trends Mach. Learn.* vol. 3, no. 1, pp. 1–122, Jan. 2011 [Online]. Available: <http://dx.doi.org/10.1561/22000000016>
- [31] X. Zhao, C. Chen, and M. Ng, "Alternating direction method of multipliers for nonlinear image restoration problems," *IEEE Trans. Image Process.*, vol. 24, no. 1, pp. 33–43, Jan. 2015.
- [32] R. Jirik and T. Taxt, "Two-dimensional blind bayesian deconvolution of medical ultrasound images," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 55, no. 10, pp. 2140–2153, Oct. 2008.
- [33] C. Yu, C. Zhang, and L. Xie, "A blind deconvolution approach to ultrasound imaging," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 59, no. 2, pp. 271–280, 2012.
- [34] A. Achim, A. Basarab, G. Tzagkarakis, P. Tsakalides, and D. Kouame, "Reconstruction of ultrasound RF echoes modeled as stable random variables," *IEEE Trans. Comput. Imag.*, vol. 1, no. 2, pp. 86–95, Jun. 2015.
- [35] O. V. Michailovich and D. Adam, "A novel approach to the 2-d blind deconvolution problem in medical ultrasound," *IEEE Trans. Med. Imag.*, vol. 24, no. 1, pp. 86–104, Jan. 2005.
- [36] J.-C. Pesquet and N. Pustelnik, "A parallel inertial proximal optimization method," *Pacific J. Optimizat.*, vol. 8, no. 2, pp. 273–305, 2012.
- [37] N. Pustelnik, C. Chau, and J.-C. Pesquet, "Parallel proximal algorithm for image restoration using hybrid regularization," *IEEE Trans. Image Process.*, vol. 20, no. 9, pp. 2450–2462, Sep. 2011.
- [38] N. Pustelnik, J. Pesquet, and C. Chau, "Relaxing tight frame condition in parallel proximal methods for signal restoration," *IEEE Trans. Signal Process.*, vol. 60, no. 2, pp. 968–973, Feb. 2012.
- [39] P. L. Combettes and J.-C. Pesquet, "Proximal splitting methods in signal processing," in *Fixed-Point Algorithms for Inverse Problems in Science and Engineering*. New York: Springer, 2011, pp. 185–212.
- [40] E. Chouzenoux, J.-C. Pesquet, and A. Repetti, "Variable metric forward-backward algorithm for minimizing the sum of a differentiable function and a convex function," *J. Optimizat. Theory Appl.*, vol. 162, no. 1, pp. 107–132, 2014.
- [41] B. He, L.-Z. Liao, D. Han, and H. Yang, "A new inexact alternating directions method for monotone variational inequalities," *Math. Program.*, vol. 92, no. 1, pp. 103–118, 2002.
- [42] J. Yang and Y. Zhang, "Alternating direction algorithms for l_1 -problems in compressive sensing," *SIAM J. Sci. Comput.*, vol. 33, no. 1, pp. 250–278, 2011.
- [43] W. Deng, W. Yin, and Y. Zhang, "Group sparse optimization by alternating direction method," vol. 8858, 2013, p. 88 580R–88 580R–15.
- [44] T. T. Do, L. Gan, N. H. Nguyen, and T. D. Tran, "Fast and efficient compressive sensing using structurally random matrices," *IEEE Trans. Signal Process.*, vol. 60, no. 1, pp. 139–154, Jan. 2012.
- [45] J. Ng, R. Prager, N. Kingsbury, G. Treece, and A. Gee, "Wavelet restoration of medical pulse-echo ultrasound images in an EM framework," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 54, no. 3, pp. 550–568, 2007.
- [46] J. A. Jensen, "A model for the propagation and scattering of ultrasound in tissue," *Acoust. Soc. Am. J.*, vol. 89, no. 1, pp. 182–190, 1991.
- [47] Z. Wang, A. C. Bovik, H. R. Sheikh, and E. P. Simoncelli, "Image quality assessment: From error visibility to structural similarity," *IEEE Trans. Image Process.*, vol. 13, no. 4, pp. 600–612, Apr. 2004.
- [48] P. L. Combettes and V. R. Wajs, "Signal recovery by proximal forward-backward splitting," *Multiscale Model. Simulat.*, vol. 4, no. 4, pp. 1168–1200, 2005.
- [49] H. Raguét, J. Fadili, and G. Peyré, "A generalized forward-backward splitting," *SIAM J. Imag. Sci.*, vol. 6, no. 3, pp. 1199–1226, 2013.
- [50] A. Lyshchik *et al.*, "Elastic moduli of thyroid tissues under compression," *Ultrason. Imag.*, vol. 27, no. 2, pp. 101–110, 2005.