

# NASA TECH BRIEF

## Marshall Space Flight Center



NASA Tech Briefs announce new technology derived from the U.S. space program. They are issued to encourage commercial application. Tech Briefs are available on a subscription basis from the National Technical Information Service, Springfield, Virginia 22151. Requests for individual copies or questions relating to the Tech Brief program may be directed to the Technology Utilization Office, NASA, Code KT, Washington, D.C. 20546.

### Improved Electrodes For Skin Contacts

#### The problem:

Measurements of body functions by the use of electrocardiograms or impedance plethysmographs are frequently distorted by the simple motion of the skin as a result of muscular movements.

#### The solution:

Thick, flexible electrodes with appropriate metal surfaces prevent unnecessary skin motion. The electrodes provide sufficient radial pressure directed toward the body surface to depress the skin a noticeable portion of its normal resilient thickness.


#### How it's done:

The electrode has a core of single-conductor, shielded cable (no insulation over shielding) with an outer diameter of 3/32 inch (2.5 mm), wrapped with a 0.25 mil (0.006 mm) thickness of Mylar. Aluminum is deposited on both sides to a total thickness of 0.40 mil (0.01 mm). Strips, 1/2-inch (1.3-cm) wide, of the aluminized Mylar are wound around the core in spiral fashion with approximately two-thirds strip-width overlap on each winding. Crimp-on connectors are attached to one end of the electrode for electrical connection to external equipment. The electrodes are held on the skin by a 2-inch wide elasticized cloth. The cloth, normally stretched 20% to 50% of the way around the limb is fastened to itself with Velcro material. To hold the electrode onto the elasticized material, loops of Teflon insulated wire are attached to the elastic, and the electrode is threaded through the loops. This allows the elastic material to be stretched over the electrodes without binding.

#### Notes:

1. Information concerning this innovation may be of interest to physicians and designers and manufacturers of medical instrumentation.

2. Requests for further information may be directed to:  
Technology Utilization Officer  
Marshall Space Flight Center  
Code A&PS-TU  
Marshall Space Flight Center, Alabama 35812  
Reference: B72-10698



#### Patent status:

Inquiries concerning rights for the commercial use of this invention should be addressed to:

Patent Counsel  
Marshall Space Flight Center  
Code A&PS-PAT  
Marshall Space Flight Center, Alabama 35812

Source: R. R. Lattanzi and  
J. G. Castle of  
University of Alabama  
under contract to  
Marshall Space Flight Center  
(MFS-21926)

Category 05