

82257

A NEW GENERATION FIBER OPTIC PROBE : CHARACTERIZATION OF BIOLOGICAL FLUIDS, PROTEIN CRYSTALS, AND OPHTHALMIC DISEASES

Rafat R. Ansari and Kwang I. Suh

NASA Lewis Research Center / CWRU

Mail Stop 333-1, 21000 Brookpark Road, Cleveland, Ohio 44135

Phone: (216) 433-5008, Fax: (216) 977-7138, e-mail: ransari@lerc.nasa.gov

ABSTRACT

A new fiber optic probe developed for determining transport properties of sub-micron particles in fluids experiments in a microgravity environment has been applied to characterize particulate dispersions/suspensions in various challenging environments which have been hitherto impossible. The probe positioned in front of a sample delivers a low power light (few nW - 3mW) from a laser and guides the light which is back scattered by the suspended particles through a receiving optical fiber to a photo detector and to a digital correlator. The probe provides rapid determination of macromolecular diffusivities and their respective size distributions. It has been applied to characterize various biological fluids, protein crystals, and ophthalmic diseases.

INTRODUCTION

In the past few years dynamic light scattering (DLS) instrumentation has embraced several new and innovative technological advances. The immediate impact of these have resulted in the miniaturization of DLS instrumentation¹. One of those advances is the use of fiber optics in DLS. Especially, the use of monomode optical fibers in designing and fabricating a DLS system was first initiated by Brown² in his 90° scattering system. Later, a lensless single fiber probe was used by Weiss and Horn³ to study highly concentrated systems (1-40% w/v) without the adverse effect of multiple light scattering. However, since the scattering volume has to be positioned right at the tip of the fiber, the probe had to be immersed in the sample, making it no longer a non-invasive measurement. Also, the heterodyning effect caused by the internal reflection at the tip of the fiber necessitated a great deal of care in fabrication, and it could not be used for dilute samples with concentrations below 1% w/v or particles smaller than 30 nm because the reflected signal dominated the scattered signal. A lensless back-scatter fiber optic probe was developed to study concentrated particulate dispersions by Dhadwal et al.⁴ Subsequently, Ansari et al.⁵ and Dhadwal et al.⁶ have shown the utility of their probe to the studies of cataractogenesis in excised bovine and cadaver eye lenses, and in live animals⁷. Unfortunately this probe has certain limitations and posed severe constraints in probing different parts of the eye, mainly due to a large scattering volume and a short penetration depth. Ansari and Suh⁸ recently developed a new DLS probe to study nucleation and aggregation phenomena during protein crystal growth in space (microgravity) experiments. The new probe is compact, rugged, and provides accurate particle size determination of variety of colloidal dispersions in 5 seconds at extremely low laser power levels. Furthermore, it does not require any vibration isolation, index matching, or optical alignment, and it can be used with many different shapes and sizes of sample containers. These include regular spectroscopic cuvettes, capillary tubes, common laboratory utensils (beakers, graduated cylinders, conical flasks, glass/clear plastic container vessels), concentric cells containing two different solutions, and hanging fluid droplets of colloidal particles. The new probe is applied to several different applications which require characterization of various suspensions/dispersions in a wide range of particle sizes and concentrations from very dilute (water-like) dispersions to highly concentrated (milk-like) suspensions in a variety of challenging situations. These applications include monitoring protein crystallization in different hanging drop configurations suitable for microgravity experiments⁹, complete eye diagnostics¹⁰⁻¹², zeolite crystal growth¹³, study of clay-polymer interactions, particle sizing in a rotating cell to avoid sedimentation due to gravity, characterization of food colloids, microemulsion and micellar systems, and tissue and skin analysis.

Protein crystals have been experimentally grown in 28 US Space Shuttle flights. One preferred method is to grow these crystals in hanging drops (~30 μ L) in a reduced gravity environment offered by the Space Shuttle orbiter or space station. In laboratory settings, DLS has been used to study protein solutions and crystal growth processes by Casay and Wilson¹⁴, Malkin and McPherson¹⁵, and references therein. However, in hanging droplets and in microgravity experiments, elaborate instrumentation and optical alignment problems have made in-situ and on-line applications difficult. Ansari et al.^{8,9} have demonstrated that such experiments are now feasible. The new probe has

been successfully used in various earth and space-bound (microgravity) protein crystallization system configurations⁹.

Cataracts remain the major cause of blindness affecting about 50 million people each year worldwide. It is estimated that over \$5 billion will be spent this year in treating cataract patients in the United States alone¹⁶. There is no medical treatment to prevent or halt the progression of a cataract; nor is there any way to reverse a cataract once it is formed. The only known treatment is surgical. However, a medical treatment could be possible if we understand how a cataract forms and what makes it grow¹⁶. In order to find a medical treatment for cataracts, first, we must be able to detect a growing cataract in early stages of formation. The ability of such a detection will be useful in patient monitoring and in the development and testing of possible "anticataract" drugs. The technique of dynamic light scattering (DLS) or quasi-elastic light scattering (QELS) was first applied to the study of cataractogenesis in the pioneering work of Tanaka and Benedek¹⁷. An extensive review of using QELS to study cataracts has been given by Bursell et al¹⁸. DLS/QELS being a non-invasive and quantitative technique seems to hold promising potential in its use as a routine ophthalmic device. But its commercial scope as an ophthalmic diagnostic tool in clinical settings hitherto has not been materialized. Since the original application of QELS, some twenty years ago by Tanaka and Benedek¹⁶, every subsequent study focused on rather easily accessible part of the eye i.e. the aqueous humor and the lens. Most recently Ansari et al¹², in the USA and Rovati et al¹⁹, in Switzerland reported the first experiments directed to the studies of the vitreous humor. Ansari et al^{10,11}, have also shown recently that the new probe alleviates many major concerns discussed earlier when used as an ophthalmic diagnostics device and is much superior in performance when compared with the earlier reported work⁵⁻⁷. Furthermore, the major problem of probing different parts of the eye has been solved. In this paper we report the characterization of several solutions at different particle sizes (7-800nm in diameter) and concentrations (0.0001-10.0% w/v), and the characterization of bovine lenses and the eyes of live animals (rabbits and mice).

EXPERIMENTAL PROCEDURE AND SETUP

A fiber optic probe comprising two monomode optical fibers and two GRIN micro lenses, as illustrated in

Figure 1, provides a compact and remote means of studying the dynamical characteristics of the macromolecules in a sample. A 13 mm diameter fiber optic probe contains the necessary optics to perform DLS measurements at a scattering angle of 161.5°. Two monomode optical fibers, each housed in a stainless steel ferrule, are mounted into a separate stainless steel housing. An air gap (0-0.5 mm) is intentionally left between the fiber housing and the lens housing in order to produce a tightly focused spot in the scattering volume. The two optical fibers in their housings are aligned and fixed into position off-axis with the GRIN lens. The two housings are placed inside





Figure 1. A new fiber optic DLS probe.

a third (outer) housing made of stainless steel, and the back end of the housing is covered with a heat-shrink tubing. The two free ends of the optical fibers were terminated with FC/PC-type male connectors for easy mating with the laser and an avalanche photodiode detector (APD). The DLS data was analyzed using the commercial software provided by the Brookhaven Instruments Company of New York.

RESULTS

First, we report a series of DLS experiments on polystyrene solutions at several different concentrations (0.0001 - 10.0% w/v) and particle sizes (32 - 800nm in diameter), and bovine serum albumin (BSA) solutions at five different concentrations (2 - 10 %), all with 30 second experiment duration. The resulting average particle sizes are shown in figure 2. The new probe accurately and reliably measures the particle size from 7nm to 800nm in diameter with concentrations ranging from 0.0001 to 10.0% w/v. Figure 3 shows DLS experiments at two different experiment durations (5 seconds and 30 seconds) using 1% BSA solution. The resulting time correlation functions (TCF) and particle size distributions (see inset) display the excellent quality of data collected even with experiment duration as short as 5 seconds.

Figure 4 shows a 3-D DLS scan of an intact bovine eye lens, which shows detailed structure of the lens in terms of α crystalline size and its distribution in the lens tissue. In a normal eye lens, α , β , and γ crystallines (proteins) concentration is ~35 wt.%. Since α crystallines are the largest (molecular weight $\sim 1 \times 10^6$ daltons) in size they scatter most of the incident laser light. In the resulting average particle sizes, we find a specific trend as we move inside the lens and away from the anterior cortex. We see gradual increase in the particle size as we move from the anterior cortex to the nucleus of the lens. The same trend can also be observed as we move from the peripheral region to the nucleus. The size gradually decreases as we move out of the nucleus region and into the posterior cortex.




Figure 2. Calibration result with suspensions/dispersions of polystyrene and bovine serum albumin (BSA).




Figure 3. Normalized time correlation function (TCF) and particle size distribution for a 2% BSA solution using a laser power of 1.0mW at 5 and 30 second experiment duration.




Figure 4. 3-D scanning of a bovine eye lens. The Z(PD)-axis represents the distance (penetration depth) from the front surface of the lens to the measurement point inside the lens.




Figure 5. In-vivo cataractogenesis (DLS) measurements in Philly mice.

The Philly mouse is a good animal model of hereditary cataract²⁰. Philly mice develop cataracts spontaneously between day 26 and 33 after their birth. This mouse model can be useful in the screening of anti-cataract drugs. We report a series of DLS experiments performed in Philly mice in collaboration with National Eye Institute (NEI) of National Institutes of Health (NIH). We applied our DLS probe to study the progression of cataractogenesis in this mouse model and to show the feasibility and safe use of our device in laboratory animals. Figure 5 shows DLS time correlation data on three Philly mice. It is quite challenging to perform DLS measurements in these animals because these animals have very small eyes. The lens in these animals has a diameter of less than 2 mm. The data includes a 45 day old normal mouse of the control FVB/N strain and two Philly mice roughly 26-29 days old. The eye examinations of these mice conducted with a slit-lamp apparatus concluded a normal (transparent) and two other eyes having trace and mild cataracts. These examinations were conducted within few minutes of the DLS measurements. Each measurement only took 5 seconds at a laser power of 50 μ W. The changing slope of the time correlation functions is an indication of cataractogenesis. We can quantitatively monitor




Figure 6. Size distribution for a 6 month old rabbit vitreous humor. Penetration depth from the corneal surface is 11mm.

cataractogenesis with reasonable reliability and reproducibility (5%-10%)²¹.

The vitreous humor of the eye described by Sebag²² is the least understood part of an eye. We present preliminary DLS measurements made in vivo on a six month old live rabbit. The vitreous body shows strongly two-exponential behavior, i.e. a fast and a slowly diffusing component, consistent with its gel-like properties. We see almost constant value for the fast component in this region. We ascribe this fast component due to the diffusion of hyaluronate molecular coils in water and the slow diffusion component due to the collagen-fibril network. A size distribution of the vitreous is presented in Figure 6. We calculated the molecular weight for the fast component to be 3.78×10^6 daltons. This is consistent with the range ($2-4.5 \times 10^6$) of values given by Balazs and Delinger²³. More in-vivo experimental work is being performed at this time in our laboratory to fully understand the structure of the vitreous humor. We have recently concluded three dimensional (3-D) scans of the bovine vitreous which is reported elsewhere¹².

CONCLUSION

In this report we have briefly described a new DLS probe to characterize various biological fluids, protein crystals, and ophthalmic diseases. The new probe is compact, rugged, non-invasive, and is free of optical alignment. Furthermore, it uses very low laser power, produces data with high spatial coherence, and exhibits no adverse effect of multiple light scattering even with highly concentrated systems (up to 10% w/v with particle size of 800nm in diameter). The probe has been and is being successfully used for several different applications in various challenging environments including monitoring protein nucleation, aggregation, and crystallization in various space-bound (microgravity) hanging drop apparatus⁹, eye diagnostics experiment on live animals as well as excised eyes¹⁰⁻¹², a zeolite crystal growth experiment¹³, experiments in a rotating cell to minimize the sedimentation due to the gravitational effect, and clay-polymer interaction in a graduated cylinder.

ACKNOWLEDGMENTS

The authors are grateful to the Microgravity Science and Applications Division (MSAD), code UG of the NASA headquarters for supporting this research. This work was completed under a NASA grant NCC 166-3/CWRU. The Philly mice were made available by S. Zigler and P. Russell of the National Eye Institute, Bethesda, MD. Kwang Suh would like to thank NASA Lewis Research Center and the National Research Council for the award of a postdoctoral research fellowship.

REFERENCES

1. R. G. W. Brown, J.G. Burnett, J. Mansbridge, and C. I. Moir, "Miniature Laser Light Scattering Instrumentation for Particle Size Analysis," *Appl. Opt.* **29**, 4159-4169, 1990.
2. R. G. W. Brown, "Optical fibre sensing using light scattering techniques," *J. Phys. E: Sci. Instrum.*, **20**, 1312-1320 (1987).
3. Weiss and D. Horn, "Single-mode fibers in fiber-optic quasielastic light scattering: A study of the dynamics of concentrated latex dispersions," *J. Chem. Phys.*, **94**, 6429-6443 (1991).
4. Dhadwal, H.S., Ansari, R.R., Meyer, W.V., "A fiber Optic Probe for Particle Sizing in Concentrated Suspensions", *Rev. Sci. Instrum.* **62** (12), December 1991.
5. Ansari, R.R., Dhadwal, H.S., Campbell, M.C.W., and DellaVecchia, M.A., "A Fiber Optic Sensor for Ophthalmic Refractive Diagnostics", *Proc. of Fiber Optic Medical and Fluorescent Sensors and Applications*, January 23-24, 1992, Los Angeles, CA., SPIE Vol. 1648.
6. Dhadwal, H.S., Ansari, R.R., DellaVecchia, M.A., "Coherent Fiber Optic Sensor for Early Detection of Cataractogenesis in a Human Eye Lens", *J. Opt. Engineering*, Vol. 32, No. 2, February 1993.
7. H.S. Dhadwal, R.R. Ansari, M.A. DellaVecchia, S. Dubin, "Fiber optic system for in vivo sizing of proteins in animal eye lenses", *proc. Ophthalmic Technologies V*, SPIE, vol. 2393, pp. 227-236, 4-5 February 1995, San Jose, CA.
8. R. R. Ansari and K. I. Suh, "Sizing of colloidal particles and protein molecules in a hanging fluid drop," *Proc. Biomedical Optoelectronics in Clinical Chemistry and Biotechnology*, SPIE, Bios Europe '95, September 12-16, 1995, **2629** (23), Barcelona, Spain.
9. R. R. Ansari, K. I. Suh, A. Arabshahi, W. W. Wilson, T. L. Bray, and L. J. DeLucas, "A fiber optic probe for monitoring protein aggregation, nucleation, and crystallization," accepted for publication, *J. Crystal Growth*, 1996.

10. R. R. Ansari, K. I. Suh, M. A. DellaVecchia, and S. Dubin, "Early diagnosis of cataracts using a fiber optic system," in audio/video Proc. Cataract and Refractive Surgery: Current Issues, Session 2-B, ASCRS Symposium on Cataract, IOL and refractive surgery, April 1-5, 1995, San Diego, CA.
11. R. R. Ansari, K. I. Suh, M. A. DellaVecchia, and S. Dubin, "Ophthalmic diagnostics using a new dynamic light scattering fiber optic probe," Proc. Medical Application, Conference on laser in Ophthalmology III, SPIE, Bios Europe '95, September 12-16, 1995, **2632** (18), Barcelona, Spain.
12. R. R. Ansari and K. I. Suh, "Three-dimensional scanning of eye (lens and vitreous) using a newly developed dynamic light scattering probe," Proc. Ophthalmic Technologies VI, SPIE, Bios '96, January 27 - February 2, 1996, **2673**, San Jose, CA.
13. K. S. N. Reddy, L. M. Salvati, P. K. Dutta, P. Abel, K. Suh, and R. R. Ansari, "Reverse micelle based growth of zincophosphate sodalite : examination of crystal growth," accepted for publication, J. Phys. Chem., 1996.
14. G. A. Casay and W. W. Wilson, "Laser scattering in a hanging drop vapor diffusion apparatus for protein crystal growth in a microgravity environment," J. Crystal Growth, **122**, 95-101, 1992.
15. A. J. Malkin and A. McPherson, "Light-scattering investigations of nucleation processes and kinetics of crystallization in macromolecular systems," Acta Cryst. **D50**, 385-395, 1994.
16. J. Shulman, "Cataracts : The complete guide from diagnostics to recovery for patients and families," St. Martins Press, New York, 1993.
17. T. Tanaka and G. B. Benedek, "Observation of protein diffusivity in intact human and bovine lenses with application to cataract," Investigative ophthalmology, **14(6)**, 449-456 (1976)
18. S. E. Bursell, P. C. Magnante and L. T. Chylack Jr., "In vivo uses of quasielastic light scattering spectroscopy as a molecular probe in the anterior segment of the eye," in Noninvasive Diagnostic Techniques in Ophthalmology, editor Barry R. Masters, Springer-Verlag, New York (1990)
19. L. Rovati, F. Frankhauser II, and J. Ricka, "Dynamic light scattering spectroscopy of in-vivo human vitreous," Proc. Medical Application, Conference on laser in Ophthalmology III, SPIE, Bios Europe '95, September 12-16, 1995, **2632** (18), Barcelona, Spain.
20. F. Kador, H. N. Fukui, S. Fukushi, H. M. Jernigan Jr., and J. H. Kinoshita, "Philly Mouse: A New Model of Hereditary Cataract", Exp. Eye Res. **30**, 59-68, 1980.
21. R. Ansari and K. I. Suh, "In situ and in vivo diagnostics of biological fluids," Proc. Optical Diagnostics of Biological Fluids, SPIE, Bios '96, January 27 - February 2, 1996, **2678C**, San Jose, CA.
22. J. Sebag, "The Vitreous; structure, function, and pathobiology", Springer-Verlag, 1989.
23. Balazs, J.L. Denlinger, "The Vitreous", in: The Eye, vol. IA, editor, H. Davson, Academic Press, pp. 533-589, 1984.

