

posed to the fast recirculation flow, which washes away newly settled sludge. A dynamic equilibrium is quickly reached between the deposition and removal of sludge.

- Unlike what would happen in a conventional filter, the buildup of sludge in the pipe-cleaner flow field does not continue indefinitely until excessive clogging occurs. Instead, the SRS design utilizes the fact that sludge is a living organism that digests its own dead parts stagnating in the pipe-cleaner flow

field. The digestion process continuously opens up new filter pores that make it possible for water to continue to flow. This process gives rise to a second dynamic equilibrium — between clogging and the generation of new flow paths in the pipe-cleaner flow field.

Only the first pipe cleaner is needed as a filter element; the other pipe cleaners are needed only for support of the tubes in their coaxial alignment. Prior to testing, it was expected that sludge would slowly migrate through the pipe-cleaner

flow field and reach the effluent port, where it would reduce the quality of the filtrate. Instead, it was found that the sludge remained stationary at the first pipe cleaner and no migration was observed, even after two weeks of operation.

This work was done by Shivaun Archer, G. Duncan Hitchens, Harry Jabs, Jennifer Cross, Michelle Pilkinton, and Michael Taylor of Lynntech, Inc. for Johnson Space Center. For further information, contact the Johnson Innovative Partnerships Office at (281) 483-3809. Refer to MSC-23293

Portable Unit for Metabolic Analysis

Respiratory signals can be temporally resolved within respiratory cycles.

John H. Glenn Research Center, Cleveland, Ohio

The Portable Unit for Metabolic Analysis (PUMA) is an instrument that measures several quantities indicative of human metabolic function. Specifically, this instrument makes time-resolved measurements of temperature, pressure, flow, and the partial pressures of oxygen and carbon dioxide in breath during both inhalation and exhalation.

Portable instruments for measuring these quantities have been commercially available, but the response times of those instruments are too long to enable temporal resolution of phenomena on the time scales of human respiration cycles. In contrast, the response time of the PUMA is significantly shorter than characteristic times of human respiration phenomena, making it possible to analyze varying metabolic parameters, not only on sequential breath cycles but also at successive phases of inhalation and exhalation within the same breath cycle.

In operation, the PUMA is positioned to sample breath near the subject's mouth. Commercial off-the-shelf sensors

are used for three of the measurements: a miniature pressure transducer for pressure, a thermistor for temperature, and an ultrasonic sensor for flow. Sensors developed at Glenn Research Center are used for measuring the partial pressures of oxygen and carbon dioxide:

- The carbon dioxide sensor exploits the relatively strong absorption of infrared light by carbon dioxide. Light from an infrared source passes through the stream of inhaled or exhaled gas and is focused on an infrared-sensitive photodetector.
- The oxygen sensor exploits the effect of oxygen in quenching the fluorescence of ruthenium-doped organic molecules in a dye on the tip of an optical fiber. A blue laser diode is used to excite the fluorescence, and the optical fiber carries the fluorescent light to a photodiode, the temporal variation of the output of which bears a known relationship with the rate of quenching of fluorescence and, hence, with the partial pressure of oxygen.

The outputs of the sensors are digitized, preprocessed by a small onboard computer, and then sent wirelessly to a desktop computer, where the collected data are analyzed and displayed. In addition to the raw data on temperature, pressure, flow, and mole fractions of oxygen and carbon dioxide, the display can include volumetric oxygen consumption, volumetric carbon dioxide production, respiratory equivalent ratio, and volumetric flow rate of exhaled gas.

This work was done by Daniel L. Dietrich, Nancy D. Piltch, Mark E. Lewis, Jeffrey R. Juergens, Michael J. Lichter, Peter M. Struk, and Dale M. Diedrick of Glenn Research Center; Russell W. Valentine of Case Western Reserve University; and Richard D. Pettigrew of the National Center for Microgravity Research.

Inquiries concerning rights for the commercial use of this invention should be addressed to NASA Glenn Research Center, Innovative Partnerships Office, Attn: Steve Fedor, Mail Stop 4-8, 21000 Brookpark Road, Cleveland, Ohio 44135. Refer to LEW-17945-1.