

the air-intake manifold of which is modified for coupling to both the outlet of the sampling tube and the sensory tip of the humidity stick. The total cost of these and ancillary components was about \$1,300 in 2003.

At the beginning of operation, the inlet end of the sampling hose is positioned to collect ambient air and the humidity stick and the sampling pump are turned on. After allowing about 20 seconds for the humidity stick to equilibrate with the sampled ambient air, the temperature and humidity readings of the humidity stick are recorded. Next, the suction cup is placed over the access hole to withdraw air from the enclosed

volume. If water drops are observed in the sampling tube, then there is no need for further sampling, and the sampling pump is stopped immediately to avoid drawing liquid water into the humidity stick and pump. If water drops are not observed in the sampling tube, then the relative-humidity reading is monitored until it reaches a maximum (usually after about 20 seconds), at which time the relative-humidity and temperature readings are recorded.

The suction cup is removed from the access hole and after about 30 seconds for equilibration, the temperature and humidity readings for ambient air are taken again. The suction cup is again

placed over the access hole and the air from the enclosed volume sampled again to obtain second temperature and humidity readings to confirm the first readings. Because some ambient (presumably drier) air could have entered the enclosed volume between the first and second humidity readings, the second enclosed-air humidity reading could be lower than the first one.

This work was done by Robert C. Youngquist of Kennedy Space Center and Jan Surma and Steve Parks of ASRC Aerospace. For further information, contact the Kennedy Innovative Partnerships Office at (321) 867-1463. KSC-12593

This method is superior to cleaning by baking.

Marshall Space Flight Center, Alabama




Figure 1. The **Ring Electrodes** used in the operation of a modified Penning-Malmberg trap can also be used in RF generation of plasma for cleaning the trap.

Radio-frequency-generated plasma has been demonstrated to be a promising means of cleaning the interior surfaces of a Penning-Malmberg trap that is used in experiments on the confinement of antimatter. {Such a trap was reported in "Modified Penning-Malmberg Trap for Storing Antiprotons" (MFS-31780), *NASA Tech Briefs*, Vol. 29, No. 3 (March 2005), page 66.} Cleaning of the interior surfaces is necessary to minimize numbers of contaminant atoms and molecules, which reduce confinement times by engaging in matter/antimatter-annihilation reactions with confined antimatter particles.

A modified Penning-Malmberg trap like the one described in the cited prior article includes several collinear ring electrodes (some of which are segmented) inside a tubular vacuum chamber, as illustrated in Figure 1. During operation of the trap, a small cloud of charged antiparticles (e.g., antiprotons or positrons) is confined to a spheroidal central region by means of a magnetic field in combination with DC and radio-frequency (RF) electric fields applied via the electrodes.

In the present developmental method of cleaning by use of RF-generated plasma, one evacuates the vacuum cham-




Figure 2. This **View Along the Axis of a Penning-Malmberg Trap** shows a plasma discharge being used for cleaning.

ber, backfills the chamber with hydrogen at a suitable low pressure, and uses an RF-signal generator and baluns to apply RF voltages to the ring electrodes. Each ring is excited in the polarity opposite that of the adjacent ring. The electric field generated by the RF signal creates a discharge in the low-pressure gas. The RF power and gas pressure are adjusted so that the plasma generated in the discharge (see Figure 2) physically and chemically attacks any solid, liquid, and gaseous contaminant layers on the electrode surfaces. The products of the physical and chemical cleaning reactions are gaseous and are removed by the vacuum pumps.

This cleaning method is much more aggressive than is the standard baking of ultrahigh-vacuum systems; adsorbed

gases are removed much faster and more nearly completely.

The cleaning is also superior to that of a system in which plasma is generated outside the apparatus to be cleaned and made to flow through the apparatus. In contemplated further development, the method would be extended to afford a capability for plasma cleaning of, not only the electrodes, but also the interior

wall of the vacuum chamber. For the purpose of cleaning the wall, it would likely be necessary to modify the electrical connections and electrical operating parameters to optimize the array of electrodes as an antenna for generating plasma between the electrodes and the wall.

This work was done by William Herbert Sims III, James Martin, and J. Boise Pearson

of Marshall Space Flight Center and Raymond Lewis of RLewis Co., and Wallace E. Fant of Cortez III. Further information is contained in a TSP (see page 1).

This invention is owned by NASA, and a patent application has been filed. For further information, contact Sammy Nabors, MSFC Commercialization Assistance Lead, at sammy.a.nabors@nasa.gov. Refer to MFS-31825.