

This is the author-manuscript version of this paper. Published in:

Frossard, L and Beck, J and Dillon, M and Evans, J (2001) Comparison Between the Forces and Moments Applied on the Residuum of Above-Knee Amputees During Daily Life Activities and Walking. In *Proceedings 10th International Society for Prosthetics and Orthotics World Congress*, pages MO10.3-MO10.3, Glasgow, United Kingdom.

<b>Title:</b>	<b>Comparison Between the Forces and Moments Applied on the Residuum of Above-Knee Amputees During Daily Life Activities and Walking</b>
<b>Presenter:</b>	Frossard, P, Ph.D
<b>Contact Address:</b>	Centre for Rehabilitation Science & Engineering – Queensland University of Technology Kelvin Grove Campus – “O” Block, Room A225 Victoria Park Road. Kelvin Grove Q 4059 AUSTRALIA Phone : 61 (0)7 38 64 58 56 Fax : 61 (0)7 38 64 96 87 E-mail : L.Frossard@qut.edu.au
<b>Other Authors:</b>	Beck J (Research Engineer) Dillon M (Prosthetist) Evans J (Senior clinical researcher)

**Background and purpose:** One significant problem experienced with the direct attachment of external prostheses to the skeleton of above-knee amputees (AKA) is early loosening of the boney anchorage. The forces and moments transmitted through this attachment have been estimated for walking but it has been conjectured that potential damaging loading might occur during other daily activities (Table 1).

**Methods:** Forces from a transducer (JR3 Inc) mounted between the socket and the knee mechanism are recorded at 150 Hz on a remote PC using a telemodem link for one AKA.

**Results and conclusion:** Only the values of torque are presented as this is arguably the most critical component of force in for the design of skeletal attachment device. In table 1, all the activities produced substantially larger external moments during the first half of the support phase than during level walking. The highest moments were developed in the latter half of the stance phase but only during walking in a circle and descending an incline was the peak moment higher than during walking. Such data will be essential in re-evaluating the current prosthesis design.

**Table 1 : Mean values of the moment along the long axis of femur produced at 25% and 75 % of the support phase during daily life activities**

Activity	Support phase	
	25 %	75 %
Walking (14 steps)	-0.01±1.32	-6.54±0.94
Ascending stairs (12 steps)	-2.82±1.27	-3.14±1.33
Descending stairs (12 steps)	-1.30±1.12	-1.79±1.07
Ascending slope (5 steps)	-1.94±1.03	-4.28±1.28
Descending slope (5 steps)	-3.39±1.03	-7.00±0.54
Around a circle (7 steps)	-2.24±0.86	-7.50±1.09