

University of Groningen

Development and evaluation of metal artifact reduction and image segmentation techniques in PET/CT

Abdoli, Mehrsima

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version

Publisher's PDF, also known as Version of record

Publication date:

2013

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Abdoli, M. (2013). Development and evaluation of metal artifact reduction and image segmentation techniques in PET/CT Groningen: s.n.

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

1. The recognition of all possible physiologic variation and technical artifacts are important in order to avoid interpretation pitfalls that could affect the diagnostic accuracy of PET. *S. Heiba et al, Clinical Positron Imaging, 2000*
2. Sinogram-based metal artifact reduction methods have more attractive properties owing to the fact that the footprints of the streaks are more precisely traceable in the sinogram domain. *This thesis, chapter 2*
3. Misalignment between PET images and attenuation maps should be avoided at all costs since it produces more severe metal-related pseudo-uptake. *This thesis, chapter 3*
4. The limitations of the current generation of PET scanners make the accurate determination of tumor shape and volume from FDG PET images a challenging task. *This thesis, chapter 7*
5. The contourlet transform of the image enhances the tumor-to-background ratio, which boosts the delineation of the object boundaries. *This thesis, chapter 8*
6. The contourlet-based active contour segmentation algorithm more successfully achieves a smooth contour over the segmented volume, than the classical active contour models do, when the image contains a high level of noise. *This thesis, chapter 8*
7. The highest forms of understanding we can achieve are laughter and human compassion. *Richard P. Feynman*
8. Men love to wonder, and that is the seed of science. *Ralph Waldo Emerson*
9. The greatest discoveries of science have always been those that forced us to rethink our beliefs about the universe and our place in it. *Robert L. Park, in The New York Times, 7 December 1999*
10. You are searching the world for treasure, but the real treasure is yourself. *Rumi*
11. The greatness of a people is not its past but its present.