



New estimators and guidelines for better use of fetal heart rate estimators with Doppler ultrasound devices

Iulian Voicu, Sébastien Ménigot, Denis Kouamé, Jean-Marc Girault

► To cite this version:

Iulian Voicu, Sébastien Ménigot, Denis Kouamé, Jean-Marc Girault. New estimators and guidelines for better use of fetal heart rate estimators with Doppler ultrasound devices. Computational and Mathematical Methods in Medicine, Hindawi Publishing Corporation, 2014, vol. 2014, pp. 1-10. <10.1155/2014/784862>. <hal-00881484v2>

HAL Id: hal-00881484

<https://hal.archives-ouvertes.fr/hal-00881484v2>

Submitted on 4 Mar 2015

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Open Archive TOULOUSE Archive Ouverte (OATAO)

OATAO is an open access repository that collects the work of Toulouse researchers and makes it freely available over the web where possible.

This is an author-deposited version published in : <http://oatao.univ-toulouse.fr/>
Eprints ID : 12757

To link to this article : DOI :10.1155/2014/784862
URL : <http://dx.doi.org/10.1155/2014/784862>

To cite this version : Voicu, Iulian and Ménigot, Sébastien and Kouamé, Denis and Girault, Jean-Marc *[New estimators and guidelines for better use of fetal heart rate estimators with Doppler ultrasound devices.](#)* (2014) Computational and Mathematical Methods in Medicine, vol. 2014. pp. 1-10. ISSN 1748-670X

Any correspondance concerning this service should be sent to the repository administrator: staff-oatao@listes-diff.inp-toulouse.fr

Research Article

New Estimators and Guidelines for Better Use of Fetal Heart Rate Estimators with Doppler Ultrasound Devices

Iulian Voicu,^{1,2} Sébastien Ménigot,^{1,2} Denis Kouamé,³ and Jean-Marc Girault^{1,2}

¹ *Signal & Imaging Group, University François Rabelais of Tours, PRES Loire Valley University, UMR INSERM U930, 7 Avenue Marcel Dassault, 37200 Tours Cedex, France*

² *Inserm, U930, 10 Boulevard Tonnelé, BP 3223, 37032 Tours Cedex, France*

³ *IRIT UMR 5505, Université Paul Sabatier Toulouse 3, 118 Route de Narbonne, 31062 Toulouse Cedex 9, France*

Correspondence should be addressed to Jean-Marc Girault; jean-marc.girault@univ-tours.fr

Characterizing fetal wellbeing with a Doppler ultrasound device requires computation of a score based on fetal parameters. In order to analyze the parameters derived from the fetal heart rate correctly, an accuracy of 0.25 beats per minute is needed. Simultaneously with the lowest false negative rate and the highest sensitivity, we investigated whether various Doppler techniques ensure this accuracy. We found that the accuracy was ensured if directional Doppler signals and autocorrelation estimation were used. Our best estimator provided sensitivity of 95.5%, corresponding to an improvement of 14% compared to the standard estimator.

1. Introduction

Continuous monitoring of fetal parameters has shown their advantages in estimating fetal wellbeing [1]. According to the report of the Society for Maternal-Fetal Medicine [2], continuous fetal heart rate monitoring has reduced infant mortality. The development or the improvement of noninvasive methods dedicated to continuous fetal monitoring is therefore of major interest.

One important parameter in assessing fetal wellbeing is the variability of the fetal heart rate. This parameter, which corresponds to the variation between intervals of two consecutive heart beats, is an indicator of central nervous system development [3–5]. It characterizes fetal behavior states [6–8] and can be an indicator of further neurological evolution [9]. Variability analysis provides a good indication of fetal distress [10] and identifies fetuses with intrauterine growth retardation [11]. According to Dawes criteria [12], variability of 4 ms is a predictor of the lack of acidosis, while a value of 2.6 ms is critical for the fetus. In the normal fetal heart rate range (110–160 bpm), a time variability of 4 ms corresponds to a cardiac frequency variability of 0.81 bpm, while a time variability of 2.6 ms corresponds to a cardiac

frequency variability of 0.53 bpm (the values of 0.53 bpm and 0.81 bpm are obtained as follows: $60/(60/110 - 2.6 \text{ ms}) - 110 = 0.56 \text{ bpm}$ and $60/(60/110 - 4.0 \text{ ms}) - 110 = 0.81 \text{ bpm}$, resp.). Other authors [13] suggest that it is necessary to estimate the heart rate with an accuracy of 0.25 bpm in order to analyze fetal heart rate variability correctly. Reliable estimation of fetal heart rate and hence of heart rate variability is therefore essential.

Several methods are available to assess the fetal heart rate. These methods differ both at the Doppler signal level (directional or nondirectional) and at the level of the algorithm that estimates the heart rate. For example, existing devices on the market such as Sonicaid Oxford (Oxford Sonicaid Instruments, Abington, UK) [14], Hewlett-Packard 8030A (Hewlett-Packard, Palo Alto, CA, USA) [15], and Philips Avalon F40 (Philips, Amsterdam, Netherlands) use the envelope of the nondirectional Doppler signal. Other authors [16, 17] have used the envelope of the directional Doppler signal. Several algorithms based on autocorrelation are commonly used to estimate the heart rate (Oxford Sonicaid, Hewlett-Packard 8030A, and Philips Avalon F40). These algorithms have been applied either directly to the Doppler signal envelope (directional or nondirectional) or to

the signals resulting from discrete wavelet decomposition of the envelope. In the latter case, the final estimated heart rate involves the combination of different estimates [16].

In this study, we first verified whether the pulsed Doppler techniques currently used in commercial devices ensure such an accuracy of 0.25 bpm, and we propose here some recommendations regarding parameter settings. We also compared the techniques used in commercial devices with other pulsed Doppler techniques that use directional Doppler signals. We evaluated the efficacy of all these techniques empirically (error of estimation of the fetal heart rate, sensitivity, and false negative rate).

The originality of this study lies in the recommendations on the parameter settings of the system and in the description of the individual limitations of each technique. Finally, to improve detection probability, a new method based on the combination of heart rates obtained from directional signals is proposed.

2. Materials and Methods

In this section, we describe the Doppler system we developed, the patients, and the synthetic signals used for the comparison of each estimator. The synthetic signals were inferred from real signals. Finally, we present the various existing techniques for estimation of fetal heart rate and we present a new technique based on a combined procedure.

2.1. The Doppler System. In order to evaluate fetal wellbeing objectively and to classify the fetus, we codeveloped the pulsed, multitransducer, multichannel Doppler Actifoetus unit with Althais Technologies (Tours, France).

Our system comprised a personal computer (PC) and our Actifoetus unit. The Actifoetus unit contained three groups of four transducers and a Doppler acquisition board. The detailed operating functions of the acquisition board were presented in [18].

The transducers exploring the fetal heart were non-focused and monoelement. They were circular in shape, with a diameter of 13.5 mm and an acoustic power of 1 mW/cm^2 . Geometrically, the transducers were located at the center of gravity and at the top of an equilateral triangle with sides measuring 40.7 mm.

The transducers were placed on the mother's abdomen. They transmitted a sinusoidal pulse at 2.25 MHz with a pulse repetition frequency (PRF) of 1 kHz. Note that a theoretical accuracy of $60/2/1000 = 0.03 \text{ bpm}$ can be achieved with this value of 1 kHz and accuracy can be still further improved by performing interpolation of the correlation function. The wave was propagated through the mother's abdomen towards the fetal heart. The backscattered signal was recorded from five different depths, annotated D_1, \dots, D_5 . Note that only one channel was considered in the present study.

The ultrasound signal received was converted into an electrical signal and amplified to compensate for the attenuation of 1 dB/cm/MHz . The signal was then demodulated in phase (I) and quadrature (Q) [19]. After demodulation,




FIGURE 1: A real Doppler signal (PRF = 1 kHz) of 1000 ms, recorded with the second transducer in the fourth-channel: (a) Doppler signal (dashed line) and its envelope (solid line); (b) the envelopes of directional signals corresponding to ultrasound scatters approaching to the transducer (solid line) and moving away from the transducer (dash-dot line), respectively.

the signals were digitized. The digital outputs of the converters represented the digital Doppler signal.

2.2. Patients. The Doppler signals were acquired at the CHRU “Bretonneau” Tours, France. The consent of each patient was obtained and the study was approved by the Ethics Committee of the Clinical Investigation Centre for Innovative Technology of Tours (CIC-IT 806 CHRU of Tours). Patients were older than eighteen years and all pregnancies were single. The recordings were made during the twenty-fifth and fortieth gestational weeks. Evolution during pregnancy was normal for all fetuses.

2.3. Simulation. Because it was difficult to quantify the effectiveness of the estimation techniques directly on real signals and because there was no suitable model, we generated synthetic signals. These synthetic signals were used as a ground-truth to evaluate the effectiveness of each estimator. To make these signals as realistic as possible, we proceeded in two stages: an analyzing stage deducing the characteristics of the real Doppler signal envelope and a synthetic phase providing realistic simulated signals. Figures 1(a) and 1(b) show 1000 ms of the envelope of a real nondirectional Doppler signal




FIGURE 2: Envelope of a real directional Doppler signal of 2000 ms. The parameters defining the synthetic signal are the amplitudes and durations of peaks, the lag, and the differences in amplitude of two consecutive peaks over time.

and the two envelopes of corresponding directional signals obtained from the I and Q signals [19, 20]. The synthetic envelope signal was calculated as follows:

$$x_e(t) = x_B(t) + x_F(t), \quad (1)$$

where $x_B(t)$ and $x_F(t)$ are the envelopes of directional Doppler signals produced by the scatters that approach and move away from the transducer, respectively.

We verified (Figures 1(a) and 1(b)) that the envelope of the real nondirectional Doppler signal had the signatures of both envelopes of the directional signals. For example, at around 500 ms, the envelope of the nondirectional signal was mainly influenced by scatters that approached the transducer, while at around 400 ms, we observed the influence of movements away from the transducer. The alternating influence of these two movements determined the envelope of the nondirectional signal.

2.3.1. Analysis of Real Directional Signals. Figure 2 shows 2000 ms of the envelope of a real directional Doppler signal. In order to find the important parameters required for the synthesis of this signal, we extracted its intrinsic features (the number and amplitudes of the peaks, the lags between the peaks, and the differences in amplitude between the peaks). The values of these parameters were evaluated by considering a quasiconstant fetal heart rate.

Figure 2 represents a sequence of several patterns. These patterns were made up of peaks that corresponded to cardiac wall and valve movements of the fetal heart. As suggested by Shakespeare et al. [21], although six peaks (atrial contraction, ventricular contraction, opening and closing of the mitral valves, and opening and closing of the aortic valves) could be detected theoretically, only a few peaks were in practice detected in the nondirectional Doppler signal. From our

analysis, it appeared that the most likely pattern was that with four peaks. Note that this signature composed of four peaks could vary considerably from one beat to another, and it was similar to that identified by Jezewski et al. [13]. Among all these patterns, the most likely was the pattern with peaks in the order 2143; that is, the highest peak M_1 was in second position, the second highest peak M_2 was in first position, and so forth. For the 2143-pattern, we evaluated the amplitudes of each peak (M_1 , M_2 , M_3 , and M_4), the peak durations (T_1 , T_2 , T_3 , and T_4), the lags between two consecutive peaks (dP_1P_2 , dP_2P_3 , and dP_3P_4), and the differences in amplitude between two consecutive peaks (dM_2M_1 , dM_1M_4 , and dM_4M_3). The results of this statistical analysis are reported in Table 1.

As the patterns observed in Figure 2 were noisy, we decided to assess the noise level in order to simulate noisy synthetic Doppler signals. We assessed the signal to noise ratio (SNR) as follows:

$$\text{SNR} = 10 \cdot \log_{10} \left(\frac{P_A}{P_P} \right), \quad (2)$$

where P_A and P_P are the powers of the active and passive regions, respectively. We considered the active region as the area containing the pattern peaks, whereas there were none in the passive region. Using (2), we found that the SNR calculated on our real signals corresponded to a Gaussian law: $\text{SNR} \sim N(11, (\sqrt{2.5})^2(\text{dB}))$.

2.3.2. Synthesis of Synthetic Directional Signals. Analysis of the envelope of directional signals showed the presence of four peaks, which appeared in order 2143 inside the periodic patterns. The synthesis of such a signal must account for




FIGURE 3: General diagram of the Doppler data processing acquired using one transducer and one channel. x_F and x_B represent the envelopes of the directional signals, and x_e represents the envelope of the nondirectional signal. I_1, I_2 are the two autocorrelation estimators.

TABLE 1: Statistics evaluated using (2143) patterns: M_i represents the statistics of the maxima, where $i = 1, \dots, 4$, $dP_1P_2, dP_2P_3, dP_3P_4$ represent the differences between the positions of two adjacent maxima over time; $dM_2M_1, dM_1M_4, dM_4M_3$ represent statistical differences between two adjacent maxima over time, and T represents the statistics of the peak durations. We found that the statistics of the four peaks of $T_i, i = 1, \dots, 4$, were identical.

	Law	
	Gaussian	Uniform
M_1	89.06 ± 31.48	
M_2	69.70 ± 21.84	
M_3	54.80 ± 19.21	
M_4	36.28 ± 18.28	
dM_2M_1		(5–50)
dM_1M_3		(5–50)
dM_3M_4		(5–40)
dP_1P_2	41.50 ± 18.18	
dP_2P_3	92.92 ± 27.76	
dP_3P_4	47.81 ± 30.09	
T (ms)		(25–45)

these characteristics. Equation (3) shows the two possible components of such a signal:

$$x_B(t) = \begin{cases} b(t) + \sum_{i=1}^4 M_i \sin\left(2\pi f_i \left(\theta + \frac{T_i}{2}\right)\right) \text{Rect}_{T_i}(\theta), & \forall t \in [0, T_m], \\ b(t), & \forall t \in (T_m, T_s], \end{cases} \quad (3)$$

where $b(t)$ is the noise, M_i is the peak amplitude, $f_i = 1/(2T_i)$ is the peak frequency, $\text{Rect}_{T_i}(\theta)$ is the unit rectangular function centered on T_{c_i} with width T_i and $\theta = t - T_{c_i}$, T_m is the pattern duration, and T_s is the synthetic signal period. We set a constant interval T_s between the highest peaks of two consecutive patterns of the synthetic signal, as illustrated in Figure 2. We also chose the T_m pattern period as 50% of the synthetic cardiac cycle period T_s , since this period can vary between 40 and 60% [22].

Using (3), we generated two synthetic noisy envelopes corresponding to the envelopes of the directional signals. The envelope of the nondirectional synthetic signal was modeled using (1), being the sum of the envelopes of the both directional synthetic signals. In order to simplify our study, only $x_B(t)$ was calculated, $x_F(t)$ being a delayed and amplified version of $x_B(t)$. To simulate realistic signals, we introduced a lag τ between the directional components:

$$x_F(t) = \alpha x_B(t - \tau), \quad (4)$$

where $x_F(t)$ and $x_B(t)$ are the envelopes of directional signals and τ is the lag between the two envelopes. α is a factor that represents the amplitude ratio between the two types of envelope. From Figure 1, $\tau \approx 40$ ms and $\alpha \approx 2$.

2.4. Estimators. In this section, we describe the different estimators used in our study. Each estimator that was based on the autocorrelation function was denoted by $E_i, i = 1, \dots, 8$, as illustrated in Figure 3. Each estimator operated on different signals: $x_F(t), x_B(t)$, and $x_e(t)$.

Devices existing on the market currently use the $x_e(t)$ envelope and autocorrelation. The estimators that used these configurations were E_7 and E_8 . The mathematical expression

of the two autocorrelation estimators is given in [23] and is represented thereafter by

$$I_1(t, k) = \frac{1}{W} \sum_{n=0}^{W-|k|-1} x(t, n) \cdot x(t, n+k); \quad (5)$$

$$I_2(t, k) = \frac{1}{W} \sum_{n=0}^{W-1} x(t, n) \cdot x(t, n+k),$$

where W is the size of the analyzing window, t is the time for which the estimator is computed, and k is the lag. $x(t)$ represents one of the signals analyzed ($x_F(t)$, $x_B(t)$, or $x_e(t)$).

We tested other estimators (E_1, \dots, E_6) which used directional signals $x_F(t)$ and $x_B(t)$, together with I_1 and I_2 .

2.4.1. Algorithm. The algorithm to estimate the fetal heart rate was the same for all three signals ($x_e(t)$, $x_F(t)$, and $x_B(t)$). The steps of the algorithm were as follows.

- (1) Extract from each signal under consideration ($x_e(t)$, $x_F(t)$, or $x_B(t)$) a limited number W of samples, W being the window size.
- (2) Compute $I_1(W)$ and $I_2(W)$.
- (3) Using an empirical threshold, detect the position of N peaks in $I_1(W)$ and $I_2(W)$.
- (4) From the position of N peaks of $I_1(W)$ and $I_2(W)$, determine the durations D_i between consecutive peaks with $i = 1, 2, \dots, N - 1$.
- (5) Calculate the $N - 1$ cardiac frequencies with $CF_i = 60/D_i$, $i = 1, 2, \dots, N - 1$. This conditional test limits the number of cardiac frequencies CF_i estimated from I_1 or I_2 in the average computation. This conditional test also permits removal of cardiac frequency estimates that are half the expected value, as are sometimes observed (Shakespeare et al. [21]).
- (6) Calculate the average cardiac frequency (FHR) from CF_i not exceeding 35 bpm [13]:

$$\text{FHR} = \frac{1}{N-1} \sum_i^{N-1} CF_i. \quad (6)$$

As an illustration, consider a window of 4.096 s. Whenever the cardiac frequency was 240 bpm, 16 peaks were observed in the autocorrelation function. Using an empirically set threshold, the duration between each peak was measured ($D_1 = 0.250$ s, \dots , $D_{15} = 0.250$ s) and cardiac frequencies of $CF_1 = 60/0.250$ bpm, \dots , $CF_{15} = 60/0.250$ bpm were estimated. The average cardiac frequency was obtained by $\text{FHR} = (1/15) \sum_i^{15} CF_i = 240$ bpm. Note that 4 peaks were observed with 60 bpm and $D_1 = 60/1.0 = 1.0$ s, \dots , $D_3 = 60/1.0 = 1.0$ s were estimated. The average cardiac frequency was obtained by $\text{FHR} = (1/3) \sum_i^3 CF_i = 60$ bpm.

Thus the proposed algorithm correctly worked in the range of 60–240 bpm. However, the standard deviation of the FHR estimation was not the same for its extreme values since in one case the average was obtained with three values

whereas the average was obtained with fifteen values in the other.

Note that such an algorithm is not perfect since it is hypothesised that the FHR is constant during the process. Sometimes the second peak of the autocorrelation can be lower than the third and the FHR estimate is incorrect. A process must be performed to remove outliers.

2.4.2. Elimination of Outliers. In order to eliminate outlier estimates associated with estimator dysfunction, we introduced a postprocessing step. This postprocessing step was applied only in the case of real signals. An estimate was considered to be an outlier if it laid outside the statistic computed from 40 previous estimates, or if it differed between two consecutive analysis windows by 35 bpm.

2.4.3. Combination. In order to improve the effectiveness of the FHR estimation, we combined estimations. For E_2 and E_5 , the two values of the fetal heart rate estimated on signals $x_F(t)$ and $x_B(t)$ were combined. The estimate on the two signals was achieved using I_1 or I_2 . The combination rule we used was as follows:

- (i) if the heart rate was detected on a single signal, the combined value took this value;
- (ii) if the heart rate was detected on both signals, the combined value was the average of the two values.


Note that, in contrast to Kret's study [16], we combined the fetal heart rates estimated on both directional Doppler signals, while Kret's technique was based on combination of fetal heart rate estimations computed after discrete wavelet decomposition of the envelope of the directional Doppler signal. Since Kret's technique was applied only for continuous Doppler signals, it was not taken into consideration in our study.

3. Results


To find the best estimators and the conditions in which they could be used, we performed a series of simulations and experiments. Using simulations, we sought configurations that ensured an error of estimation, that is, the expected accuracy below 0.25 bpm, the highest sensitivity, and the lowest false negative rate. Experimentally, we sought the best configuration for optimal use of the estimator.

3.1. Simulated Signals. We present the results from two types of simulation. In the first series of simulations, we sought parameter settings that ensured the desired accuracy of 0.25 bpm. In the second series of simulations, we evaluated the effectiveness of each estimator in terms of true positive rate and false negative rate.

3.1.1. Optimal Parameter Settings. The results presented in Figures 4 and 5 were obtained for synthetic signals of 30 s. The parameters that varied in our analysis were the periodicity of the signal T_s , the SNR, the window size W , and the lag



(a)




(b)

FIGURE 4: Duration of the analyzing window W required to reach an error of at least 0.25 bpm with the autocorrelation estimator (I_1) and a SNR > 0.6 dB.


τ . We varied the signal periodicity T_s between 1000 and 250 ms, as these values corresponded to the standard range of exploration (60–240 bpm) of different fetal monitors. The SNR range varied between 0 and 14 dB, in order to include our measured SNR values on the real signals and in order to take into account the worst cases. The range of W size varied between 512 and 4096 ms. The highest fetal heart rate could be obtained with a window size of 512 ms, although we limited the maximum window to 4096 ms to reduce computation time.

Figures 4 and 5 show the smallest window size analyzed ($W = 4096$ ms) of all estimators tested that ensured the expected accuracy of 0.25 bpm in the range of 60–240 bpm and that ensured a SNR at least greater than 0.6 dB. Note that for estimator I_2 reported in Figure 5, we showed that there was no size which ensured the desired accuracy, whatever the SNR or the frequency. To test estimator robustness in relation to the increasing complexity of the simulated signals, the lag τ varied between 0 and 40 ms, this value of 40 ms being taken from Figure 1.

The results derived from Figure 4 showed that accuracy for the envelope signal (estimator E_7) was no longer achieved



(a)



(b)

FIGURE 5: Duration of the analyzing window W required to reach an error of at least 0.25 bpm with the autocorrelation estimator (I_2) and with a SNR > 0.6 dB.

for certain frequencies, but it still was for directional signals. Finally, Figure 5, shows the best estimators (E_1, E_2, E_4) and their respective best parameter settings ($W = 4096$) that ensured an accuracy of 0.25 bpm with a SNR > 0.6 dB in the 60–240 bpm range.

To summarize, these first results showed the superiority of I_1 compared to I_2 and the superiority of the envelope of directional signals compared to that of nondirectional signals. We therefore recommend the use of I_1 and the estimators (E_1, E_2 , and E_4) based on the envelope of directional signals.

3.1.2. Performance Levels of Estimators. In this study, the performance levels of estimators we wanted to compute were sensitivity and the false negative rate. Fetal heart rates were evaluated every 250 ms from noisy signals with $W = 4096$ ms. Sensitivity was computed with the equation: $S = TP/(TP + FN)$, where TP was the true positive rate and FN was the false negative rate. Estimates of simulated heart rate were considered to be false negative if they did not ensure the expected accuracy; otherwise, they were true positive. Sensitivity and the false positive rate were evaluated as the

TABLE 2: FHR error of estimation (bpm) obtained by all estimators tested for different configurations of SNR, false negative rate (FNR), and $W = 4096$ ms.

Estimators	FHR error of estimation (bpm)		
	E_1, E_2, E_4, E_7	E_3, E_5, E_6	E_8
FNR = 0%, 6 dB > SNR > 2 dB	0.8	4	6
FNR = 1.5%, SNR > 6 dB	0.25	—	—







FIGURE 6: True positive rates for I_1 and I_2 with SNR > 0.6 dB, $W = 4096$, and an error of estimation of 0.25 bpm. (a) True positive rate for I_1 . (b) True positive rate for I_2 .

FIGURE 7: False negative rates for I_1 and I_2 with SNR > 0.6 dB, $W = 4096$. Accuracy of 25 bpm was obtained only for I_1 whatever the frequency, whereas for I_2 accuracy was obtained for only 100, 150, 200, 220, and 240 bpm. (a) False negative rate for I_1 . (b) False negative rate for I_2 .

average of 30 values. Each value was determined after analysis of a noisy signal of 30 s where sensitivity and the false positive rate had converged to the highest value and to the lowest value, respectively. Convergence was reached for a minimum SNR of 6 dB.

The results of this second series of simulations are presented in Figures 6 and 7 for $W = 4096$ ms and SNR \geq 6 dB. Note that the estimation of error of 25 bpm was obtained only for I_1 whatever the frequency, whereas accuracy for I_2 was obtained only for 100, 150, 200, 220, and 240 bpm.

The results set out in Figure 7 show that estimators based on I_1 (E_1, E_2, E_4 , and E_7) had an average (average obtained from the cardiac frequency) false negative rate of 1.5%, while those based on I_2 (E_3, E_5, E_6 , and E_8) presented a higher

average false negative rate of approximately 14.8%. The 97.5% average true positive rate of I_1 was slightly lower than that of estimators based on I_2 , which was 100%. Finally, when the accuracy of 0.25 bpm was reached, we observed that the estimators based on I_1 were generally more accurate than those based on I_2 , although the average false negative rate was not zero.

Figure 8 shows the error of estimation corresponding to different values of SNR when a zero false negative rate was imposed. This zero false negative rate was obtained by modifying the detection threshold, and a direct consequence was an increase in the estimation. The results derived from Figure 8 showed that the zero false negative rate for I_1 was

TABLE 3: Sensitivity (%) of estimators for $W = 4096$ ms. I_1 and I_2 are the two autocorrelation estimators, respectively. x_B , x_F are the envelopes of directional signals, and x_e is the envelope of the nondirectional signal. Fus indicates the combined estimator.

Sensitivity (S)	Estimators	
	I_1	I_2
$S(x_B)$	88.50% (E_4)	88.43% (E_6)
$S(x_F)$	86.63% (E_1)	84.79% (E_3)
$S(x_e)$	81.79% (E_7)	75.05% (E_8)
$S(Fus)$	95.48% (E_2)	94.93% (E_5)

ensured for a $SNR \geq 2$ dB (below the SNR measured on real signals) and for an error of estimation of 0.8 bpm. In the case of I_2 and directional signals, we obtained an error of estimation of 4 bpm, whereas for a nondirectional signal it was 6 bpm.

Table 2 summarizes the effectiveness of each estimator in terms of FHR error of estimation, SNR, and average false negative rate. To reach an error of estimation, that is, the expected accuracy of 0.25 bpm, we recommend I_1 , that is, autocorrelation-based estimators (E_1 , E_2 , E_4 , and E_7), the price to be paid being an average false negative rate of 1.5%. To reach an average false negative rate of 0%, we recommend I_1 autocorrelation-based estimators (E_1 , E_2 , E_4 , and E_7), where the price to be paid is an error of estimation of 0.8 bpm far from the expected accuracy of 0.25 bpm.

3.2. Results Obtained on Real Signals. We recorded 580 minutes for the analysis of real Doppler signals. We selected areas where signals had the cardiac activity signature. The performance levels on these signals were evaluated on the envelopes of both the nondirectional and the directional signals. The FHR estimation obtained with a commercial device (Oxford Sonicaid) was used as a reference to evaluate sensitivity which was evaluated for each estimator. All estimators were evaluated using a size of $W = 4096$ ms.

The results obtained for all the signals are presented in Table 3. The estimators based on directional signals (E_1 , E_4 , E_3 , and E_6) provided a higher level of sensitivity compared to those which used nondirectional signals (E_7 , E_8). In the case of directional signals, the results of I_1 and I_2 were close but slightly better for I_1 . This result confirmed the results obtained by simulations. We therefore recommend the use of estimators based on I_1 calculated on the directional signals (E_1 , E_4).

Sensitivity was improved using the combination method. The sensitivity of estimator E_2 was 95.5% (see Table 3). Using the combination method, the sensitivity increased to about (95.5%–88.5%) $\approx 7\%$ compared to directional signals E_4 and to about (95.5%–81.8%) $\approx 14\%$ compared to nondirectional signals E_7 .

4. Discussion and Conclusion

In this study, we focused on different settings of the estimators (window size, lag) that ensured a fetal heart rate estimation




FIGURE 8: FHR error of estimation (bpm) when the false negative rate was zero, when the $SNR \in (0 \text{ dB}, 2 \text{ dB})$, and when the signals tested were x_F , x_B , and x_e .

with a maximum authorized error of 0.25 bpm. We found in simulation that only estimators based on I_1 and directional signals could ensure such an accuracy of 0.25 bpm. The size necessary in this case was $W = 4096$ ms.

Note that, although we proposed synthetic signals that were as realistic as possible, we are aware that the plotted performance levels are representative only of our simulations and not of all cases encountered in practice. It is likely that the performance levels of the algorithms tested can be reduced in the presence of artefacts. However, the 95% sensitivity obtained from real signals suggests that our proposed estimators may be trusted.

In the case of real signals, the sensitivity was quantified. Since in our study the estimated SNR on the real signals was greater than the threshold of 6 dB (deduced on simulated signals) required to reach the desired accuracy of 0.25 bpm, a denoising filter was not necessary. However, in cases of a

SNR lower than 6 dB, a denoising process (Wiener, wavelet) could be introduced.

Sensitivity was quantified using a W size of 4096 ms. We found that the estimators E_1 , E_2 , and E_4 based on I_1 had slightly greater sensitivity than those based on I_2 . We therefore recommend the use of I_1 .

Various cases were considered on the basis of this study, that is, those that do not require a precise estimate of the fetal heart rate and those for which accuracy is critical. The accuracy of fetal heart rate estimation in the first case is not important but the false negative rate should be as low as possible. For example, if the goal of a monitoring system is simply to verify that the fetal heart rate is in the normal range (110–160 bpm), very high accuracy is not needed. In this case, an error of estimation of 0.8 bpm is sufficient. Our computations showed that for an error of estimation of 0.8 bpm, a zero false negative rate in the zones when the rhythm is quasi-constant could be ensured. In the second case, an error of estimation of 0.25 bpm is required for a system in which the goal is not only to estimate the heart rate, but also to evaluate fetal wellbeing. Our study showed that for this type of system, the false negative rate may be slightly higher than zero. It is important to note that this error of estimation was guaranteed for a quasi-constant heart rate. This is not a constraint for such a system, since the variability of fetal heart rate must be evaluated in these ranges to predict fetal distress.

Applied to real signals, the estimators based on I_1 provided sensitivity close to those of I_2 , and the most efficient of these estimators were those that used directional signals (E_1 , E_2 , and E_4).

A 7% increase in sensitivity compared to estimators based on individual directional signals was possible when we combined the two heart rates calculated on the directional signals. A 14% increase in sensitivity compared to estimators based on individual nondirectional signals was possible when we combined the two heart rates calculated on the directional signals. When a combination was used, both signals were processed in parallel, thus doubling the number of operations.

The good levels of performance of our estimator based on this combination suggest first that it can be adapted to multitransducer, multichannel configurations and second that such an estimator will improve fetal diagnosis.

Conflict of Interests

The authors declare that there is no conflict of interests regarding the publication of this paper.

Acknowledgments

This study was supported by the Agence Nationale de la Recherche (Project ANR-07-TECSAN-023, Surfoetus). The authors would like to thank the Clinical Investigation Centre for Innovative Technology of Tours (CIC-IT 806 CHRU of Tours) and Professor F. Perrotin's Obstetric Department Team for their support in recording the signals. The authors

would like to thank the anonymous referees for their valuable comments.

References

- [1] F. Manning, I. Morrison, I. Lange, C. Harman, and P. Chamberlain, "Fetal assessment based on fetal biophysical profile scoring: experience in 12,620 referred high-risk pregnancies. I. Perinatal mortality by frequency and etiology," *The American Journal of Obstetrics and Gynecology*, vol. 151, no. 3, pp. 343–350, 1985.
- [2] H.-Y. Chen, S. P. Chauhan, C. V. Ananth, A. M. Vintzileos, and A. Z. Abuhamad, "Electronic fetal heart rate monitoring and infant mortality: a population-based study in the United States," *The American Journal of Obstetrics and Gynecology*, vol. 204, no. 6, pp. S43–S44, 2011.
- [3] G. G. Berntson, J. Thomas Bigger Jr., D. L. Eckberg et al., "Heart rate variability: origins methods, and interpretive caveats," *Psychophysiology*, vol. 34, no. 6, pp. 623–648, 1997.
- [4] M. Malik and A. J. Camm, "Heart rate variability," *Clinical Cardiology*, vol. 13, no. 8, pp. 570–576, 1990.
- [5] M. David, M. Hirsch, J. Karin, E. Toledo, and S. Akselrod, "An estimate of fetal autonomic state by time-frequency analysis of fetal heart rate variability," *Journal of Applied Physiology*, vol. 102, no. 3, pp. 1057–1064, 2007.
- [6] B. Arabin and S. Riedewald, "An attempt to quantify characteristics of behavioral states," *The American Journal of Perinatology*, vol. 9, no. 2, pp. 115–119, 1992.
- [7] B. Frank, B. Pompe, U. Schneider, and D. Hoyer, "Permutation entropy improves fetal behavioural state classification based on heart rate analysis from biomagnetic recordings in near term fetuses," *Medical and Biological Engineering and Computing*, vol. 44, no. 3, pp. 179–187, 2006.
- [8] S. Lange, P. Van Leeuwen, U. Schneider et al., "Heart rate features in fetal behavioural states," *Early Human Development*, vol. 85, no. 2, pp. 131–135, 2009.
- [9] J. A. DiPietro, D. M. Hodgson, K. A. Costigan, S. C. Hilton, and T. R. B. Johnson, "Fetal neurobehavioral development," *Child Development*, vol. 67, no. 5, pp. 2553–2567, 1996.
- [10] M. Ferrario, M. G. Signorini, G. Magenes, and S. Cerutti, "Comparison of entropy-based regularity estimators: application to the fetal heart rate signal for the identification of fetal distress," *IEEE Transactions on Biomedical Engineering*, vol. 53, no. 1, pp. 119–125, 2006.
- [11] A. Kikuchi, T. Shimizu, A. Hayashi et al., "Nonlinear analyses of heart rate variability in normal and growth-restricted fetuses," *Early Human Development*, vol. 82, no. 4, pp. 217–226, 2006.
- [12] G. S. Dawes, M. O. Lobbs, G. Mandruzzato, M. Moulden, C. W. G. Redman, and T. Wheeler, "Large fetal heart rate decelerations at term associated with changes in fetal heart rate variation," *The American Journal of Obstetrics and Gynecology*, vol. 188, no. 1, pp. 105–111, 1993.
- [13] J. Jezewski, J. Wrobel, and K. Horoba, "Comparison of Doppler ultrasound and direct electrocardiography acquisition techniques for quantification of fetal heart rate variability," *IEEE Transactions on Biomedical Engineering*, vol. 53, no. 5, pp. 855–864, 2006.
- [14] J. Pardey, M. Moulden, and C. W. G. Redman, "A computer system for the numerical analysis of nonstress tests," *The American Journal of Obstetrics and Gynecology*, vol. 186, no. 5, pp. 1095–1103, 2002.

- [15] E. Courtin, W. Ruchay, P. Salfeld, H. Sommer, and A. Versatile, "Semiautomatic fetal monitor for non-technical users," *Hewlett-Packard Journal*, vol. 28, no. 5, pp. 16–23, 1977.
- [16] T. Kret and K. Kaluzynski, "The fetal heart rate estimation based on continuous ultrasonic Doppler data," *Biocybernetics and Biomedical Engineering*, vol. 26, no. 3, pp. 49–56, 2006.
- [17] I. Voicu, D. Kouame, M. Fournier-Massignan, and J.-M. Girault, "Estimating fetal heart rate from multiple ultrasound signals," in *Proceedings of the International Conference on Advancements of Medicine and Health Care through Technology*, pp. 185–190, Berlin, Germany, September 2009.
- [18] A. Kribèche, F. Tranquart, D. Kouame, and L. Pourcelot, "The Actifetus system: a multidoppler sensor system for monitoring fetal movements," *Ultrasound in Medicine and Biology*, vol. 33, no. 3, pp. 430–438, 2007.
- [19] J. A. Jensen, *Estimation of Blood Velocities Using Ultrasound*, Cambridge University Press, Cambridge, UK, 1996.
- [20] N. Aydin, L. Fan, and D. H. Evans, "Quadrature-to-directional format conversion of Doppler signals using digital methods," *Physiological Measurement*, vol. 15, no. 2, article 007, pp. 181–199, 1994.
- [21] S. A. Shakespeare, J. A. Crowe, B. R. Hayes-Gill, K. Bhogal, and D. K. James, "The information content of Doppler ultrasound signals from the fetal heart," *Medical and Biological Engineering and Computing*, vol. 39, no. 6, pp. 619–626, 2001.
- [22] E. Hernandez-Andrade, H. Figueroa-Diesel, C. Kottman et al., "Gestational-age-adjusted reference values for the modified myocardial performance index for evaluation of fetal left cardiac function," *Ultrasound in Obstetrics and Gynecology*, vol. 29, no. 3, pp. 321–325, 2007.
- [23] A. V. Oppenheim and R. W. Schaffer, *Discrete-Time Signal Processing*, Prentice Hall Signal Processing, Prentice-Hall International, Englewood Cliffs, NJ, USA, 1989.