

AIP | Applied Physics
Letters

Analyzing threshold pressure limitations in microfluidic transistors for self-regulated microfluidic circuits

Sung-Jin Kim, Ryuji Yokokawa, and Shuichi Takayama

Citation: *Appl. Phys. Lett.* **101**, 234107 (2012); doi: 10.1063/1.4769985

View online: <http://dx.doi.org/10.1063/1.4769985>

View Table of Contents: <http://apl.aip.org/resource/1/APPLAB/v101/i23>

Published by the AIP Publishing LLC.

Additional information on *Appl. Phys. Lett.*


Journal Homepage: <http://apl.aip.org/>

Journal Information: http://apl.aip.org/about/about_the_journal

Top downloads: http://apl.aip.org/features/most_downloaded

Information for Authors: <http://apl.aip.org/authors>

ADVERTISEMENT



Analyzing threshold pressure limitations in microfluidic transistors for self-regulated microfluidic circuits

Sung-Jin Kim,¹ Ryuji Yokokawa,^{1,2} and Shuichi Takayama^{1,3,4,a)}

¹Department of Biomedical Engineering, University of Michigan, Ann Arbor, Michigan 48109, USA

²Department of Microengineering, Kyoto University, Yoshida-honmachi, Sakyo, Kyoto 606-8501, Japan

³Department of Macromolecular Science and Engineering, University of Michigan, Ann Arbor, Michigan 48109, USA

⁴Division of Nano-Bio and Chemical Engineering WCU Project, UNIST, Ulsan, South Korea

(Received 11 October 2012; accepted 19 November 2012; published online 7 December 2012)

This paper reveals a critical limitation in the electro-hydraulic analogy between a microfluidic membrane-valve (μ MV) and an electronic transistor. Unlike typical transistors that have similar on and off threshold voltages, in hydraulic μ MVs, the threshold pressures for opening and closing are significantly different and can change, even for the same μ MVs depending on overall circuit design and operation conditions. We explain, in particular, how the negative values of the closing threshold pressures significantly constrain operation of even simple hydraulic μ MV circuits such as autonomously switching two-valve microfluidic oscillators. These understandings have significant implications in designing self-regulated microfluidic devices. © 2012 American Institute of Physics. [<http://dx.doi.org/10.1063/1.4769985>]

Electric circuit analogy is widely used in microfluidic circuit design and analysis. For example, electric resistors correspond to microfluidic channel resistances and capacitors to flexible membranes. The analogy is based on the similarity in equations between these circuit components. Since the theory and simulation methods of electric circuits are well-established, they greatly facilitate the design and analysis of various microfluidic circuits.¹

Recently, analogy has been drawn between electronic transistors and microfluidic membrane-valves (μ MV), which are used for self-regulated microfluidic circuits such as frequency-specific flow regulators,² digital logic circuits,^{3–6} and oscillators.^{7–9} Like an electronic circuit that operates itself with only a power source and thus minimize the use of its external controllers, the self-regulated microfluidic circuits have the potential to greatly reduce reliance on expensive and complex external controllers, which are barriers to broader use of microfluidic devices. A μ MV is a crucial component that enables operation of self-regulated microfluidic devices through its on-off switching.

Similar to the electronic transistor, the μ MV's on-off switching is determined by the relative difference between the source (P_S) minus gate pressure (P_G) versus the threshold pressure (Figure 1). As depicted in Figure 1(a), when P_S is sufficiently greater than P_G , the membrane of the μ MV deflects down and the μ MV is on (open). In other words, the μ MV is on when $P_S - P_G$ is greater than opening threshold pressure ($P_{th-open}$). To turn off (close) the μ MV, $P_S - P_G$ is less than closing threshold pressure ($P_{th-close}$). In typical silicon transistors,^{10,11} the difference between on and off threshold voltage is negligible, thus providing a large parameter space for the design and operation of large-scale integrated circuits. Pneumatic μ MVs also exhibit small differences between opening ($P_{th-open}$) and closing ($P_{th-close}$) threshold

pressures.^{5,12} However, in self-regulated microfluidic devices, the detailed characteristic of μ MV's threshold pressure in the context of other microfluidic parameters such as fluidic resistor and inflow rate is largely unknown.

Here, we report that $P_{th-open}$ and $P_{th-close}$ are significantly different in *hydraulic* μ MVs, where liquid directly passes through. We further analyze implications of this threshold pressure gap for fluidic circuit design using, as a model system, a constant flow-driven oscillator that functions like an




FIG. 1. Microfluidic oscillator with two membrane valves. (a) Cross-section of a microfluidic valve showing its on- and off-conditions. P_S , P_G , and P_D are pressures at the source (S), drain (D), and gate (G) terminals, respectively. $P_{th-open}$ and $P_{th-close}$ are opening and closing threshold pressures, respectively. (b) Schematic of the oscillator consisting of two microfluidic membrane valves. The valve has S, D, and inlet and outlet gate terminals (G_1 and G_2). Gray and blue areas represent the bottom and top channels of the oscillator, respectively. R_c and R_d are connection and downstream resistance, respectively. (c) Schematic showing oscillating and non-oscillating outflows. The oscillating and non-oscillating outflows are determined by R_d/R_c and the inflow rate (Q_i). Constant Q_i is provided by a syringe pump (enhanced online) [URL: <http://dx.doi.org/10.1063/1.4769985.1>] [URL: <http://dx.doi.org/10.1063/1.4769985.2>].

^{a)} Author to whom correspondence should be addressed. E-mail: takayama@umich.edu.

electronic DC to AC converter.^{7,8} Then, we reveal the fundamental mechanisms of oscillation as well as the restrictions on operational circuit design parameters when using these hydraulic μ MVs that exhibit negative $P_{\text{th-close}}$ and a dependency of $P_{\text{th-open}}$ on external parameters.

Experimental setups and device fabrication were explained in our previous study.⁸ Briefly, each device consists of three layers of poly(dimethylsiloxane) (PDMS): top and bottom slabs for 75 μm -high channels and valves, and middle layer for thin membrane. A syringe pump was used to provide a constant inflow and two pressure sensors were connected at the inlets to measure source pressures. We used commercial software (PLECS, Plexim GmbH, Switzerland) for the numerical simulation of the microfluidic oscillator.¹³

In a hydraulic μ MV, we observed that $P_{\text{th-open}}$ and $P_{\text{th-close}}$ are significantly different. For the initial valve-on state, the hydraulic μ MV also requires a positive threshold pressure [i.e., $P_S - P_G > P_{\text{th-open}} > 0$, see Figure 1(a)]. Then, we could turn off the hydraulic μ MV, only when P_S is $< P_G$. This result means additional P_G greater than its on-state is necessary and, in turn, $P_{\text{th-close}}$ has a negative value (i.e., $P_S - P_G < P_{\text{th-close}} < 0$); through repetitive experiments, we measured $P_{\text{th-close}}$ to be -1 ± 0.2 kPa in our system.¹³ On the other hand, positive value of $P_{\text{th-open}}$ can be measured directly from the two P_S profiles of the oscillator's two μ MVs under the condition that downstream resistance [R_d , defined in Figure 1(b)] is one order of magnitude higher than the connection resistance (R_c).⁸

The origin of the positive $P_{\text{th-open}}$ comes from the elastic force of the membrane and the adhesive force between the valve seat and its membrane, but that of negative $P_{\text{th-close}}$ is different. When a μ MV is on and P_S is relatively close to P_G , the membrane of the μ MV tends to restore its off-state owing to the membrane's elastic force. In a pneumatic μ MV, its adhesive force can be easily recovered for the valve's off-state, because there is only air between the valve seat and the membrane. In a hydraulic μ MV, even when its membrane approaches the valve seat, additional pressure is necessary to squeeze out liquid and to recover the valve's adhesive force, thus requiring a negative $P_{\text{th-close}}$.

Figure 2 explains how the two threshold conditions determine the μ MV on- and off-states in the oscillator. The oscillator is an excellent system to study the effect of threshold pressures because its continued oscillation [Figure 1(c)] is possible only when hydraulic μ MV can be turned off (close) each cycle, under conditions that satisfy the $P_{\text{th-close}}$ values. In the oscillator, two microfluidic valves are connected to each other through their drain and gate terminals [Figure 1(b)]. For example, valve 2's drain terminal (D) is sequentially connected to valve 1's gate inlet terminal (G_i), gate outlet terminal (G_o), and finally to the outlet of the device. For simplicity, the pressure profile of just valve 2 is shown in Figure 2. Here, P_S and P_G of valve 2 are noted as P_{S2} and P_{G2} , respectively. Initially, valve 2 is in the off-state and $P_{S2} - P_{G2}$ of valve 2 increases because P_{S2} accumulates through constant inflow and P_{G2} is constant. After $P_{S2} - P_{G2}$ reaches $P_{\text{th-open}}$, valve 2 turns on. Then, its high P_{S2} shuts off valve 1 through valve 1's gate terminal. Then, $P_{S2} - P_{G2}$ decreases as fluid flows out from the open valve 2 (the first gray region in Figure 2). Likewise, when valve 1 turns on,




FIG. 2. Simulated source and gate pressure profiles of valve 2 at the onset of oscillation. Q_i is 2 $\mu\text{l}/\text{min}$. Gray region corresponds to valve 2-on state. At the bottom panel, $P_{\text{SG-min}}$ is a minimum value of $P_S - P_G$. If $P_{\text{SG-min}}$ is $> P_{\text{th-close}}$ at the moment of valve 1-on to valve 2 off, valve 2 stays on. As a result, both valves are on and oscillation stops.

P_{G2} of valve 2 jumps up, thereby making $P_{S2} - P_{G2} < P_{\text{th-close}}$ of valve 2. This causes valve 2 to turn off. In this way, owing to alternating on-off states of the two valves, oscillation continues.

Note that, for the oscillation, on-off state of the two valves is always opposite. If valve 2 is on but does not satisfy $P_{S2} - P_{G2} < P_{\text{th-close}} = -1$ kPa at the moment of valve 1 on, both valves are on and oscillation stops. This condition can be described as $P_{\text{SG-min}} > P_{\text{th-close}}$, where $P_{\text{SG-min}}$ is defined in Figure 2.

In the hydraulic μ MV of microfluidic oscillators, we show how negative $P_{\text{th-close}}$ constrains operational ranges of the ratio of downstream to connection resistance (R_d/R_c); R_d and R_c are depicted in Figure 1(b). Under the constraints of negative $P_{\text{th-close}}$, oscillation occurs at $R_d/R_c = 7.5$ and 13.7 but not at $R_d/R_c = 0.3$ [Figure 3(a)]. For oscillation to be maintained, valve 2 must be turned off when valve 1 turns on. At the moment valve 1 turns on, the serially connected R_d and R_c work as a pressure divider;¹³ the relation between P_S of valve 1 (P_{S1}) and P_G of valve 2 (P_{G2}) is

$$P_{G2} \approx P_{S1}r/(r+1), \quad (1)$$

where r is R_d/R_c . When this relationship is applied to valve 2's off-condition ($P_{S2} - P_{G2} < P_{\text{th-close}}$), this gives $P_{S2} - P_{S1}r/(r+1) < P_{\text{th-close}} < 0$. Because P_{S2} is $< P_{S1}$ at the moment valve 1 turns on, as long as R_d/R_c (i.e., r) is large enough valve 2's off-condition is satisfied and the outflow oscillation continues [inset of Figure 3(a)]. In contrast, in an equivalent electronic oscillator, oscillation is possible even at $R_d/R_c = 0.3$ because of positive $P_{\text{th-close}}$, thus allowing more flexibility in the range of fluidic resistances.

Interestingly, in addition to contributions from the intrinsic properties of the μ MV such as membrane elasticity and adhesion, $P_{\text{th-open}}$ of the hydraulic μ MV changes with different operating conditions such that the value increases




FIG. 3. Effect of $P_{th-close}$ on the operational range of the ratio of fluidic resistance [R_d/R_c , defined in Fig. 1(b)]. Filled and unfilled points are the experimental and simulation results, respectively. The color of the points corresponds to the R_d/R_c values shown in the inset of (a) and R_c is fixed as $1.46 \times 10^{12} \text{ N}\cdot\text{s}/\text{m}^5$. (a) Change of oscillation periods by R_d/R_c . There is no oscillation at the square points. When P_{SG-min} (defined in the bottom panel of Fig. 2) is $< P_{th-close}$, outflow oscillation is available (gray region of the inset). (b) Corresponding opening threshold pressures, $P_{th-open}$. (c) Corresponding drain minus gate pressure just before the valve-on.

with increasing Q_i and R_d (we change R_d but fix R_c); see Figure 3(b). Compared to electronic transistors where threshold voltage is an intrinsic property, $P_{th-open}$'s dependency on external parameters (Q_i and R_d) makes design of μMV circuits more complex. This dependency can be explained by the drain minus gate pressure ($P_D - P_G$). If $P_D - P_G$ of the off-valve becomes more negative, as illustrated in the valve-off state of Figure 1(a), its membrane more tightly presses its seat at the drain side thereby increasing $P_{th-open}$. To explain how increasing R_d makes $P_D - P_G$ more negative [Figure 3(c)], we consider the state of valve 1 on and valve 2 off [see Figure 1(b) with valves 1 and 2 labeled]. P_D of valve 2 is close to the outlet pressure ($\approx 0 \text{ kPa}$) because valve 2 is off and there is negligible flow between valve 2's drain and the outlet. On the other hand, P_G of valve 2 (P_{G2}) increases with increasing R_d at constant Q_i because P_{G2} follows Poiseuille's law at valve 1 on. Thus, $P_D - P_G$ of valve 2 becomes more negative at higher R_d [Figure 3(b)], thereby increasing $P_{th-open}$. In the same way, $P_{th-open}$ also increases with increasing Q_i because $P_D - P_G$ becomes more negative by Poiseuille's law. Despite the increasing $P_{th-open}$ by Q_i , the oscillation period ($\propto P_{th-open}/Q_i$, Ref. 8) decreases at higher Q_i [Figure 3(a)] because increment of $P_{th-open}$ is $\ll Q_i$. Specifically, as shown in Figure 3(b), $P_{th-open}$ changes from 9.8 to 12.7 kPa and from 8.9 to 10.4 kPa, whereas Q_i changes from 2 to 10 $\mu\text{L}/\text{min}$; this makes the increment of $P_{th-open}$ and Q_i as 30% and 16% for $P_{th-open}$ and 400% for Q_i .

Notably, negative $P_{th-close}$ also limits the operational range of Q_i . At a higher Q_i , outflow changes from oscillatory to non-oscillatory flow (Figure 4). The inset of Figure 4 shows how P_{SG-min} changes with increasing Q_i ; in the case of the equivalent electronic transistor having positive $P_{th-close}$, it is clear that the oscillator will have higher operational Q_i .




FIG. 4. Effect of $P_{th-close}$ on the operational range of the inflow rate. Filled and unfilled points are the experimental and simulation results, respectively. $R_d + R_c$ is $1.7 \times 10^{13} \text{ N}\cdot\text{s}/\text{m}^5$. Gray region ($P_{SG-min} < P_{th-close}$) of the inset is oscillation region.

Interestingly, P_{SG-min} initially decreases; then it increases and finally becomes greater than $P_{th-close}$ (i.e., no oscillation). The initial decrease of P_{SG-min} comes from increasing $P_{th-open}$ with increasing Q_i . This is because the switching-on valve's increased $P_{th-open}$ directly raise the switching-off valve's P_G and thus P_{SG-min} of the switching-off valve decreases. However, as the switching becomes faster, the on-valve, which is to be switched off, cannot faithfully follow the change of $P_{th-open}$. To explain, we compare off- and on-valves' P_S : in the off-valve, the accumulation rate of P_S is proportional to Q_i , whereas in the on-valve, regardless of Q_i , the release of P_S follows characteristic time constant (τ) that is approximated as $(R_c + R_d)C_{tot}$; see Figure S3.¹³ When the valve's on-to-off switching is faster than its τ , the on-valve's P_S discharges insufficiently, thereby raising P_{SG-min} . τ is calculated as 13 s for one on-valve and twofold of that τ for a two valve system. This approximated critical period (26 s) is in relatively good agreement with the base of P_{SG-min} (inset of Figure 4) obtained from computational simulation. The on-off condition of valves and oscillation conditions are summarized in Table S1.¹³

In conclusion, we show that hydraulic membrane-valve have significantly different closing and opening threshold pressures and that the opening threshold pressure changes depending on circuit design parameters such as channel resistances and operation conditions such as flow rates. The negative closing threshold pressure constrains the values of resistors and flow rates that can be used to achieve oscillation. A comprehensive analysis of how the microfluidic resistors and inflow rates affect the oscillator circuit's ability to turn valves off clarifies the conditions under which oscillation can occur. This study provides a fundamental understanding and theoretical framework for a broad range of microfluidic valve-based devices.

This work was supported by the NIH (GM096040-01 and HL084370-05) and Institutional Program for Young Researcher Overseas Visits, Japan Society for the Promotion of Science (JSPS). Devices were partially fabricated in the Lurie Nanofabrication Facility, a member of the National Nanotechnology Infrastructure Network, which is supported by the National Science Foundation.

¹K. W. Oh, K. Lee, B. Ahn, and E. P. Furlani, *Lab Chip* **12**, 515 (2012).

²D. C. Leslie, C. J. Easley, E. Seker, J. M. Karlinsey, M. Utz, M. R. Begley, and J. P. Landers, *Nature Phys.* **5**, 231(2009).

- ³W. H. Grover, R. H. C. Ivester, E. C. Jensen, and R. Mathies, *Lab Chip* **6**, 623 (2006).
- ⁴E. C. Jensen, W. H. Grover, and R. Mathies, *J. Microelectromech. Syst.* **16**, 1378 (2007).
- ⁵M. Rhee and M. A. Burns, *Lab Chip* **9**, 3131 (2009).
- ⁶J. Weaver, J. Melin, D. Stark, S. R. Quake, and M. Horowitz, *Nature Phys.* **6**, 218 (2010).
- ⁷B. Mosadegh, C.-H. Kuo, Y.-C. Tung, Y.-S. Torisawa, T. Bersano-Begey, H. Taviana, and S. Takayama, *Nature Phys.* **6**, 433 (2010).
- ⁸S.-J. Kim, R. Yokokawa, S. C. Leshner-Perez, and S. Takayama, *Anal. Chem.* **84**, 1152 (2012).
- ⁹H. M. Xia, Z. P. Wang, W. Fan, A. Wijaya, W. Wang, and Z. F. Wang, *Lab Chip* **12**, 60 (2012).
- ¹⁰C.-T. Sah, *Proc. IEEE* **76**, 1280 (1988).
- ¹¹R. A. Street, *Adv. Mater.* **21**, 2007 (2009).
- ¹²H. Takao, M. Ishida, and K. Sawada, *J. Microelectromech. Syst.* **11**, 421 (2002).
- ¹³See supplementary material at <http://dx.doi.org/10.1063/1.4769985> for details on close threshold pressure, the oscillator circuit models, and movies.