



UNIVERSITY OF LEEDS

This is a repository copy of *Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology*.

White Rose Research Online URL for this paper:
<http://eprints.whiterose.ac.uk/670/>

Article:

Berry, E., Handley, J.W., Fitzgerald, A.J. et al. (6 more authors) (2004) Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology. *Medical Engineering and Physics*, 26 (5). pp. 423-430. ISSN 1350-4533

<https://doi.org/10.1016/j.medengphy.2004.02.011>

Reuse

See Attached

Takedown

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by emailing eprints@whiterose.ac.uk including the URL of the record and the reason for the withdrawal request.



eprints@whiterose.ac.uk
<https://eprints.whiterose.ac.uk/>



White Rose
university consortium
Universities of Leeds, Sheffield & York

White Rose Consortium ePrints Repository

<http://eprints.whiterose.ac.uk/>

This is an author produced version of a paper published in Medical Engineering and Physics. This paper has been peer-reviewed but does not include the final publisher proof-corrections or journal pagination.

White Rose Repository URL for this paper:
<http://eprints.whiterose.ac.uk/archive/00000670/>

Citation for the published paper

Berry, E. and Handley, J.W. and Fitzgerald, A.J. and Merchant, W.J. and Boyle, R.D. and Zinov'ev, N.N. and Miles, R.E. and Chamberlain, J.M. and Smith, M.A. (2004) *Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology*. Medical Engineering and Physics, 26 (5). pp. 423-430.

Citation for this paper

To refer to the repository paper, the following format may be used:

Berry, E. and Handley, J.W. and Fitzgerald, A.J. and Merchant, W.J. and Boyle, R.D. and Zinov'ev, N.N. and Miles, R.E. and Chamberlain, J.M. and Smith, M.A. *Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology*. Author manuscript available at:
<http://eprints.whiterose.ac.uk/archive/00000670/> [Accessed date].

Published in final edited form as:

Berry, E. and Handley, J.W. and Fitzgerald, A.J. and Merchant, W.J. and Boyle, R.D. and Zinov'ev, N.N. and Miles, R.E. and Chamberlain, J.M. and Smith, M.A. (2004) *Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology*. Medical Engineering and Physics, 26 (5). pp. 423-430.

Multispectral classification techniques for terahertz pulsed imaging: an example in histopathology

Classification of article: Communication

Journal subject area: Imaging and Image Processing

¹Elizabeth Berry, ²James W Handley, ^{1*}Anthony J Fitzgerald, ³WJ Merchant, ²Roger D Boyle, ^{4†}NN Zinov'ev, ⁴Robert E Miles, ⁴J Martyn Chamberlain, ¹Michael A Smith

¹Academic Unit of Medical Physics, University of Leeds, Leeds, LS1 3EX.

²School of Computing, University of Leeds, Leeds, LS2 9JT

³Department of Histopathology, Leeds Teaching Hospitals NHS Trust, Leeds, LS1 3EX

⁴School of Electronic and Electrical Engineering, University of Leeds, Leeds, LS2 9JT

*Present address Teraview Ltd, Cambridge

† Present address National Physical Laboratory, Teddington, Middlesex

Correspondence: Dr E Berry, Academic Unit of Medical Physics, University of Leeds, Wellcome Wing, Leeds General Infirmary, Great George Street, Leeds, LS1 3EX. Tel: 0113 392 2438 Fax: 0113 234 7619 email: e.berry@leeds.ac.uk

Keywords: terahertz pulsed imaging, multispectral classification, clustering, histopathology, basal cell carcinoma, melanoma

Abstract

Terahertz pulsed imaging is a spectroscopic imaging modality using pulses of electromagnetic radiation (100 GHz to 10 THz), and there has been recent interest in studying biomedical specimens. It is usual to display parametric images derived from the measured pulses. In this work, classification was achieved by applying multispectral clustering techniques to sets of parametric images. It was hypothesised that adequate information for clustering was carried in a small number of parametric images, providing these were weighted by complementary physical properties. Materials prepared for histopathological examination were chosen because their condition remained stable during long imaging periods and because their dehydrated state led to greater penetration of the radiation. Two specimens were examined in this pilot study, one of basal cell carcinoma and one of melanoma. Unsupervised ISODATA classification using three selected parametric terahertz pulsed images was compared qualitatively with k-means classification using the shape of the whole time series, and with conventional stained microscope slides. There was good qualitative agreement between the classifications. Classifications were consistent with the morphological appearances expected, but further work is required to determine if tumour discrimination is possible. The results have implications for the future development of the technique as the need for only a small number of features could lead to considerably reduced acquisition times.

Introduction

Terahertz radiation is the part of the electromagnetic spectrum between the far infrared and microwave regions, often defined as 100 GHz to 10 THz (or 30 μm to 3 mm in wavelength).

There has been a recent surge of interest in using such radiation because of the development of new, relatively compact, methods of generating and detecting coherent terahertz radiation^{1,2}. Pulsed sources of radiation are ideal for spectroscopy^{3,4}, and the use of terahertz frequency radiation for these investigations is attractive because in molecules the separations between rotational energy levels, and between vibrational energy levels, correspond to this frequency band⁵. It is also possible to form images using pulsed sources of terahertz radiation^{6,7}. Although such images have potential to display information about the spatial distribution of the spectroscopic data, there is a reduced frequency resolution compared with purely spectroscopic investigations.

In intact samples, or for tissue *in vivo* however, the reduction in spectral resolution is not such a concern as would be the case for specially prepared samples. This is because there is a possibility that strong water absorption may mask signature features from other molecules. Additionally, it is expected that smoothing of spectral features will arise from the presence of a mixture of several molecules, and because different chemical environments will exist even for a single molecule.

The absence of strong spectral features means that an appropriate way to display image data is by using parametric images generated from the acquired data that consists of a time series at each pixel. Parametric images are used in other areas of medical imaging. For example in nuclear medicine, parametric images of amplitude and phase of ejection from the left ventricle can be obtained from cardiac multiple gated acquisitions (MUGA). In magnetic resonance imaging (MRI), parametric images are used to illustrate uptake of contrast medium in dynamic studies. In terahertz pulsed imaging, a parametric image might show a parameter calculated in the time domain, such as the time delay for the peak of the pulse relative to a

reference pulse that has traversed the same path but with no sample present. Parameters may also be calculated in the frequency domain, following Fourier transformation of the pulse. An example would be the transmittance, or relative intensities of sample and reference at a particular frequency, which is commonly used in spectroscopy and optics. Typical parameters, with explanations of how they relate to physical properties of a material, are described by Löffler et al.⁸. As many parameters may be used, each acquired data set can lead to a family of parametric images.

If multiple parametric images are generated from each data set, it is then necessary to combine the information from the images in order to maximise the use of the information present. This is a common problem in image analysis, and one solution is to use multispectral classification methods. It is assumed that a particular class will be made up of members that have similar values for each of a number of features. In a multi-dimensional feature space, all the members of a particular class will be clustered together. In the case of parametric images, the features are the parameters and the feature values are the pixel values. Multispectral classification techniques work by assigning pixels to classes based on their similarity to other pixels already assigned to that class. Such classifiers have been used very successfully with MRI data^{9,10} for classification in the brain. In MRI, several different pulse sequences are used to obtain sets of images with different contrast properties, and this means that a registration step is often required before multispectral classification, to compensate for movement of the subject between the acquisitions. Terahertz pulsed imaging data is even better suited to this sort of analysis. There is no need for a registration step because the sets of images to be used come from a single scanning sequence and are thus in spatial registration. For the optimisation of clustering in feature space, well separated clusters are desirable, and so ideally the images used should each be related to different physical

properties of the material. Although several physical properties can affect the values of the parameters⁸, particular parametric images have greater dependence on a particular physical property, and such images should be valuable for clustering. The choice of appropriate parametric images is discussed in the theory section below.

Theory

Multispectral classification

In this article we develop the use of the parametric images in terahertz imaging by proposing that the information from separate parametric images be combined using multispectral classification. Image analysis tools that can perform multispectral classification are readily available, and for these it is necessary to have generated two or more images that are spatially registered to one another. Note that, more generally, clustering classification analysis can be performed without the explicit generation and display of images. For example, to ensure that the entire shape of the pulse at each pixel was taken into account, the parameters used might be the set of time series amplitudes or the coefficients of the Fourier transform. This would lead to a feature space where the number of dimensions was equal to the number of points or coefficients.

Automated classifiers fall into two groups, supervised and unsupervised. In the former, pixels known to belong to particular classes are used for initialisation, and all other pixels are assigned to classes on the basis of their similarity to members of this initial training set. In contrast, unsupervised classifiers do not use a training set, though they are not entirely free of user input as it is necessary to define input parameters, to set stopping criteria for example. No training data were available for this work, so unsupervised classifiers were used.

Two similar algorithms were used. The first of these was the ISODATA algorithm, originally described by Duda and Hart¹¹. It was chosen as it is available for unsupervised classification in the AnalyzeTM biomedical image analysis software (AnalyzeDirect.com, Lenexa, Kansas, USA), and so could be readily applied to a small number of parametric images. In the ISODATA algorithm, the starting locations for the class centroids are set to cover a broad range of feature space, and each pixel is put into the class whose centroid is closest (in Euclidean distance) in feature space. Once each pixel has been assigned, the centroid locations are updated and the algorithm is repeated. The stopping criterion is a set percentage of pixels that have changed class from the previous iteration. The merge criterion allows small classes to be discarded and merged with the nearest class centroid, based on a user defined minimum separation for classes. This minimum separation is called the threshold for the merge and is expressed as a distance in feature space. Definition of clusters using Euclidean distance is justified where the feature space is isotropic with an even spread of data.

In order to allow comparison with classifications based on information from a large number of features, such as the time series amplitudes, the widely used k-means clustering algorithm was employed. The algorithm is unsupervised and broadly similar to the ISODATA approach, the main difference being that a pre-defined number of classes is found. In the k-means algorithm, pixels are assigned to one of k classes, whose centres are defined by the mean value of the feature vector for that class. The initial k cluster centres are chosen randomly, then each pixel is assigned to the class to which it is nearest. The mean value of the feature vector is re-calculated for each class and pixels re-assigned. The algorithm is repeated until the calculated means remain unchanged.

Choice of parametric images for ISODATA algorithm

Although the terahertz frequency band includes regions conventionally approached using the optical properties (300 GHz to 115 THz) or the dielectric properties (15 GHz to 300 GHz) of materials, in this work the behaviour was considered to be entirely optical in character. This is consistent with guidance on radiofrequency electromagnetic fields¹², in which exposures are treated as quasi-optical for all frequencies over 6 GHz. Attenuation of terahertz radiation takes place by reflection, scattering and absorption. If scattering may be neglected, then the refractive indices and linear absorption coefficients of the materials along the beam path will determine the degree of attenuation, as these properties determine the degree of reflection and absorption as defined by the Fresnel coefficients¹³. For uniform thickness, single layer, geometry the pulse delay will be determined solely by the refractive index. The pulse height will be strongly dependent on absorption coefficient, particularly if reflection losses are small compared with attenuation from absorption as is expected in this case. In analogy to MRI, the two parametric images could be described as refractive index-weighted and absorption coefficient-weighted images – whilst the images do not map absolute values of refractive index and absorption coefficient, these are the parameters that affect the contrast most strongly. A further time domain parameter that is simple to calculate is the full width at half maximum (FWHM) of the pulse. This will be affected by both the refractive index (especially where the medium is strongly dispersive) and the absorption coefficient, and was chosen as the third image parameter because of its potential to provide complementary information.

The purpose of this pilot study was to investigate the feasibility of using multispectral classifiers on pulsed terahertz datasets, to determine if classifications consistent with

expected tumour morphology could be generated from multispectral classification using the three complementary time domain features described above. It was hypothesised that adequate information for clustering is carried in a small number of parametric images, providing these are weighted by complementary physical properties. As no reference classification was available, the validity of the classifications made using only three features was qualitatively assessed by comparison with those produced using a larger number of features. The use of a small number of features is desirable as with suitably adjusted acquisition methods it could lead to considerably reduced acquisition times.

Methods

In this preliminary work materials prepared for histopathological examination, but sliced less thinly, were chosen. This choice of prepared tissue was primarily because basic terahertz imaging equipment, as used in this work, is slow. Image acquisition could extend well over 24 hours, a period during which the condition and hydration of untreated biological materials would be expected to change. Tissue already prepared for microscopy had been treated by dehydration and fixed with formalin, so it was not expected that the physical properties would change during imaging. The stability of the samples would also allow for repeat imaging on another date. An associated benefit was that de-hydration of the tissue would lead to an increased depth of penetration of terahertz radiation compared with its untreated state, allowing the transmission mode of imaging to be used, which is simpler to analyse than the alternative reflection mode. Diagnosis in histopathology is based on morphological appearance under the microscope. Once samples have been stained with haematoxylin and eosin, which are taken up in varying degrees by different types of cell, different tumours have characteristic appearances. It is known that contrast in terahertz images between tissues in the living body is strongly dependent upon the degree of hydration¹⁴. However, if the water

is removed, as is the case in the samples used here, then the contrast may be more closely associated with the features identified by the pathologist, which are associated with the distribution of cell types. This makes the samples particularly attractive for the application of classification techniques.

Local research ethics committee approval was given to use human tissue in this study. Two biopsy samples were chosen. In each case they had previously been prepared in the routine manner by dehydration, formalin fixation and mounting in paraffin wax. Diagnosis was made by the pathologist from a thin slice taken from the block of prepared tissue and the remaining tissue was then stored in case further samples were required for clinical purposes. One sample was diagnosed as a nodular basal cell carcinoma (BCC), the other as melanoma. In both cases a large group of cells within the dermis, displacing the normal structures of the skin, was seen (Figure 1). In the case of the BCC a characteristic low cellularity, fibromyxoid stroma was seen separating the tumour from normal structures (Figure 1a arrowed).




Figure 1. Stained slides, magnification 2.5, used for diagnosis of (a) BCC and (b) melanoma; these are from the same tissue samples but do not represent the same slice or area as the sample used to obtain the terahertz images. In each case normal tissue appears lighter

than the tumour, and the epidermis is at the top of the image with the direction into the skin running from top to bottom. The arrow indicates a characteristic low cellularity, fibromyxoid stroma separating the tumour from normal structures in the BCC.

For this study a further, thicker slice (1 mm) was removed from the block and mounted in wax, for terahertz pulsed imaging. As this was done at some time after the initial diagnosis it was not possible to spatially register the new slice to the microscope slides used for diagnosis.

Pulsed terahertz imaging¹⁵ was performed using imaging equipment shown schematically in Figure 2. The method used is the pump and probe technique, which is commonly applied in optical spectroscopy. The beam of an ultrafast pulsed infrared laser is divided into two, and one part is used as the pump to generate terahertz pulses, while the other is part of a coherent detection scheme used to measure the amplitude of the terahertz electric field after transmission through the subject. The optical delay line allows sampling in time of the terahertz pulse at each pixel, meaning that the total acquisition time increases as more time points are acquired, whilst raster scanning of the sample leads to an array of time series ready for the production of parametric images. In each case 64 time points were acquired at 0.15 ps intervals, and a reference pulse was acquired in the same way through an area containing only the mounting wax. For the melanoma a 7 x 7 mm area was imaged, with 39 pixels in each direction. For the basal cell carcinoma, a 10 x 10 mm area was imaged with 56 pixels in each direction. The pixel size was 0.18 mm in both cases.




Figure 2. Schematic layout of a terahertz pulsed imaging system in transmission mode. The terahertz beam generated from an antenna or non-linear crystal is collimated and then focused on the sample using off axis parabola mirrors¹⁵.

Three time domain parametric images were generated: (i) Pulse time delay relative to reference pulse (refractive index weighted), (ii) pulse amplitude relative to reference pulse (absorption coefficient weighted) and (iii) the full width at half maximum of the pulse. The feature space for these images was made suitable for the ISODATA method by re-scaling into the range 0 to 255. This meant that any merge threshold (distance in feature space) chosen would also be expressed as a value in the range 0 to 255.

Unsupervised ISODATA clustering with merging was applied to the groups of three images, using the AnalyzeTM software. For the first classification, a single merge threshold of 70 was used for the whole image. A second set of classifications was obtained using the two stage merge process described below, so that small changes in composition within the tumour were identified without multiple classes also being defined in the background regions. The merge threshold of 70 was retained for the entire image except for the innermost region defined at this threshold. For this region the merge threshold distance was halved to a value of 35 to highlight structural differences that may be of value for tumour discrimination. For comparison, a classification using many features was performed using unsupervised k-means clustering (without any merging) on the full set of time series amplitudes of the two datasets; 64 features were used and 8 classes were defined.

Results

The three time domain parametric images for the BCC and melanoma samples are shown in Figure 3. The terahertz images show the tumour mass as an area of higher attenuation (Figure 3a), with particularly strong attenuation around the borders especially noticeable on the BCC image. Further investigation will be needed to determine if the bright regions in the FWHM image of the BCC (Figure 3c), are associated with the characteristic clefts seen around BCC tumours.





Figure 3. Time domain parametric images for the BCC (top row) and melanoma (bottom row) samples. The arrows indicate the location of the epidermis and direction into the skin. The reference pulse was transmitted through the wax surrounding the sample. (a) Relative pulse amplitude, dark values represent high attenuation compared with the reference, which appears bright. (b) Time delay relative to reference pulse, dark values represent small delay times compared with that of the reference pulse, brighter areas represent longer time delays (c) FWHM of pulse peak, dark values represent a pulse width similar to that of the reference pulse, brighter areas have larger FWHM.


The classified images from the ISODATA algorithm are shown in Figures 4 and 5. Boundary regions, shown in blue, have been classified in both cases. These are several pixels in width and are unlikely to be fully explained by partial volume effects. For the merge threshold used here, two regions were defined for the melanoma and three for the BCC. Figure 5 shows the sub-classifications achieved within the innermost regions shown in Figure 4. The internal

structure of the BCC is more organised than that of the melanoma, which is consistent with the appearances of the stained slides seen in Figure 1.

Results from a k-means classification using time series features appear in Figure 6. These show the same shapes as the ISODATA classification using three features, and for the BCC the well-defined internal structures are also seen. There are more classes defined in the background regions of these images because the number of classes was fixed at eight, without the merging options used for the ISODATA algorithm.




(a)




(b)

Figure 4. Classifications resulting from three parametric images, using the unsupervised ISODATA algorithm with a single merge step (threshold distance 70) (a) BCC (b) melanoma.




(a)



(b)

Figure 5. Classifications resulting from three parametric images, using the unsupervised ISODATA algorithm with two merge steps. A threshold of 35 was used to sub-classify the innermost region of each classification in Figure 4. (a) BCC (b) melanoma.



(a)



(b)

Figure 6. Results of k-means clustering into eight classes using the 64 time series values. (a) BCC (b) melanoma

Discussion

In spite of a surge of interest in biomedical applications for terahertz imaging and although such data have appeared in conference presentations, at the time of writing only one image of histopathologically prepared human tissue has appeared in the literature¹⁶. This work demonstrated terahertz imaging of a human liver sample containing several tumours. The appearances were consistent with our findings, with the tumours seen as regions of higher absorption in both a pulsed terahertz image (integrated from 0.2 to 0.5 THz) and continuous wave imaging at 230 GHz. As the density of the metastases and normal tissue were measured to differ by only 5% to 10%, the authors concluded that the higher absorption was caused by differing chemical composition. Only one parametric image was presented and no classification methods were applied. Veterinary histopathology samples were used by Löffler et al⁸ to demonstrate the various parameters that can be used for imaging. They hypothesised that the enhanced boundary seen on certain parametric images of a mast cell tumour was the result of diffraction or scattering, and investigated the phenomenon further using the technique of dark-field imaging¹⁷. The boundary regions seen in our classified images are consistent with these findings. To establish if the feature is an artefact that may be ignored, or if in some cases it carries information that may help in tumour discrimination, will require further studies. These may use a range of spatial resolutions and boundary properties, together with detection of radiation that has been scattered out of the incident direction.

The larger of the two datasets presented here took 34 hours to acquire, using the optical delay line shown in Figure 2. Developments in instrumentation have already led to reduced acquisition times¹, but further reductions to allow real time imaging could be achieved by development of rapid imaging techniques that avoid the need to acquire the full time series at every pixel. In the work presented here, there was good qualitative agreement between

classification using only the three features and classification using multiple features to represent the whole pulse. This supports the hypothesis that adequate information for clustering is carried in a small number of parametric images, if these are weighted by complementary physical properties. The results are encouraging for the development of a custom designed acquisition to determine values for the three parameters alone. Furthermore, the concept of using weighted images for classification reduces the computational demands of the technique, compared with a method using absolute values of the physical properties. While it is possible to extract physical properties from image data^{18,19}, this is computationally intensive.

The absence of training data meant that it was necessary to use unsupervised classifiers in this work. Supervised classifiers have previously been successfully applied to terahertz pulsed imaging data²⁰, capitalising on the availability of large amounts of training data from chicken and beef. Although such training data are not yet available for human tissue²¹, that work taken together with the results presented here, indicate that the methods show much promise, providing that clinically relevant differences between tissues exist.

Effort was concentrated on three complementary time domain images, as they were suited to the direct application of multispectral techniques. There is also potential for using frequency domain features, but pre-processing steps will be required to avoid artefacts occurring because of the differing levels of noise and differing spatial resolution in the images^{22,16}.

Terahertz images offer the possibility of providing information from deeper structures than currently available to the pathologist from stained slides. Whilst the edges of the tumours appear to have been defined, it is necessary to confirm this experimentally. Similarly, the

differences between tumour types that have been suggested in this pilot study require further investigation, and in particular it is essential to include location-specific diagnoses using one or more reference techniques, and samples of healthy tissue for comparison.

Conclusions

These results confirm that the families of images that can be generated from terahertz pulsed imaging data sets are amenable to multispectral classification, and that classification can be made using a subset of the total information available. Further investigation is indicated.

Acknowledgements

We are grateful for financial support from the Engineering and Physical Sciences Research Council and from the European Union under the Teravision Project (IST-1999-10154). The help of Peter Jackson with sample preparation is acknowledged with thanks.

References

- [1] Grischkowsky DR and Mittleman DM. Introduction in 'Sensing with terahertz radiation', Ed. Mittleman DM. Springer, Berlin, Heidelberg, New York, 2003: 1-38
- [2] Auston DH and Nuss MC. Electrooptic generation and detection of femtosecond electrical transients. *IEEE Journal of Quantum Electronics* 1988; 24: 184-197
- [3] Grischkowsky D, Keiding S, van Exter M, and Fattinger C. Far-infrared time-domain spectroscopy with terahertz beams of dielectrics and semiconductors. *Journal of the Optical Society of America B-Optical Physics* 1990; 7: 2006-2015
- [4] Kindt JT and Schmuttenmaer CA. Far-infrared dielectric properties of polar liquids probed by femtosecond terahertz pulse spectroscopy. *Journal of Physical Chemistry* 1996; 100: 10373-10379
- [5] Smye SW, Chamberlain JM, Fitzgerald AJ, and Berry E. The interaction between Terahertz radiation and biological tissue. *Physics in Medicine and Biology* 2001; 46: R101-R112
- [6] Hu BB and Nuss MC. Imaging with Terahertz Waves. *Optics Letters* 1995; 20: 1716-1719
- [7] Wu Q, Hewitt TD, and Zhang XC. Two-dimensional electro-optic imaging of THz beams. *Applied Physics Letters* 1996; 69: 1026-1028

- [8] Löffler T, Siebert K, Czasch S, Bauer T, and Roskos HG. Visualization and classification in biomedical terahertz pulsed imaging. *Physics in Medicine and Biology* 2002; 47: 3847-3852
- [9] Lundervold A, Taxt T, Ersland L, and Fenstad AM. Volume distribution of cerebrospinal fluid using multispectral MR imaging. *Medical Image Analysis* 2000; 4: 123-136
- [10] Vinitiski S, Gonzalez C, Mohamed F, Iwanaga T, Knobler RL, Khalili K, and Mack J. Improved intracranial lesion characterization by tissue segmentation based on a 3D feature map. *Magnetic Resonance in Medicine* 1997; 37: 457-469
- [11] Duda RO and Hart PE. 'Pattern classification and scene analysis', John Wiley and Sons Inc, New York , 1973.
- [12] Institute of Electrical and Electronic Engineers. 'Standard for safety levels with respect to human exposure to radiofrequency electromagnetic fields, 3 kHz to 300 GHz (IEEE C95.1 1999 Edition)', Institute of Electrical and Electronic Engineers, New York , 1999.
- [13] Hecht E and Zajac A. 'Optics', Addison-Wesley Pub. Co, Reading, MA , 1974.
- [14] Mittleman DM, Gupta M, Neelamani R, Baraniuk RG, Rudd JV, and Koch M. Recent advances in terahertz imaging. *Applied Physics B-Lasers and Optics* 1999; 68: 1085-1094

- [15] Mittleman DM. Terahertz Imaging in 'Sensing with Terahertz Radiation', Ed. Mittleman DM. Springer-Verlag, Berlin Heidelberg New York, 2003: 117-154
- [16] Knobloch P, Schildknecht C, Kleine-Ostmann T, Koch M, Hoffmann S, Hofmann M, Rehberg E, Sperling M, Donhuijsen K, Hein G, and Pierz K. Medical THz imaging: an investigation of histo-pathological samples. *Physics in Medicine and Biology* 2002; 47: 3875-3884
- [17] Löffler T, Bauer T, Siebert KJ, Roskos HG, Fitzgerald A, and Czasch S. Terahertz dark-field imaging of biomedical tissue. *Optics Express* 2001; 9: 616-621
- [18] Dorney TD, Baraniuk RG, and Mittleman DM. Material parameter estimation with terahertz time-domain spectroscopy. *Journal of the Optical Society of America A-Optics Image Science and Vision* 2001; 18: 1562-1571
- [19] Duvillaret L, Garet F, and Coutaz JL. Highly precise determination of optical constants and sample thickness in terahertz time-domain spectroscopy. *Applied Optics* 1999; 38: 409-415
- [20] Ferguson B, Wang S, Gray D, Abbott D, and Zhang XC. Identification of biological tissue using chirped probe THz imaging. *Microelectronics Journal* 2002; 33: 1043-1051
- [21] Berry E, Fitzgerald AJ, Zinovev NN, Walker GC, Homer-Vanniasinkam S, Sudworth CD, Miles RE, Chamberlain JM, and Smith MA. Optical properties of

tissue measured using terahertz pulsed imaging. *Proceedings of SPIE: Medical Imaging* 2003; 5030: 459-470

- [22] Fitzgerald AJ, Berry E, Miles RE, Zinovev NN, Smith MA, and Chamberlain JM. Evaluation of image quality in terahertz pulsed imaging using test objects. *Physics in Medicine and Biology* 2002; 47: 3865-3873

Tables

No tables included

Abbreviated title

Multispectral classification of THz images

Figure captions

Figure 1. Stained slides, magnification 2.5, used for diagnosis of (a) BCC and (b) melanoma; these are from the same tissue samples but do not represent the same slice or area as the sample used to obtain the terahertz images. In each case normal tissue appears lighter than the tumour, and the epidermis is at the top of the image with the direction into the skin running from top to bottom. The arrow indicates a characteristic low cellularity, fibromyxoid stroma separating the tumour from normal structures in the BCC.

Figure 2. Schematic layout of a terahertz pulsed imaging system in transmission mode. The terahertz beam generated from an antenna or non-linear crystal is collimated and then focused on the sample using off axis parabola mirrors¹⁵.

Figure 3. Time domain parametric images for the BCC (top row) and melanoma (bottom row) samples. The arrows indicate the location of the epidermis and direction into the skin. The reference pulse was transmitted through the wax surrounding the sample. (a) Relative pulse amplitude, dark values represent high attenuation compared with the reference, which appears bright. (b) Time delay relative to reference pulse, dark values represent small delay times compared with that of the reference pulse, brighter areas represent longer time delays

(c) FWHM of pulse peak, dark values represent a pulse width similar to that of the reference pulse, brighter areas have larger FWHM.

Figure 4. Classifications resulting from three parametric images, using the unsupervised ISODATA algorithm with a single merge step (threshold distance 70) (a) BCC (b) melanoma.

Figure 5. Classifications resulting from three parametric images, using the unsupervised ISODATA algorithm with two merge steps. A threshold of 35 was used to sub-classify the innermost region of each classification in Figure 5. (a) BCC (b) melanoma.

Figure 6. Results of k-means clustering into eight classes using the 64 time series values. (a) BCC (b) melanoma