

Research

Open Access

An active electrode for biopotential recording from small localized bio-sources

Emil S Valchinov and Nicolas E Pallikarakis*

Address: Department of Medical Physics, University of Patras, Patras 26500, Greece

Email: Emil S Valchinov - emil@bit.med.upatras.gr; Nicolas E Pallikarakis* - nipa@bme.med.upatras.gr

* Corresponding author

Published: 22 July 2004

Received: 02 June 2004

BioMedical Engineering OnLine 2004, **3**:25 doi:10.1186/1475-925X-3-25

Accepted: 22 July 2004

This article is available from: <http://www.biomedical-engineering-online.com/content/3/1/25>

© 2004 Valchinov and Pallikarakis; licensee BioMed Central Ltd. This is an open-access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/2.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Abstract

Background: Laser bio-stimulation is a well-established procedure in Medical Acupuncture. Nevertheless there is still a confusion as to whether it works or the effect is just placebo. Although a plethora of scientific papers published, showing positive clinical results, there is still a lack of objective scientific proofs about the bio-stimulation effect of lasers used in Acupuncture. The objective of this work was to design and build a body surface electrode and an amplifier for biopotential recording from acupuncture points, considered here as small localized bio-sources (SLB). The design is aimed for studying SLB potentials provoked by laser stimulus, in search for objective proofs of the bio-stimulation effect of lasers used in Medical Acupuncture.

Methods: The active electrode presented features a new adjustable anchoring system and fractionation of the biopotential amplifier between the electrode and the cabinet's location. The new adjustable electrode anchoring system is designed to reduce the electrode-skin contact impedance, its variation and motion artifacts. That is achieved by increasing the electrode-skin tension and decreasing its relative movement. Additionally the sensing element provides local constant skin stretching thus eliminating the contribution of the skin potential artifact. The electrode is attached to the skin by a double-sided adhesive pad, where the sensing element is a stainless steel, 4 mm in diameter. The fractionation of the biopotential amplifier is done by incorporating the amplifier's front-end op-amps at the electrodes, thus avoiding the use of extra buffers. The biopotential amplifier features two selectable modes of operation: semi-AC-mode with a -3 dB bandwidth of 0.32–1000 Hz and AC-mode with a bandwidth of 0.16–1000 Hz.

Results: The average measured DC electrode-skin contact impedance of the proposed electrode was 450 k Ω , with electrode tension of 0.3 kg/cm² on an unprepared skin of the inner forearm. The peak-to-peak noise voltage measured at the amplifier output, with input terminals connected to common, was 10 mV_{p-p}, or 2 μ V_{p-p} referred to the input. The common-mode rejection ratio of the amplifier was 96 dB at 50 Hz, measured with imbalanced electrodes' impedances. The prototype was also tested practically and sample records were obtained after a low intensity SLB laser stimulation. All measurements showed almost a complete absence of 50 Hz interference, although no electrolyte gel or skin preparation was applied.

Conclusion: The results showed that the new active electrode presented significantly reduced the electrode-skin impedance, its variation and motion artifact influences. This allowed SLB signals with relatively high quality to be recorded without skin preparation. The design offers low noise and major reduction in parts, size and power consumption. The active electrode specifications were found to be better or at least comparable to those of other existing designs.

Background

The non-invasive nature of laser bio-stimulation have made lasers an attractive alternative in Medical Acupuncture at the last 25 years. Unfortunately there is still a confusion as to whether they work or their effect is just placebo. Although a plethora of scientific papers published, showing positive clinical results, there is still a lack of objective scientific proofs about the bio-stimulation effect of lasers used in Acupuncture. The objective of this work was to design and built a body surface electrode and an amplifier for biopotential recording from acupuncture points. The latter are considered here as small localized bio-sources (SLB). As discussed by other authors, SLB are small area body regions with specific electrical, physiological and anatomical properties (e.g. high density of gap junctions, relatively low impedance etc.) [1-4]. They appear to be highly sensitive to mechanical, thermal, electrical or electromagnetic stimulation and are found to take place from the epidermis (stratum granulosum) to a maximum depth of 2 cm [5-8].

The active electrode is aimed for studying SLB potentials provoked by laser stimulus, in search for objective proofs of the bio-stimulation effect of lasers used in Medical Acupuncture.

Methods

Electrode design

The attempt to define the optimal parameters of the active electrode was based on a set of preliminary measurements performed in our laboratory. Anatomical, physiological and electrical characteristics of the signal source were considered. The SLB AC signal level, after stimulation, varied from subject to subject, but did not exceed 1 mV peak-to-peak (p-p). Additionally SLB occasionally manifested a high DC potential up to 200 mV, implying the use of differential amplifier with optional DC coupling and wide DC input range. The frequency band of the signal of interest was found to be in the range 0–200 Hz. The preliminary experiments showed that SLB potentials were best recorded with small pasteless electrodes although their contact impedance depends strongly on sweat gland secretion. The application of electrolytic gel resulted in significant reduction of the SLB signal amplitude, probably due to smoothing of the potential caused by saturation of the epidermis with electrolyte. Moreover, potentials between closely spaced SLB might be shortened by the application of excessive gel or large surface electrodes. An additional difficulty is that the SLB are often situated at convex or concave body surface areas where large flat electrodes could not be easily affixed. Skin abrasion with sandpaper is also not recommended since it can cause skin irritation and SLB potential changes. However, the use of small passive dry electrodes on an unprepared skin results in high electrode-skin impedance, motion artifacts, high power-

line cable interference and noise. When the electrodes are DC coupled to the amplifier, a motion induced interfering signal appears at the amplifier input, mainly due to:

- Electrokinetic effect – the disturbance of the double layer of charge at the electrode-skin interface causes variations of the DC polarization potential [9].
- Skin potential or skin stretch artifact – stretching of the skin causes a change of the potential of the barrier layer between the epidermis and the dermis [10].
- Variation in the electrode-skin contact resistance – caused by the amplifier input bias current and the current flowing due to the polarization potential.

The complex electrochemical interactions that take place at the electrode-skin interface have been subject to much study in order an equivalent electrical model to be developed [10-12]. The simple but adequate electrical model used, is shown in Fig. 1, where $C_d//R_d$ is the coupling impedance of the double layer at the electrode-skin interface, $C_i//R_i$ is the amplifier input impedance, R_s is the minimum series contact resistance and V_{pol} is the DC polarization potential. Then the motion artifact signal at the amplifier input can be expressed as

$$V_{mot} = \Delta V_{pol} + \Delta V_{skin} + (\Delta R_d + \Delta R_s) (V_{POL} / R_i + i_b) \quad (1)$$

where ΔV_{skin} is the skin stretch artifact and i_b is the amplifier input bias current. It was deduced that in order to keep the resistive interfering component less than 10 μV when DC coupling is employed and with both currents contributing equally to it, then $i_b < 50$ pA and $R_s > 1$ G Ω [11]. If an AC coupling is used then the resistive component of the motion artifact is eliminated.

For dry electrodes the motion artifacts are mainly caused by changes of the polarization potential and the contact impedance due to the poor electrolyte layer at the electrode-skin interface. Therefore a firm electrode-skin contact is of primary importance. Thus a new adjustable electrode anchoring system was designed for the purpose, as shown in Fig. 2. Turning the electrode cap clockwise pushes the sensing element against the skin. Thus electrode-skin pressure is increased, leading to reduction of the contact impedance and its variation. The electrode-skin relative movement is also reduced, making the noise contribution of the electrokinetic effect insignificant. Additionally, the sensing element provides constant skin stretching that lowers the contribution of the skin potential artifact. Turning the electrode cap counter-clockwise, releases the spring, which pushes back the sensing element, resulting in a lower electrode-skin pressure. Titanium, stainless steel and aluminum were considered as




Figure 1
Equivalent electrical model of the electrode-skin interface.

electrode sensing materials. Stainless steel was chosen because it is more commonly available than titanium and preferable to aluminum, which has been shown to have problems due to chemical response of its oxide to perspiration [13]. The use of stainless steel electrode material ensures low noise, minimal offset potentials and excellent DC stability suitable for low-drift DC measurements [14]. The sensing element is a 4 mm diameter, selected according to the average SLB size. Although the obtained contact impedances with the prototype electrode were relatively low for dry electrodes of that size, impedance matching at the electrode site was still needed to cope with the power-line cable interference.

Basic amplifier circuit

Electrodes with impedance matching at the sensing site are referred to as active electrodes and have been designed since 1960's [15-17]. The electronic part of these transducers mostly consists of a buffer amplifier, but some have been designed to need only two lead connection wire [18,19]. However, as the signal is not amplified, buffers introduce significant noise and a low noise amplifier is still needed at the front-end. In order to avoid this drawback we used a two-op-amp biopotential amplifier [20] shown in Fig. 3, where op-amps A_0 and A_1 were integrated at the electrodes (Fig. 4), instead of using extra buffers. This resulted in lower noise and less parts, at the expense of increased number of electrode leads. The amplifier is based on the two-op-amp instrumentation amplifier

shown in Fig. 5. The output voltage of the basic two-op-amp amplifier is

$$U_1(s) = A_{d1}(s)(E_1 - E_0) + A_{c1}(s)(E_1 + E_0)/2 \quad (2)$$

where $A_{d1}(s)$ is the differential-mode (DM) gain and $A_{c1}(s)$ is the common-mode (CM) gain of op-amp A_1 . If we take the usual definitions for the DM input signal, $V_d=(E_1-E_0)$, and for the CM input signal, $V_c=(E_1 + E_0)/2$, then the output voltage can be also written as

$$U_1(s) = A_d(s)V_d + A_c(s)V_c \quad (3)$$

It can be shown that the respective expressions for the DM gain $A_d(s)$, and the CM gain $A_c(s)$, are given by

$$A_d(s) = \frac{1}{2} \frac{\left(1 + \frac{Z_1(s)}{Z_1(s)+R_2} A_{d0}(s)\right) \left(A_{d1}(s) + \frac{A_{c1}(s)}{4}\right) + A_{d0}(s) \left(\frac{Z_E(s)}{R_3 + Z_E(s)}\right) \left(A_{d1}(s) - \frac{A_{c1}(s)}{4}\right)}{\left(1 + \frac{Z_1(s)}{Z_1(s)+R_2} A_{d0}(s)\right) \left[1 + \frac{R_3}{R_3 + Z_E(s)} \left(A_{d1}(s) - \frac{A_{c1}(s)}{4}\right)\right]} \quad (4)$$

$$A_c(s) = \frac{\left(1 + \frac{Z_1(s)}{Z_1(s)+R_2} A_{d0}(s)\right) \left(A_{d1}(s) + \frac{A_{c1}(s)}{4}\right) - A_{d0}(s) \left(\frac{Z_E(s)}{R_3 + Z_E(s)}\right) \left(A_{d1}(s) - \frac{A_{c1}(s)}{4}\right)}{\left(1 + \frac{Z_1(s)}{Z_1(s)+R_2} A_{d0}(s)\right) \left[1 + \frac{R_3}{R_3 + Z_E(s)} \left(A_{d1}(s) - \frac{A_{c1}(s)}{4}\right)\right]} \quad (5)$$

where $A_{d0}(s) = \frac{A_{d0}}{1 + s\tau_0}$, $A_{d1}(s) = \frac{A_{d1}}{1 + s\tau_1}$, τ_0 and τ_1 are the respective DM open loop gains, and the time constants of the first poles of op-amps A_0 and A_1 (assumed to be internally compensated), and $A_{c1}(s)$ is the CM gain of op-amp




Figure 2
Dry active electrode with adjustable anchoring system.

A_1 . It should be noted that the CM gain of op-amp A_0 is omitted from (4) and (5), since its influence on both gains is insignificant.

Considering (4) and (5), then the common-mode rejection ratio $CMRR(s)$ is

$$CMRR(s) = \frac{A_d(s)}{A_c(s)} = \frac{1}{2} \frac{A_{00}(s)A_{01}(s) \left(\frac{Z_1(s)}{Z_1(s)+R_2} + \frac{Z_E(s)}{R_3+Z_E(s)} \right) + A_{00}(s) \frac{A_1(s)}{4} \left(\frac{Z_1(s)}{Z_1(s)+R_2} - \frac{Z_E(s)}{R_3+Z_E(s)} \right) + A_{01}(s) + \frac{\Delta_1(s)}{4}}{A_{00}(s)A_{01}(s) \left(\frac{Z_1(s)}{Z_1(s)+R_2} - \frac{Z_E(s)}{R_3+Z_E(s)} \right) + A_{00}(s) \frac{A_1(s)}{4} \left(\frac{Z_1(s)}{Z_1(s)+R_2} + \frac{Z_E(s)}{R_3+Z_E(s)} \right) + A_{01}(s) + \frac{\Delta_1(s)}{4}} \quad (6)$$

Assuming op-amps A_0 and A_1 are ideal then the only factor contributing to the $CMRR$ is the mismatching of the resistors. Thus we can define a common-mode rejection ratio for the resistors, $CMRR_R$. By taking $1/A_{d0}(s) = A_c(s) = 0$ in (6) we obtain

$$CMRR_R(s) = \frac{1}{2} \frac{2Z_1(s)Z_E(s) + R_2Z_E(s) + Z_1(s)R_3}{R_2Z_E(s) - Z_1(s)R_3} \quad (7)$$

Therefore $CMRR_R(s)$ approaches infinity if the relevant impedances are chosen according to

$$\frac{Z_1(s)}{R_2} = \frac{Z_E(s)}{R_3} \quad (8)$$

If the condition in (8) is fulfilled and op-amps A_0 and A_1 are ideal, then (4) simplifies to

$$A_d(s) = \frac{Z_E(s)}{R_3} + 1 \quad (9)$$

Considering (7), equation (6) can be written as




Figure 3
Schematic of the biopotential amplifier with active DC rejection/suppression.

$$\frac{1}{CMRR(s)} = \frac{1}{2CMRR_{A1}(s)} + \frac{1}{A_{d0}(s)/2} + \frac{1}{CMRR_R(s)} \quad (10)$$

where $CMRR_{A1}(s)$ is the CMRR of op-amp A_1 . Further if we assume that Z_1, R_2, R_3 and Z_E have tolerance t , then from (7) and (8) we can deduce that the worst case condition will be when $Z_1 = Z_{10}(1-t), R_2 = R_{20}(1+t), R_3 = R_{30}(1-t)$ and $Z_E = Z_{E0}(1+t)$ where Z_{10}, R_{20}, R_{30} , and Z_{E0} are the respective nominal values. Equation (7) then can be written as

$$CMRR_R(s) = \frac{p(s)(1-t^2) + 1 + t^2}{4t} \approx \frac{p(s) + 1}{4t} = \frac{A_d(s)}{4t} \quad (11)$$

where $p(s) = \frac{Z_{10}(s)}{R_{20}} = \frac{Z_{E0}(s)}{R_{30}}$. This means that large differential gain is desirable since very small tolerance com-

ponents are expensive. Therefore, considering (11), equation (10) can be written as

$$\frac{1}{CMRR(s)} = \frac{1}{2CMRR_{A1}(s)} + \frac{1}{A_{d0}(s)/2} + \frac{1}{A_d(s)/4t} \quad (12)$$

The $CMRR_{A1}$ has the form

$$CMRR_{A1}(s) = \frac{CMRR_{A1}}{1 + s/\omega_r} \quad (13)$$

where ω_r is the frequency where $CMRR_{A1}$ has decreased by 3 dB and is usually between 100 Hz and 1 kHz. The open loop gain $A_{d0}(s)$, also decreases at higher frequencies with a corner frequency $\omega_0 = 1/\tau_0$ which is usually lower than ω_r if A_0 and A_1 are of the same type. Therefore the CMRR is mainly determent at low frequency by the matching of the resistors and the DM gain, and at high frequencies by




Figure 4
Simplified schematic of the biopotential amplifier with active electrodes.

the open loop response of op-amp A_0 , rather than its CMRR. If we take the advantage of the fact implicit in (10), and achieve

$$\frac{1}{2CMRR_{A1}(s)} + \frac{1}{A_{d0}(s)/2} = -\frac{1}{CMRR_R(s)} \quad (14)$$

then theoretically the CMRR becomes infinite. In part this can be achieved by the use of a capacitor and resistor in parallel for the impedance Z_1 (Fig. 3), and then trimming R_2 . Thus the need of low-tolerance components is eliminated. Therefore $Z_1(s)$ will have the form

$$Z_1(s) = \frac{R_1}{1 + sR_1C_1} \quad (15)$$

It has been shown [20] that a good approximation for the optimal value of the capacitor C_1 is

$$C_1(s) = \frac{A_d}{R_1 2\pi GBP_{A0}} \quad (16)$$

where GBP_{A0} is the gain bandwidth product of op-am A_0 . Trimming R_2 is a good solution for achieving an ultra high CMRR for demanding application, however it is not practical since the trimmer has to be incorporated in the electrode. Alternatively, Z_E or R_3 can be trimmed, which however will alter also the amplifier DM gain. Considering equation (12) it can be shown that for application with relatively high DM gain and proper op-amp selection, both trimming and compensation (C_1) can be omitted, without significantly degrading the CMRR. For example, if the usual 1% tolerance resistors are used and op-amps with CMRR of 100 dB and DM open loop gain of 120 dB at 50 Hz, then for an amplifier with DM gain of 5000, a CMRR of 96 dB can be achieved without trimming.

In the amplifier circuit shown in Fig. 3, Z_E is replaced with an active DC rejection/suppression circuit [20]. It includes an integrator (A_2, R_i, C_i) and two potential dividers (R_6, R_5 and R_4, R_3). The amplifier can operate in AC-mode or in semi-AC-mode. The two modes are selectable by the




Figure 5
Schematic of the basic two-op-amp instrumentation amplifier.

switch S_1 : AC-mode with S_1 open and semi-AC-mode with S_1 closed. In AC-mode the DC signals are rejected, where in semi-AC-mode they are suppressed. If $R_6 = R_6^*$, $R_5 = R_5^*$ and $R_i = R_i^*$, then the respective expressions for the equivalent impedance $Z_E(s)$ for the two modes are given by

$$Z_{ESAC}(s) = \frac{(R_3 + R_4)(R_5 + R_6) \left[2k \left(1 + s \frac{\tau_i}{2} \right) + (1 + s k \tau_i) A_{d2}(s) \right] - R_3}{2k R_5 \left(1 + s \frac{\tau_i}{2} \right) A_{d2}(s)} \quad (17)$$

$$Z_{EAC}(s) = \frac{(R_3 + R_4)(R_5 + R_6) [1 + s \tau_i A_{d2}(s)] - R_3}{R_5 (1 + s \tau_i) A_{d2}(s)} \quad (18)$$

where $\tau_i = R_i C_i$ is the time constant of the integrator,

$A_{d2}(s) = \frac{A_{d2}}{1 + s \tau_2}$ and τ_2 are the respectively the DM open

loop gain and the time constant of the first pole of op-

amp A_2 . Whenever $R_i \gg R_5$ then $k \approx (R_6/R_5 + 1)$, which is true with the time constants and voltage gains, typical in biopotential recordings. For signals bellow the amplifier high-pass cut-off frequency, $Z_E(s)$ decreases due to the active DC rejection/suppression circuit. For DC signals equation (8) is maximally imbalanced and thus $CMRR_R(0) \approx A_d(0)$. Since for biopotential amplifiers $A_d(0)$ is much lower than $CMRR_{A1}(0)$ and $A_{d0}(0)$, therefore $CMRR(0) \approx A_d(0)$, which represents the worst case.

If we consider only the -3 dB bandwidth and assume that op-amp A_2 is ideal, then (17) and (18) simplify to

$$Z_E = \frac{R_4 R_5 + R_4 R_6 + R_3 R_6}{R_5} \quad (19)$$

Therefore, in this case (9) can be written as

$$A_d = \frac{Z_E}{R_3} + 1 = \left(\frac{R_4}{R_3} + 1 \right) \left(\frac{R_6}{R_5} + 1 \right) \quad (20)$$

which represents the mid band DM gain for both modes. After substituting (17) and (18) in (9), it can be shown that the respective DC differential gains for the two modes are given by

$$A_{dAC}(0) = \frac{A_d}{A_{d2}}; \quad A_{dsAC}(0) = \frac{A_d}{2k} \quad (21)$$

where $2k$ is approximately the DC gain of the stopped integrator ($A_2, R_i, C_i, R_i^*, R_5^*, R_6^*$) in semi-AC-mode. Thus DC signals meet lower gain, in order to prevent saturation from large electrode offsets or other high DC potentials.

The active electrodes' input resistances R_{iE0} and R_{iE1} , are not equal due to the different closed loop gains of op-amps A_0 and A_1 , and can be expressed as

$$R_{iE0} = R_{iA0} \left(1 + \frac{A_{d0}Z_1(0)}{R_1 + R_2} \right); \quad R_{iE1} = R_{iA1} \left(1 + \frac{A_{d1}R_3}{Z_E(0) + R_3} \right) \quad (22)$$

where R_{iA0} and R_{iA1} are the input resistances of op-amps A_0 and A_1 . However, at higher frequencies, the electrodes' input impedances are much lower and about the same (assumed that A_0 and A_1 are of the same type), due to the op-amps' input and additional stray capacitance, being in parallel to the high op-amps' input resistance.

The output noise spectral density for the -3 dB bandwidth is approximately the same for both modes and can be written as

$$v_{no} = A_d \sqrt{\left(1 + \frac{R_2}{Z_1} \right)^2 e_{n0}^2 + e_{n1}^2 + \left(\frac{R_3}{R_3 + R_4} \right)^2 e_{n2}^2} \quad (23)$$

where e_{n0} , e_{n1} and e_{n2} are the respective voltage noises of op-amps A_0 , A_1 and A_2 .

Assuming E_0 is connected to common (Fig. 3), then the amplifier transfer function $H(s)$ is given by

$$H(s) = \frac{U_1(s)}{E_1(s)} = \frac{A_{d1}(s)(R_3 + Z_E(s))}{R_3 + Z_E(s) + R_3A_{d1}(s)} \quad (24)$$

After substituting $Z_E(s)$ and $A_{d1}(s)$ in (24), it can be shown that $H(s)$ has three poles and two zeros for both modes. However, with the time constants and voltage gain used in the current application, one pole almost coincides with one zero. Therefore, $H(s)$ can be approximated very well by a transfer function with two poles and one zero. The

respective approximations for AC-mode and semi-AC-mode are given by

$$H_{AC}(s) = \frac{A_d(1 + s\tau_i A_{d2})}{A_{d2}(1 + s\tau_i) \left(1 + s\tau_1 \frac{A_d}{A_{d1}} \right)} \quad (25)$$

$$H_{sAC}(s) = \frac{A_d(1 + s\tau_i k)}{2k \left(1 + s \frac{\tau_i}{2} \right) \left[1 + s\tau_1 \frac{A_d}{A_{d1}} \right]} \quad (26)$$

The circuits described by the transfer functions $H_{AC}(s)$ and $H_{sAC}(s)$ are stable because all the poles are situated in the left half of the complex s-plane and there are no resonance effects as the poles are on the real s-axis.

Practical amplifier circuit

The schematic of a multichannel amplifier with active electrodes, built according to the design discussed is shown in Fig. 6. Each channel amplifies the signal between its input ($E_1 \dots E_N$) and the reference input E_0 (monopolar configuration). The output voltage of op-amp A_2 is equal to the DC input voltage, multiplied by the ratio $(R_3 + R_4)/R_3$ for both modes. The choice of the resistor ratio $(R_3 + R_4)/R_3$ is a trade-off between DC input range and noise, since a low ratio enhances the noise contribution of op-amp A_2 . The ratio R_4/R_3 was chosen so that to allow a DC input voltage range of ± 370 mV, without saturating op-amp A_2 . The offset voltage at the amplifier output (U_1) is the input offset voltage of op-amp A_2 , times the resistor ratio $(R_5 + R_6)/R_5$. In case of high DM gain, the output offset voltage would become unacceptably high. Thus, op-amp A_2 was selected for its ultra-low offset voltage of 1 μ V, low noise and high CM input range to prevent latch-up. Moreover A_2 is a low input bias current type, which allows the use of high value input resistances without producing large offset voltages between its inputs. Op-amp A_0 , was selected with a relatively high GBP, in order to confine its influence on the CMRR at higher frequencies. Combining high DM gain, with high GBP op-amps, allowed Z_1 to be implemented only with its active part R_1 , as thus the CMRR at 50 Hz was not significantly degraded. As shown in (23), the equivalent input noise is mainly determent by the noise of op-amps A_0 and A_1 . They were implemented with the low-noise CMOS op-amp LMV751 in a SOT23-5 package.

Because of the large integrator's time constant, the amplifier has a very slow response after overloads ($\leq 10\tau_i$), caused by large signal disturbances. Thus a deblocking circuit was added at the cabinet's location, for temporary reduction of the time constant during overload [24]. It is controlled by the output voltage U_1 through the low pass




Figure 6
Multichannel biopotential amplifier with active electrodes and DRL circuit.

filter (R_{13}, C_2). The filter output controls two threshold triggers (A_3, A_4), which through D_1, D_2 control the MOS transistor T_1 , acting like a switch. When the output signal reaches its range limits (defined by R_{14}, R_{15}, R_{16}), T_1 opens and the new reduced time constant $\tau_1^* = (R_7/R_{11})C_2$, pulls the output signal to the zero level. This state is maintained for additional hundred milliseconds ($R_{13}C_2$) and then is switched back to its original value.

The connection between the amplifier common and the signal source is implemented by a driven right-leg (DRL) circuit. The CM voltage at the output of A_0 is reduced by a factor equal to the DRL circuit gain ($A_{DRL} = 314$ at 50 Hz), which theoretically should give a 50 dB extra CMRR at 50 Hz. In addition, in case of a faulty op-amp, the DRL circuit will limit the maximum patient current to a safe level of 50 μ A.

Results

The contact impedance of the proposed electrode, measured and averaged over five subjects, and its calculated model impedance are shown in Fig. 7. The values of the model elements were determined to give the closest agree-

ment between the measured electrode-skin impedance and that of the model. The values are: $R_s = 300 \Omega$, $R_d = 450 \text{ k}\Omega$ and $C_d = 3 \text{ nF}$. The measuring technique is described by Bergey et al. [21] and was performed five minutes after the application of the electrodes to allow their impedance to settle to a constant value [13]. The electrodes were applied with moderate tension of 0.3 kg/cm² on an unprepared skin of the inner forearm. The impedance was lower when higher tension was applied or when sweat was present on the skin. The applied current density was 0.01 mA/cm² and no current density impedance dependence was observed [22]. The electrode-skin impedance showed two-decade spread between different subjects, which was also reported in other works [23].

Simulations of the amplifier circuit were carried out using PSPICE. The op-amps used in the model were with gain-bandwidth product (GBP) of 5 MHz, DM open loop gain of 120 dB and CMRR of 100 dB. The integrator time constant and the resistor ratios of the feedback loop were: $\tau_i = R_i C_i = 1$, $R_4/R_3 = 6$ and $R_6/R_5 = 700$. The amplifier frequency response plots, for both operating modes, are shown in Fig. 8. In semi-AC-mode the high-pass response




Figure 7

The average electrode-skin contact and the calculated model impedance against frequency. The plot shows also the minimum and the maximum data sets for five subjects.

is with 1st order pole at 0.32 Hz and zero at 0.2 mHz. In AC-mode the high-pass response is with 1st order pole at 0.16 Hz and zero at 0.16 μ Hz. The mid band DM gain is the same for both modes. Simulations were also carried out for the estimation of the power line interference due to induced displacement currents into the electrode leads. The increased stray capacitance between the power line and the amplifier, caused by the increased number of electrode leads, was taken into consideration in the simulation model. The results showed that the interference caused by the active electrode unshielded leads was insignificant.

Table 1 shows the amplifier specifications, measured with battery powered prototype and test equipment. All the parameters were in close agreement with those of the simulations. The peak-to-peak noise voltage measured at the amplifier output, with input terminals connected to common, was 10 mV_{p-p} or 2 μ V_{p-p} when referred to the input. The CMRR of the amplifier was 96 dB at 50 Hz, measured with imbalanced electrode impedances ($\Delta Z_e = 47$ k Ω). The maximum measured CMRR with DRL and a CM input signal of 4 V_{p-p} was 126 dB at 50 Hz, where the output signal level was approximately equal to the amplifier output voltage noise.




Figure 8
A plot of the amplifier frequency response.

Table 1: Active electrode specifications

Parameter	semi-AC-mode	AC-mode
Bandwidth (-3 dB)	0.32–1000 Hz	0.16–1000 Hz
DC gain	3.22	≈ 0
AC mid band gain	74 dB	
Differential mode AC input range	0.005–1 mV _{p-p}	
Differential mode DC input range	± 370 mV	
Common mode input range	± 2 V	
Input noise current	1 pA _{rms} @ 0.1–200 Hz	
Input bias current	1.5 pA	
Input impedance, Active Electrode	320 MΩ @ 50 Hz (1000 GΩ //10 pF)	
CMRR	96 dB @ 50 Hz	
Output offset	0.7 mV	
Input noise voltage	2 μV _{p-p} (0.33 μV _{rms}) @ 0.1–200 Hz	
Power consumption	11 mW @ one channel	

A sample record from the practical application of the active electrode, obtained after a low intensity SLB laser stimulation, is shown in Fig. 9. The amplifier was battery powered and optically isolated by linear optocouplers. The bandwidth was limited to 200 Hz, by 6th order low pass Bessel filter, and the signal was sampled with 1 kHz. The electrodes were connected with a high-density unshielded ribbon cable. One electrode (E₁) was placed on SLB TH-23, near the eyebrow, where the reference (E₀) was placed on the ear lobe. No electrolyte gel or skin preparation was applied. The measurements were performed in a typical laboratory room. All measurements showed almost complete absence of 50 Hz interference. The noise present in the signal is mainly compounded of artifacts from eye movements, electromyographic signals, and noise from the electrode-skin interface.

Discussion

The best solution for an active electrode would be to perform the entire analog signal processing at the electrode site. This could be achieved with a custom made integrated circuit, but the cost would be much higher. We found a good alternative in using SMD technology and integrating only the front-end of the amplifier into the electrode.

The ultra high input resistance of the electrode is degraded at higher frequencies by the op-amp's input capacitance in parallel with the stray capacitance due to the electrode Printed Circuit Board (PCB). Nevertheless, combining an op-amp with low input capacitance and a proper PCB design, allowed a relatively high input impedance to be achieved at 50 Hz. That decreased the amplifier sensitivity to high electrode-skin impedance imbalances, by reducing the transformation of the CM interference signal into unwanted DM signal. Unfortunately, most data sheets do not properly specify op-amp's input capacitance, neither DM nor CM.

The active electrode presented is not suitable for applications requiring a low differential gain and large signal bandwidth due to the decreasing CMRR at higher frequencies, if not properly compensated. On the other hand, below the high-pass cut-off frequency, the CMRR is degraded by the active feedback circuit, and reaches its minimum value for DC signals, equal to the DM gain. The circuit can accept high value input filter resistances, which will also limit the patient auxiliary current in case of fault condition of op-amps A₀ and A₁. Because of the limited electrode space, it is preferable that the front-end op-amps feature internal electrostatic discharge protection circuitry, rather than building an external one.




Figure 9
Biopotential acquired in semi-AC-mode from SLB TH-23 after low intensity laser stimulation.

Conclusions

The new electrode anchoring system significantly reduced the electrode-skin impedance, its variation and motion artifact influences. The proposed amplifier fractionation resulted in lower noise and less parts. Moreover splitting the amplifier between the electrodes and the cabinet's location allowed the use of an automatic DC deblocking system and mode switching. The prototype tests showed that with the active electrode presented, SLB signals with relatively high quality could be recorded without skin preparation. The 50 Hz interference pickup by the electrode leads was practically eliminated. Because high electrode-skin impedances are tolerated, no electrolytic gel is needed. This allows fast application of the electrodes,

minimizes patient discomfort and eliminates the risk of infection.

With proper op-amps selection, the active electrode specifications were found to be better or at least comparable to those of other existing designs. The design offers low noise and major reduction in parts, size and power consumption. It is currently used in studying laser provoked SLB potentials and their propagation, aiming to gain a better insight into the bio-stimulation effect of lasers in Medical Acupuncture.

Authors' contributions

The authors contributed equally to this work

References

- Mashansky VF, Markov UV: **Topography of the gap junctions in the human skin and their possible role in the non-neural signal transduction.** *Arch Anat Histol Embryol* 1983, **84**:53-60.
- Reichmanis M, Marino AA, Becker RO: **Electrical correlates of acupuncture points.** *IEEE Trans Biomed Eng* 1975, **22**:533-535.
- Reichmanis M, Marino AA, Becker RO: **Laplace plane analysis of transient impedance between acupuncture points Li-4 and Li-12.** *IEEE Trans Biomed Eng* 1977, **24**:402-405.
- Johng HM, Cho JH, Shin HS, Soh KS, Koo TH, Choi SY, Koo HS, Park MS: **Frequency dependence of impedances at the acupuncture point Quze (PC3).** *IEEE Eng Med Biol Mag* 2002, **21**:33-36.
- Litscher G: **Effects of acupressure, manual acupuncture and Laserneedle acupuncture on EEG bispectral index and spectral edge frequency in healthy volunteers.** *Eur J Anaesthesiol* 2004, **21**:13-19.
- Litscher G: **Cerebral and peripheral effects of laser needle-stimulation.** *Neurol Res* 2003, **25**:722-728.
- Niboyet J: *Traite d'acupuncture* Maisonneuve: Sainte-Ruffine; 1979.
- Stux G, Pomeranz B: *Basics of Acupuncture* New York: Springer-Verlag; 1998.
- Khan A, Greatbatch W: **Physiologic electrodes.** In *Medical Engineering* Edited by: Ray CD. Chicago: Year Book Medical Publishers; 1974:1073-1082.
- Travis C: **Nonlinear aspects of the bioelectrode-electrolyte interface.** In *Biomedical electrode technology* Edited by: Miller HA. New York: Academic Press; 1974:143-159.
- Zipp P, Ahrens H: **A model of bioelectrode motion artifact and reduction of artifact by amplifier input stage design.** *J Biomed Eng* 1979, **1**:273-276.
- Neuman MR: **Biopotential electrodes.** In *Medical Instrumentation, Application and Design* 2nd edition. Edited by: Webster JG. New York: John Wiley and Sons; 1995:227-287.
- Searle A, Kilrup L: **A direct comparison of wet, dry and insulating bioelectric recording electrodes.** *Physiol Meas* 2000, **21**:271-283.
- Godin DT, Parker PA, Scott RN: **Noise characteristics of stainless-steel surface electrodes.** *Med Biol Eng Comput* 1991, **29**:585-590.
- Richardson PC: **Some new electrode techniques for long-term physiological monitoring.** *Aerosp Med* 1968, **39**:745-750.
- Ko WH, Hyneczek J: **Dry electrodes and electrode amplifiers.** In *Biomedical electrode technology* Edited by: Miller HA, Harrison DC. New York: Academic Press; 1974:169-181.
- MettingVanRijn AC, Kuiper AP, Dankers TE, Grimbergen CA: **Low cost active electrode improves the resolution in biopotential recordings.** In *Proceedings of the 18th Ann Int Conf IEEE Eng Med Biol, Amsterdam* 1:101-102. 31 October-3 November 1996
- Hagemann B, Luhede G, Luczak H: **Improved active electrode for recording bioelectric signals in work physiology.** *Eur J Appl Physiol Occup Physiol* 1985, **54**:95-98.
- Padmadinata FZ, Veerhoek JJ, Van Dijk GJA, Huijsing JH: **Microelectronic skin electrode.** *Sensors and Actuators* 1990, **B1**:491-494.
- MettingVanRijn AC, Peper A, Grimbergen CA: **Amplifiers for bioelectric events: a design with a minimal number of parts.** *Med Biol Eng Comput* 1994, **32**:305-310.
- Bergey GE, Squires RD, Sipple WC: **Electrocardiogram recording with pasteless electrodes.** *IEEE Trans Biomed Eng* 1971, **18**:206-211.
- Geddes LA: *Electrodes and the measurement of bioelectric events* New York: Wiley-Interscience; 1972.
- Rosell J, Colominas J, Riu P, Pallas-Areny R, Webster JG: **Skin impedance from 1 Hz to 1 MHz.** *IEEE Trans Biomed Eng* 1988, **35**:649-651.
- MettingVanRijn AC, Kuiper AP, Honsbeek RH, Speijer K, Peper A: **DC rejection and deblocking in multichannel bioelectric recordings.** In *Proceedings of the 17th Ann Int Conf IEEE Eng Med Biol, Montreal* 2:1665-1666. 20-23 September 1995

Publish with **BioMed Central** and every scientist can read your work free of charge

"BioMed Central will be the most significant development for disseminating the results of biomedical research in our lifetime."

Sir Paul Nurse, Cancer Research UK

Your research papers will be:

- available free of charge to the entire biomedical community
- peer reviewed and published immediately upon acceptance
- cited in PubMed and archived on PubMed Central
- yours — you keep the copyright

Submit your manuscript here:
http://www.biomedcentral.com/info/publishing_adv.asp

