


Contents lists available at [ScienceDirect](http://www.elsevier.com/locate/aott)

Acta Orthopaedica et Traumatologica Turcica

journal homepage: <https://www.elsevier.com/locate/aott>

Dual plating for fixation of humeral shaft fractures: A mechanical comparison of various combinations of plate lengths

Ahmet Karakasli ^{a,*}, Onur Basci ^a, Fatih Ertem ^b, Eyad Sekik ^c, Hasan Havitcioglu ^a^a Dokuz Eylul University, Faculty of Medicine, Dept. Orthopaedics and Traumatology, Izmir, Turkey^b Dokuz Eylul University, Institute of Health Sciences, Dept. Biomechanics, Izmir, Turkey^c Karatas Hospital, Izmir, Turkey

ARTICLE INFO

Article history:

Received 10 June 2015

Received in revised form

7 August 2015

Accepted 29 September 2015

Available online 29 July 2016

Keywords:

Humeral shaft fractures

Dual plate

Fixation

ABSTRACT

Objective: The role of plate configuration was found inconclusive on the biomechanical effects of the plate size and hole number for dual plate constructions in humeral shaft fractures. The purpose of this study was to test the biomechanical stability of various dual plate constructions.

Methods: Twenty-four left humeri (4th Generation Composite Humerus, Sawbones, Malmö, Sweden) with comminuted midshaft humeral fracture were used. Four groups of plate constructs were tested: laterally fixed 8-hole locking plate and screws were combined with anteriorly locking plates containing 0, 4, 6, or 8 holes in groups I, II, III, and IV, respectively. The alterations in axial, bending, and torsional angles were recorded.

Results: There were no fixation failures during axial, bending, or torsional stiffness testing within the elastic behavior limits. Axial stiffness was highest in Group IV. Torsional stiffness, posterior-to-anterior bending stiffness, lateral-to-medial bending stiffness, and medial-to-lateral bending stiffness were lowest in Group I.

Conclusion: The similar stiffness values for the 8-to-4 hole and 8-to-6 hole plate constructions indicate that the 8-to-4 hole construction is an option in young adults, while the stiffest 8-to-8 hole combination may be an option for osteoporotic patients.

© 2016 Turkish Association of Orthopaedics and Traumatology. Publishing services by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

Diaphyseal humeral fractures are seen relatively more frequently in the elderly population.¹ Even though nonoperative treatment is preferable, osteopenia as a result of lack of use leads to the need for options for internal fixation to avoid high levels of disability associated with humeral shaft nonunion.²

In humeral shaft fractures managed by surgery, the conventional manner for internal fixation is the use of large fragment plates. However, the variable size and shape of the humerus creates difficulties during the procedure in determining the appropriate combination of plate size and screw number.³ The recent literature

indicates that the use of dual plate yields better results in terms of mechanical properties than does the use of large fragment plate.⁴ In the use of dual plate, the layout angles of the plates relative to the humeral shaft is controversial. Placement of the anterior and lateral plates at 90° was found to be best configuration for dual plating.⁵

Despite the increased usage of locking plates in osteoporotic humeral shaft fractures, the few studies on the plate configuration were inconclusive regarding the biomechanical effects of plate size and hole number in dual plate constructions. The purpose of this study was to test the biomechanical stability of various dual plate constructions.

Patients and methods

Twenty-four left humeri (4th Generation Composite Humerus, Sawbones, Malmö, Sweden) were used in the present study. The specimens were embedded in cement at both ends, which were cut into a cylindrical shape to facilitate insertion into the testing grips. The center of the bone was determined by vernier calipers, and a

* Corresponding author.

E-mail addresses: karakasliahmet@gmail.com (A. Karakasli), dronurbasci@gmail.com (O. Basci), fatih_ertem@hotmail.com (F. Ertem), hasan.havitcioglu@deu.edu.tr (H. Havitcioglu).

Peer review under responsibility of Turkish Association of Orthopaedics and Traumatology.

comminuted midshaft humeral fracture was modeled with a 1-cm midshaft gap created with a surgical reciprocating saw.^{6–8} All specimens were prepared by the same 2 orthopedic surgeons. Standard technique for plate fixation was performed, placing all of the screws bicortically. The osteotomy provided a noncontact situation, allowing for isolated testing of the plate constructs.

Four groups of plate constructs were tested. Group I specimens were fixed laterally by an 8-hole 3.5-mm locking plate (8hLP) (all locking plates used in this study were produced by Med Tip Medical Device Company A.S., Izmir, Turkey), Group II specimens were fixed laterally with an 8hLP and anteriorly with a 4hLP, Group III were fixed laterally with an 8hLP and anteriorly with a 6hLP, and Group IV were fixed both laterally and anteriorly with an 8hLP (Fig. 1).

All tests were performed with a mechanical test machine (AG-1S 10 kN, Shimadzu, Kyoto, Japan). The humeral bone models were fixed to the load cell of the test machine. The axial load was applied to the cylindrical embedded end (Fig. 2). In all tests, the alterations in axial, bending, and torsional angles were recorded both in loaded and unloaded states.

The bone-plate constructs were tested under axial loading with the embedded humeral head. While simultaneously recording the vertical displacement and strain, 500 N for 5000 cycles at 3 Hz were applied. Displacement was recorded.⁴

A 4-point bending model was used for the anterior-posterior, posterior-anterior (sagittal plane), medial-lateral, and lateral-medial (coronal plane) testing. In each bending test, a maximal load of 250 N was applied at 10 mm/min. Bending moment was applied to the same point by centering the device on the midpoint of the fracture gap. Load versus displacement values were recorded to calculate the bending stiffness and flexibility.

Torsion test was performed with a servo sync torque machine (SQM132, 245 Nm 100 rpm, ELSIM Elektrotechnik A.S, Istanbul, Turkey). The torsion tests were conducted in the displacement control mode with a maximum moment of 4.5 Nm in both directions; the premoment was 0 Nm, and the test velocity was 0.3°/second. The testing cycle was applied from 0 to 4.5 Nm. Torque versus the degree of angle deformation values were recorded.^{7–9}




Fig. 1. Schematic illustration of the plate configurations.



Fig. 2. Axial load was applied to the cylindrical potted end.

Each specimen was tested 3 times in bending and torsion tests to ensure reproducibility of the results. All tests were performed within the elastic behavior limits of the construct; the load–deflection data did not show any sign of plastic or permanent deformation for any of the constructs in any orientation. The testing was performed in the same order for each sample. Statistical analysis was conducted with Mann–Whitney U test by using SPSS software (version 15.0, SPSS Inc., Chicago, IL, USA). The level for significance was defined as $p < 0.05$.

Results

There were no fixation failures during axial, bending, or torsional stiffness testing within the elastic behavior limits. Mean stiffness values of all groups are presented in Table 1. Axial stiffness in Group IV was 706.2 N/mm, which was significantly higher than in Group I (475.6 N/mm, $p = 0.004$), Group II (516.6 N/mm, $p = 0.025$), and Group III (543.5 N/mm, $p = 0.006$) (Fig. 3).

For torsional stiffness measurements, stiffness in Group I was 6.31 N/degree, which was significantly lower than in Group II (12.16 N/degree, $p = 0.01$), Group III (11.51 N/degree, $p = 0.01$), and Group IV (15.10 N/degree, $p = 0.006$) (Fig. 4).

Bending stiffness was also compatible with the previously described results. By all of the measured parameters of stiffness (posterior-anterior bending, lateral-medial bending, medial-lateral bending) results for Group I were significantly lower than for all of the other groups (Fig. 5).

Discussion

Plate fixation is the gold standard for treatment of humeral nonunion. It enables compression, correction of malalignment, and stimulation of osteogenesis (grafting) in a single procedure.¹⁰ Humeral nonunion can be severely disabling. Although several authors have recommended plate fixation for the management of nonunion at midshaft level, inappropriate plate fixation techniques are one of the main reasons that fractures fail to heal.^{11–16} Foster et al reported that the most common indication for surgical management of a humeral shaft fracture is a concurrent multiple injury, and the second is nonunion of humeral shaft fracture. They reported a 96% success rate for union in their study, using both single- and dual-plate constructs either with or without lag screws.¹¹ Murray et al pioneered the use of double-plate constructs for

Table 1
Mean stiffness values of all testing groups.

	Axial stiffness (N/mm)	Bending stiffness (N/mm)				Rotational stiffness (N/°)
		A-P	P-A	L-M	M-L	
Group I	475,6 ± 63	340,4 ± 37	285,8 ± 24	134,5 ± 4	196,55 ± 28	6,3 ± 2
Group II	516,6 ± 124	372,7 ± 41	424,8 ± 39	413,1 ± 23	382,3 ± 17	12,1 ± 3,9
Group III	543,5 ± 102	330,4 ± 39	430,8 ± 125	373,3 ± 46	411,9 ± 100	11,5 ± 4,3
Group IV	706,2 ± 78	414,6 ± 96	509,1 ± 135	410,3 ± 69	496,9 ± 174	15,1 ± 5,7




Fig. 3. Axial stiffness according to groups.




Fig. 5. Bending stiffness according to groups.

nonunion in 1964. Using an anterolateral approach, they placed 1 plate on the lateral side of the humerus and the other on the posterior aspect. They suggested that this method provides a more stable construct than that achieved with single plates because of the biplanar control provided by dual plates at right angles to each other.¹⁴

In our study, we investigated the in vitro mechanical stability of different dual plate constructs. It is apparent in the literature that use of the dual plate construct in the treatment of humeral fractures has a significantly higher mechanical stability compared to the single plate application.⁶ Similar to the published results, our study confirmed that the dual plate construct had a significantly higher stability than single plate fixation.

In the treatment of nonunions, which arise from simple transverse fractures of long bones like the humerus and ulna, the increased stability with dual plates may lead to union.¹⁷ Egol et al suggested a theoretical concern for the amount of soft-tissue dissection and revascularization needed to place a second plate at a 90° angle to the first.^{18,19} Although this is a concern in the treatment of acute fractures, in the case of nonunions, the dissection needed to correct the deformity and debride the fibrous tissue

allows application of the second plate without additional exposure. Furthermore, a study by Rubel et al showed that the addition of a second plate did not increase the rate of treatment failure.²⁰ The results of our biomechanical tests in the gap model suggest that higher axial and rotational stiffness is achieved with a dual plate construct. It has been shown that a dual plate construct is sometimes helpful when bone quality is poor and pseudoarthrosis is present or when the stability achieved by the first plate is not clinically sufficient.¹¹ The use of dual plate provided healing in nonunion patients in such situations. In our study, performed on the humerus transverse fracture model, 8-hole single plate fixation (Group I) showed statistically significant lower stiffness than the dual plate (Group II, Group III, Group IV) constructions. There were no statistical differences between the 8-hole lateral plate fixation combined with a second 90 anterior plate with 4 holes (Group II) or with 6 holes (Group III). The plate combinations with 8-hole plate lateral and 8-hole plate anterior construct (Group IV) had significantly higher axial stiffness. However, Group IV showed excessive stiffness, in which the fixation construct reduces stress and strain on the bone and may lead to bone resorption.²⁰

Construct stiffness is highly important. It governs the performance of the fixation system, using the other interrelated outcomes measures reported (bone stress shielding, hardware stress, and interfragmentary strain),^{4,7} which eventually causes screw loosening and construct failure.^{21–23} However, excessive reductions in stiffness may lead to increased screw and plate stress and early fatigue failure of the construct.^{22–24} Stiffness resulting from the compressive loading simulations of the humerus may be especially important during crutch weight-bearing.^{8,22} Moreover, torsional loading also is of interest in the analysis of humeral fracture fixation constructs because it has been reported as a predominant loading mode and possible cause for nonunion of humeral fractures.^{22,24,25} Hertel et al recommended obtaining a minimum of 3 cortices per segment.²⁶ In good quality bone, Gautier and Sommer recommended using a minimum of 2 screws per segment, with at least 3 cortices for simple fractures and 4 cortices for comminuted




Fig. 4. Torsional stiffness according to groups.

fractures. In other cases such as osteoporotic bone, they recommended a minimum of 3 screws per segment.²² Stoffel et al found that axial stiffness and torsional rigidity were mainly influenced by the distance between the fracture site and the closest screw. After moving the screw 1 hole farther from the fracture, the construct became almost twice as flexible in compression and torsion.²³ The development of locking plates has provided an alternative to standard compression plates. Locking plates can provide fracture fixation without the undesirable effects on periosteal vascularization and mechanical drawbacks that are encountered with standard compression plates.²⁴

Biomechanics of the fixation device is only one of many factors to be considered in the operative treatment of humeral fractures. Although numerous factors affect operative treatment decisions, it is helpful for the orthopedic surgeon to have comparative biomechanical information of different plating constructs. Because the amount of motion at the fracture site influences the biologic reaction of bone, the biomechanical stability of a fracture fixation implant plays an important role in fracture healing.^{21,27–29}

This biomechanical study might have a critical role in determining the appropriate construction in the treatment of humeral shaft nonunion. Our results suggest that, depending on the fracture type, in young adult patients, lateral 8-hole and anterior 4-hole plate constructions may be used, while in osteoporotic patients, lateral 8-hole and anterior 6-hole plate constructions may be utilized. Aziz et al reported that the strengths of the cortical screws are significantly lower according to their pullout tests on osteoporotic bone.³⁰ In osteoporotic bone, insufficient screw purchase leads to loosening of nonlocking screws and levering of the plate away from the bone. Loosening of nonlocking screws in osteoporotic bone is an important factor in implant failure in humeral shaft fracture fixation.³¹ The stability of the fixation system is influenced by hardware factors including the number of screws, type of screws (i.e., bicortical, unicortical), working length, plate offset from the bone cortex, and placement of the hardware.^{8,22,32} Ideal plate fixation in shaft fractures should have at least 6 (or preferably 8) cortex fixations above and below the fracture line.³³ In addition to plate placement, the hardware variables were controlled in this study based on recommendations from the literature.^{17,21} In our study, all of the lateral locked plates used 4 screws per fragment, based on the literature findings that additional screws did not show a significant increase in torsional stiffness.^{5,23}

In conclusion, in the treatment of nonunion or pseudoarthrosis of humeral fractures, dual plate fixation is a technique that increases the stability and healing rates. As shown in our results of similar stiffness values for the 8-to-4 hole and 8-to-6 hole plate construction, an 8-to-4 hole construction may be suggested as an option in young adults, while an 8-to-8 hole combination, the overall stiffest combination, may be an option for osteoporotic patients. Further studies are required to understand the biomechanical behavior of osteoporotic bones on dual plate fixation.

Conflict of interest

None declared.

References

- Baron JA, Karagas M, Barrett J, et al. Basic epidemiology of fractures of the upper end and lower limb among Americans over 65 years of age. *Epidemiology*. 1996;7:612–618.
- Ring D, Perey BH, Jupiter JB. The functional outcome of ununited fractures of the humeral diaphysis in older patients. *J Bone Jt Surg Am*. 1999;81:177–190.
- Patel R, Neu CP, Curtiss S, Fyhrie DP, Yoo B. Crutch weightbearing on comminuted humeral shaft fractures: a biomechanical comparison of large versus small fragment fixation for humeral shaft fractures. *J Orthop Trauma*. 2011;25:300–305.
- Watts A, Weinhold P, Kesler W, Dahners L. A biomechanical comparison of short segment long bone fracture fixation techniques: single large fragment plate versus 2 small fragment plates. *J Orthop Trauma*. 2012;26:528–532.
- Kosmopoulos V, Nana AD. Dual plating of humeral shaft fractures: orthogonal plates biomechanically outperform side-by-side plates. *Clin Orthop Relat Res*. 2014 Apr;472(4):1310–1317. <http://dx.doi.org/10.1007/s11999-013-3379-7>. Epub 2013 Nov 12. PubMed PMID: 24218163; PubMed Central PMCID: PMC3940765.
- Tejwani NC1, Murthy A, Park J, McLaurin TM, Egol KA, Kummer FJ. Fixation of extra-articular distal humerus fractures using one locking plate versus two reconstruction plates: a laboratory study. *J Trauma*. 2009 Mar;66(3):795–799. <http://dx.doi.org/10.1097/TA.0b013e318181e53c>.
- Fitzpatrick DC, Doornink J, Madey SM, Bottlang M. Relative stability of conventional and locked plating fixation in a model of the osteoporotic femoral diaphysis. *Clin Biomech*. 2009;24:203–209.
- O'Toole RV, Andersen RC, Vesnovsky O, et al. Are locking screws advantageous with plate fixation of humeral shaft fractures? a biomechanical analysis of synthetic and cadaveric bone. *J Orthop Trauma*. 2008 Nov–Dec;22(10):709–715. <http://dx.doi.org/10.1097/BOT.0b013e31818df8cb>.
- Blum J, Macherer H, Baumgart F, Schlegel U, Wahl D, Rommens PM. Biomechanical comparison of bending and torsional properties in retrograde intramedullary nailing of humeral shaft fractures. *J Orthop Trauma*. 1999 Jun–Jul;13(5):344–350. PubMed PMID: 10406701.
- Padhye KP, Kulkarni VS, Kulkarni GS, et al. Plating, nailing, external fixation, and fibular strut grafting for non-union of humeral shaft fractures. *J Orthop Surg (Hong Kong)*. 2013 Dec;21(3):327–331. PubMed PMID: 24366794.
- Foster RJ, Dixon Jr GL, Bach AW, Appleyard RW, Green TM. Internal fixation of fractures and non-unions of the humeral shaft. Indications and results in a multi-center study. *J Bone Jt Surg Am*. 1985 Jul;67(6):857–864. PubMed PMID: 4019533.
- Lin CL, Fang CK, Chiu FY, Chen CM, Chen TH. Revision with dynamic compression plate and cancellous bone graft for aseptic nonunion after surgical treatment of humeral shaft fracture. *J Trauma*. 2009 Dec;67(6):1393–1396. <http://dx.doi.org/10.1097/TA.0b013e31818c1595>. PubMed PMID: 20009693.
- Marti RK, Verheyen CC, Besselaar PP. Humeral shaft nonunion: evaluation of uniform repair in fifty-one patients. *J Orthop Trauma*. 2002 Feb;16(2):108–115. PubMed PMID: 11818806.
- Murray WR, Lucas DB, Inman VT. Treatment of non-union of fractures of the long bones by the two plate method. *J Bone Jt Surg Am*. 1964 Jul;46:1027–1048. PubMed PMID: 14192498.
- Spitzer AB, Davidovitch RI, Egol KA. Use of a "hybrid" locking plate for complex metaphyseal fractures and nonunions about the humerus. *Injury*. 2009 Mar;40(3):240–244. <http://dx.doi.org/10.1016/j.injury.2008.07.019>. Epub 2009 Feb 4. PubMed PMID: 19195653.
- Van Houwelingen AP, McKee MD. Treatment of osteopenic humeral shaft nonunion with compression plating, humeral cortical allograft struts, and bone grafting. *J Orthop Trauma*. 2005 Jan;19(1):36–42. PubMed PMID: 15668582.
- Egol KA, Kubiak EN, Fulkerson E, Kummer FJ, Koval KJ. Biomechanics of locked plates and screws. *J Orthop Trauma*. 2004 Sep;18(8):488–493. Review. PubMed PMID: 15475843.
- Zhiqian A, Bingfang Z, Yeming W, Chi Z, Peiyang H. Minimally invasive plating osteosynthesis (MIPO) of middle and distal third humeral shaft fractures. *J Orthop Trauma*. 2007 Oct;21(9):628–633. PubMed PMID: 17921838.
- Lee HJ, Oh CW, Oh JK, et al. Minimally invasive plate osteosynthesis for humeral shaft fracture: a reproducible technique with the assistance of an external fixator. *Arch Orthop Trauma Surg*. 2013 May;133(5):649–657. <http://dx.doi.org/10.1007/s00402-013-1708-7>. Epub 2013 Mar 5. PubMed PMID: 23463256.
- Woo SL, Akesson WH, Coutts RD, et al. A comparison of cortical bone atrophy secondary to fixation with plates with large differences in bending stiffness. *J Bone Jt Surg Am*. 1976 Mar;58(2):190–195.
- Rubel IF, Kloen P, Campbell D, et al. Open reduction and internal fixation of humeral nonunions: a biomechanical and clinical study. *J Bone Jt Surg Am*. 2002 Aug;84-A(8):1315–1322. PubMed PMID: 12177259.
- Gautier E, Sommer C. Guidelines for the clinical application of the LCP. *Injury*. 2003 Nov;34(Suppl 2):B63–B76. Review. PubMed PMID: 14580987.
- Stoffel K, Dieter U, Stachowiak G, Gächter A, Kuster MS. Biomechanical testing of the LCP—how can stability in locked internal fixators be controlled? *Injury*. 2003 Nov;34(Suppl 2):B11–B19. PubMed PMID: 14580982.
- Tan SL, Balogh ZJ. Indications and limitations of locked plating. *Injury*. 2009 Jul;40(7):683–691. <http://dx.doi.org/10.1016/j.injury.2009.01.003>. Epub 2009 May 22. Review. PubMed PMID: 19464682.
- Verbruggen JP, Sternstein W, Blum J, Rommens PM, Stapert JW. Compression-locked nailing of the humerus: a mechanical analysis. *Acta Orthop*. 2007 Feb;78(1):143–150. PubMed PMID: 17453406.
- Hertel R, Eijer H, Meisser A, Hauke C, Perren SM. Biomechanical and biological considerations relating to the clinical use of the Point Contact-Fixator—evaluation of the device handling test in the treatment of diaphyseal fractures of the radius and/or ulna. *Injury*. 2001 Sep;32(Suppl 2):B10–B14. PubMed PMID: 11718734.
- Mehling I, Schmidt-Horlohé K, Müller LP, Sternstein W, Korner J, Rommens PM. Locking reconstruction double plating of distal humeral fractures: how many screws in the distal ulnar column segment in A3 fracture provide superior stability? A comparative biomechanical in vitro study. *J Orthop Trauma*. 2009

- Sep;23(8):581–587. <http://dx.doi.org/10.1097/BOT.0b013e3181a87725>. PubMed PMID: 19704274.
28. Hak DJ, Althausen P, Hazelwood SJ. Locked plate fixation of osteoporotic humeral shaft fractures: are two locking screws per segment enough? *J Orthop Trauma*. 2010 Apr;24(4):207–211. <http://dx.doi.org/10.1097/BOT.0b013e3181bdd1da>. PubMed PMID:20335752.
 29. Prasarn ML, Ahn J, Paul O, et al. Dual plating for fractures of the distal third of the humeral shaft. *J Orthop Trauma*. 2011 Jan;25(1):57–63. <http://dx.doi.org/10.1097/BOT.0b013e3181df96a7>. Erratum in: *J Orthop Trauma*. 2013 Sep;27(9):541. PubMed PMID: 21085023.
 30. Aziz MS, Nicayenzi B, Crookshank MC, Bougherara H, Schemitsch EH, Zdero R. Biomechanical measurements of cortical screw purchase in five types of human and artificial humeri. *J Mech Behav Biomed Mater*. 2014 Feb;30:159–167. <http://dx.doi.org/10.1016/j.jmbbm.2013.11.007>. Epub 2013 Nov 19. PubMed PMID: 24295967.
 31. Frigg R. Locking compression plate (LCP). An osteosynthesis plate based on the dynamic compression plate and the point contact fixator (PC-Fix). *Injury*. 2001 Sep;32(Suppl 2):63–66. PubMed PMID: 11718740.
 32. Wagner M. General principles for the clinical use of the LCP. *Injury*. 2003 Nov;34(Suppl 2):B31–B42. Review. PubMed PMID: 14580984.
 33. Walker M, Palumbo B, Badman B, Brooks J, Van Gelderen J, Mighell M. Humeral shaft fractures: a review. *J Shoulder Elb Surg*. 2011 Jul;20(5):833–844. <http://dx.doi.org/10.1016/j.jse.2010.11.030>. Epub 2011 Mar 9. Review. PubMed PMID: 21393016.