



Optical and dielectric sensors based on antimicrobial peptides for microorganism diagnosis

Rafael R. Silva¹, Karen Y. P. S. Avelino², Kalline L. Ribeiro¹, Octavio L. Franco³, Maria D. L. Oliveira² and Cesar A. S. Andrade^{1,2*}

¹ Programa de Pós-Graduação em Inovação Terapêutica, Universidade Federal de Pernambuco, Recife, Brasil

² Departamento de Bioquímica, Universidade Federal de Pernambuco, Recife, Brasil

³ Centro de Análises Proteômicas e Bioquímicas, Universidade Católica de Brasília, Brasília-DF, Brasil

Edited by:

Luis Cláudio Nascimento Da Silva,
University of Copenhagen, Denmark

Reviewed by:

Margaret E. Bauer, Indiana University
School of Medicine, USA

Noton Kumar Dutta, Johns Hopkins
University, USA

Stefano Sforza, University of Parma,
Italy

Etinosa Igbino, University of Benin,
Nigeria

*Correspondence:

Cesar A. S. Andrade, Departamento
de Bioquímica, Universidade Federal
de Pernambuco, Avenida Professor
Moraes Rêgo S/N, 50670-901 Recife,
PE, Brasil
e-mail: cstrandrade@gmail.com

Antimicrobial peptides (AMPs) are natural compounds isolated from a wide variety of organisms that include microorganisms, insects, amphibians, plants, and humans. These biomolecules are considered as part of the innate immune system and are known as natural antibiotics, presenting a broad spectrum of activities against bacteria, fungi, and/or viruses. Technological innovations have enabled AMPs to be utilized for the development of novel biodetection devices. Advances in nanotechnology, such as the synthesis of nanocomposites, nanoparticles, and nanotubes have permitted the development of nanostructured platforms with biocompatibility and greater surface areas for the immobilization of biocomponents, arising as additional tools for obtaining more efficient biosensors. Diverse AMPs have been used as biological recognition elements for obtaining biosensors with more specificity and lower detection limits, whose analytical response can be evaluated through electrochemical impedance and fluorescence spectroscopies. AMP-based biosensors have shown potential for applications such as supplementary tools for conventional diagnosis methods of microorganisms. In this review, conventional methods for microorganism diagnosis as well as new strategies using AMPs for the development of impedimetric and fluorescent biosensors are highlighted. AMP-based biosensors show promise as methods for diagnosing infections and bacterial contaminations as well as applications in quality control for clinical analyses and microbiological laboratories.

Keywords: antimicrobial peptides, biosensors, bacterial infections, impedance spectroscopy, fluorescence spectroscopy

ANTIMICROBIAL PEPTIDES

Currently, the search for novel compounds with antibiotic ability to overcome bacterial resistance has increased. Antimicrobial peptides (AMPs) appear to be an excellent alternative, since more effective therapeutic approaches are required for many types of pathogens (Brogden et al., 2005; Hancock and Sahl, 2006; Schmidtchen et al., 2014). AMPs are components of the innate immune system, acting in defense against multiple pathogens (Schmidtchen et al., 2014). In general, AMPs may also show other different activities such as antiviral or antitumor properties, making them excellent candidates as new therapeutic drugs (Reddy et al., 2004).


Antimicrobial peptides are biomolecules present in diverse organisms, such as insects, plants, and animals (Mendez et al., 1990; Schnapp and Harris, 1998). In general, AMPs are mainly cationic small molecules composed of 10–50 amino acids residues in length, with molecular masses ranging from 1 to 5 kDa. In addition, AMPs comprise a chemically and structurally heterogeneous family (Andreu and Rivas, 1998; Costa et al., 2011).

In general, AMPs only assume an amphipathic structure after interacting with the bacterial membrane, since it is not a favorable structure in an aqueous environment, allowing lethal permeabilization of the bacterial membrane (van't Hof et al., 2001; Costa et al., 2011). The classical modes of

action described previously in the literature are not necessarily exclusive of one other (van't Hof et al., 2001; Brogden et al., 2005; Toke, 2005; Bechinger and Lohner, 2006). In the case of the barrel-stave model, a lipid bilayer disruption by AMPs occurs, until the peptides reach a threshold concentration and insert themselves across the bilayer to obtain peptide-lined pores. Subsequently, a membrane solubilization into micellar structures occurs, resulting in the carpet model or in forming peptide-and-lipid lined pores (toroidal pore model; Rapaport and Shai, 1991).

The mechanisms of action of AMPs (Figure 1) were reviewed and described by Nguyen et al. (2011). In this new proposal, pore formation in the disordered toroidal pore model is more stochastic, involving low peptide quantity. The presence of peptides can directly affect the bilayer thickness and, therefore, the membrane is remodeled to form rich domains of anionic lipids on the peptide surface. In addition, AMPs can form non-bilayer intermediates in the membrane coupled to small anions. On the other hand, in the molecular electroporation model, a peptide accumulation occurs on the outer lipid membrane leaflet, making the membrane permeable to various molecules including the AMPs (Nguyen et al., 2011).

Nanotechnology has been widely introduced for multidisciplinary applications, especially associated with the use of AMPs



in biology and biomedicine (Barkalina et al., 2014). Among the diverse uses of AMPs have been highlighted quality controls of foods, raising animals, controlled release systems and biosensors (Li et al., 2012; Sobczak et al., 2013). In this context, the peptide nisin has been applied as a milk-derived preservative for quality control in foods (Bi et al., 2011). In animal husbandry, peptides can be used in ruminants as an alternative to common antibiotics, avoiding bacterial resistance or assisting in the immunization of animals for the prevention of diseases (Cheema et al., 2011).

Infectious diseases have been part of humanity for centuries, some of them killing millions of people worldwide. Different types of pathogens (bacteria, viruses, fungi, and parasites) can cause infections from either exogenous (acquired from the environment, animals, or people) or endogenous (from the normal flora) origins (Mazzoni et al., 1987; Murthy et al., 2013). Furthermore, in agribusiness, the use of purified peptides, as well as the development of transgenic plants, aims to minimize plant diseases, avoiding the use of fungicides (Keymanesh et al., 2009). In clinical applications, some AMPs are in the early stages of clinical research to prevent inflammation and sepsis (Schuerholz et al., 2012).

Identification and quantification of pathogenic agents are two important parameters for the diagnosis of bacterial infections and

the implementation of effective drug therapy (Lazcka et al., 2007). The most common methods are direct examination by counting bacterial colonies in culture plates and serodiagnosis tests (van Pelt et al., 1999). Direct examination by optical microscopy can identify the morphology using a simple and specific Gram staining. Culture media allow the growth as well as the isolation and identification of specific types of bacteria (Szakal and Pal, 2003; Bragulat et al., 2004). However, these methods are time-consuming and require specialized technicians. Of note, it may require weeks to obtain a correct diagnosis of some pathogens from samples of mucosa, skin, or blood (Benes et al., 1996). Unfortunately, a few days may be sufficient to significantly worsen the clinical status of the patient before receiving appropriate treatment.

Traditional methods of diagnosis also consist of molecular tests or nucleic acid amplification test (NAAT). Techniques based on NAAT include polymerase chain reaction (PCR), ligase chain reaction, transcription mediated amplification, strand displacement amplification, and loop-mediated isothermal amplification (Andersen et al., 2000; Yang et al., 2002; Palka-Santini et al., 2009). Techniques based on nucleic acids enable specific molecular detection through hybridization between a DNA molecule and its complementary strand (Junhui et al., 1997; Wang et al., 1997; Wang, 1998). Although effective, these techniques may present

some limitations, such as the need for sample enrichment and purification prior to analysis, prolonged experimentation time, false-positive results due to cross-reactions, and high cost (Caygill et al., 2010; Singh et al., 2014).

In order to reduce or overcome these restrictions, new detection methods are required, and biosensors are considered promising tools for clinical diagnosis (Abdel-Hamid et al., 1999; Deisingh and Thompson, 2004). In this context, AMPs can be an excellent alternative for the development of biosensors, since their potential use for specific detection of pathogenic bacteria has been demonstrated (Kulagina et al., 2005).

STRATEGIES FOR THE DEVELOPMENT OF AMP-BASED BIOSENSORS

Advances in nanotechnology have contributed to the improvement of biotechnological research associated with the rapid progression of biodetection devices (Jianrong et al., 2004; Fournier-Wirth and Coste, 2010; Yeom et al., 2011). Nanostructures have been extensively used in the biosensor development due to their unique physicochemical characteristics such as quantum size effect, elevated ratio between surface area and volume, and analytical signal amplification capacity (Jianrong et al., 2004; Li et al., 2010). The utilization of AMPs as recognition elements in biosensors is a relatively new concept, but their development is of great relevance due to numerous potential areas for application (Etayash et al., 2014).

Biosensors are chemical devices comprising two basic functional units (Figure 2). The first one is a biomolecule responsible for recognition of the target substance through specific intermolecular binding or by means of catalytic reactions (D'Orazio, 2003). The second is the transducer that converts the biochemical response into a measurable electric signal, which is mainly proportional to analyte concentration (Higgins and Lowe, 1987; Hulanicki et al., 1991; Thévenot et al., 2001). In addition, diverse transducers can be used for conversion of the biochemical response into a quantifiable analytical signal, being classified according to their physical principles as electrochemical, electrical, optical, piezoelectric, calorimetric, acoustic, and magnetic

(Hulanicki et al., 1991; Sethi, 1994; Marco and Barcelo, 1996; Thévenot et al., 2001).

Electrochemical impedance spectroscopy (EIS) comprises the current main strategy used for the development of biodevices based on AMPs, as shown in Table 1. EIS is an effective method for the investigation of modified surfaces and monitoring of interfacial processes, allowing the characterization of electrochemical systems (Stoynov and Savova-Stoynova, 1986; Bard and Faulkner, 2000; Lasia, 2002). In the EIS technique, a signal of perturbation of low amplitude, in the form of sinusoidal potential and different frequency values, is applied to the transducer, forming sinusoidal alternating current (Macdonald, 1990, 1992; Katz and Willner, 2003; Chang and Park, 2010). EIS is a valuable tool for interfacial phenomena analyses occurring between electrode and solution. Dielectric analyses allows for the evaluation of electron charge-transfer parameters, double-layer, electrical layer, and modeling the experimental data to equivalent circuits (Stoynov and Savova-Stoynova, 1986; Macdonald, 1990; Bott, 2001; Oliveira et al., 2011). EIS is a technique of great relevance due to numerous technological applications such as the development of biosensors, studies of dielectric materials, corrosion processes, biofuel cells and rechargeable batteries (Macdonald, 1992; Guan et al., 2004). EIS is a new experimental approach for understanding AMP mechanisms of action through lipid membrane properties by monitoring changes such as thickness, ion permeability and homogeneity after peptide exposure (Chang et al., 2008; Oliveira et al., 2011; Nascimento et al., 2012; Sugihara et al., 2012).

Several biological elements can be used for the manufacturing of biosensitive systems, such as cell receptors, enzymes, antibodies, antigens, nucleic acids, aptamers, lectins, and peptides (Byfield and Abuknesha, 1994; Guan et al., 2004; Caygill et al., 2010). In particular, AMPs are outstanding molecules due to their wide biotechnological applications (Dawson and Liu, 2008; Meng et al., 2010; Seo et al., 2012). In addition, AMPs are promising molecules for the development of biosensors due to the possibility of recognizing a wide variety of pathogenic agents, including bacteria,




Table 1 | Antimicrobial peptides, their immobilization substrates and molecular targets with the respective detection limits obtained through different techniques.

AMPs	Substrate immobilization	Target recognition	Detection limit	Technique	Reference
Magainin I	Gold surface – lipoic acid <i>N</i> -hydroxysuccinimide – ferrocene	<i>Escherichia coli</i> O157:H7	10 ³ CFU · mL ⁻¹	Electrochemical impedance	Li et al. (2014)
Magainin I	Glass microbeads – <i>N</i> -[γ-maleimidobutyryloxy] succinimide ester (GMBS) – cysteine	Non-pathogenic <i>Escherichia coli</i>	10 ³ cells · mL ⁻¹	Fluorescence microscopy	Yoo et al. (2014)
Magainin I	Gold surface – cysteine	<i>Salmonella</i> <i>typhimurium</i>	10 ³ CFU · mL ⁻¹	Electrical impedance	Mannoor et al. (2010)
Leucocin A	Gold surface – cysteamine	<i>Listeria</i> <i>monocytogenes</i>	10 ³ CFU · mL ⁻¹	Electrical impedance	Etayash et al. (2014)
Bactenecin	Glass slide – 3-mercaptopropyl triethoxysilane (MPTES) – GMBS	<i>Brucella melitensis</i> Vaccinia virus Venezuelan equine encephalitis virus <i>Coxiella burnetti</i>	5 × 10 ⁴ cells · mL ⁻¹ <5 × 10 ⁵ PFU · mL ⁻¹ <5 × 10 ⁵ PFU · mL ⁻¹ 5 × 10 ⁵ cells · mL ⁻¹	Fluorescence spectroscopy	Kulagina et al. (2007)
G10KHc	Gold surface – cysteine	<i>Pseudomonas</i> <i>aeruginosa</i>	10 ⁵ CFU · mL ⁻¹	Electrical impedance	Lillehoj et al. (2014)
C16G2cys	Gold surface – cysteine	<i>Streptococcus</i> <i>mutans</i>	10 ⁵ CFU · mL ⁻¹	Electrical impedance	Lillehoj et al. (2014)
Cecoprin A	Glass slide – poly(dimethyl) siloxane (PDMS)	Botulinum toxin A	1 ng · mL ⁻¹	Fluorescence spectroscopy	Kulagina et al. (2006)
Melittin	Glass slide – PDMS	Botulinum toxin B	10 ng · mL ⁻¹	Fluorescence spectroscopy	Kulagina et al. (2006)

*CFU, colony forming units; PFU, plaque forming units.

fungi, toxins, and viruses with lipoprotein envelope (Zasloff, 2002; De Smet and Contreras, 2005; Kulagina et al., 2007).

The basic principle that enables the use of AMPs in biosensors is its ability to selectively interact with the cell membrane components of pathogens (Kulagina et al., 2005). The (bio) interaction is mainly driven by electrostatic forces, hydrogen bonds, and hydrophobic interactions (Seelig, 2004). Biological recognition thermodynamics depends on lipid membrane composition and peptide properties such as hydrophobicity, amphipathicity, molecular charge, and degree of secondary structure angle (Mozsolits et al., 2001; Papo and Shai, 2003; Fernandes et al., 2012; Oliveira et al., 2013). Therefore, AMP specificity is based on the different molecular affinities, allowing their employment in diagnostic tests (Kulagina et al., 2005; Etayash et al., 2014; Li et al., 2014).

Interdigitated electrodes are an excellent alternative for biosensor development. Mannoor et al. (2010) reported the development of an interdigitated capacitive biosensor for detection of *Escherichia coli* and *Salmonella typhimurium*. In this study, magainin I was immobilized on gold microelectrodes via its C-terminal cysteine residue, and the binding capacity to bacterial cells was evaluated by EIS. The changes in the dielectric properties of the modified electrode surface allowed the determination of selectivity

for pathogenic Gram-negative bacteria *E. coli* and *Salmonella typhimurium* in relation to non-pathogenic *E. coli* and the Gram-positive species *Listeria monocytogenes* (Mannoor et al., 2010).

Recently, Li et al. (2014) developed an impedimetric biosensor with magainin I conjugated to a structured film of ferrocene for bacterial detection. The biosensor showed a preferential selectivity for *E. coli* O157: H7 followed by non-pathogenic *E. coli* K12, *Bacillus subtilis* and *Staphylococcus epidermidis*. In addition, a detection limit was observed for *E. coli* O157: H7 of 10³ CFU mL⁻¹.

A microelectromechanical sensor using two synthetic peptides (C16G2cys or G10KHc) was developed for the detection of *Streptococcus mutans* and *Pseudomonas aeruginosa* (Lillehoj et al., 2014). It is important to note that C16G2cys and G10KHc have a cysteine amino acid residue in the C-terminus, and the -SH group of the cysteine can be used to promote binding on the gold surface, resulting in the vertical orientation of the recognition molecules.

Among the optical properties explored for the construction of biosensors, one of the most prevalent is fluorescence, evaluated by a significant number of techniques that allow the analysis of the systems conjugated with fluorophores and molecular targets quantification (Gauglitz, 2005; Altschuh et al., 2006; Lazcka et al., 2007; Fan et al., 2008). Fluorescence is a phenomenon of photon emissions, resulting from the passage of a valence electron in

an orbital of ground state to an orbital of higher energy to the absorption of radiation of appropriate wavelength. Upon returning to the state of origin, photons are liberated with bottom energy to absorbed light. Fluorescence spectroscopy allows for the quantification of the analyte with sensibility, low cost, and easy implementation (Eaton, 1990; Lazcka et al., 2007). Thus, the molecular recognition is assessed through variations in fluorescence properties after the interaction of the bioreceptor with the specific target (McFarland and Finney, 2001; Altschuh et al., 2006).

In this context, magainin I was immobilized on glass slides modified with 3-mercaptopropyl triethoxysilane (MPTES) and *N*-(γ -maleimidobutyryloxy) succinimide ester (GMBS) through direct covalent binding or by avidin–biotin coupling, being used as a recognition molecule for *E. coli* O157: H7 and *Salmonella typhimurium* (Kulagina et al., 2005). The immobilization method interferes with biosensor sensitivity, and the direct binding of magainin I reduces non-specific interactions, resulting in detection limits of 1.6×10^5 and 6.5×10^4 cells mL⁻¹ for *E. coli* and *Salmonella typhimurium* labeled with Cy5, respectively. Through the indirect method values of 6.8×10^5 and 5.6×10^5 cells mL⁻¹ were obtained (Kulagina et al., 2005). Moreover, cecropin A, parasin, magainin I, and polymyxin B and E, immobilized on a glass slide modified with MPTES and GMBS, were also used in screening fluorescent assays for detection of *E. coli* O157:H7 and *Salmonella typhimurium* (Kulagina et al., 2006). The assay in “sandwich” format, with the employment of Cy3-labeled anti-*E. coli* or anti-*Salmonella*, demonstrated different AMP affinities for pathogenic species (Kulagina et al., 2006). In another study, cecropin (A, B, and P), parasin, magainin I, polymyxin (B and E), melittin, and batenecin were evaluated for the biodetection of viral particles and bacterial cells using Cy3 (Kulagina et al., 2007).

Cecropin P1 (CP1), SMAP-29, and PGQ were used as alternative molecules for detection of *E. coli* O157:H7 in substitution of the anti-*E. coli* O157:H7 antibody (Arcidiacono et al., 2008). Through screening in solution and quantification of fluorescence, it was verified that the detection limits for Cy5 CP1, Cy5 SMAP, and Cy5 PGQ were 10^4 , 10^5 e 10^6 CFU mL⁻¹, respectively, in comparison to 10^5 CFU mL⁻¹ for Cy5 anti-*E. coli* O157 antibody. Because of the high sensitivity of the CP1 and specificity of the anti-*E. coli* O157:H7 antibody, a prototype immuno-magnetic bead biosensor was developed, resulting in a 10-fold improvement in sensitivity (Arcidiacono et al., 2008).

Glass microspheres (GMs) based on microfluidic chip were used for magainin I immobilization as a new method for *E. coli* biodetection (Yoo et al., 2014). GMs provided greater detection efficiency through the increase of superficial area of adsorption relative to the volume, enabling a greater number of bonds between microorganism-AMPs. The biodevice presented a good detection efficiency of 87% in a limit of 10^3 cells mL⁻¹ obtained by image analysis under fluorescence microscopy (Yoo et al., 2014).

The most common application of AMP-based biosensors is the identification of diverse bacteria such as *E. coli*, *Salmonella typhimurium*, *L. monocytogenes*, *Streptococcus mutans*, *P. aeruginosa*, and others (Mannoor et al., 2010; Li et al., 2014). In this scenario, AMPs stand out for presenting differential characteristics

and large applicability potential, since AMPs are highly stable under unsuitable conditions and can be continuously exposed to natural surroundings (Casteels et al., 1989; Mannoor et al., 2010; Laverty et al., 2011; Hassan et al., 2012; Dawgul et al., 2014). AMPs are capable of interacting with invariant components of the target surfaces (Kulagina et al., 2006), providing each peptide with the possibility of recognizing a variety of pathogens (Zasloff, 2002; De Smet and Contreras, 2005; Kulagina et al., 2007). Despite the importance of bacterial diagnosis, diverse studies have undertaken to develop detection systems for other microorganisms and toxins (Zasloff, 2002; De Smet and Contreras, 2005; Kulagina et al., 2007).

Therefore, AMPs can be used in nanostructured platforms for the detection of pathogenic agents due to their singular properties, ease of acquisition, and relevant biological activity (Kulagina et al., 2005; Yonekita et al., 2013). AMPs are an excellent alternative for obtaining new, sensitive, inexpensive, portable, versatile, and fast methods of diagnostics for analyte identification and quantification (Caygill et al., 2010; Yoo et al., 2014). Thus, nanobiosensors based on AMPs can be used in basic and applied research, clinical analysis, commercial applications, microbiological quality control, and environmental monitoring (Higgins and Lowe, 1987; D’Orazio, 2003; Mehrvar and Abdi, 2004).

CONCLUSION

Antimicrobial peptides are considered promising molecules in the development of biodetection devices. Nanotechnology has enabled the construction of biosensors based on AMPs with more specificity and sensitivity for detection of pathogenic agents. Biosensors based on AMPs are a useful tool since it is possible to investigate a wide variety of molecular targets in real time, providing quantitative or semi-quantitative analytical information that is both specific and selective. In spite of AMP-based biosensors being prototypes and their remaining challenges to commercialization such as reproducibility, validation and proper standardization, they may be considered valuable alternatives for obtaining new diagnostic methods.

ACKNOWLEDGMENTS

The authors are grateful for the support provided by the FACEPE and Rede de Nanobiotecnologia/CAPES. Andrade and Oliveira are also gratefully for CNPq financial support (grant 310305/2012-8 and 310361/2012-5, respectively). Ribeiro and da Silva would like to thank FACEPE for M.Sc. and Ph.D. scholarships, respectively.

REFERENCES

- Abdel-Hamid, I., Ivnitski, D., Atanasov, P., and Wilkins, E. (1999). Flow-through immunofiltration assay system for rapid detection of *E. coli* O157:H7. *Biosens. Bioelectron.* 14, 309–316. doi: 10.1016/S0956-5663(99)00004-4
- Altschuh, D., Oncul, S., and Demchenko, A. P. (2006). Fluorescence sensing of intermolecular interactions and development of direct molecular biosensors. *J. Mol. Recognit.* 19, 459–477. doi: 10.1002/jmr.807
- Andersen, P., Munk, M. E., Pollock, J. M., and Doherty, T. M. (2000). Specific immune-based diagnosis of tuberculosis. *Lancet* 356, 1099–1104. doi: 10.1016/S0140-6736(00)02742-2
- Andreu, D., and Rivas, L. (1998). Animal antimicrobial peptides: an overview. *Biopolymers* 47, 415–433. doi: 10.1002/(SICI)1097-0282(1998)47:6<415::AID-BIP2>3.0.CO;2-D

- Arcidiacono, S., Pivarnik, P., Mello, C. M., and Senecal, A. (2008). Cy5 labeled antimicrobial peptides for enhanced detection of *Escherichia coli* O157:H7. *Biosens. Bioelectron.* 23, 1721–1727. doi: 10.1016/j.bios.2008.02.005
- Bard, J. B. F., and Faulkner, L. R. (2000). *Electrochemical Methods: Fundamentals and Applications*. New York, NY: Wiley.
- Barkalina, N., Charalambous, C., Jones, C., and Coward, K. (2014). Nanotechnology in reproductive medicine: emerging applications of nanomaterials. *Nanomedicine* 14, 1549–9634. doi: 10.1016/j.nano.2014.01.001
- Bechinger, B., and Lohner, K. (2006). Detergent-like actions of linear amphipathic cationic antimicrobial peptides. *Biochim. Biophys. Acta* 1758, 1529–1539. doi: 10.1016/j.bbamem.2006.07.001
- Benes, J., Gabrielova, A., Horova, B., Zitko, K., Rozsypal, H., Dzupova, O., et al. (1996). [Comparison of various methods of hemoculture tests]. *Epidemiol. Mikrobiol. Immunol.* 45, 101–107. doi: 10.4103/0974-2727.129804
- Bi, L., Yang, L., Narsimhan, G., Bhunia, A. K., and Yao, Y. (2011). Designing carbohydrate nanoparticles for prolonged efficacy of antimicrobial peptide. *J. Control. Release* 150, 150–156. doi: 10.1016/j.jconrel.2010.11.024
- Bott, A. W. (2001). Electrochemical techniques for the characterization of redox polymers. *Curr. Sep.* 19, 71–75.
- Bragulat, M. R., Martinez, E., Castella, G., and Cabanes, F. J. (2004). Selective efficacy of culture media recommended for isolation and enumeration of *Fusarium* spp. *J. Food Prot.* 67, 207–211.
- Brogden, K. A., Guthmiller, J. M., Salzet, M., and Zasloff, M. (2005). The nervous system and innate immunity: the neuropeptide connection. *Nat. Immunol.* 6, 558–564. doi: 10.1038/ni1209
- Byfield, M. P., and Abuknesha, R. A. (1994). Biochemical aspects of biosensors. *Biosens. Bioelectron.* 9, 373–400. doi: 10.1016/0956-5663(94)80038-3
- Casteels, P., Ampe, C., Jacobs, F., Vaecq, M., and Tempst, P. (1989). Apidaecins: antibacterial peptides from honeybees. *EMBO J.* 8, 2387–2391.
- Caygill, R. L., Blair, G. E., and Millner, P. A. (2010). A review on viral biosensors to detect human pathogens. *Anal. Chim. Acta* 681, 8–15. doi: 10.1016/j.aca.2010.09.038
- Chang, B. Y., and Park, S. M. (2010). Electrochemical impedance spectroscopy. *Annu. Rev. Anal. Chem. (Palo Alto Calif)* 3, 207–229. doi: 10.1146/annurev.anchem.012809.102211
- Chang, W. K., Wimley, W. C., Searson, P. C., Hristova, K., and Merzlyakov, M. (2008). Characterization of antimicrobial peptide activity by electrochemical impedance spectroscopy. *Biochim. Biophys. Acta* 1778, 2430–2436. doi: 10.1016/j.bbamem.2008.06.016
- Cheema, U. B., Younas, M., Sultan, J. I., Iqbal, A., Tariq, M., and Waheed, A. (2011). Antimicrobial peptides: an alternative of antibiotics in ruminants. *Adv. Agric. Biotechnol.* 2, 15–21.
- Costa, F., Carvalho, I. E., Montelaro, R. C., Gomes, P., and Martins, M. C. (2011). Covalent immobilization of antimicrobial peptides (AMPs) onto biomaterial surfaces. *Acta Biomater.* 7, 1431–1440. doi: 10.1016/j.actbio.2010.11.005
- Dawgul, M., Maciejewska, M., Jaskiewicz, M., Karafova, A., and Kamysz, W. (2014). Antimicrobial peptides as potential tool to fight bacterial biofilm. *Acta Pol. Pharm.* 71, 39–47.
- Dawson, R. M., and Liu, C. Q. (2008). Properties and applications of antimicrobial peptides in biodefense against biological warfare threat agents. *Crit. Rev. Microbiol.* 34, 89–107. doi: 10.1080/10408410802143808
- Deisingh, A. K., and Thompson, M. (2004). Biosensors for the detection of bacteria. *Can. J. Microbiol.* 50, 69–77. doi: 10.1139/w03-095
- De Smet, K., and Contreras, R. (2005). Human antimicrobial peptides: defensins, cathelicidins and histatins. *Biotechnol. Lett.* 27, 1337–1347. doi: 10.1007/s10529-005-0936-5
- D’Orazio, P. (2003). Biosensors in clinical chemistry. *Clin. Chim. Acta* 334, 41–69. doi: 10.1016/S0009-8981(03)00241-9
- Eaton, D. F. (1990). Recommended methods for fluorescence decay analysis. *Pure Appl. Chem.* 62, 1631–1648. doi: 10.1351/pac199062081631
- Etayash, H., Jiang, K., Thundat, T., and Kaur, K. (2014). Impedimetric detection of pathogenic gram-positive bacteria using an antimicrobial peptide from class IIa bacteriocins. *Anal. Chem.* 86, 1693–1700. doi: 10.1021/ac4034938
- Fan, X. D., White, I. M., Shopova, S. I., Zhu, H. Y., Suter, J. D., and Sun, Y. Z. (2008). Sensitive optical biosensors for unlabeled targets: a review. *Anal. Chim. Acta* 620, 8–26. doi: 10.1016/j.aca.2008.05.022
- Fernandes, F. C., Rigden, D. J., and Franco, O. L. (2012). Prediction of antimicrobial peptides based on the adaptive neuro-fuzzy inference system application. *Biopolymers* 98, 280–287. doi: 10.1002/bip.22066
- Fournier-Wirth, C., and Coste, J. (2010). Nanotechnologies for pathogen detection: future alternatives? *Biologicals* 38, 9–13. doi: 10.1016/j.biologics.2009.10.010
- Gauglitz, G. (2005). Direct optical sensors: principles and selected applications. *Anal. Bioanal. Chem.* 381, 141–155. doi: 10.1007/s00216-004-2895-4
- Guan, J. G., Miao, Y. Q., and Zhang, Q. J. (2004). Impedimetric biosensors. *J. Biosci. Bioeng.* 97, 219–226. doi: 10.1016/S1389-1723(04)70195-4
- Hancock, R. E. W., and Sahl, H. G. (2006). Antimicrobial and host-defense peptides as new anti-infective therapeutic strategies. *Nat. Biotechnol.* 24, 1551–1557. doi: 10.1038/nbt1267
- Hassan, M., Kjos, M., Nes, I. F., Diep, D. B., and Lotfipour, F. (2012). Natural antimicrobial peptides from bacteria: characteristics and potential applications to fight against antibiotic resistance. *J. Appl. Microbiol.* 113, 723–736. doi: 10.1111/j.1365-2672.2012.05338.x
- Higgins, I. J., and Lowe, C. R. (1987). Introduction to the principles and applications of biosensors. *Philos. Trans. R. Soc. Lond. B Biol. Sci.* 316, 3–11. doi: 10.1098/rstb.1987.0013
- Hulanicki, A., Glab, S., and Ingman, F. (1991). Chemical sensors definitions and classification. *Pure Appl. Chem.* 63, 1247–1250. doi: 10.1351/pac199163091247
- Jianrong, C., Yuqing, M., Nongyue, H., Xiaohua, W., and Sijiao, L. (2004). Nanotechnology and biosensors. *Biotechnol. Adv.* 22, 505–518. doi: 10.1016/j.biotechadv.2004.03.004
- Junhui, Z., Hong, C., and Ruifu, Y. (1997). DNA based biosensors. *Biotechnol. Adv.* 15, 43–58. doi: 10.1016/S0734-9750(97)00003-7
- Katz, E., and Willner, I. (2003). Probing biomolecular interactions at conductive and semiconductive surfaces by impedance spectroscopy: routes to impedimetric immunosensors, DNA-Sensors, and enzyme biosensors. *Electroanalysis* 15, 913–947. doi: 10.1002/elan.200390114
- Keymanesh, K., Soltani, S., and Sardari, S. (2009). Application of antimicrobial peptides in agriculture and food industry. *World J. Microb. Biotechnol.* 25, 933–944. doi: 10.1007/s11274-009-9984-7
- Kulagina, N. V., Lassman, M. E., Ligler, F. S., and Taitt, C. R. (2005). Antimicrobial peptides for detection of bacteria in biosensor assays. *Anal. Chem.* 77, 6504–6508. doi: 10.1021/ac050639r
- Kulagina, N. V., Shaffer, K. M., Anderson, G. P., Ligler, F. S., and Taitt, C. R. (2006). Antimicrobial peptide-based array for *Escherichia coli* and *Salmonella* screening. *Anal. Chim. Acta* 575, 9–15. doi: 10.1016/j.aca.2006.05.082
- Kulagina, N. V., Shaffer, K. M., Ligler, F. S., and Taitt, C. R. (2007). Antimicrobial peptides as new recognition molecules for screening challenging species. *Sens. Actuators B Chem.* 121, 150–157. doi: 10.1016/j.snb.2006.09.044
- Lasia, A. (2002). “Electrochemical impedance spectroscopy and its applications,” in *Modern Aspects of Electrochemistry*, 1st Edn, eds B. E. Conway, J. O. M. Bockris, and R. White (New York: Springer), 143–248.
- Laverty, G., Gorman, S. P., and Gilmore, B. F. (2011). The potential of antimicrobial peptides as biosensors. *Int. J. Mol. Sci.* 12, 6566–6596. doi: 10.3390/ijms12106566
- Lazcka, O., Del Campo, F. J., and Munoz, F. X. (2007). Pathogen detection: a perspective of traditional methods and biosensors. *Biosens. Bioelectron.* 22, 1205–1217. doi: 10.1016/j.bios.2006.06.036
- Lillehoj, P. B., Kaplan, C. W., He, J., Shi, W. Y., and Ho, C. M. (2014). Rapid, electrical impedance detection of bacterial pathogens using immobilized antimicrobial peptides. *J. Lab. Autom.* 19, 42–49. doi: 10.1177/2211068213495207
- Li, Y., Afrasiabi, R., Fathi, F., Wang, N., Xiang, C., Love, R., et al. (2014). Impedance based detection of pathogenic *E. coli* O157:H7 using a ferrocene-antimicrobial peptide modified biosensor. *Biosens. Bioelectron.* 58, 193–199. doi: 10.1016/j.bios.2014.02.045
- Li, Y., Schluessener, H. J., and Xu, S. (2010). Gold nanoparticle-based biosensors. *Gold Bull.* 43, 29–41. doi: 10.1007/BF03214964
- Li, Y., Xiang, Q., Zhang, Q., Huang, Y., and Su, Z. (2012). Overview on the recent study of antimicrobial peptides: origins, functions, relative mechanisms and application. *Peptides* 37, 207–215. doi: 10.1016/j.peptides.2012.07.001
- Macdonald, D. D. (1990). Review of mechanistic analysis by electrochemical impedance spectroscopy. *Electrochim. Acta* 35, 1509–1525. doi: 10.1016/0013-4686(90)80005-9
- Macdonald, J. R. (1992). Impedance spectroscopy. *Ann. Biomed. Eng.* 20, 289–305. doi: 10.1007/BF02368532

- Mannoor, M. S., Zhang, S., Link, A. J., and McAlpine, M. C. (2010). Electrical detection of pathogenic bacteria via immobilized antimicrobial peptides. *Proc. Natl. Acad. Sci. U.S.A.* 107, 19207–19212. doi: 10.1073/pnas.1008768107
- Marco, M. P., and Barcelo, D. (1996). Environmental applications of analytical biosensors. *Meas. Sci. Technol.* 7, 1547–1562. doi: 10.1088/0957-0233/7/11/002
- Mazzoni, A., Fiacchini, M., Cevenini, R., Gatti, M., and Nanetti, A. (1987). Incidence of endogenous and exogenous opportunistic fungal infections in relation to different prophylactic measures: clinical and microbiological study of 431 immunocompromised patients. *Chemioterapia* 6, 374–376.
- McFarland, S. A., and Finney, N. S. (2001). Fluorescent chemosensors based on conformational restriction of a biaryl fluorophore. *J. Am. Chem. Soc.* 123, 1260–1261. doi: 10.1021/ja005701a
- Mehrvar, M., and Abdi, M. (2004). Recent developments, characteristics, and potential applications of electrochemical biosensors. *Anal. Sci.* 20, 1113–1126. doi: 10.2116/analsci.20.1113
- Mendez, E., Moreno, A., Colilla, F., Pelaez, F., Limas, G. G., Mendez, R., et al. (1990). Primary structure and inhibition of protein synthesis in eukaryotic cell-free system of a novel thionin, gamma-hordothionin, from barley endosperm. *Eur. J. Biochem.* 194, 533–539. doi: 10.1111/j.1432-1033.1990.tb15649.x
- Meng, S., Xu, H., and Wang, F. (2010). Research advances of antimicrobial peptides and applications in food industry and agriculture. *Curr. Prot. Pept. Sci.* 11, 264–273. doi: 10.2174/138920310791233369
- Mozsolits, H., Wirth, H. J., Werkmeister, J., and Aguilar, M. I. (2001). Analysis of antimicrobial peptide interactions with hybrid bilayer membrane systems using surface plasmon resonance. *Biochim. Biophys. Acta* 1512, 64–76. doi: 10.1016/S0005-2736(01)00303-0
- Murthy, S., Keystone, J., and Kisson, N. (2013). Infections of the developing world. *Crit. Care Clin.* 29, 485–507. doi: 10.1016/j.ccc.2013.03.005
- Nascimento, J. M., Franco, O. L., Oliveira, M. D. L., and Andrade, C. A. S. (2012). Evaluation of magainin I interactions with lipid membranes: an optical and electrochemical study. *Chem. Phys. Lipids* 165, 537–544. doi: 10.1016/j.chemphyslip.2012.05.002
- Nguyen, L. T., Haney, E. E., and Vogel, H. J. (2011). The expanding scope of antimicrobial peptide structures and their modes of action. *Trends Biotechnol.* 29, 464–472. doi: 10.1016/j.tibtech.2011.05.001
- Oliveira, M. D. L., Abdalla, D. S. P., Guilherme, D. F., Faulin, T. E. S., and Andrade, C. A. S. (2011). Impedimetric immunosensor for electronegative low density lipoprotein (LDL-) based on monoclonal antibody adsorbed on (polyvinyl formal)-gold nanoparticles matrix. *Sens. Actuators B Chem.* 155, 775–781. doi: 10.1016/j.snb.2011.01.046
- Oliveira, M. D. L., Franco, O. L., Nascimento, J. M., Melo, C. P., and Andrade, C. A. S. (2013). Mechanistic aspects of peptide-membrane interactions determined by optical, dielectric and piezoelectric techniques: an overview. *Curr. Prot. Pept. Sci.* 14, 543–555. doi: 10.2174/13892037113149990070
- Palka-Santini, M., Cleven, B. E., Eichinger, L., Krönke, M., and Krut, O. (2009). Large scale multiplex PCR improves pathogen detection by DNA microarrays. *BMC Microbiol.* 9:1. doi: 10.1186/1471-2180-9-1
- Papo, N., and Shai, Y. (2003). Exploring peptide membrane interaction using surface plasmon resonance: differentiation between pore formation versus membrane disruption by lytic peptides. *Biochemistry* 42, 458–466. doi: 10.1021/bi0267846
- Rapaport, D., and Shai, Y. (1991). Interaction of fluorescently labeled pardaxin and its analogs with lipid bilayers. *J. Biol. Chem.* 266, 23769–23775.
- Reddy, K. V. R., Yedery, R. D., and Aranha, C. (2004). Antimicrobial peptides: premises and promises. *Int. J. Antimicrob. Agents* 24, 536–547. doi: 10.1016/j.ijantimicag.2004.09.005
- Schmidtchen, A., Pasupuleti, M., and Malmsten, M. (2014). Effect of hydrophobic modifications in antimicrobial peptides. *Adv. Colloid Interface Sci.* 205, 265–274. doi: 10.1016/j.cis.2013.06.009
- Schnapp, D., and Harris, A. (1998). Antibacterial peptides in bronchoalveolar lavage fluid. *Am. J. Respir. Cell Mol. Biol.* 19, 352–356. doi: 10.1165/ajrcmb.19.3.3384
- Schuerholz, T., Brandenburg, K., and Marx, G. (2012). Antimicrobial peptides and their potential application in inflammation and sepsis. *Crit. Care* 16:207. doi: 10.1186/cc11220
- Seelig, J. (2004). Thermodynamics of lipid-peptide interactions. *Biochim. Biophys. Acta* 1666, 40–50. doi: 10.1016/j.bbame.2004.08.004
- Seo, M. D., Won, H. S., Kim, J. H., Mishig-Ochir, T., and Lee, B. J. (2012). Antimicrobial peptides for therapeutic applications: a review. *Molecules* 17, 12276–12286. doi: 10.3390/molecules171012276
- Sethi, R. S. (1994). Transducer aspects of biosensors. *Biosens. Bioelectron.* 9, 243–264. doi: 10.1016/0956-5663(94)80127-4
- Singh, P. K., Shiha, M. J., and Kumar, A. (2014). Antibacterial responses of retinal Muller glia: production of antimicrobial peptides, oxidative burst and phagocytosis. *J. Neuroinflammation* 11:33. doi: 10.1186/1742-2094-11-33
- Sobczak, M., Debek, C., Oledzka, E., and Kozłowski, R. (2013). Polymeric systems of antimicrobial peptides -strategies and potential applications. *Molecules* 18, 14122–14137. doi: 10.3390/molecules181114122
- Stoynov, Z. B., and Savova-Stoynova, B. (1986). Structural simulation of electrochemical impedances. *J. Electroanal. Chem.* 209, 11–18. doi: 10.1016/0022-0728(86)80181-4
- Sugihara, K., Delaia, M., Szendro, I., Guillaume-Gentil, O., Vörös, J., and Zambelli, T. (2012). Simultaneous OWLS and EIS monitoring of supported lipid bilayers with the pore forming peptide melittin. *Sens. Actuators B Chem.* 161, 600–606. doi: 10.1016/j.snb.2011.11.007
- Szkal, D., and Pal, T. (2003). Comparison of media for the selective culture of enteroinvasive *Escherichia coli*. *Eur. J. Clin. Microbiol. Infect. Dis.* 22, 235–241.
- Thévenot, D. R., Toth, K., Durst, R. A., and Wilson, G. S. (2001). Electrochemical biosensors: recommended definitions and classification. *Biosens. Bioelectron.* 16, 121–131. doi: 10.1016/S0956-5663(01)00115-4
- Toke, O. (2005). Antimicrobial peptides: new candidates in the fight against bacterial infections. *Biopolymers* 80, 717–735. doi: 10.1002/bip.20286
- van't Hof, W., Veerman, E. C., Helmerhorst, E. J., and Amerongen, A. V. (2001). Antimicrobial peptides: properties and applicability. *Biol. Chem.* 382, 597–619. doi: 10.1515/BC.2001.072
- van Pelt, C., Verduin, C. M., Goessens, W. H., Vos, M. C., Tümmler, B., Segonds, C., et al. (1999). Identification of *Burkholderia* spp. in the clinical microbiology laboratory: comparison of conventional and molecular methods. *J. Clin. Microbiol.* 37, 2158–2164.
- Wang, J. (1998). DNA biosensors based on Peptide Nucleic Acid (PNA) recognition layers. A review. *Biosens. Bioelectron.* 13, 757–762. doi: 10.1016/S0956-5663(98)00039-6
- Wang, J., Rivas, G., Cai, X., Palecek, E., Nielsen, P., Shiraish, H., et al. (1997). DNA electrochemical biosensors for environmental monitoring. A review. *Anal. Chim. Acta* 347, 1–8. doi: 10.1016/S0003-2670(96)00598-3
- Yang, S., Lin, S., Kelen, G. D., Quinn, T. C., Dick, J. D., Gaydos, C. A., et al. (2002). Quantitative multiprobe PCR assay for simultaneous detection and identification to species level of bacterial pathogens. *J. Clin. Microbiol.* 40, 3449–3454. doi: 10.1128/JCM.40.9.3449-3454.2002
- Yeom, S. H., Kang, B. H., Kim, K. J., and Kang, S. W. (2011). Nanostructures in biosensor-a review. *Front. Biosci.* 16:997–1023. doi: 10.2741/3731
- Yonekita, T., Ohtsuki, R., Hojo, E., Morishita, N., Matsumoto, T., Aizawa, T., et al. (2013). Development of a novel multiplex lateral flow assay using an antimicrobial peptide for the detection of Shiga toxin-producing *Escherichia coli*. *J. Microbiol. Methods* 93, 251–256. doi: 10.1016/j.mimet.2013.03.006
- Yoo, J. H., Woo, D. H., Chang, M.-S., and Chun, M.-S. (2014). Microfluidic based biosensing for *Escherichia coli* detection by embedding antimicrobial peptide-labeled beads. *Sens. Actuators B Chem.* 191, 211–218. doi: 10.1016/j.snb.2013.09.105
- Zaslhoff, M. (2002). Antimicrobial peptides of multicellular organisms. *Nature* 415, 389–395. doi: 10.1038/415389a

Conflict of Interest Statement: The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Received: 30 May 2014; accepted: 04 August 2014; published online: 20 August 2014.
 Citation: Silva RR, Avelino KYPS, Ribeiro KL, Franco OL, Oliveira MDL and Andrade CAS (2014) Optical and dielectric sensors based on antimicrobial peptides for microorganism diagnosis. *Front. Microbiol.* 5:443. doi: 10.3389/fmicb.2014.00443
 This article was submitted to *Antimicrobials, Resistance and Chemotherapy*, a section of the journal *Frontiers in Microbiology*.
 Copyright © 2014 Silva, Avelino, Ribeiro, Franco, Oliveira and Andrade. This is an open-access article distributed under the terms of the Creative Commons Attribution License (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) or licensor are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.